

Service Request Number _____

PM Checklist

Site Name: _____

SwMR MR Planned Maintenance Checklist

FE Name: _____

Month: _____

Procedure

OK
Cleaned
Adjusted
Needs
Correction

Pre PM Checklist

Talk With Technologist About System Operation..... ☐ ☐ ☐ ☐

Look at Image Quality of Previous Exams..... ☐ ☐ ☐ ☐

Check and Set System Date and Time..... ☐ ☐ ☐ ☐

SAFETY AND REGULATORY

SYSTEM - Check Gating Cable/ Respiratory Bellows / PG Probe..... ☐ ☐ ☐ ☐

SYSTEM - Check Pneumatic Patient Alert System..... ☐ ☐ ☐ ☐

SYSTEM - Check Emergency Release of Cradle and Patient Transport..... ☐ ☐ ☐ ☐

SYSTEM - Check Patient Transport Casters and Arm board Set Screws..... ☐ ☐ ☐ ☐

MAGNET - Check Magnet Bore for Metal Objects and Remove..... ☐ ☐ ☐ ☐

MAGNET - Test Magnet Emergency Rundown (Test Battery, Htr LED)..... ☐ ☐ ☐ ☐

MAGNET - Check for Condensation or Leaks on Top and Around Magnet..... ☐ ☐ ☐ ☐

MAGNET - Record Cryostat Pressure and Flow Rates..... ☐ ☐ ☐ ☐

He Level 1 _____ Pressure _____ Flow One _____ Flow Two _____ (HFO ONLY) He Level 2 _____

MAGNET - Compressor Run-Time / Pressure (Dynamic - Static)..... ☐ ☐ ☐ ☐

Compressor One -Hrs _____ DY _____ -ST _____ = _____ Compressor Two - Hrs _____ DY _____ -ST _____ = _____

MAGNET - Coldhead and Magnet Temps..... ☐ ☐ ☐ ☐

Inner/RuO _____ Outer/Coldhead _____ Magnet RuO _____ Magnet Si410 _____ Water Temp _____ Water flow _____

Clean, Test and Check

Check Correlated Noise (RF Integrity / RF Door Seal, top hatch, clean, Check)..... ☐ ☐ ☐ ☐

Clean Light-Weight Cradle Wheels..... ☐ ☐ ☐ ☐

Check Table Functionality (Clutch, Alignment, Undock)..... ☐ ☐ ☐ ☐

Check Body Chiller Temp and Fluid..... ☐ ☐ ☐ ☐

Water Temp and pressure _____

Clean Patient Blower Filter..... ☐ ☐ ☐ ☐

Clean All Cabinet Filters..... ☐ ☐ ☐ ☐

Clean Console area (octane or GOC)..... ☐ ☐ ☐ ☐

Annual - January/February and First PM Safety Checks

Month: _____

Test Emergency off Buttons..... ☐ ☐ ☐ ☐

Test Emergency Stop Buttons..... ☐ ☐ ☐ ☐

Inspect all System Cables and Connections for Tightness and Security..... ☐ ☐ ☐ ☐

Check Condition and Placement of Warning Labels..... ☐ ☐ ☐ ☐

Check Condition and Placement of 0.5 mT Exclusion Warnings..... ☐ ☐ ☐ ☐

Check Magnet Exhaust Vent and Vent Glass..... ☐ ☐ ☐ ☐

Clean Water Chiller Filter - If filter is really dirty then need to clean on next PM and note status below..... ☐ ☐ ☐ ☐

Tip Seal and Solenoid Valve Replacement (Twinspeed)..... ☐ ☐ ☐ ☐

Change Out All Patient Cradle Wheels..... ☐ ☐ ☐ ☐

Inspect and Clean Erbttec Interior..... ☐ ☐ ☐ ☐

Supercon Shim Currents

Ax1 _____ Ax2 _____ Ax3 _____ Ax4 _____ Ax5 _____ Ax6 _____

T1-1 _____ T1-2 _____ T1-3 _____ T1-4 _____ T1-5 _____ T1-6 _____

Center Frequency _____ T2-1 _____ T2-2 _____ T2-3 _____ T2-4 _____ T2-5 _____ T2-6 _____

Service Request Number _____

PM Checklist

Site Name: _____

SwMR MR Planned Maintenance Checklist

FE Name: _____

Month: _____

Procedure

IMAGE QUALITY

OK
Cleaned
Adjusted
Needs
Correction

Check ISO-Center (Check Alignment Lights)..... ☐ ☐ ☐ ☐

APS-R1 _____ R2 _____ TG _____ AX _____ iso Z _____

Check Gradient Calibration: X,Y,Z-axis..... ☐ ☐ ☐ ☐

Cfxfull - Old _____ New _____ Cfyfull - Old _____ New _____ Czfzfull - Old _____ New _____

Check Gradient Calibration zoom mode (TRM ONLY): X,Y,Z-axis..... ☐ ☐ ☐ ☐

Cfxfull - Old _____ New _____ Cfyfull - Old _____ New _____ Czfzfull - Old _____ New _____

Check SNR and RF output- Head..... ☐ ☐ ☐ ☐

APS-R1 _____ R2 _____ TG _____ AX _____

Check SNR (CTLTOP)..... ☐ ☐ ☐ ☐

APS-R1 _____ R2 _____ TG _____ AX _____

Check SNR (CTLBOT)..... ☐ ☐ ☐ ☐

APS-R1 _____ R2 _____ TG _____ AX _____

Check SNR and RF output- Body..... ☐ ☐ ☐ ☐

APS-R1 _____ R2 _____ TG _____ AX _____

Check Shim (45 DSV / 22 DSV)(make sure to run auto L-coil and DC off-sets)..... ☐ ☐ ☐ ☐

APS-R1 _____ R2 _____ TG _____ AX _____

Shim Values Before X _____ Y _____ Z _____ Z2 _____

With Gradient Shim - 22cm DSV

Zero Gradient Shim - (Test mode) 45cm DSV

Shim Values After X _____ Y _____ Z _____ Z2 _____

Inhomogeneity _____ Z1(1,0) _____ X(1,1) _____ Z2X(3,1) _____
Spine _____ Z2(2,0) _____ Y(1,-1) _____ Z2Y(3,-1) _____
Shoulder _____ Z3(3,0) _____ ZX(2,1) _____ Zxy2(3,2) _____
Average B field _____ Z4(4,0) _____ ZY(2,-1) _____ ZXY(3,-2) _____
Std B field _____ Z5(5,0) _____ X2-Y2(2,2) _____ X3(3,3) _____
Std B field _____ Z6(6,0) _____ XY(2,-2) _____ Y3(3,-3) _____

Inhomogeneity _____ Z1(1,0) _____ X(1,1) _____ Z2X(3,1) _____
Spine _____ Z2(2,0) _____ Y(1,-1) _____ Z2Y(3,-1) _____
Shoulder _____ Z3(3,0) _____ ZX(2,1) _____ Zxy2(3,2) _____
Average B field _____ Z4(4,0) _____ ZY(2,-1) _____ ZXY(3,-2) _____
Std B field _____ Z5(5,0) _____ X2-Y2(2,2) _____ X3(3,3) _____
Std B field _____ Z6(6,0) _____ XY(2,-2) _____ Y3(3,-3) _____

T R M O N L Y	Shim Values Before				Zoom Mode (After)			
	Zoom Mode (zero)45DSV				Zoom Mode (After)			
	X	Y	Z	Z2	X	Y	Z	Z2
	Inhomogeneity	Z1(1,0)	X(1,1)	Z2X(3,1)	Inhomogeneity	Z1(1,0)	X(1,1)	Z2X(3,1)
	Spine	Z2(2,0)	Y(1,-1)	Z2Y(3,-1)	Spine	Z2(2,0)	Y(1,-1)	Z2Y(3,-1)
	Shoulder	Z3(3,0)	ZX(2,1)	Zxy2(3,2)	Shoulder	Z3(3,0)	ZX(2,1)	Zxy2(3,2)
	Average B field	Z4(4,0)	ZY(2,-1)	ZXY(3,-2)	Average B field	Z4(4,0)	ZY(2,-1)	ZXY(3,-2)
	Std B field	Z5(5,0)	X2-Y2(2,2)	X3(3,3)	Std B field	Z5(5,0)	X2-Y2(2,2)	X3(3,3)
	Std B field	Z6(6,0)	XY(2,-2)	Y3(3,-3)	Std B field	Z6(6,0)	XY(2,-2)	Y3(3,-3)

FINAL

Check and Delete Error Message Log..... ☐ ☐ ☐ ☐

Do Save INFO, Save GUI and Save Protocol..... ☐ ☐ ☐ ☐

RESTORATION - DQA Phantom scan, Remove Test Scans, and Check Cabinet Doors / Covers..... ☐ ☐ ☐ ☐

RESTORATION - Shut down system pre site procedure..... ☐ ☐ ☐ ☐

RESTORATION - Update site log book, E-mail update with any follow up's and status..... ☐ ☐ ☐ ☐

N
o
t
e
s