

TABLE OF CONTENTS

TABLE OF CONTENTS	1
DESCRIPTION	2
REVISION HISTORY	3

DESCRIPTION

The Full-Size PDU is a two-bay cabinet, and the Compact PDU is a single-bay cabinet. Procedures for the Full-Size PDU do not apply to the Compact PDU.



FATAL ELECTRIC SHOCK HAZARD!! LETHAL VOLTAGES ARE PRESENT WITHIN THE PDU EVEN WHEN ALL PDU BREAKERS ARE OFF. ENSURE THAT POWER AT MAIN DISCONNECT IS OFF, LOCKED, AND TAGGED BEFORE REPLACING PARTS.

1. Lock out and tag out facility circuit breaker using appropriate OSHA procedure. (Refer to CD-ROM *Dir. 2187583-3 [or -2], MR Release Signa 5x/8x Service Methods, Renewal Parts and Service Tools*, Procedure for Safety, Section 6, OSHA LOCKOUT/TAGOUT REQUIREMENTS.)
2. Verify that the three Phase indicator lights on the PDU are off.

Refer to *Direction 15404, Signa Advantage 1.5T, 1.0T, & 0.5T Renewal Parts* for illustrated parts breakdown of Full-Size PDU components.

REVISION HISTORY

REV	DATE	AUTHOR	PRIMARY REASONS FOR CHANGE
0	June17, 1998	M. Whitlow	Initial Conversion from Toolbook to Word.