

HLA

SCHEDULE D (HIGH USAGE SCANNER)

SITE:

DATE:

SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
COMPUTER			
Clean/Vacuum all exhaust vent holes	_____	_____	_____
Check that all fans work	_____	_____	_____
Clean air intake filter screen	_____	_____	_____
Clean media drive heads	_____	_____	_____
SYSTEM			
Verify emergency off buttons (table/console)	_____	_____	_____
Check operation of final scan abort control	_____	_____	_____
Check X-Ray on lights	_____	_____	_____
Update Site Log	_____	_____	_____
Check PM manual revision	_____	_____	_____
TABLE			
Inspect cradle for adhesive tape/verify operation	_____	_____	_____
Inspect encoder cable	_____	_____	_____
Check I/S, cradle, elevation pots/belt	_____	_____	_____
Check limit switches	_____	_____	_____
Inspect TGGC fans	_____	_____	_____
Clean table covers and cradle	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: