

IMAGE DATA SHEET

QA Phantom #2

QA#2 CONTRAST FACTOR

QA#2/S/10mm/120KV/170mA/2 sec .

Scan at 35mm. PERISTALTIC OFF

Exam # # of Slice	Visible Holes (View at window of 20)	Contrast factor	Comments/ Artifacts	Scan Tech Aprov
SPECS	See below	4.0 to 12.0	---	---

Contrast Factor	Lower limit number of visible holes *	Upper limit number of visible holes*	Smallest visible hole size
4.00 to 7.99	3	5	5mm
8.00 to 12.00	4	5	3mm

* Number of holes required to be visible , depending on contrast factor. Two out of four slices taken must meet specs.

Image Acceptance Date: _____ Image Reviewer: _____