

9800

SCHEDULE H (HIGH USAGE SCANNER)

SITE:

DATE:
SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
COMPUTER			
Clean magnetic tape unit	_____	_____	_____
GANTRY			
Clean tube radiator/ check fans	_____	_____	_____
Inspect for leaky tube wells	_____	_____	_____
Inspect collimator filter age	_____	_____	_____
GENERATOR			
Check generator lamps	_____	_____	_____
Check mA and kV meter accuracy	_____	_____	_____
Verify mA overload circuit	_____	_____	_____
Verify exposure time	_____	_____	_____
Calibrate 80 kw internal bleeder	_____	_____	_____
Clean generator filter/check fans	_____	_____	_____
Check PDU fans/ clean filters	_____	_____	_____
SYSTEM			
Update site log	_____	_____	_____
Check PM manual revision	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: