

SITE:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
COMPUTER			
Clean/Vacuum all exhaust vent holes	_____	_____	_____
Check that all fans work	_____	_____	_____
CONSOLE			
Check display monitor for raster burns	_____	_____	_____
Check for smooth rolling trackball/Clean	_____	_____	_____
Clean/ vacuum console interior	_____	_____	_____
GANTRY			
Check the number of gantry revolutions	_____	_____	_____
Remove and measure POWER brushes	_____	_____	_____
Remove and measure SIGNAL brushes	_____	_____	_____
Check ventilation holes and check fans	_____	_____	_____
Handling and removal of slip ring brush debris	_____	_____	_____
Gantry Slip Ring maintenance	_____	_____	_____
Check scan and laser windows	_____	_____	_____
HIGH VOLTAGE			
mA meter verification (HHS)	_____	_____	_____
KV test (HHS) or	_____	_____	_____
HV tank feedback resistors calibration	_____	_____	_____
Total KV measurement	_____	_____	_____
KV meter verification	_____	_____	_____
Auto mA cal	_____	_____	_____
Create kV Baseline	_____	_____	_____
mA & kV Verification (HHS)	_____	_____	_____
SYSTEM			
Check Web for Latest PM Manual Revision	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: