

LS2002 Software Installation Procedures

The information in this manual applies to the following CT Systems:

- *LightSpeed 2.X (LightSpeed Plus w/SDAS: 4-slice)*
- *LightSpeed 3.X (LightSpeed Ultra & LightSpeed Plus w/MDAS: 8-slice) & 4-slice, respectively)*
- *LightSpeed 4.X (LightSpeed 16 s/MDAS: 16-slice)*

Operating Documentation



LEGAL NOTES

TRADEMARKS

Adobe, the Adobe logo, Acrobat, the Acrobat logo, Exchange, and PostScript are trademarks of Adobe Systems Incorporated or its subsidiaries and may be registered in certain jurisdictions.

Microsoft is a registered trademark and Windows is a trademark of Microsoft Corporation.

All other products and their name brands are trademarks of their respective holders.

COPYRIGHTS

All Material Copyright © 2002-2005 by General Electric Company. All rights reserved.

IMPORTANT PRECAUTIONS

LANGUAGE

WARNING

- THIS SERVICE MANUAL IS AVAILABLE IN ENGLISH ONLY.
- IF A CUSTOMER'S SERVICE PROVIDER REQUIRES A LANGUAGE OTHER THAN ENGLISH, IT IS THE CUSTOMER'S RESPONSIBILITY TO PROVIDE TRANSLATION SERVICES.
- DO NOT ATTEMPT TO SERVICE THE EQUIPMENT UNLESS THIS SERVICE MANUAL HAS BEEN CONSULTED AND IS UNDERSTOOD.
- FAILURE TO HEED THIS WARNING MAY RESULT IN INJURY TO THE SERVICE PROVIDER, OPERATOR OR PATIENT FROM ELECTRIC SHOCK, MECHANICAL OR OTHER HAZARDS.

AVERTISSEMENT

- CE MANUEL DE MAINTENANCE N'EST DISPONIBLE QU'EN ANGLAIS.
- SI LE TECHNICIEN DU CLIENT A BESOIN DE CE MANUEL DANS UNE AUTRE LANGUE QUE L'ANGLAIS, C'EST AU CLIENT QU'IL INCOMBE DE LE FAIRE TRADUIRE.
- NE PAS TENTER D'INTERVENTION SUR LES ÉQUIPEMENTS TANT QUE LE MANUEL SERVICE N'A PAS ÉTÉ CONSULTÉ ET COMPRIS.
- LE NON-RESPECT DE CET AVERTISSEMENT PEUT ENTRAÎNER CHEZ LE TECHNICIEN, L'OPÉRATEUR OU LE PATIENT DES BLESSURES DUES À DES DANGERS ÉLECTRIQUES, MÉCANIQUES OU AUTRES.

WARNUNG

- DIESES KUNDENDIENST-HANDBUCH EXISTIERT NUR IN ENGLISCHER SPRACHE.
- FALLS EIN FREMDER KUNDENDIENST EINE ANDERE SPRACHE BENÖTIGT, IST ES AUFGABE DES KUNDEN FÜR EINE ENTSPRECHENDE ÜBERSETZUNG ZU SORGEN.
- VERSUCHEN SIE NICHT, DAS GERÄT ZU REPARIEREN, BEVOR DIESES KUNDENDIENST-HANDBUCH NICHT ZU RATE GEZOGEN UND VERSTANDEN WURDE.
- WIRD DIESE WARNUNG NICHT BEACHTET, SO KANN ES ZU VERLETZUNGEN DES KUNDENDIENSTTECHNIKERS, DES BEDIENERS ODER DES PATIENTEN DURCH ELEKTRISCHE SCHLÄGE, MECHANISCHE ODER SONSTIGE GEFAHREN KOMMEN.

AVISO

- ESTE MANUAL DE SERVICIO SÓLO EXISTE EN INGLÉS.
- SI ALGÚN PROVEEDOR DE SERVICIOS AJENO A GEMS SOLICITA UN IDIOMA QUE NO SEA EL INGLÉS, ES RESPONSABILIDAD DEL CLIENTE OFRECER UN SERVICIO DE TRADUCCIÓN.
- NO SE DEBERÁ DAR SERVICIO TÉCNICO AL EQUIPO, SIN HABER CONSULTADO Y COMPRENDIDO ESTE MANUAL DE SERVICIO.
- LA NO OBSERVANCIA DEL PRESENTE AVISO PUEDE DAR LUGAR A QUE EL PROVEEDOR DE SERVICIOS, EL OPERADOR O EL PACIENTE SUFRAN LESIONES PROVOCADAS POR CAUSAS ELÉCTRICAS, MECÁNICAS O DE OTRA NATURALEZA.

ATENÇÃO

- ESTE MANUAL DE ASSISTÊNCIA TÉCNICA SÓ SE ENCONTRA DISPONÍVEL EM INGLÊS.
- SE QUALQUER OUTRO SERVIÇO DE ASSISTÊNCIA TÉCNICA, QUE NÃO A GEMS, SOLICITAR ESTES MANUAIS NOUTRO IDIOMA, É DA RESPONSABILIDADE DO CLIENTE FORNECER OS SERVIÇOS DE TRADUÇÃO.
- NÃO TENHA TENTADO REPARAR O EQUIPAMENTO SEM TER CONSULTADO E COMPREENDIDO ESTE MANUAL DE ASSISTÊNCIA TÉCNICA.
- O NÃO CUMPRIMENTO DESTA AVISO PODE POR EM PERIGO A SEGURANÇA DO TÉCNICO, OPERADOR OU PACIENTE DEVIDO A CHOQUES ELÉTRICOS, MECÂNICOS OU OUTROS.

AVVERTENZA

- IL PRESENTE MANUALE DI MANUTENZIONE È DISPONIBILE SOLTANTO IN INGLESE.
- SE UN ADDETTO ALLA MANUTENZIONE ESTERNO ALLA GEMS RICHIEDE IL MANUALE IN UNA LINGUA DIVERSA, IL CLIENTE È TENUTO A PROVVEDERE DIRETTAMENTE ALLA TRADUZIONE.
- SI PROCEDA ALLA MANUTENZIONE DELL'APPARECCHIATURA SOLO DOPO AVER CONSULTATO IL PRESENTE MANUALE ED AVERNE COMPRESO IL CONTENUTO.
- NON TENERE CONTO DELLA PRESENTE AVVERTENZA POTREBBE FAR COMPIERE OPERAZIONI DA CUI DERIVINO LESIONI ALL'ADDETTO ALLA MANUTENZIONE, ALL'UTILIZZATORE ED AL PAZIENTE PER FOLGORAZIONE ELETTRICA, PER URTI MECCANICI OD ALTRI RISCHI.

警告

このサービスマニュアルには英語版しかありません。

GEMS 以外でサービスを担当される業者が英語以外の言語を要求される場合、翻訳作業はその業者の責任で行うものとさせていただきます。

このサービスマニュアルを熟読し理解せずに、装置のサービスを行わないで下さい。

この警告に従わない場合、サービスを担当される方、操作員あるいは患者さんが、感電や機械的又はその他の危険により負傷する可能性があります。

注意:

本维修手册仅存有英文本。

非 GEMS 公司的维修员要求非英文本的维修手册时，客户需自行负责翻译。

未详细阅读和完全了解本手册之前，不得进行维修。

忽略本注意事项会对维修员，操作员或病人造成触电，机械伤害或其他伤害。

DAMAGE IN TRANSPORTATION

Check for damage to property that may have occurred at the site during delivery, such as damage to floors, door frames or walls. If damage is found, notify the install specialist.

All packages should be closely examined at time of delivery. If damage is apparent, have notation "damage in shipment" written on all copies of the freight or express bill before delivery is accepted or "signed for" by a General Electric representative or a hospital receiving agent. Whether noted or concealed, damage MUST be reported to the carrier immediately upon discovery, or in any event, within 14 days after receipt, and the contents and containers held for inspection by the carrier. A transportation company will not pay a claim for damage if an inspection is not requested within this 14-day period.

To file a report:

- Call 1-800-548-3366 and use option 8.
- Fill out a report on <http://us44hdd21/sctq/InstallFulfill/InstalFulfillment.htm>
- Contact the local service coordinator for more information on this process.

Rev. Jan. 5, 2005

CERTIFIED ELECTRICAL CONTRACTOR STATEMENT

All electrical Installations that are preliminary to positioning of the equipment at the site prepared for the equipment shall be performed by licensed electrical contractors. In addition, electrical feeds into the Power Distribution Unit shall be performed by licensed electrical contractors. Other connections between pieces of electrical equipment, calibrations and testing shall be performed by qualified GE Healthcare personnel. The products involved (and the accompanying electrical installations) are highly sophisticated, and special engineering competence is required. In performing all electrical work on these products, GE Healthcare uses its own specially trained field engineers. All of GE's electrical work on these products will comply with the requirements of the applicable electrical codes.

The purchaser of GE Healthcare equipment shall only utilize qualified personnel (i.e., GE's field engineers, personnel of third-party service companies with equivalent training, or licensed electricians) to perform electrical servicing on the equipment.

IMPORTANT...X-RAY PROTECTION

X-ray equipment if not properly used may cause injury. Accordingly, the instructions herein contained should be thoroughly read and understood by everyone who will use the equipment before you attempt to place this equipment in operation. The General Electric Company, **GE Healthcare** Group, will be glad to assist and cooperate in placing this equipment in use.

Although this apparatus incorporates a high degree of protection against x-radiation other than the useful beam, no practical design of equipment can provide complete protection. Nor can any practical design compel the operator to take adequate precautions to prevent the possibility of any persons carelessly exposing themselves or others to radiation.

It is important that anyone having anything to do with x-radiation be properly trained and fully acquainted with the recommendations of the National Council on Radiation Protection and Measurements as published in NCRP Reports available from NCRP Publications, 7910 Woodmont Avenue, Room 1016, Bethesda, Maryland 20814, and of the International Commission on Radiation Protection, and take adequate steps to protect against injury.

The equipment is sold with the understanding that the General Electric Company, **GE Healthcare** Group, its agents, and representatives have no responsibility for injury or damage which may result from improper use of the equipment.

Various protective materials and devices are available. It is urged that such materials or devices be used.

IMPORTANT...RADIOACTIVE MATERIAL HANDLING

Only employees formally trained in radioactive materials handling and this equipment are authorized by the **GE Healthcare** Radiation Safety Officer to use radioactive materials to service this equipment.

GE Healthcare Services is required to notify the applicable U.S. state agency PRIOR to any source service event involving pin source handling. See NUC/PET Radioactive material guides for specific instruction or contact your EHS Specialist.

A radiation survey must be performed when a pin source has been removed and replaced. See Radiation Survey Form Instructions or contact your EHS Specialist.

Rev 2 (July 21, 2005)

LITHIUM BATTERY CAUTIONARY STATEMENTS



CAUTION
Risk of
Explosion

Danger of explosion if battery is incorrectly replaced. Replace only with the same or equivalent type recommended by the manufacturer. Discard used batteries according to the manufacturer's instructions.



ATTENTION
Danger
d'Explosion

Il y a danger d'explosion s'il y a remplacement incorrect de la batterie. Remplacer uniquement avec une batterie du même type ou d'un type recommandé par le constructeur. Mettre au rebut les batteries usagées conformément aux instructions du fabricant.

OMISSIONS & ERRORS

Customers, please contact your GE Healthcare Sales or Service representatives.

GE personnel, please use the **GE Healthcare** iTrak Process to report all omissions, errors, and defects in this publication.

Revision History

Revision	Date	Reason for change
0	06/25/01	Initial Release of Document, for LS2002
1	07/17/02	M3 updates.
2	10/31/02	Updated to include LightSpeed 16 product. Miscellaneous, General updates.
3	11/19/02	Corrected CD-ROM part numbers
4	12/17/02	M4 updates: Added Section 3.6 Load the LS2002M5/M5.1 Patch Updates to Chapter 1
5	12/20/02	Corrections to Patch C load instructions; added Patch C CD P/N.
6	11/19/03	M5 updates: Added Section 2.2.1 - LS2002 M5 Software, Section 3.6.1 - LS2002 M5 Patch A CD Installation, and Appendix B - Un-Install Instructions - LS2002 M5 Patch A.
7	7/27/04	Updated Damage in Transportation Statement and Chapter 1 Section 2.2.1 Patch CD information.
8	4/19/05	PQR13027166: In 3.5.2.4 - Configure the Network Settings, revised the rules for setting host names.
9	11/17/06	Renamed and replaced information in Chapter 1, section 3.6 regarding the LS2002 patch, now the G.6 and GraphicRxFile Transfer. Removed Appendix B. Updated Chapter 1, section 2.2.1; removed references to LS2002M4 from section 2.2.2 Replaced front and back covers; replaced the Important Precautions section.
10	3/20/08	PQRs 13109077, 13114503, and 13119713: Moved G.6 Patch and GraphicRx File Transfer sections to end of software installation instructions (New section 13) FCTge34603: In the GraphicRx File Transfer section, step 5 - Change the word "gin" to "bin" (seen in 2 places). In the Verify System Operation section, step 4.b - Change "TOOLBOX/UTILITIES" to "UTILITIES".

Table of Contents

Preface

Publication Conventions 17

Section 1.0

Safety & Hazard Information..... 17

- 1.1 Text and Character Representation..... 17
- 1.2 Graphical Representation 18

Section 2.0

Publication Conventions..... 19

- 2.1 General Paragraph and Character Styles..... 19
- 2.2 Page Layout..... 19
- 2.3 Computer Screen Output/Input Character Styles 20
- 2.4 Buttons, Switches and Keyboard Inputs (Hard & Soft Keys)..... 20

Chapter 1

Software Load Procedure 21

Section 1.0

Introduction..... 21

- 1.1 Document Overview 21
- 1.2 Software Load Procedure (LFC) Overview..... 21

Section 2.0

Before You Begin..... 22

- 2.1 Check System Software Version & Verify DAS Type 22
- 2.2 Gather Software CD-ROMs 22
 - 2.2.1 LS2002 M5 Software 22
- 2.3 Check CD-ROM Drive Operation..... 22
- 2.4 Save System State to MOD..... 23
- 2.5 Record KEY System Information 23

Section 3.0

Load Procedure 27

- 3.1 Shutdown the System..... 27
- 3.2 Boot Disk Partitioning Utility..... 27
- 3.3 Select Keyboard Language 28
- 3.4 Install Operating System..... 29
- 3.5 Install CT Application Software 32
 - 3.5.1 Starting the Installation 32
 - 3.5.2 Configure Application Software 33
 - 3.5.2.1 Configure System Settings..... 33
 - 3.5.2.2 Configure Preferences 34

3.5.2.3	Configure Hardware Settings.....	35
3.5.2.4	Configure the Network Settings	36
3.5.3	Finishing Up	37
3.6	Set Time and Date	37
3.7	Start CT Application Software	38
3.8	Restore System State	38
3.9	Download Flash Firmware	39
3.10	Adding Software Options	40
3.11	Install Cameras	42
3.12	Shutdown System	42
3.13	Load Patch G.6 and Graphic Rx File Transfer	43
3.13.1	Patch G.6 Load	43
3.13.2	GraphicRx File Transfer	43
3.14	Verify System Operation	44

Chapter 2

Partitioning Disks..... 47

Section 1.0	
Overview	47

Section 2.0	
Procedure	47

Chapter 3

Re-configuration of CT Application SW 49

Section 1.0	
Overview	49

Section 2.0	
Procedure	49

2.1	Overview	49
2.2	Starting "Reconfig"	50
2.3	System Configuration Settings	50
2.4	Finishing System Reconfiguration.....	51

Chapter 4

Laser & DICOM Camera Setup..... 53

Section 1.0	
Camera Setup	53

Section 2.0	
Data Sheets.....	60

Chapter 5

Changing System Time Zone 65

Chapter 6

Regenerating a Scan Database 67

Appendix A

Error Messages And Troubleshooting 69

Preface

Publication Conventions

Purpose: This section means to inform the reader on publication conventions used. So that the reader can identify safety and general material that is considered important by its format. This includes the interpretation of computer screen text as either input or output. There are a number of specific text and paragraph styles/conventions used within this section to accomplish this task. Please become familiar with the conventions used within this publication before proceeding.

Section 1.0

Safety & Hazard Information

1.1 Text and Character Representation

Within this publication, different paragraph and character styles have been used to indicate potential hazards. Paragraph prefixes, such as hazard, caution, danger and warning, are used to identify important safety information. Text (Hazard) styles are applied to the paragraph contents that is applicable to each specific safety statement. Words describe the type of potential hazard that may be encountered and are placed immediately before the paragraph it modifies. Safety information will normally include:

- Type of potential Hazard
- Nature of potential injury
- Causative condition
- How to avoid or correct the causative condition

EXAMPLES OF HAZARD STATEMENTS USED

A few examples are provided that have been adapted from GEMS' global document standard (2119696-100). They include paragraph prefixes and modified text styles.



CAUTION
Pinch Points
Loss of Data
Sharp Objects

Caution is used when a hazard exists which can or could cause minor injury to self or others if instructions are ignored. They include for example:

- Loss of critical patient data
- Crush or pinch points
- Sharp objects



DANGER
EXCESSIVE
VOLTAGE
CRUSH
POINT

DANGER IS USED WHEN A HAZARD EXISTS WHICH WILL CAUSE SEVERE PERSONAL INJURY OR DEATH IF INSTRUCTIONS ARE IGNORED. THEY CAN INCLUDE:

- ELECTROCUTION
- CRUSHING
- RADIATION

 WARNING
ROTATING
EQUIPMENT
BARE WIRES

WARNING IS USED WHEN A HAZARD EXISTS WHICH COULD OR CAN CAUSE SERIOUS PERSONAL INJURY OR DEATH IF INSTRUCTIONS ARE IGNORED. THEY CAN INCLUDE:

- Potential for shock
- Exposed wires
- Failure to Tag and lockout system power could allow for un-command motion.

 NOTICE
Equipment
Damage
Possible

Notice is used when a hazard is present that can cause property damage but has absolutely no personal injury risk. They can include:

- Disk drive will crash
- Internal mechanical damage, such as to the x-ray tube
- Coasting the rotor through resonance.

It's important that the reader not ignore hazard statements in this document.

1.2 Graphical Representation

Important information will always be preceded by the exclamation point  contained within a triangle, as seen throughout this chapter. In addition to text, several different graphical icons (symbols) may be used to make you aware of specific types of hazards that could possibly cause harm.

ELECTRICAL



LASER



MECHANICAL



HEAT



RADIATION



PINCH



Some others make you aware of specific procedures that should be followed.

AVOID STATIC ELECTRICITY



TAG AND LOCK OUT



WEAR EYE PROTECTION



Section 2.0 Publication Conventions

2.1 General Paragraph and Character Styles

Prefixes are used to highlight important non-safety related information. Paragraph prefixes (such as Purpose, Example, Comment and Note) are used to identify important but non-safety related information. Text styles and shading are also applied to text within each paragraph modified by the specific prefix.

EXAMPLES OF PREFIXES USED FOR GENERAL INFORMATION

Purpose: Introduces and provides meaning as to the information contained within the chapter, section or subsection, Such as used at the beginning this chapter for example.

Note: Conveys information that should be considered important to the reader.


Example: Used to make the reader aware that the paragraph(s) that follow are examples of information possibly stated previously.

Comment: Represents “additional” information that may or may not be relevant.

2.2 Page Layout

Publication Part Number & Revision Number

Publication Title



The current section and its title are always shown in the footer of the left (even) page.

An exclamation point in a triangle is used to indicate important information to the user.

Paragraphs preceded by **Alphanumeric** (e.g. numbers) characters is information that must be followed in a **specific order**.

The current chapter and its title are always shown in the footer of the right (odd) page.

Paragraphs preceded by **symbols** is information that has **no specific order**.

Headers and footers in this publication are designed to allow you to quickly identify your location. The document's part number and revision number appears in every header on every page. Odd

numbered page footers indicate the current chapter, its title and current page number. Even page footers show the current section and its title, as well current page number.

2.3 Computer Screen Output/Input Character Styles

Within this publication different character styles are used to indicate computer input and output text. Character (input, output, and variable) styles are used and applied to the text within a paragraph so as to indicated direction. Computer screen output and input is also formatted using mono (fixed width) spaced fonts.

Example:
Fixed Output

This paragraph denotes computer screen fixed output. It's output is fixed from the sense that it does not vary from application to application. It's the most commonly used style used to indicate filenames, paths and text.

Example:
Variable Output

This paragraph denotes computer screen output that is variable. Its output varies from application to application. Variable output is sometimes found placed between greater than and lesser than operators. For example: <variable_output>

Example:
Fixed Input

This paragraph denotes fixed input. It's typed input that will not vary from application to application. Fixed text the user is required to supply as input.

Example:
Variable Input

This paragraph denotes computer input that can vary from application to application. Variable text the user is required to supply as input. Variable input sometimes is placed between greater than and lesser than operators. For example: <variable_input>. In these cases, the (<>) operators are dropped prior to input. Exceptions are noted in the text.

2.4 Buttons, Switches and Keyboard Inputs (Hard & Soft Keys)

Different character styles are used to indicate actions requiring the reader to press either a hard or soft button, switch or key. Physical hardware, such as buttons and switches, are called hard keys because they are hard wired or mechanical in nature. A keyboard or on/off switch would be a hard key. Software or computer generated buttons are called soft keys because they are software generated. Software driven menu buttons are an example of such keys. Soft and hard keys are represented differently in this publication.

Example:
Hard Keys

A power switch **ON/OFF** or a keyboard key like **ENTER** is indicated by applying a character style that uses both over and under-lined bold text that is bold. This is a hard key.

Example:
Soft Keys

Whereas the computer **MENU** button that you would click with your mouse or touch with your hand uses over and under-lined regular text. This is a soft key.

Chapter 1

Software Load Procedure

Section 1.0 Introduction

1.1 Document Overview

- 1.) Before you begin, become familiar with the representation of computer text and character used in the load from cold procedure. Please review the [Preface - "Publication Conventions"](#), [Section 2.3, on page 20](#). Chapter 1 is organized as follows:
 - [Section 1.0 - "Introduction"](#): Overview of the SW installation procedure.
 - [Section 2.0 - "Before You Begin"](#): Things you should know and do before you begin.
 - [Section 3.0 - "Load Procedure"](#) for the operating system and CT applications software.
- 2.) It may be necessary to reference the following information in this document during the installation procedure, pending your specific needs:
 - [Chapter 2 - "Partitioning Disks"](#)
 - [Chapter 3 - "Re-configuration of CT Application SW"](#)
 - [Chapter 5 - "Changing System Time Zone"](#)
 - [Chapter 6 - "Regenerating a Scan Database"](#)
 - [Chapter 4 - "Laser & DICOM Camera Setup"](#)
 - [Appendix A - "Error Messages And Troubleshooting"](#)

1.2 Software Load Procedure (LFC) Overview

CT software load procedure, also referred to as Load From Cold (LFC), prepares the system disk, installs and configures the operating system, and installs and configures applications software on the OC computer.




Figure 1-1 Software Overview

Section 2.0

Before You Begin

2.1 Check System Software Version & Verify DAS Type

- 1.) To view version information for the system's currently running software, use the following command:

```
showprods | grep -i apps
```

- 2.) Verify and record (write it down) system DAS type:
Open a shell, and type: `cat usr/g/config/INFO` ENTER

2.2 Gather Software CD-ROMs

2.2.1 LS2002 M5 Software

Operating and applications software is contained on the first five (5) CD-ROMs listed below. The sixth CD is Patch G.6, which should also be loaded to bring software up to latest revision. To load system software, you must have the following CD-ROMs:

- 1.) IRIX Operating System CD-ROM Disk #1 (P/N 2356015)
- 2.) IRIX Operating System CD-ROM Disk #2 (P/N 2356016)
- 3.) IRIX Operating System CD-ROM Disk #3 (P/N 2356017)
- 4.) M5 or M5.1 Applications Software CD-ROM Disk #1
- 5.) M5 or M5.1 Applications Software CD-ROM Disk #2 LS2002M5 Software Patch G.6 (P/N 5142334)

Note: Patch G.6 contains numerous safety fixes and quality improvements, including the SmarTube Features. (The Patch can be obtained from Coakley or by contacting GE Service.)

2.3 Check CD-ROM Drive Operation

Since the CD-ROM Drive is seldom used except for loading system software, it is advisable at this time to perform a quick test on the ability of the CD-ROM drive to read data from a CD before you begin the LFC procedure. This test is accomplished by inserting and mounting one of the OS CD-ROM disks, and performing a checksum (read check) on the files in one of the directories on the CD. If any errors occur while the system is reading this CD-ROM, you must stop to investigate and correct the CD problem (for example the SCSI Bus, CD-ROM Drive, etc. problems).

- 1.) Open a UNIX shell.
- 2.) Type: `tail -f /usr/adm/SYSLOG` ENTER
- 3.) Open a second UNIX shell.
- 4.) Insert the IRIX Operating System CD-ROM Disk #1 into the CD-ROM Drive and type:
`mountCD /mnt` ENTER
`cd /mnt/dist` ENTER
`sum *` ENTER
- 5.) While the system is reading the CD-ROM and calculating checksums, you need to monitor the SYSLOG in the first UNIX shell window for any error messages that may occur that are related to SCSI or CD-ROM problems (timeouts, read errors, SCSI Bus resets, etc.).
- 6.) If checks are successful (no errors produced), you can continue with the LFC procedure. Continue to [Section 3.0](#).
- 7.) If you encounter any CD-ROM or SCSI related errors during this check, stop and troubleshoot/

repair the problem before attempting to perform an LFC.

- 8.) Terminate the 'tail' command by typing `ctrl+c` in the first UNIX shell window where the 'tail' command is running.
- 9.) Unmount the CD-ROM by typing the following commands in the second UNIX shell window:
`cd /` ENTER
`unmountCD` ENTER

2.4 Save System State to MOD

Before beginning the software install, you should save the system configuration information, characterization, calibration, protocols, etc. to a System State MOD.

- 1.) Bring the system up if it is not already up.
- 2.) Insert the System State MOD.
- 3.) Click on the SERVICE DESKTOP.
- 4.) Click on the PROACTIVE/PREVENTIVE/PLANNED MAINTENANCE icon.
- 5.) Click on SYSTEM STATE.
- 6.) Click on ALL. This will highlight Cals, Characterization, Reconfig Info, etc.
- 7.) Click on SAVE.

The save will take a few minutes. Review the output for errors or missing files; the scroll bar on the right works only when the tool isn't busy performing some task, it may take a little while. If you see any missing files or failures, then refer to the note below.

- 8.) Click on DISMISS.

If System State reported no problems, then you may proceed to [Section 2.5](#)

Note: **IF YOU SHOULD GET SAVE/RESTORE SYSTEM STATE ERRORS:** The Save/Restore System State process reports status on each file specification listed in the `/usr/g/config/SysState.cfg` file. The status may be:

- Save (or Restore) of {filename} succeeded
- Save (or Restore) of {filename} skipped. File not found
- FAILED

A status of `FAILED`, indicates potential problems, `skipped` means the file not found was not required. Please review the following for each file you see an error reported.

Some files simply may not exist on the system. If certain tools were never run then certain files may never get created. You may view `/usr/g/config/SysState.cfg` and look for a required field. `req=Y` means the file in question is required. You may also wish to check the system and MOD to verify whether it was actually missing or if there were permission or owner problems with the file.

2.5 Record KEY System Information

Keep a record of system information in a safe place. System information is saved on the System State MOD, but to be safe it should also be recorded somewhere else at your site in case the State MOD becomes lost or damaged.

- 1.) Record the most important Reconfig INFO for the scanner. Use the following procedure to help gather and record key system information from the INFO file and other system files.
 - a.) Open or go to an UNIX shell on the OC.
 - b.) View the following files to find the information for the chart on the following page ([Table 1-1](#)); type the following:

`more /usr/g/config/INFO` ENTER

Gets all data except Exam Number and Camera information. HIS/RIS information will not be in the INFO file if the customer does not have this option.

more /usr/g/bin/scanRx.db ENTER

Gets the next Exam Number (Add 1 to the number to record the "next" exam number.)

For each of the fields below look at the more file output to find the associated **setenv** value and record the value in [Table 1-1](#).

Note: IP Addresses and Net/Broadcast Masks have the following format: **www.xxx.yyy.zzz**

SYSTEM NETWORK CONFIGURATION			
	FIELD NAME:	"setenv" NAME:	VALUE:
System Settings:	Service ID	SERVICE_ID	
	Hospital Name	HOSPITAL_NAME	
	Exam Number**	Use procedure above to determine Exam Number.	
	DASM Type	DASMTYPE	See Chapter 4
	Camera Type	CAMERATYPE	See Chapter 4
Network Settings:	Host Name	GATEWAY_HOSTNAME	
	Gateway IP Address	GATEWAY_IP	
	Gateway Net Mask	GATEWAY_NETMASK	
	Gateway Broadcast Mask	GATEWAY_BROADCAST	
	Default Gateway	GATEWAY_DEFAULT	
Optional - Network Printer	Suite Name	SUITEID	
	Printer Hostname	PRINTER_HOSTNAME	
	Printer IP Address	PRINTER_IPADDRESS	
Optional - HIS/RIS Systems*	CT AE Title	HRCTTITLE	
	Port Number	HRPORTNUM	
	H/R Server AE Title	HRAETITLE	
	IP Address	HRIPADDR	

Table 1-1 Re-configuration Information

Note:

- * HIS/RIS information will not be in the INFO file if the customer does not have the HIS/RIS option.
- ** Do not record the Exam Number from the data contained in the INFO file, because the value will be from the last time reconfig was performed. (See the procedure on the previous page to determine the current Exam Number.)

- c.) Go to [Chapter 4 - "Laser & DICOM Camera Setup"](#), [Table 4-1 on page 60](#) or [Table 4-4 on page 61](#) to record camera information in the appropriate table provided.
- 2.) Now record the Networking Application (Image transfer) Configuration information, see [Table 1-2](#). On the ImageWorks Desktop, in the Browser under Network, for each host configured for networking do the following and record the information in the table provided:

a.) Click on IMAGE WORKS

b.) Click on NETWORK

For each host configured for networking, perform the following steps, and record the information in [Table 1-2 on page 26](#)

c.) Click on SELECT REMOTE HOST.

d.) Click on the HOSTNAME.

e.) Click on UPDATE.

f.) Write the information in the table provided on the next page.

g.) Click on CANCEL to exit Select Remote Host.

If you find this process tedious, you can look at the file by typing:

more /usr/g/ctuser/Prefs/SdCRHosts on the system to get this information. The 1 in the file means Advantage Net, the 2 means DICOM. If you see the pound sign (#), it's used as a separator and means nothing.

TYPICAL OUTPUT WOULD APPEAR AS THE FOLLOWING:

A	B	C	D	E	F	G	H	I	J
3.7.52.135	engbay04	#2	4006	engbay04	No comment	0	1	1	0
3.7.52.151	engbay26	#2	4006	engbay26	No comment	0	1	1	0
3.7.52.121	engbay16	#1	4005	engbay16	No comment	0	1	0	1

The fields in the above file relate to the Network GUI settings.

A - IP Address

B - Hostname

C - # is a space, 1 is Advantage Net Protocol, 2 is DICOM Protocol

D - Port Number

E - AE Title

F - Comment Fields

G - Archive Node - 1 = Yes, 0 = No

H - Send Images - 1 = Yes, 0 = No

I - Query/Retrieve - 1 = Yes, 0 = No

J - Custom Search - 1 = Yes, 0 = No

3.) Check the memory to ensure it is recognized.

Open a UNIX shell and type: **hinv**

The output for the OC memory should be at least 512MB. If it is not, check the Host computer to make sure the proper amount of memory is installed.

NETWORKING APPLICATION (IMAGE TRANSFER) CONFIGURATION					**	**	**	**	**
Host name	Network address	Network protocol	Port number	AE Title	Archive Node	Send Images	Query/ Retrieve	Custom Search	Comments

Table 1-2 Network Application Configuration

Section 3.0

Load Procedure



NOTICE
Potential for
Data Loss

Make sure that all Images have been reconstructed and archived before performing this procedure. This procedure will re-initialize all system data disks, erasing all images and scan data.

3.1 Shutdown the System

Click on the SHUTDOWN button (shown in [Figure 1-2](#)). An attention box will pop up indicating "Shutdown the System?"; press OK to shutdown the system. This will shutdown both Applications and UNIX, and put you at the boot prompt



Figure 1-2 Shutdown Icon

The following message should now come up:

Okay to power off the system now. Press any key to restart.

3.2 Boot Disk Partitioning Utility

- 1.) Power off the Console and wait for 15-20 seconds to reset everything properly.
- 2.) Please read carefully before proceeding.

You may notice a message about power up diagnostics running. Then a window will come up about starting the system. There will be a small button in the right hand corner named STOP FOR MAINTENANCE for 10 seconds after the console is powered "ON". Using the mouse, quickly press this button when it appears.



Figure 1-3 Starting System Pop-up Window

Now, power on the console and select STOP FOR MAINTENANCE.

- 3.) A screen will pop up with several icons displayed. The selections on the System Maintenance Menu are:
 - a.) Start System
 - b.) Install System Software
 - c.) Run Diagnostics
 - d.) Recover System
 - e.) Enter Command Monitor
 - f.) Select Keyboard Layout
- 4.) Insert the LS2002 IRIX Operating System CD-ROM Disk #1 into the CD-ROM drive.

3.3 Select Keyboard Language

If you have a non-US keyboard, you should follow the procedure that follows. Else you may proceed to [Section 3.4, on page 29](#).

- 1.) Using the mouse, click on SELECT KEYBOARD LAYOUT




Figure 1-4 Keyboard Layout Window

- 2.) Select the appropriate keyboard type. Supported keyboards are:
 - English - US (default)
 - German - DE
 - Swedish - SE
 - French - FR
- 3.) Select APPLY to reconfigure.

3.4 Install Operating System

This procedure assumes you have already created default partitions. You need to follow the procedures for creating default partitions in [Chapter 2 - "Partitioning Disks"](#) *ONLY IF*: the disk is new, the disk has just been formatted, or if the disk has been corrupted. Otherwise continue on to [step 1](#).

- 1.) Click on INSTALL SYSTEM SOFTWARE.

Make sure that the IRIX Operating System CD-ROM Disk #1 is in the CD-ROM drive and *the CD-ROM drive is not busy before proceeding to the next step*.

- 2.) Click on LOCAL CD ROM. (It may be LOCAL SCSI CDROM DRIVE 6 ON CONTROLLER 1)

- 3.) Click on INSTALL to continue installation.

A message appears: Insert the installation CD-ROM now

- 4.) Click on CONTINUE.

The following message appears: Obtaining installation tools

Then another message appears: Copying installation tools to disk

The screen goes blank and another message appears: Creating miniroot devices.
please wait ...

Note: Error Possibilities

If you do not get one of the following 4 error messages below, continue on to [step 5](#). If you receive an error that is not listed, please see [Appendix A - "Error Messages And Troubleshooting"](#), on [page 69](#).

- **If the message Unable to continue; press <Enter> to return to the menu appears on screen, do the following:**
 - a.) Eject the CD from the CD-ROM drive, and re-insert the CD into the drive.
 - b.) Click on the RETURN button to continue.
 - c.) Click on the STOP FOR MAINTENANCE button.
 - d.) Return to [step 1](#) of this section.
- **If the message make new file system on /dev/dsk/dks0d1s0 [yes/no/sh/help]; appears on the screen, do the following:**
 - a.) Type: **yes** and press ENTER.
 - b.) When message Are you sure ? [y/n] (n) appears, type: **yes** and press ENTER.
 - c.) When message Blocksize Of filesystem 512 or 4096 bytes ? appears, type: **512** and press ENTER.
- **If the message Unable to mount partition /dev/dsk/dks0d1s0 appears on screen do the following:**
 - a.) Press ENTER to go to a csh shell.
 - b.) Press ENTER and in the shell, type: **mkfs /dev/dsk/dks0d1s0**
 - c.) Type: **exit** ENTER and proceed to [step 5](#) on [page 30](#).
- **If the disk has never been partitioned, an error condition is reported when starting to partition the disk: swap partition (1) is not a valid swap area .**
Refer to [Chapter 2 - "Partitioning Disks"](#) before continuing.

Example: 5.) The install menu should appear: Go to the next step if/when the menu appears.
Screen Output

```
Default distribution to install from: /CDROM/dist
For help on inst commands, type help overview.
Inst Main Menu
 1. from [source]      Specify location of ...
 2. open [source]     Specify additional ...
 3. close [source]     Close distributions
 4. list [keywords] [names] Display information ...
 5. go                 Perform software ...
 6. install[keywords] [names] Select subsystems ...
 7. remove[keywords] [names] Select subsystems ...
 8. keep [keywords] [names] Do not install or ...
 9. step [keywords] [names] Interactive mode ...
10. conflicts [choice ...] List or resolve ...
11. help [topic]       Get help in general ...
12. view ...           Go to the view ...
13. admin ...          Go to the Admin ...
14. quit               Terminate software ...
inst>
```

6.) Type: **sh** and press ENTER.

Ignore the message: Can't open /etc/sys_id

a.) At the prompt: # type: **mount /CDROM** and press ENTER.

b.) At the prompt: # type: **/CDROM/ct_prep** and press ENTER.

Messages appear: (Several pages of scrolling text appear)

Starting automatic disk partition script

Ensuring the unnecessary partitions are unmounted

Copying Miniroot Image to the new partition location

It may take a few minutes...

Note: Ignore the message: UX:dd: ERROR:Read error: No space left on device

Copying contents of the first OS Disk (Disk #1) to working directory
...

After a wait of about 5 minutes, you will see the message:

Eject the first OS Disk #1

Please insert the second OS Disk (Disk #2), then press ENTER.

c.) Replace the IRIX CD-ROM Disk #1 in the drive with the IRIX CD-ROM Disk #2 and press ENTER.

You will get a number of messages:

Copying Copying packages from the second OS Disk #2 to working directory...

After approx. 5 minutes, you will see the message:

Eject the second OS Disk #2

Please insert the third OS Disk (Disk #3), then press ENTER

d.) Replace the IRIX CD-ROM Disk #2 in the drive with the IRIX CD-ROM Disk #3 and press ENTER

A message appears:

Copying the disk, Please Wait...

Then, a second message appears:

Copying packages from the third OS Disk #3 to working directory...

The packages from the last OS disk will be copied and the installation process will continue. This process will take approximately 30-40 minutes and no user input is required.

Note:

During this time, files will be uncompressed, installations and removals will take place, but no errors should be generated. If any errors do occur during this process, refer to [Appendix A - "Error Messages And Troubleshooting", page 69](#).

After the OS installation is complete the user will be instructed to type in the following commands to complete the IRIX OS(OC) install.

***** Done installing the OC System *****

- e.) Type:'**exit**' then **ENTER** to exit back into the original Inst session.
- f.) Type:'**exit**' then **ENTER** again at the Inst> prompt, restart the system.
- g.) Type:'**yes** ' then **ENTER** when you are prompted to restart the system.

The system monitors will go blank and restart the OC Operating system, and automatically log you into the system as the root user on the SGI standalone workstation desktop.

Note:
CD-ROM errors

Ignore the following messages during OC bootup:

Failed to configure efl as gate-ct01.oc

exportfs: /CDROM: No such file or directory

After this message appears, there will be a delay during which the system may appear to be hung. This is normal. Please wait until the system again responds.

3.5 Install CT Application Software

Note: If you have problems, see [Appendix A - "Error Messages And Troubleshooting"](#), beginning on [page 69](#) for possible error messages and their recovery.

3.5.1 Starting the Installation

- 1.) On the OC, open a "UNIX Shell" window by using the "Desktop" pull down menu.
 - 2.) Change to super-user by typing:
`su -` ENTER
Enter the root password to login as root, if required.
 - 3.) Eject the IRIX Operating System CD-ROM Disk #3 by typing:
`eject 1 6` ENTER
 - 4.) Insert the LS2002 Applications Software CD-ROM Disk #1 into the CD-ROM drive.
- Note: Confirm that the DAS type listed on the Applications CD label matches the system's DAS type (per [Section 2.1, on page 22](#)).
- 5.) Make sure the System State MOD is in the MOD drive and is not busy.
 - 6.) Wait until the CD-ROM drive is not busy. Then type:
`/CDROM/install/OC/do_install` ENTER.
The system will take 3-5 minutes to copy the contents of disk 1 to the system disk.
 - 7.) Wait for the prompt (approximately three minutes):
Please insert Applications install disk #2. Then press enter to continue.
The system will eject Applications Disk #1.
 - 8.) Insert the LS2002 Applications Software CD-ROM Disk #2 into the CD-ROM drive. Then press ENTER to continue.

Note: CD-ROM errors If the message `/mnt/install - no such file or directory` appears, or if the CD-ROM fails to mount, proceed as follows:

- a.) Type: `umount /mnt` and press ENTER
- b.) Eject and then re-insert the CD-ROM media into the CD-ROM drive
- c.) Return to [step 6](#) of this section.

- 9.) Wait for the window shown in [Figure 1-5](#) to appear, before proceeding to the next step.



Figure 1-5 Install Utility Window

- 10.) In the Install Utility window, click on LOAD. A pop up message asks if you wish to load the INFO file from the System State MOD. If you have a System State MOD, select YES. Another pop-up message appears. Make sure the System State MOD is in the drive and then select OK.

3.5.2 Configure Application Software

In the steps that follow, go through each of the configuration screens and verify the information is correct for your system. Make changes as necessary. If you do not have a System State MOD, enter the appropriate information (refer to the data you recorded in [Section 2.5, on page 23](#)).

3.5.2.1 Configure System Settings

Figure 1-6 System Settings Sample Screen

- 1.) Change the **Hospital Name** to the site's preferred name.
- 2.) Enter the **Service ID** issued by the service organization.
- 3.) Select the proper **Time Zone** values.

Use the scrollbar at the bottom of the timezone selection list to view the entire description of the timezone you are about to select, to ensure that you are selecting the correct timezone for your location.

If the time zone of your location is not in the list above, select one of the universal times in the selection menu. In this case, automatic changes for daylight savings time will not take effect.

Note: The **Next MOD#** and **Next Diagnostic Exam #** screens are currently not implemented.

- 4.) Verify and/or change as necessary the **Next Patient Exam #**. The value should be one larger than the "last exam number" recorded. The value entered configures the next Exam number the scan user interface will use. Get this information from the customer's patient log, or by examining information recorded previously.
- 5.) If you have a mobile CT system, select YES, else the default in NO.
- 6.) Select PREFERENCES to proceed to the next screen.

3.5.2.2 Configure Preferences

LightSpeed16 Configuration

System | **Preferences** | Hardware | Network | Accept | Quit

File | File | File | File

Preferences Settings

System Defaults

Doctors Title: radiologist

Units for patient weight: ☐ Pounds ☒ Kilograms

Language: ☒ English ☐ French ☐ German ☐ Italian ☐ Portuguese ☐ Spanish

Selected Preferred Fast Cal KV: ☒ 80 ☒ 100 ☒ 120 ☒ 140

Target Noise Index Table: ☐ Table 1 ☒ Table 2 ☐ Table 3

Date Format: ☒ MM/DD/YY ☐ DD/MM/YY ☐ YY/MM/DD

Time Format: ☒ 12:00:00 AM/PM ☐ 24:00:00 Military

Default Archive Device: ☒ DICOM

Modified In Room Start: ☐ On ☒ Off

Dose Information Display: ☒ On ☐ On without Total DLP ☐ Off

1. This option only appears during Reconfig.
2. This option is only available in Japan. All other localities should select "Off"

Figure 1-7 Preferences Setup Sample Screen

- 1.) Select the Language (ENGLISH, FRENCH, GERMAN, ITALIAN) to display on the system.
- 2.) Select Preferred FastCal KV - select the kV to be calibrated during FastCal. (Default selections are 120 and 140)
These kVs should include all kVs which the site uses for patient scanning. Deselecting All Preferred FastCal KVs is the same as selecting ALL the Preferred FastCal KVs
- 3.) Select the desired Date Format and the desired Time Format.
- 4.) Default Archive Device is DICOM.
- 5.) Select the site preferred Dose Information Display option for the site to use in monitoring calculated Patient Dose:
 - Select ON (full CTDiw Display)
 - Select ON WITHOUT TOTAL DLP (no Dose Length Product Display), or
 - Select OFF (no CTDiw Display)
- 6.) Select Target Noise Index Table. The default is TABLE 2.
- 7.) Select HARDWARE to proceed to the screen in the next step.

3.5.2.3 Configure Hardware Settings

Note: Verify hardware Verify your scan recon hardware and gantry type now. If you choose incorrectly now, you will have to re-load software completely.

The screenshot shows the 'Hardware Settings' window. At the top is a menu bar with 'System', 'Preferences', 'Hardware', 'Network', 'Accept', and 'Quit'. Below the menu bar, the 'Hardware Settings' section is active. It contains several configuration options: 'Das Type' with buttons for 'SDAS', 'MDAS4', 'MDAS8', and 'MDAS_H16_16'; 'Detector Type'; 'Tube Type' with 'MX_200MCT'; 'PDU Type' with 'CT COMPACT' and 'NG PDU'; 'Collimator/Filter Type' with 'Helios G1'; and a 'Network Image Printer' checkbox which is unchecked. On the right side, there is a 'Camera Setup' section with a note: 'After the Software Installation has been completed and the system state has been restored, cameras must be installed through the Utilities Menu in the Service Desktop.' Below that is the 'Scan Recon Hardware Configuration' section with buttons for 'O2', 'NO O2 PEGASUS', and 'NO O2 SDC/IG'. At the bottom right is the 'GANTRY TYPE' section with buttons for 'H1 GANTRY', 'H2 GANTRY', and 'HiSpeed QXi'.

Figure 1-8 Hardware Settings Sample Screen

- 8.) Select MX_200MCT for the Tube Type for the system.
- 9.) Select CT_COMPACT for the PDU Type for the system.
- 10.) Select HELIOS G1 for the Collimator/Filter Type.
- 11.) Select for GANTRY TYPE.
- 12.) Select NO O2 PEGASUS for the Scan Recon Hardware Configuration.
- 13.) Do not enable Network Image Printer option. Network Image printers must be configured using the **reconfig** utility, following system software installation. Please see note below.

Note: Print Servers

If system software has already been installed, and your network image printer meets the following criteria, you may enable the option.

- The device connected cannot be a print server. For example, a computer with a printer attached. Printers must be connected directly to the network.
- The connected device is a true postscript printer (such as Codonics printers) only.

When the option is selected, you are asked to supply the following information:

- Enter the printer Hostname in the Name field.
- Enter the printer's IP Address in the IP Address field.

- 14.) Camera setup: Cameras setup is preformed using utilities accessed through the Service Desktop. See [Section 3.11](#).
- 15.) Select NETWORK to proceed to the screen in the next step.

3.5.2.4 Configure the Network Settings

This screen provides the ability to declare the CT system on a hospital network. Key information such as Host Name, IP Address, Net Mask, and Default Gateway should be obtained from the hospital's network administrator prior to beginning.

Figure 1-9 Network Settings Sample Screen

Refer to the data you recorded in [Section 2.5](#) to verify these entries.

- 1.) Enter the Suite Name.
 - The Suite Name must start with a letter, followed by 3 alphanumeric characters
 - Total must be four characters long.
 - Setting the Suite Name to CT01 is suggested, unless a different name is required.
 - The name of the OC interface is <Suite Name>_oc within the scanner's subnet.
- 2.) Enter the Host Name. The Host Name identifies the hostname and AE Title of the scanner. The Host Name:
 - **MUST NOT** contain the letters "oc" anywhere, in any combination of lowercase or uppercase.
 - **MUST NOT** exceed 16 characters, but may be truncated by the system to 8 characters
 - **MUST** only contain the following characters:
 - * A through z
 - * a through z
 - * 0 through 9
 - * - and _
 - **MUST** be unique in the local system.
- 3.) Enter the Host System's IP Address.

If possible consult the site's network administrator before configuring the interface to the site's network. The site administrator should provide the System's Host Name, IP Address, Net Mask, Broadcast Address, and Default Gateway (if a Default Gateway is used at the site).

-
- 4.) If you changed the Host System's IP address from the default then the Broadcast Address also needs to be updated. The Broadcast Address should be the Gateway IP address with the bits of the host id portion of the address either set to 1s or 0s depending on the configuration of the network. The standard default is 1's but older SunOS machines used 0's.

If the Host System's IP address is 192.100.9.17 the broadcast address should be 192.100.9.255 if the network is configured to use 1's to specify the broadcast address. If the network contains genesis based scanners or other SunOS 3.5 or 4.1 machines the broadcast address should be 192.100.9.0.

- 5.) Enter the Default Gateway IP Address (optional).

If there is a Default Gateway for the site's network, enter the IP Address of the gateway. Otherwise, leave the field blank.

- 6.) Check/select ADVANCED OPTIONS. The Advanced networking options provide support for NIS. Do Not turn on NIS unless requested by your site's network administrator. Turning on NIS by selecting the Use NIS? option allows the scanner to use the site's NIS (a.k.a. Yellow Pages) database. The Domain Name must be provided by the site's network administrator

-
- To turn ON NIS:
 - A.) Highlight ADVANCED OPTIONS.
 - B.) Highlight USE NIS?.
 - C.) Enter *Domain Name*
 - To turn OFF NIS:
 - A.) Highlight ADVANCED OPTIONS.
 - B.) De-select USE NIS.
-

System is now ready to initiate installation.

3.5.3 Finishing Up

- 1.) Select ACCEPT to utilize the configuration and continue installation.
- 2.) The system will load the CT software, OS patches, kernel changes and configure the system on the OC.

This loading process will take about 15 minutes. While the load is going on, the results will be displayed in a shell window (which closes when the load is complete). All the window output is logged to a file named `/var/adm/install.log.YYYYMMDDWWHHMMSS`. (Where `YYYYMMDDWWHHMMSS` is the Date/Time that the process was started.)

Note: During an applications load, the system automatically performs a firmware flash and reconfigures the Pflash Failure or Scan Data Disk. If either should fail, the system logs (but does not display) the event. IG Flash Failure Prior to the reboot prompt in the next step, a pop-up window will be displayed, indicating that pflash, igflash or re-configuration has failed and must be manually performed. Press OK to continue.

To perform the manual pflash or manual ig flash, please see [Appendix A - "Error Messages And Troubleshooting"](#).

- 3.) When the loading process and configuration changes are complete, the system will prompt you to reboot. Click on YES.

3.6 Set Time and Date

You must always set date and time on the OC.

- 1.) Open a Unix Shell on the OC from the toolchest.
- 2.) Log in as superuser:
`su -` ENTER
password: #bigguy ENTER
- 3.) Type: `setdate mmddhhmmYYYY` (e.g.: `setdate 050409001999`) and press ENTER.

Note: Alternate method exists

You may also type: `setdate` ENTER and you will be prompted through the individual entries. Where:

`mm` is month (01-12)
`dd` is day (01-31)
`hh` is hour (00-23)
`mm` is minutes (00-59)
`YYYY` is year (1980-2030)

- 4.) The `setdate` program will set and display the time and date for on the OC. Verify the date and time on the OC is correct.
- 5.) `exit` ENTER (to exit the root session).

3.7 Start CT Application Software

To start up the system, open a unix shell window and type: `startup` ENTER

3.8 Restore System State

This procedure restores Characterization Data, Calibration Data, Protocols, etc., from your System State MOD. If you don't have a System State MOD, skip this section (however, understand that you will have to perform ALL Characterizations and Calibrations in order to make the system operational).

Comment: Image Works should not be displaying any images when you restore System State (to ensure preferences are restored correctly). If you were in Image Works earlier, you should go back and exit to the Browser.

- 1.) Insert the Save System State MOD into the MOD drive.
- 2.) Click the SERVICE DESKTOP button.

Comment: If the Service Key is installed, the user has to press the (1), (2), (3) keys in order to proceed.

- 3.) Press the PM option from the list.
- 4.) Press SYSTEM STATE.
- 5.) Click on ALL, which will highlight the cals, characterization, etc.
- 6.) Press RESTORE; the restore takes a few minutes.

Note:

If you have problems, see [Appendix A - "Error Messages And Troubleshooting"](#), beginning on [page 69](#) for possible error messages and their recovery.

- 7.) Review the screen output for errors or missing files.
The scroll bar on the right works only when the tool is not busy performing some task. Please be patient, it takes a little while.

Note:

If you should get SAVE/RESTORE SYSTEM STATE ERRORS:

The Save/Restore System State process reports status on each file specification listed in the /usr/g/config/SysState.cfg file. The status may be:

- Save (or Restore) of {filename} succeeded
- Save (or Restore) of {filename} skipped. File not found
- FAILED

A status of FAILED, indicates potential problems, skipped means the file not found was not required. Please review the following for each file for which you see an error reported.

Some files simply may not exist on the system. If certain tools were never run then certain files may never get created. You may view /usr/g/config/SysState.cfg and look for a required field. req=Y means the file in question is required. You may also wish to check the system and MOD to verify whether it was actually missing or if there was a permission or owner problem with the file.

- 8.) Click NO for Reset Scan hardware pop up message.
- 9.) Click on DISMISS.

3.9 Download Flash Firmware




Figure 1-10 Flash Download Tool Graphical User Interface (GUI)

- 1.) If not already in the Service Desktop, click on SERVICE DESKTOP.
- 2.) Proceed to Flash Download Tool GUI:
 - a.) Select TOOLBOX/UTILITIES menu.
 - b.) Select INSTALL from the menu.

- c.) Select FLASH DOWNLOAD TOOL from the list.
- 3.) Click UPDATE button to start download.

Note: Ignore all pop-up messages by choosing CANCEL.

This will take approximately 20 minutes. Any invalid files should be re-downloaded by the software, and the output can be seen on screen.

- 4.) Click DISMISS to exit the graphical user interface.

3.10 Adding Software Options

Software Options must be loaded every time the OC software is loaded. INSTALL OPTIONS must be executed once, for each option MOD. For example, if you have two options MODs, you must run INSTALL OPTIONS twice.

Remember, MOD's cannot be interchanged between systems. The first time the MOD is used to install a software option, it must not be write protected. Prior to loading software, the MOD is checked for a valid system ID. If none exists, the system writes a serial number to the MOD. The MOD becomes serialized to specific CT systems.

- 1.) On the Service Desktop, select TOOLBOX/UTILITIES -> INSTALL -> INSTALL OPTIONS. The Software Options Window is displayed (see [Figure 1-11](#)):
- 2.) Click on the INSTALL button. (see [Figure 1-11](#)).




Figure 1-11 Software Options Screen

- 3.) Click on the PERMANENT button. (See [Figure 1-13](#)).



Figure 1-12 Select Mechanism Screen

- 4.) Click on the MOD button ([Figure 1-13](#)).



Figure 1-13 Select Device Screen

- 5.) Click on OK. Software options available for installation are displayed on the options screen.




Figure 1-14 SW Options MOD screen

Example:
Options screen
with Permanent
& FlexTrial
options.




Figure 1-15 Example Options Screen

- 6.) Select the option(s) to be installed, using your mouse. You can click, drag downwards to select multiple options. Use the scroll bar to reveal options not visible in the window.
- 7.) Click on the INSTALL button. When the option appears in the "Installed Options" list, the installation of that option is complete. Please be patient, installation time varies by option.
- 8.) After installing the options, select QUIT twice, then OK. Do not reboot.
- 9.) Remove the MOD & write protect the side of the MOD with the option installed. The initial install requires the MOD to NOT be write protected; subsequent installs can be done with the MOD write protected). Repeat step this section ([Section 3.10](#)) if you have additional options to load.

3.11 Install Cameras

Camera installation is required in the following cases:

- **Loading a new system**
- **When a system state MOD is unavailable, unreadable or does not contain camera preference information**
- **If changes need to be made to existing camera configurations**

To install cameras, proceed to [Chapter 4 - "Laser & DICOM Camera Setup"](#).

3.12 Shutdown System

- 1.) Press the pink SHUTDOWN button to bring down the system.
- 2.) Once the Applications and the OS have shutdown, press RESTART to bring the system back up again.

3.13 Load Patch G.6 and Graphic Rx File Transfer

Note: If Patch G.6 is not loaded, then omit the GraphicRx File Transfer.

3.13.1 Patch G.6 Load

- 1.) Shut down all CT applications.
- 2.) Open a UNIX shell.
- 3.) Type: `su -` ENTER
- 4.) Type password: `#bigguy` ENTER
- 5.) Insert the LS2002 M5 Patch G.6 CD into the CT CD drive (CD #5142334).
- 6.) Type `mountCD /mnt` ENTER.
- 7.) Type `cd /mnt` ENTER.
- 8.) Type `./install_patch -p ls2002_M5_patch_G_6` ENTER.
The patch starts loading and information is displayed regarding the patch loading. This process takes approximately 2 minutes. The text ends with POST-INSTALL.
- 9.) Type: `cd /` ENTER.
- 10.) Type: `umount /mnt` ENTER.
- 11.) Type: `eject 1 6` ENTER.
- 12.) Remove the CD from the disk drive.
- 13.) Type `reboot` ENTER.
The system goes down and then comes back to applications.
- 14.) When the system is back to Applications level, open a UNIX shell.
- 15.) As `ctuser`, type: `index_img_uid` ENTER. (This step indexes the image table.)
- 16.) Confirm that the patch was successfully installed by typing:
`showprods | grep ls2002_M5_patch_G_6` ENTER.
The output will be similar to:
`I ls2002_M5_patch_G.6 02/05/2005 Patch G.6 for ls2002 M5`
`I ls2002_M5_patch_G.6.img 02/05/2005 Patch files for ls2002 m5`
`I ls2002_M5_patch_G.6.img.ocroot 02/05/2005 Files that go onto ocroot.`
- 17.) Close the UNIX shell.
- 18.) Open the Common Service Desk Top (CSD) and run the Flash Download:
Select Utilities>Install Folder>Flash Down Load Tool>Update.
- 19.) Select the (pink) Shutdown button>Restart.
- 20.) When the system returns to Applications, please set Tube Identification on the Common Service Desktop, by selecting Configuration>Tube Identification Certification Tool>Tube Identity.

3.13.2 GraphicRx File Transfer

- 1.) Change directory to the location of the fix:
As `ctuser`, type: `cd /usr/g/ctuser/display/arch/irix65` ENTER.
- 2.) Save the existing graphicRx file that is in the bin directory:
As `ctuser`, type: `mv bin/graphicRx bin/graphicRx.OLD` ENTER.
- 3.) Copy the correct graphicRx file to the bin directory:
As `ctuser`, type: `cp graphicRx bin/graphicRx` ENTER.
- 4.) Set the file permissions of the graphic Rx file that is now in the bin directory.
As `ctuser`, type: `chmod 555 bin/graphicRx` ENTER.

- 5.) Verify that the file changes were correctly made:
As ctuser, type: `ls -al bin/graphicRx*` **ENTER.**
The shell should display the following (with appropriate date and order for your system):
`-r-xr-xr-x 1 ctuser ctuser 1315232 Apr 8 2006
bin/graphicRx.OLD*
-r-xr-xr-x 1 ctuser ctuser 1315368 Sep 18 2006 bin/graphicRx*`
The size of the bin/graphicRx file should be 1315368.
This verifies that the copy worked correctly.
- 6.) Reboot the system to activate the fix.

3.14 Verify System Operation

The following procedure makes sure the system is operating properly after upgrading or reloading the software.

- 1.) Open a UNIX shell and type: `hinv`
- 2.) ExamRx Desktop
 - a.) Start the system and make sure you are on the ExamRx desktop.
 - b.) Verify that the date and time shown in the system status area is correct.
 - c.) Perform Tube Warm-up and Fastcal.
 - d.) Set ExamRx Display up to Autoview using your favorite Autoview layout
 - e.) Scan your favorite image quality phantoms to ensure the system is making good images. Make more than one image in a series so you can play with the images below.
 - f.) Verify that the date and time shown on the images and in the browser are correct.
 - g.) Watch to make sure the images autoview.
 - h.) Use List/Select to bring an image up in an ExamRx free viewport.
 - i.) Try each of the Window/Level Preset keys to make sure they are set correctly.
 - j.) Set the window level via the mouse.
 - k.) Set the window level via the trackball.
 - l.) Use the Trackball to Page through the images.
 - m.) Film at least 1 sheet of film and verify the image quality.
 - n.) If the site has a Network Image Printer, verify the print quality.
- 3.) ImageWorks Desktop
 - a.) Go to the Image Works Desktop.
 - b.) Select Network and make sure all of the Remote hosts were restored when you restored the system state. If they were not restored, use the information you saved earlier ([Section 2.5, on page 23](#)) to configure the site's networking.
 - c.) If you have any Genesis based systems configured for networking, and they are up and running, select one of the remote systems and **Query** to see if you can see those systems. *(If the Hospital and the user on a remote system grants you permission, try pushing images to the remote system.)* Use both genesis and non-genesis images if available.
 - d.) If you can see those systems, try pulling one or more exams onto this system.
 - e.) If it is all right with your site, try pushing images to all Genesis and DICOM systems that you have in your network configuration.
 - f.) Use Archive to restore one or more of the most recent Exams performed by the site.
 - g.) Select an interesting Exam in the Browser and click on CT Viewer to see the images.
 - h.) Try each of the Window/Level Presets to make sure they are set correctly.

-
- i.) Set the window level via the mouse.
 - j.) Try a reformat of an interesting series.
 - k.) If you have the 3D option, try doing a 3D rendering of an interesting series.
 - l.) Print at least one film and check the image quality.
 - m.) Remove one or more of the Exams you pulled onto the system.
- 4.) Service Desktop
- a.) Go to the Service Desktop.
 - b.) Click on UTILITIES to see the list of Utility Applications.
 - c.) Click on TOOLS, then VERIFY SECURITY, to see if the system can see a security key.
 - d.) Click on INSTALL and then VERIFY OPTIONS to see if the right options are configured.
 - e.) Click on SCAN ANALYSIS to see if you can see the scan files you made earlier.
 - f.) Click on SYSTEM RESETS, perform a SCAN HARDWARE RESET, and verify that it completes normally.
 - g.) When the Scan Hardware Reset is complete, click on CLEANUP to exit all the Service applications.

SOFTWARE LOAD COMPLETE

Chapter 2

Partitioning Disks

Section 1.0 Overview

Partitioning is the process of preparing a disk to work with an operating system. In doing this, the process destroys all data stored on the disk. Normally, it's only necessary to partition a disk when a hard drive has been replaced, reformatted or has become damaged.

Do not attempt to partition the Scan Data Disk. The Scan Data Disk is prepared when you run **reconfig**, and select "Re-generate Database" (See ["Chapter 6" on page 67](#)).

Note: Ensure the LS2002 Operating System CD-ROM Disk #1 is in the drive before continuing. If the disk has never been partitioned, swap partition (1) is not a valid swap area is reported. This is normal.

If you need to format the System Disk, see [Appendix A - "Error Messages And Troubleshooting" page 71](#) for details.

Section 2.0 Procedure



NOTICE
Potential for
Data Loss

Make sure that all Images have been reconstructed and archived before performing this procedure.

- 1.) Select the ENTER COMMAND MONITOR, selection 5. A boot monitor screen will be displayed. Check to make sure the CD-ROM drive is NOT busy before you proceed. Start the standalone disk partitioning utility.

For the Octane Computer (OC), type:

```
boot -f dksc(1,6,8) sash64 dksc(1,6,7) stand/fx.64 ENTER
```

- 2.) Enter the responses as indicated, in the following:

a.) Do you require extended mode with all options available(no)? **yes**

b.) fx: device-name = (dksc) ENTER

c.) fx: ctrl # = (0) ENTER

d.) fx: drive # = (1) ENTER

A message appears: SGI drive type = <Disk Model No>.

please choose one (? for help, .. to quit this menu) ---

[exit] [d]ebug/[l]abel/

[b]adblock/[exe]rcise/[r]epartition/

e.) fx> **r** ENTER

--- please choose one (? for help, .. to quit this menu) ---

- [ro]otdrive[o]ptiondrive[e]xpert
[u]srrootdrive[re]size
- f.) Select: [ro]otdrive
fx/repartition> **ro** ENTER
- g.) fx/repartition/rootdrive: type of data partition = (xfs) ... ENTER
- h.) Continue? **y** ENTER
- i.) A menu appears asking: Please Choose One.
- j.) Go up a menu: **..** ENTER (two periods)
- k.) Type: **d/fi** ENTER (There are no spaces in this command)
- l.) fx/debug/fillbuf: buf offset = (0)
Press ENTER
- m.) fx/debug/fillbuf: "fill string" = ()
Type: **a** and press ENTER
- n.) fx/debug/fillbuf: nbytes = (524288)
Type: **0x10000** and press ENTER
- o.) A menu appears asking: Please Choose One.
Type: **d/see** and press ENTER (No spaces in this command.)
- p.) fx/debug/seek:blocknum = (0)
Type: **266240** and press ENTER
- q.) A menu appears asking: Please Choose One.
Type: **d/w** and press ENTER (No spaces in command.)
- r.) fx/debug/writebuf: buf offset = (0)
Press ENTER
- s.) fx/debug/writebuf: nblocks = (4294967296)
Type: **128** and press ENTER
- t.) A menu appears asking: Please Choose One.
Type: **1/sy** and press ENTER (No spaces and lower case "L")
- u.) You are Done
Type: **exit** and press ENTER.
- 3.) You may now continue installation. After performing disk partitioning on OC, [See "Install Operating System" on page 29](#), to continue system software installation.

Chapter 3

Re-configuration of CT Application SW

Section 1.0 Overview

This chapter describes how to modify an existing system configuration. Many of the parameters entered during a LFC can be changed without a complete re-installation of software. However, to change computer type or image generator hardware requires a re-load of CT applications.

The procedure consists of running the software utility **reconfig**, changing system values and accepting the changes. The following table indicates the parameter settings that can be changed using this **reconfig**.

SYSTEM SETTINGS	PREFERENCES	HARDWARE SETTINGS	NETWORK SETTINGS
Hospital Name	Screen Display Language	Network Image Printer	Suite Name
Service ID	Preferred Fastcal kV		Host Name/AE Title
Time Zone	Date Display Format		Host IP Address
Next Exam #	Time Display Format		Net Mask
Scan Database Regeneration	Dose Information Display		Broadcast Address
			Default Gateway Address
			NIS (Yellow Pages)

Table 3-1 System Re-configuration Settings

Section 2.0 Procedure

2.1 Overview

In the procedure that follows, you will initialize the reconfiguration utility, modify as needed system setting and commit changes. You can quit the reconfiguration procedure at any time by pressing the **QUIT** button. Quitting exits the utility without changing the any system setting. There is **NO** safe way to abort the reconfiguration **after** pressing the **ACCEPT** button. If the entries made in the screens were incorrect, **DO NOT** try to stop the reconfiguration, instead wait for it to complete, and rerun reconfig, entering the correct parameters.

While the reconfiguration is going on, messages are displayed in a shell window, which closes when reconfiguration is complete. Should you later want to review the reconfiguration output, it is logged to the file `/var/adm/install.log.YYYYMMDDWWHHMMSS`

Where `YYYYMMDDWWHHMMSS` is the Date/Time that the reconfiguration was started. To view the file, type: `more /var/adm/install.log.YYYYMMDDWWHHMMSS`

2.2 Starting “Reconfig”

- 1.) If CT Applications is up and running, shutdown system applications to the `ctuser` level.
 - a.) Click on the SERVICE DESKTOP button.
 - b.) On the desktop toolbar select UTILITIES.
 - c.) Select APPLICATIONS SHUTDOWN (to bring down applications only).
- 2.) On the OC, open a UNIX Shell window.
 - a.) Type: `su -` ENTER at the prompt.
 - b.) Enter the root (super user) password: `#bigguy`
- 3.) Change directory to scripts:
Type: `cd /usr/g/scripts` ENTER at the prompt.
- 4.) Launch the Install utility:
Type: `reconfig` ENTER at the prompt.
The OC displays the Install Utility Window as shown in [Figure 3-1](#).




Figure 3-1 Install Utility Window

- 5.) Using the mouse, click on the CONFIG button.
The System Configuration “System” Settings Screen is displayed.

2.3 System Configuration Settings

- 1.) When the System Settings Screen appears ([Figure 1-6, on page 33](#)), select either YES or NO to Regenerate Database. Select NO to Regenerate Database, unless it's absolutely necessary to rebuild the scan database. This selection determines whether the Scan Database will be regenerated. To avoid accidental “regeneration”, NO is the default selection.

NOTICE Potential for Data Loss	Select YES if you want to regenerate the database. If <u>YES</u> is selected, then all scan data is deleted (erased) on the Scan Data Disk. Regenerating the Scan Database destroys all the scan and calibration files on the Scan Data Disk. Make sure that all images have been reconstructed before regenerating the Scan Database. This procedure is normally only needed when a Scan Data Disk is replaced or reformatted.
---	---

- 2.) For setting up the remaining parameters, follow the instructions in [Chapter 1](#), beginning at [Section 3.5.2.2](#), and then return to step 3 when complete.
- 3.) After accepting the changes, application software, OS patches, kernel changes take place. The configuration process takes approximately 5 minutes. Status and results are displayed in

a shell window, which closes when the loading process is complete.

Output is logged to a file named: `/var/adm/install.log.YYYYMMDDWWHHMMSS`

(Where `YYYYMMDDWWHHMMSS` is the Date/Time that the loading process was started).

2.4 Finishing System Reconfiguration

- 1.) When the loading process and configuration changes are complete, the system displays a prompt to reboot. Click on YES. (See [Figure 3-2](#).)




Figure 3-2 Reboot Screen

- 2.) The system will automatically login as `ctuser` after the reboot. Select OK on the Autostart Disabled popup message.

APPLICATIONS START-UP

- 3.) In the Console shell, type: `st` ENTER

Chapter 4

Laser & DICOM Camera Setup

Purpose: This Chapter illustrates the steps involved to setup a camera. In addition, data sheets are provided to record the information you will need to complete setup.

Section 1.0 Camera Setup



NOTICE
Potential For
Data Loss

You should empty all filming queues before modifying any camera parameter.

- 1.) Click on the SERVICE DESKTOP button.
- 2.) On the Desktop Toolbar select UTILITIES -> INSTALL -> INSTALL CAMERA. The Install Camera window appears, along with a warning message pop-up box. To remind you that all filming queues must be empty before you begin to update or delete a camera.
- 3.) The Graphical User Interface displayed shows a list of cameras installed (See Figure 4-2). First, you must click OK in the warning message box. See Figure 4-1.




Figure 4-1 Warning Screen

- 4.) Now you are asked a series of questions.
 - a.) To add a new camera, click the ADD button (See Figure 4-2).
 - b.) Now a dialog window for the camera type (DASM/DICOM) appears. If no DASM is detected during the OC boot, the DASM button will be disabled (See Figure 4-3). If a DASM is present and has not been detected, re-boot the OC and run the camera configuration tool again.




Figure 4-2 Printer Configuration GUI




Figure 4-3 Dialog Box for camera Type (with and without DASM)

- 5.) To add a new laser camera, click DASM in the camera type dialog box. This brings up a list of available camera models. Select the appropriate model from the list and click SELECT (See [Figure 4-4](#)). Now configure it.




Figure 4-4 Camera Model Dialog (with DASM)

- a.) DASM Interface is automatically detected as either Analogue or Digital
- b.) Two Laser Options are available for laser cameras: SLIDES and ZOOM. Set this option only if the camera being installed supports slides and zoom. Setting the option allows it to be enabled or disabled at the application level.
- c.) Camera manufactures provide two (2) Magnification Type options for cameras. The SMOOTH resolution blurs the image, while the SHARP resolution makes the image pixels more pronounced. The default is smooth.

To film good images, and have them look like images filmed by other GE CT products, the following camera settings are suggested:

Kodak: SMOOTH

Dupont/Sterling: SMOOTH

3M/Imation (Laser Camera): SHARP

3M/Imation (Dry View): SMOOTH

Agfa: SMOOTH

- d.) Select the appropriate File Format. Select ON from the drop down list boxes on the menu. Valid film formats are determined by the camera manufacture. IMATION for example, doesn't support 4x4, 2x4 or 1x2 and AGFA does not support 2x4) The DICOM print convention designates film formats by column and row (e.g. 12 on 1 film is 3x4). When finished setting parameters, click on DONE and proceed to step 8.

Laser Camera Configuration

DASM Interface : Analogue

Laser Options : ☐ Slide
☐ Zoom

Magnification Type : ☒ Smooth
☐ Sharp

Film Format

F_1x1 On	F_2x1 On	F_2x2 On	F_3x2 On
F_3x3 On	F_4x2 On	F_4x3 On	F_4x4 On
F_5x3 On	F_5x4 On	F_6x4 On	

Done Cancel

Figure 4-5 Laser Camera Configuration

- 6.) To add a new DICOM camera, click on ADD and then DICOM in the dialog box that appears.




Figure 4-6 Camera Model Dialog (DICOM)

- a.) A list of camera models appears (See Figure 4-6). Select the appropriate model from the list and click SELECT. Clicking SELECT presets all the parameters to that models except the Network parameters.

Comment:

It's advised to re-check the preset information with the camera manufacturer's representative.

Selection of a different camera model clears the Image Quality parameters, because these are camera manufacture dependent.

- b.) Enter the Network Parameters (See Figure 4-7)

>Device NameA unique name used to identify the camera.

>Host NameDICOM print server host name, as defined by the hospital.

>IP AddressDICOM print server IP address, as defined by the hospital.

>AE TitleDICOM print server application entity title, as defined by the print server.


*You should consult the manufacturer's DICOM
Conformance Statement.*

Note:

The Application Entity Title for the Camera may be site specific. Make sure that you check with the Camera Manufacturer's Representative and the hospital network administrator to ensure you are using the correct AE Title for the destination DICOM Print Camera.

>TCP/IP Listen PortDICOM print server TCP/IP listen port, as defined by the server. *You should consult the manufacturer's DICOM
Conformance Statement.*

>CommentsOptional comments used by the DICOM print server.



The image shows a 'DICOM printer configuration parameters' dialog box. It contains several input fields and dropdown menus. The fields are: Device Name, Host Name, IP address, Application title, TCP listen port, and Comments. Below these are two columns of dropdown menus: Medium Type (with 'Clear Film' selected) and Destination (with 'Magazine' selected). Below these are two more dropdown menus: Film orientation (with 'Portrait' selected) and Magnification type (with 'Bilinear' selected). At the bottom, there is a section for 'Standard film formats' with a grid of 12 'On' buttons for various film sizes: F_1x1, F_2x1, F_2x2, F_3x2, F_3x3, F_4x2, F_4x3, F_4x4, F_5x3, F_5x4, F_6x4, and F_6x5. At the very bottom are three buttons: 'OK', 'Cancel', and 'Advanced'.

Figure 4-7 DICOM Camera Configuration

- c.) Medium Type specifies the type of film being used. Currently, only BLUE FILM and CLEAR FILM are supported.
- d.) Set Destination to the final location for film output. This is either MAGAZINE or PROCESSOR.
- e.) Orientation selects film orientation. Only PORTRAIT is currently supported.
- f.) Set the Magnification Type. This parameter selects the algorithm used to interpolate pixels for proper film resolution. Set this parameter after consulting the camera manufacture to ensure optimal image quality. Choices are describe below:

>NoneNo interpolation. This option is not supported by all cameras.

>ReplicateAdjacent pixels are interpolated. This can result in "pixelized" images. *This algorithm is not normal preferred.*

>BilinearA 1st order interpolation of pixels. Results in images usually described as blurred. *This algorithm is not normal preferred.*

Comment: For most Camera Manufacturers, the preferred selection is CUBIC.>CubicA 3rd order interpolation. Used with a large number of possible formulations. Camera manufactures define parameters called "smoothing type" to set co-efficients used in this algorithm. The implementation of these "smoothing coefficients" is camera dependent.

- g.) Select the Standard Film Formats. Select the film format by choosing ON in the pull-down menu box located below each selection. See Figure 4-7. Valid film formats are set by the camera manufacture. IMATION for example, doesn't support 4x4, 2x4 or 1x2 and AGFA does not support 2x4) The DICOM print convention designates film formats by column and row (e.g. 12 on 1 film is 3x4).
- h.) After the camera has been configured, click the ADVANCED button. This creates the camera device file for you automatically and pops up the Advanced Parameters screen. See Figure 4-8.

Figure 4-8 Advanced Parameters (Camera Attribute Dialog)

- 7.) Advanced camera parameters control the image quality of films.

Note: For more information on the proper settings for these parameters, please see the Camera's DICOM Print Device Conformance Statement or the Camera Manufacturer Representative. A detailed DICOM Conformance Statement for is available. You may need to refer to this document as you are working with the Camera Manufacturer's Representative, to correctly set up the DICOM Print Camera I/Q and Time-out Settings.

- a.) **Configuration** - This parameter is camera manufacturer dependent as is typically used to specify the image contrast. The Configuration field may be up to 1024 characters long. The field will scroll automatically as text is entered. To review your entry, simply click and hold the middle mouse button, while the cursor is in the field, and drag the mouse towards the right (or left) as needed.

Note: Typical Configuration Setting Values:

Agfa Drystar (MG3000)	PERCEPTION_LUT=KANAMORI (100)
Imation Dryview (8700)	LUT=0, 7
Kodak Laser Printer 190	CS434\CN0\PD1.20

- b.) Smoothing Type - Set Smoothing Type to ON, and input the selected value. This parameter is used when the magnification type is CUBIC. It represents the coefficient for the image resolution algorithm. This parameter is camera manufacturer dependent, and should be re-verified with your radiology department.

Note: Recommended Smoothing Type Starting Values and Ranges:
Agfa DryStar (MG3000) Start Value: 140 Range: 137 - 150
Imation Dryview (8700) Start Value: 3 Range: 3 - 13
Kodak Laser Printer 190 Start Value: Enhanced Range: Normal

- c.) Minimum and Maximum Density - Used to set brightness of the images on film. The range of values is 0-4095, although the valid range for a specific camera is manufacture dependent. For Maximum Density, input the correct value into the text box. For Minimum Density, set it to ON and input the correct value in the text box.

Note: Recommended Minimum and Maximum Density Starting Values:
Agfa Drystar (MG3000) Min.: 20 or 23 Max: 300
Imation Dryview (8700) Min.: (Blank) Max: 300
Kodak Laser Printer 190 Min.: 20 Max: 300

- d.) Empty Image Density - This parameter sets the density for empty film viewports. Typically, BLACK is used but WHITE is an available option. The minimum and maximum density values are used as the representation.
- e.) Border Density - This sets the density for the border used around the film viewports. Typically, BLACK is used but WHITE is an available option. The minimum and maximum density values are used as the representation.
- f.) Film Size - Allows the system to specify a particular film size, if selected.
- g.) Trim - YES produces a white (clear) box surrounding each image.
- h.) Priority - This sets the print priority.
- i.) If you have completed entry of advanced parameters, click DONE.

- 8.) Go to [section 3.12 on page 42](#) for continuing installation if you are performing a LFC.
- 9.) **Camera Data Tables:** To locate Install Camera Information: Click on the SERVICE DESKTOP button. On the Desktop Toolbar select UTILITIES -> INSTALL -> INSTALL CAMERA. The Install Camera window appears. Select each of the cameras that are installed from the list of installed cameras, and click on UPDATE to view the camera's settings. Record the values used to setup each camera in the tables that follow. Extra tables are provided for multiple cameras

Note: You can determine this information by looking at the contents of the following files:
For a DASM Camera: `/usr/g/ctuser/app-defaults/devices/camera.dev`
For a DICOM Print Camera: `/usr/g/ctuser/app-defaults/devices/name.cfg`
where, `name.cfg` is the camera device name from the printer configuration GUI.
Example: `more <filename from above> ENTER`

End of Procedure

Section 2.0 Data Sheets

Use the following tables to record information about your cameras. Keep for future reference.

DASM CAMERA #1

GUI SETTING	SELECTIONS	VALUE
Camera Type	Model Type of Camera	
DASM Type	Digital or Analog	☐ Digital ☐ Analog
Options	Slides or Zoom	☐ Slides ☐ Analog
Film	Smooth or Sharp	☐ Smooth ☐ Sharp
Film Format Available	1x1, 2x1, 2x2, 3x2, etc.	
Film Format Default	1x1, 2x1, 2x2, 3x2, etc.	

Table 4-1 DASM Camera #1

DASM CAMERA #2

GUI SETTING	SELECTIONS	VALUE
Camera Type	Model Type of Camera	
DASM Type	Digital or Analog	☐ Digital ☐ Analog
Options	Slides or Zoom	☐ Slides ☐ Analog
Film	Smooth or Sharp	☐ Smooth ☐ Sharp
Film Format Available	1x1, 2x1, 2x2, 3x2, etc.	
Film Format Default	1x1, 2x1, 2x2, 3x2, etc.	

Table 4-2 DASM Camera #1

DASM CAMERA #3

GUI SETTING	SELECTIONS	VALUE
Camera Type	Model Type of Camera	
DASM Type	Digital or Analog	☐ Digital ☐ Analog
Options	Slides or Zoom	☐ Slides ☐ Analog
Film	Smooth or Sharp	☐ Smooth ☐ Sharp
Film Format Available	1x1, 2x1, 2x2, 3x2, etc.	
Film Format Default	1x1, 2x1, 2x2, 3x2, etc.	

Table 4-3 DASM Camera #1

DICOM CAMERA #1

GUI SETTING	SELECTIONS	VALUES
DICOM Camera Type	Model Type of Camera	
Film Format Available	1x1, 2x1, 2x2, 3x2, etc.	
Network Parameters	Host Name	
	IP Address	
	AE Title	
	TCP Listen Port	
	Comments	
	Destination	o Magazineo Processor
Special Set Up	Orientation	o Portraito Landscape
	Medium Type	o Blue o Clear
	Magnification Type	o None o Replicate o Bilinearo Cubic
*Advanced Parameters - IQ	Smoothing Type	o ON o OFF Value:
	Configuration	
	Minimum Density	o ON o OFF Value:
	Maximum Density	
	Empty Density (Black/White)	o Black o White
	Border Density (Black/White)	o Black o White
	TRIM	o YES o NO
	Priority	o HI o MEDo LOW
	Film Size	

NOTES

*To view Advanced DICOM Camera Settings, you must click on ADVANCED.

Table 4-4 DICOM Camera #1

DICOM CAMERA #2

GUI SETTING	SELECTIONS	VALUES
DICOM Camera Type	Model Type of Camera	
Film Format Available	1x1, 2x1, 2x2, 3x2, etc.	
Network Parameters	Host Name	
	IP Address	
	AE Title	
	TCP Listen Port	
	Comments	
	Destination	o Magazineo Processor
Special Set Up	Orientation	o Portraito Landscape
	Medium Type	o Blue o Clear
	Magnification Type	o None o Replicate o Bilinearo Cubic
*Advanced Parameters - IQ	Smoothing Type	o ON o OFF Value:
	Configuration	
	Minimum Density	o ON o OFF Value:
	Maximum Density	
	Empty Density (Black/White)	o Black o White
	Border Density (Black/White)	o Black o White
	TRIM	o YES o NO
	Priority	o HI o MEDo LOW
	Film Size	

NOTES

*To view Advanced DICOM Camera Settings, you must click on ADVANCED.

Table 4-5 DICOM Camera #1

DICOM CAMERA #3

GUI SETTING	SELECTIONS	VALUES
DICOM Camera Type	Model Type of Camera	
Film Format Available	1x1, 2x1, 2x2, 3x2, etc.	
Network Parameters	Host Name	
	IP Address	
	AE Title	
	TCP Listen Port	
	Comments	
	Destination	☐ Magazine ☐ Processor
Special Set Up	Orientation	☐ Portrait ☐ Landscape
	Medium Type	☐ Blue ☐ Clear
	Magnification Type	☐ None ☐ Replicate ☐ Bilinear ☐ Cubic
*Advanced Parameters - IQ	Smoothing Type	☐ ON ☐ OFF Value:
	Configuration	
	Minimum Density	☐ ON ☐ OFF Value:
	Maximum Density	
	Empty Density (Black/White)	☐ Black ☐ White
	Border Density (Black/White)	☐ Black ☐ White
	TRIM	☐ YES ☐ NO
	Priority	☐ HI ☐ MED ☐ LOW
	Film Size	

NOTES

*To view Advanced DICOM Camera Settings, you must click on ADVANCED.

Table 4-6 DICOM Camera #1

Chapter 5

Changing System Time Zone

Reconfig has the ability to change the system's time zone setting without having to perform a complete LFC. The following procedure can be followed when the system is installed in a location where there is no local time zone selection available. (See "Set Time and Date" on page 37.)

- 1.) If Applications is running, shutdown Applications first by selecting UTILITIES -> SHUTDOWN APPLICATIONS from the SERVICE DESKTOP.
- 2.) Open a Unix Shell.
- 3.) Change to superuser by typing:
`su -` ENTER
password: #bigguy ENTER
- 4.) Now type: `reconfig` ENTER
- 5.) Select CONFIG from the installation title screen
- 6.) On the System Configuration screen, select the Time zone for this site. It's suggested that you record it below for future use:

Region / Time Zone

Table 5-1 Time Zone Setting

- 7.) Press ACCEPT to accept re-configuration
- 8.) Now click YES to reboot the system.
- 9.) Proceed to Set Time and Date on page 37 to setup time on OC.

Chapter 6

Regenerating a Scan Database



NOTICE
Potential for
Data Loss

This procedure removes all scan and calibration data from the system disks. Make sure that all Images have been reconstructed before performing this procedure.

This procedure will re-initialize the Scan Data Disk, erasing all scan and calibration data. You will need to restore Calibrations (from System State), or recalibrate the system once you have performed this procedure.

Reconfig has the ability to regenerate the Scan Database. The initial OC reconfiguration screen (Figure Figure 1-6 on page 33) defaults the Regenerate Database selection to NO.

This procedure will most likely be needed/used when the Scan Data Disk has been replaced or reformatted. Using this procedure will eliminate the need to perform a complete LFC in these circumstances.

- 1.) If you need to regenerate the Scan Database, perform the [Re-configuration of CT Application SW on page 49](#), and select the YES button for Database Regeneration on the System Settings Screen.
- 2.) Select ACCEPT, and allow the Reconfiguration to complete.
- 3.) Once the Reconfiguration has completed, and the system is rebooted, reload System Calibrations from the System State MOD, or if the State MOD is not available, perform all system Calibrations.

Appendix A

Error Messages And Troubleshooting

A table of potential error messages and potential solutions is provided below.

SECTION	ERROR	POSSIBLE CAUSE	SUGGESTED RECOVERY SOLUTION
Install Operating System on page 29 (step 4)	dks1d6s8: drive not ready	CD-ROM busy when the OS tried to mount it.	Exit menu and try again. Type: <u>ENTER</u> Continue from beginning of Section 3.4
Install Operating System step 6b, on page 30.	OS Hangs. Does not accept user input	IRIX6.5 problem	Restart system and restart installation at Section 3.2
Install CT Application Software on page 32	Detection of Motorola Board failed. ICE may be down. Manually reset the VME chassis or check if /etc/hosts has all the entries. Press OK after you are done with the reset. PLEASE NOTE: Click OK after the reset.	ICE may be down. Could be other hardware related problems.	Reset the VME chassis or check if <code>.etc/hosts</code> has all the entries in the file.
Install CT Application Software on page 32	Specific Errors unknown.		Review the installation log file: <code>more /var/adm/install.log.YYYYMMDDWWHHMMSS</code> (where YYYYMMDDWWHHMMSS is the date/time that the software installation was started. Restart the Applications Software Installation at Section 3.5, page 32 .

Table 6-1 Potential LFC Error Messages and Recovery Solutions

SECTION	ERROR	POSSIBLE CAUSE	SUGGESTED RECOVERY SOLUTION
Install CT Application Software on page 37	"pflash" failed to download.	Hardware or communications problem.	A.) The system must be shut down and restarted prior to manual flash or reconfig. B.) Manually flash the RIP Board firmware. At the <code>ctuser</code> prompt, type: <code>/usr/g/ice/bin/pflash</code> ENTER If the RIP Board firmware flash routine operates correctly, the screen displays firmware loading type text in approximately 12 lines. If the routine continues to fail: - Shutdown the system. - Open the Console to gain access to the VME Chassis (right side of console). - Check the seating of the RIP Board. - Power up the system. - Repeat the manual RIP Board pflash process. C.) Manually reconfigure the Scan Data Disk. At the <code>ctuser</code> prompt, type: <code>/usr/g/scripts/reconfigScanDisk</code> ENTER D.) If above process fails, identify hardware failure.
	IG Flash failure	Communications time-out problem	A.) The system must be shut down and restarted prior to manual IG Flash. B.) Open a UNIX shell. At the <code>ctuser</code> prompt, type: <code>/usr/g/httpd/cgi-bin/ig_diags.ex</code> <code>-connect pig -menu flash_appsRom</code> ENTER
Restore System State on page 38 (step 6)	Corruption detected on the MOD in the drive. MOD cannot be mounted. MOD initialization completed with failure Save/Restore System State: Completed with failures.	Bad or corrupted MOD.	Attempt to repair MOD with <code>fsck</code> and reissue command. Open a new UNIX window, insert MOD in drive and type (as root): <code>fsck -n /dev/dsk/dks1d3s7</code> ENTER If <code>fsck</code> fails or system hangs, the system state cannot be loaded from the MOD and the system must be re-calibrated and manually reconfigured. Use re-configuration data recorded in Section 2.5 If <code>fsck</code> completes successfully, restart procedure at Section 3.8

Table 6-1 Potential LFC Error Messages and Recovery Solutions (Continued)

SECTION	ERROR	POSSIBLE CAUSE	SUGGESTED RECOVERY SOLUTION
Chapter 2	FORMATTING PROCEDURE for Octane Disk		Issue commands in Chapter 2 up to and including step 2, step d. For the next step, select [f]ormat by typing: <u>f</u> <u>ENTER</u> fx/format: Drive parameters to use in formatting = (current) Type: <u>ENTER</u> format will take approximately 90 minutes .. This is seldom necessary and may cause drive problems * * * * * WARNING * * * * * about to destroy data on disk dksc(0,1,0)! Ok? Type: <u>y</u> <u>ENTER</u>

Table 6-1 Potential LFC Error Messages and Recovery Solutions (Continued)

(c) 2003-2006, General Electric Company.
GE Medical Systems, a General Electric Company, going to market as GE Healthcare.
3200 N. Grandview Boulevard
Waukesha, Wisconsin 53149
USA

www.gehealthcare.com

