

GANTRY REVOLUTIONS:

DATE:

SYSTEM ID:

SITE:

| <u>Procedure</u>                             | <u>Inspect</u> | <u>Clean</u> | <u>Split</u> |
|----------------------------------------------|----------------|--------------|--------------|
| <b>CONSOLE</b>                               |                |              |              |
| Console Inspection                           | _____          | _____        | _____        |
| <b>GANTRY</b>                                |                |              |              |
| Grease Main Bearing                          | _____          | _____        | _____        |
| Removal of Slip Ring Brush Debris            | _____          | _____        | _____        |
| Schleifring Maintenance and brush inspection | _____          | _____        | _____        |
| Clean DAS, DUCT Fan Filters                  | _____          | _____        | _____        |
| <b>HIGH VOLTAGE</b>                          |                |              |              |
| Meter Verification                           | _____          | _____        | _____        |
| HV Tank Feedback Resistor Verification       | _____          | _____        | _____        |
| Clean Tube Heat Exchanger                    | _____          | _____        | _____        |
| HHS Scans                                    | _____          | _____        | _____        |
| <b>PDU</b>                                   |                |              |              |
| PM Checks                                    | _____          | _____        | _____        |
| <b>SYSTEM</b>                                |                |              |              |
| Verify Emergency Off Buttons                 | _____          | _____        | _____        |
| Check Scan Control Pushbuttons               | _____          | _____        | _____        |
| Update Site Log                              | _____          | _____        | _____        |
| Check PM Manual                              | _____          | _____        | _____        |
| <b>TABLE</b>                                 |                |              |              |
| Tighten Motor Coupling Screws                | _____          | _____        | _____        |

**Comments:**

---



---



---



---



---



---



---

Is Follow Up Work Required ? ☐ Yes ☐ No

If yes, indicate the reschedule date: \_\_\_\_\_

Field Engineer:

Id#:

Dispatch#:

**Procedure**

**Inspect**

**Clean**

**Split**

This Page Intentionally Left Blank.

---