

GANTRY REVOLUTIONS:

DATE:

SITE:

SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
CONSOLE			
Console Inspection	_____	_____	_____
GANTRY			
Grease main bearing	_____	_____	_____
Handling & Removal of Slip Ring Brush Debris	_____	_____	_____
Schleifring Maintenance and brush inspection	_____	_____	_____
MDAS DAS/Duct Cleaning	_____	_____	_____
Allocation for Gantry cover handling	_____	_____	_____
HIGH VOLTAGE			
Meter Verification	_____	_____	_____
HV Tank Feedback Resistors Verification	_____	_____	_____
Clean Heat Exchanger	_____	_____	_____
HHS Scans Verification	_____	_____	_____
PDU			
PM Checks	_____	_____	_____
SYSTEM			
Verify emergency off buttons (table/console)	_____	_____	_____
Emergency Off - Powerware 9155-10 GE UPS Unit (Option)	_____	_____	_____
Check operation of final scan abort pushbuttons	_____	_____	_____
Update site log	_____	_____	_____
Check PM Manual	_____	_____	_____
TABLE			
General Table PM	_____	_____	_____
Cradle bottom inspection	_____	_____	_____

Comments:

Is Follow Up Work Required? ☐ Yes ☐ No If yes, indicate the reschedule date: _____

Field Engineer:

Id#:

Dispatch#:
