

FCT SERVICE NOTE SN06052311a

Class A Service information enclosed.

Date: May 11, 2006

Subject: Bypassing Tube Warm up may cause tube damage

Effectivity: CT systems using the Performix Pro VCT 100 tube (Hercules -D3194T).

Details: Bypassing the Tube warm-up is very damaging to the tube and will cause unnecessary stress on the internal components of the tube which may lead to early failure and possible oil leakage.

Resolution: Remind customers that they must run tube warm-up according to the Operators manual before starting an exam when they receive the pink Pop-Up Warning message. The warning normally occurs if there is a significant length of time between scans. The exception is due to an emergency driven by the need of a patient (example: Trauma Patient). Even then a 2 minute tube warm-up should be considered to mitigate the risk of tube damage/failure.

Tom Hurdis
Tubes Engineering