

**MI&CT ESERVICE NOTE
SN08091016a**

Class M information is Property of GE. Disclosure to Third Parties prohibited.

This short-term informational note is invalid 6 months after the issue date.

Date: September 22, 2008

Subject: Handling Non-standard Parts in the Field

Effectivity: X-Ray Tube part number D3186T

Details:

A D3186T x-ray tube was received not fully assembled at the factory. Assembly was completed in the field after procuring the right tools and specifications.

Resolution:

Any non-standard product that is received for which there is no approved documented process or processes for correction shall be returned as DOA and a replacement part ordered immediately. Any unapproved field modification of a product creates a product that is deemed adulterated per the FDA and is not acceptable for use.

Affected Service Publication:

None

Tracking Number (PQR/SPR/ECR): 13186668

Paul Lukezich
Sr. Engineering Technician- Tubes Engineering