Gradient – Pulse

Equipment Room

Main Disconnect Panel CB3 TB1-15-16 CB2
Remote MRU
E-Off Switch
O_2 Monitor Control Unit: Meter, Ctrl Board, Alarm, Light
Customer-Supplied Outlet
Site E-Net
Facility Power Filters
Facility Utilities: Facility Coolant Supply, Facility Coolant Return

Mag Mon J5 J10 J8 J9 J7
HEC: HEC Power Box (Cryo, Power Supply, GC, PE, Blower); Signal Box J24 J11 J12; Grad HtXchr; Grad Res; Grad Pump; Flow Meter; Filter; Blower HtXchgr; PGR Pump; PGR Res; PGR HtXchr; Grad/RF Coil Blower
JR Power Cryo-Cooler Compressor JC Water Return, Water Supply, Gas Return, Gas Supply

PGR
PDM J29 J25 L1/L2/L3 J27 J28 J24
VRE 2 (32-ch) J5 Net 3 Net 2 Net Mgmt J1
VRE 1 J1 Net Mgmt Net 2 Net 3 J5
ETH 1 2 3 4 Term Server
Aux Board J4 J6 J1 J2/J3 J9 J11 J10
XRFD LS1-A LS1-B LS2-A LS2-B LS3-A LS3-B LS4-A Term; PDU Top 1; PDU Top 1; PDU Front
J4 XGD Fan x3
Unused
CB1 J1 J4 J3 PGR PDU Host J6/J7/J8 J11/J12/J13 Recon RF/MNS J5 J2 PEN
CB2 J14 NB Amp J9 J10 J20 J5 J11 J7 J15 J4 J3 J18 J1 J2 J19 J16 J8 J12 J6
16 Port E-net Sw
J8 NB AIF J3
UPM1 Det J1 J2 J4 J5 J3 Proc J3
UPM2 J1 Det J2 J4 J5 J3 Proc J3
CAM J2 ETH SCP2 PMC J3
10/100 Base AGP2
STIF J4 J3 J18/J19
J23 I/O Panel J10 IRF3 J8 J14 J13 J12 J11
Cronus Chassis: XPS CB J13/J16 J12/J15 J11/J14; XGA CB J13/J16 J12/J15 J11/J14; J10/J11 J6 J9 GMC J8 J7; Cronus backplane (circuit modules, connectors); P1 J3 J4 J5 J2 J1
Pwr XPS x3 J2 J10 J1 J11 J8 J9 J3 J4/J5 J4/J5
XGA x3 J8 J7 J16 J13 J15 J14 J5 J12 J6 J17 J18 J1/J2 J1/J2
Manifold Supply
SMA Coax 9-pin LEM Icoil x3 Grad Pulse In Grad Pulse Out
Term Block x3 E3317 E3318 E3319
Manifold Return
J3 J6 J2 J102 J19 J18 J10 J9 J17 J4 J14 J15 J13 J12 J8 J20 J204 J201 J200 J203 J202 J205 J22 J11 J21 J1 J31 J5 J16 J103

SPW
J115 J116 J117 J16 J106 J108 J107
Term Conn F1 F2 F3 F4
Gradient Filter X Y Z M3317 M3318 M3319
J128
J121 J122 Xmit Sw1 J120 J123
J104 Wvgde
J105 Wvgde
J118 J17
J100 Wvgde J101 Wvgde J111 Wvgde J110 Wvgde Patient Alert Wvgde
Primary MRU
E-Off Switch
O_2 Monitor Sensor Alarm Light
Mag Enclosure Chimney
Gradient Bus Bar X Y Z
Mag I/O J1 J3 J6 J4 J2 J5 J14 J15 J12 J11 J13 J8 J7 J9 J10 J17 J16
FO I/F J28 J20 J21

Mag Bore
J6 J7 Body Coil J2 J3 J9
Bore-PAL L R
Left Arc Cover
Ambient Temp Snsr
Bore Temp Sensors
Front Mic
Airflow Sensors
X Y Z XRMB Gradient Temp Sensor
J101 MAG-Grad Coil
J100 MAG-Grad Coil
J104 Grad Coil Air
Patient Air

PEN
J14 J1 J13 J11 J12 J10 F/O Wvgde
J19 J18 J15 J17 J4 J3 J9 J5 J6 J16
NB Exciter J19 J20 J21 J29 J32 J30 J15 J14 J17 J1
Power Distribution J1 J4 J2 J3
Driver Module J14 J8 J19 J17 J11 J20 J12 J13 J2 J1 J7 J6 J23 J25 J21
DD Filter J94 J96 J92 J74 J76 J72
CFB J18 J19 J7/J8 J1/J2 J5/J6 J3/J4 J17
PHPS J12 J9 J37/J38 J2 J43 J11 J39 J54 J1 J5 J41 J10 J40 J42 J44 J35 J8 J36
SRPS J6 J5 J1 J51 J52 J46 J48 J45 J50 J47 J3
J39 J54 J41 J40 J42 J35 J36 J51 J52 J46 J48 J45 J50 J47 J29 J32 J30 J34 F11 F10 F9 F8 F7 F6 F5 J31 J33
Temp/Hum Sensor
Scan Room Blower Box
Patient Comfort
Hub

Door Switch
Coolant Removal Kit

Operator Workspace

Temp/Hum Sensor
Server (Option)
Service Laptop
Optional Speaker
GOC
IRDC (Option) J5 J4 J3 J2 J1
USB-Serial Converter
Ethernet Switch 1 2
GOCAA J12 J14 J17 J18 J15 J7 J6
Host PC: Ethernet B, Lwr Graphic Card Port 1, USB, RS232, USB, USB, USB, Upr Graphic Card Port 1, Ethernet A, Line In, Line Out
J10 J4 J6 J9 J2 GOC J5 J1 PDU J3
Global Modem (Option)
Key-board
Mouse
Primary Display
Pneumatic Patient Alert System

Scan Room

Magnet
Terminal Box J3 P403
Pressure Transducer
Reed Switch
RUO PreAmp J2 J1
Coldhead Pwr Sleeve
Coldhead Supply Return
Airflow Sensor Module
J9/J10 J2 PAC
SRI4 J23 J10 J34 J3 J20 J32 J24 J17 J1 J4 J5 J11/J12 J13/J14 J33 J21 J31 J2 J22
PreAmp & Combiner J11 J10 J9
J1 J5 J2 J6 J3 J7 BHYB J4
LCP J11 J3
RCP J11 J3
IRD J1
TDM J11 J2 J9 J1 J10 J12
RRx1 J9 J7 J1 J4/J5 J8 J2
RRx2 (32-ch) J7 J1 J4/J5 J2
MAG-PAC J4
PAC I/F
Squeeze Bulb
Speaker
Left E-Stop Button
Rt E-Stop Button

Patient Table
Coil Port P3
Coil Port P4
Touch N Go_R
Touch N Go_L
Table Hydraulics

Dock
J2 Mag J1 Dock J3 Left
ODU

Rear Pedestal
J14 J13 Dummy Load J12HYB
PED Antenna
Lights
RF Hub
Hub Backplane (bias, power, status/control)
J102 J101 J104 J103
J1 J3 J2 J4 PDB J5 J6
MCDB
J9 J13 J16 J17 J10 J14/J15 J7 J6 J1 J2 RFCB J12 Debug Port J8 J3 J4
J1–J6 (ch 17-32)
RFSB (ch 1-16) J1–J6
J4 J6 J3 TIF J1 J2 J5
Long Drive J1 ASM (MTR) J2
Load A
POS Enc
J3 EOT Sensor HOME Sensor EGRESS Sensor
LPCA J4 J2 Port A J5 J1 Port P2 J4 Port P1 J5
Rear Mic
J1 PED I/O J2
Rear Pedestal Air
Reroute Box J10 J7 P2A P3A P2B P3B P1A P4A P1B P4B J8 J9
Load B

Key

TBD = To Be Determined (Future Enhancement)
Cabinet
Chassis/ Sub-system
Part
Data Cables
Coolant
Power
Pulse
Data Paths

© 2009, GE Healthcare