

SITE:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
COMPUTER			
Check fans/Clean/Vacuum all exhaust vent holes	_____	_____	_____
CONSOLE			
Check for smooth rolling trackball/mouse	_____	_____	_____
Check Fans/Clean/Vacuum console interior	_____	_____	_____
GANTRY			
Check the number of gantry revolutions	_____	_____	_____
Grease main bearing	_____	_____	_____
Check drive belt for wear	_____	_____	_____
Check scan and laser windows	_____	_____	_____
Clean ventilation holes and check fans	_____	_____	_____
Handling & Removal of Slip Ring Brush debris	_____	_____	_____
Clean or Replace tube fan filter	_____	_____	_____
Slip Ring Vacuuming	_____	_____	_____
SDAS DAS/Duct Cleaning	_____	_____	_____
PDU			
Inspect filter	_____	_____	_____
Inspect 7 green indicator lamps on GPDU	_____	_____	_____
Inspect 4 green lamps (L1, L2, L3, DCRGS status)	_____	_____	_____
Inspect 6 green lamps on transformer assembly	_____	_____	_____
SYSTEM			
Verify emergency off buttons (table/console)	_____	_____	_____
Check X-Ray on lights (HHS)	_____	_____	_____
Check operation of final scan abort pushbuttons	_____	_____	_____
Monitor Sys Logs	_____	_____	_____
General Cleaning	_____	_____	_____
Update site log	_____	_____	_____
Check PM Manual	_____	_____	_____
TABLE			
Check for wear or damage on table extensions	_____	_____	_____

Comments:

Field Engineer:

ID#:

Dispatch#: