

GE Medical Systems

BrightSpeed Select 16, 8, 4

PM REPORT

SCHEDULE ☐ A ☐ B ☐ C
GANTRY REVOLUTIONS:
DATE:
SYSTEM ID:

SITE:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
CONSOLE Console Inspection	_____	_____	_____
GANTRY Removal of Slip Ring Brush Debris	_____	_____	_____
Schleifring Maintenance and brush inspection	_____	_____	_____

Comments:

Is Follow Up Work Required ? ☐ Yes ☐ No

If yes, indicate the reschedule date: _____

Field Engineer:

Id#:

Dispatch#:

SCHEDULE ☐ A ☐ B ☐ C
GANTRY REVOLUTIONS:
DATE:
SYSTEM ID:

SITE:

Procedure

Inspect

Clean

Split

This Page Intentionally Left Blank.
