



GE Medical Systems

Technical Publication

Direction 2346035-100

Revision 00

Mobile Ultra LS Service Intro Planner

THIS MATERIAL IS THE PROPERTY OF GEMS AND CONTAINS
INFORMATION WHICH IS RESTRICTED TO USE BY GEMS
MANUFACTURING, ENGINEERING AND SERVICE PERSONNEL. ANY
REPRODUCTION IN WHOLE OR IN PART IS STRICTLY PROHIBITED.
NEITHER THIS MATERIAL NOR ANY INFORMATION CONTAINED
HEREIN SHALL BE DISCLOSED TO ANYONE OUTSIDE OF GEMS
MANUFACTURING, ENGINEERING AND
SERVICE PERSONNEL.

*Copyright© 2002 by General Electric Company, Inc.
All Rights Reserved.*



Page Left Intentionally Blank

LEGAL NOTES

GE EXCLUSIVE USE ONLY

Restricted To General Electric Medical Systems (GEMS)

This Material is the property of GEMS and contains information that is restricted to use by GEMS Manufacturing, Engineering and Service personnel. Any reproduction in whole or in part is strictly prohibited. Neither this material nor any information contained herein shall be disclosed to anyone outside of GWMS Manufacturing, Engineering and Service personnel.

If you are not a GE Medical Systems Employee, you must exit this material now.

TRADEMARKS

All products and their name brands are trademarks of their respective holders.

OMISSIONS & ERRORS

GE personnel, please use the GEMS CQA Process to report all omissions, errors, and defects in this publication.

COPYRIGHTS

All Material, Copyright (c) 2002 by General Electric Company, Inc. All rights reserved.

REVISION HISTORY

REVISION HISTORY

Revision	Date	Reason for change
0	05/21/02	Initial Release

TABLE OF CONTENTS

Legal Notes	3
GE EXCLUSIVE USE ONLY	3
TRADEMARKS.....	3
OMISSIONS & ERRORS.....	3
COPYRIGHTS.....	3
REVISION HISTORY	3
Table of Contents.....	4
OVERVIEW.....	5
1.1 PRODUCT STRUCTURE.....	5
1.2 SYSTEM SPECIFICATIONS.....	7
1.3 KEY FEATURES OF NON-FIXED OFFERINGS.....	7
1.3.1 <i>Service Features</i>	7
1.3.2 <i>Customer Features</i>	7
1.4 SERVICE AND OPERATOR DOCUMENTATION.....	7
1.5 SUMMARY OF HARDWARE (MAIN DIFFERENCES VS. LIGHTSPEED FIXED SITE HARDWARE)	8
2.0 INTEGRATION	8
3.0 INSTALLATION	9
3.1 <i>Installation time</i>	9
3.2 <i>Documents & Tools</i>	9
3.3 <i>On-Site Set up Procedure</i>	10
3.4 <i>Insite Connection</i>	10
3.5 <i>Software</i>	10
4.0 SERVICE STRATEGY.....	10
4.1 <i>Service Responsibilities</i>	10
4.2 MAINTENANCE	10
4.3 <i>Preventive Maintenance</i>	10
4.4 <i>Corrective Maintenance</i>	11
4.5 <i>Spare Parts – additions for Mobile application</i>	11
5.0 TRAINING.....	11

Service Introduction Planner (iSERV) LightSpeed Ultra Mobile

Issuing Organization:

CT Engineering, GEMS-Am

OVERVIEW

This Introduction Planner provides information on the LightSpeed Ultra 4 and 8 slice products to be installed in non-fixed site environments.

This planner is intended to provide field service personnel with information regarding the LightSpeed ultra CT non-fixed site system going into forward production. This Service Introduction Planner is **NOT** a service manual, direction, or user guide. It is a collection of information taken from several source documents, ALL of which are available and should be used to install, maintain, and operate the CT LightSpeed Ultra non-fixed site system.

A **GE Medical System Engineering approved** non-fixed environment is any prefabricated, stand-alone environment, which is designed to hold a CT scanner. Non-fixed environments are shipped with the CT scanner preinstalled prior to the environment/scanner package being delivered to the customer site. Examples of non-fixed environments are mobile trailers, transportable trailers, relocatable buildings, and ISO containers.

The LightSpeed Ultra system for non-fixed environments is very similar to that for fixed site installations. Minor changes have been incorporated into the non-fixed version to accommodate the mobile environment needs (vibration, multi-hospital users on the same equipment, etc.)

1.1 PRODUCT STRUCTURE

Product Name	Engineering Program Name	Catalog Number	Description
Non-fixed LightSpeed ultra 4- SLICE MOBILE	Helios 3 Mobile	B7854MB	Contains the following: B7854NG 4 slice GANTRY B7850NP PDU B7850NT H2 TABLE B7850MS SHIP COLLECTOR B7858MT H3 MOBILE SFT & TECH PUBS B7800KE ENGLISH KEYBOARD B7858CCH3 OCTANE II CONSOLE B7850LL LONG CABLE B7999PY UPS+HW KIT, MOBILE B7850MK LTSPD MOBILE KIT B7858LC LCD MONITOR 2PC/KIT B7530SP SMART PREP

Non-fixed
LightSpeed
ultra 8-
SLICE
MOBILE

B7858MB

B7810WS VARISPEED
B7810THLTSPD THIN TWIN
HELICAL
B7810TL LTSPD HELICAL TILT
B7810RPLTSPD RECON PLUS
K9050SF 8PT 10/100TX SWITCH
K9000JR 10BASET PATCH
CABLE KIT
K9000KP 10M CAT V UTP PATCH
CABLE
B7858TE LTSPDUULTRA ENGL
TECH PUBS

Contains the following:
B7858NG 4 slice GANTRY
B7850NP PDU
B7850NT H2 TABLE
B7850MS SHIP COLLECTOR
B7858MT H3 MOBILE SFT &
TECH PUBS
B7800KE ENGLISH KEYBOARD
B7858CDH3 OCTANE II CONSOLE
W/D3D
B7850LL LONG CABLE
B7999PY UPS+HW KIT, MOBILE
B7850MK LTSPD MOBILE KIT
B7858 8 SLICE SOFTWARE FOR
OC

B7810WS VARISPEED
B7810TH LTSPD THIN TWIN
HELICAL
B7810TL LTSPD HELICAL TILT
B7810RP LTSPD RECON PLUS
B7858LC LCD MONITOR 2PC/KIT
K9050SF 8PT 10/100TX SWITCH
K9000JR 10BASET PATCH
CABLE KIT
K9000KP 10M CAT V UTP PATCH
CABLE
B7530SP SMART PREP
B7810SS D3D SOFTWARE KIT
FOR LS CONSOLE
B7858TE LTSPDUULTRA ENGL
TECH PUBS

1.2 SYSTEM SPECIFICATIONS

The Non-fixed LightSpeed Ultra configuration will have the same system performance specifications as the equivalent fixed site LightSpeed Ultra. All options specified in Non – Fixed CT LS system spec & Install requirements would be available for LightSpeed Ultra non-fixed site installations with the following exceptions:

1.3 KEY FEATURES OF NON-FIXED OFFERINGS

1.3.1 SERVICE FEATURES

The service features for the non-fixed version are virtually identical to the H2.1 fixed site features. There are no changes to the diagnostics or service procedures. Software changes are documented elsewhere within this ISERV.

1.3.2 CUSTOMER FEATURES

The customer features of the non-fixed LightSpeed Ultra systems are the same as for the fixed site equivalent. There are some additional features built into the customer software which will accommodate multi-hospital users on the same scanner, a modified shut down procedure which automatically tilts the gantry back for transport, the addition of flat panel monitors as standard monitors in non-fixed environments, and hardware to support the table for transport.

1.4 SERVICE AND OPERATOR DOCUMENTATION

All service documents are available at the CT Service Engineering Web-site
http://www.ct.med.ge.com/tech_pubs/welcome.htm

Direction	Description	Classification
2296438-100	Pre-Installation/Site Planning Manual, Rev 01	General Class
2296439-100	Installation Manual, Rev 03	General Class
2296434-100	System Service Manual, Rev 03	General Class
2296442-100	Illustrated Parts Manual, Rev 04	General Class
2296441-100	Schematics, Rev 03	General Class
2296436-100	SW Load Procedure, Rev 03	General Class
2296447-100	Applications Tips and Work-Arounds, Rev 03	General Class
2302180-100	Technical Reference Manual, Rev 01	General Class
2304119-100	Learning and Reference Guide, Rev 00	General Class
2319916-100	Ultra 8 slice upgrade procedure, Rev 00	
2320439-100	Conformance statement for DICOM v3.0, rev 00	

2337856-100	Plus/ Ultra v303 s/w patch Load Instructions Rev 00	
2296435-100	System Service Manual (Advanced), Rev 03	Advanced
2320440-100	Mobile Start Up /Shutdown	General Class
2315243-100	UPS Installation Mobile LightSpeed	General Class
2305587-100	Non-Fixed CT LS Systems Spec & Install Requirements, Rev ??	General Class
2305588-100	Non-Fixed CT LS Mechanical Integration Requirements, Rev ??	General Class
2318573-100	Non-Fixed CT LS Electrical Integration Procedures, Rev ??	General Class
2318669-100	On-Site Mobile Set-up Procedure, Rev ??	General Class
2320359-100	Service Introduction Planner (Restricted), Rev ??	Restricted

1.5 SUMMARY OF HARDWARE (MAIN DIFFERENCES VS. LIGHTSPEED FIXED SITE HARDWARE)

- *UPS and custom A1*
- *Flat Screen LCDs*
- *Net Work Switch*
- *No DASM (DICOM Print Only)*
- *Service Panel + Display for Gantry*
- *POGO Stick (table Transport)*
- *Modified Gantry Front cover Dollies*

Customer–Driven User Interface Features

Flat panel monitors as standard

Transport warning switch: To trigger the transport lamp . Switch is physically mounted on STC side backstop

2.0 INTEGRATION

- *Reconfig to enable mobile features*
- *Multi-Site set up*
- *Physical tilt limit: Can be found in Non– Fixed CT LS Electrical Integration Procedure 2318573-100 Rev01*
- *Mechanical and Electrical integration requirement for the non-fixed environment manufacturer can be found in the Non-Fixed CT LS Mechanical Integration requirement ref: 2305588-100 Rev01 and Non-Fixed CT LS Electrical Integration Procedures, 2318573-100 rev 01*

The respective environment manufacturer provides site-specific pre-installation information. This information will include the pad requirements, power requirements and connector requirements, phone and network

requirements, utility requirements (water, drains, etc. as required), and radiation safety zones as required at a minimum.

3.0 INSTALLATION

3.1 INSTALLATION TIME

Turnover time at customer delivery will be no longer than 6 hours at the customer site. There will be no further service involvement at subsequent customer locations. Power ground audit is required for all sites.

3.2 DOCUMENTS & TOOLS

Changes required to the LightSpeed Ultra Service documentation and Operators documentation will be made to the existing fixed site operators and service documentation. Separate documents will be developed for the non-fixed Mechanical, Electrical and installation requirements for non-fixed environment. The part numbers are as follows:

2305588-100 Non-Fixed CT LS Mechanical Integration Requirements

2318573-100 Non-Fixed CT LS Electrical Integration Procedure

2305587-100 Non-Fixed CT LS Systems Spec & Install Requirements

Gantry service tools are supplied in the install support kit. Mobile Gantry dollies may be different than fixed site dollies.

Supplied by Environment Manufacturer

F.E storage cabinet provided by the Van Company. The following tools will be stored in the Cabinet

- *QA Phantom (Water Filled)*
- *20cm Phantom*
- *48cm Phantom*
- *Phantom Holder*
- *Gantry Display*
- *FE Box Purple **All Documentation***
- *Rear Cover dollies*
- *Front Cover Dollies*
- *Install support kit (Box)*
- *Tube Hoist Kit*
- *Gantry Control Pad*

Required Tools:

Items supplied by Field - standard tool kit as required to service the current LightSpeed Ultra fixed site systems.

Items supplied by the Environment Manufacturer - any special tools or dollies required to support the van and/or the LightSpeed Ultra system

Some van models may require a cabinet to be removed for major changes

3.3 ON-SITE SET UP PROCEDURE

On - Site set up procedure can be found in On-Site Mobile Set-up Procedure, Rev 00 Ref. 2318669-100

3.4 INSITE CONNECTION

InSite connection is mandatory on this system.

MODEM: Global modem to be shipped with each system

Standard weather resistance cables are supplied with each van

3.5 SOFTWARE

H3 M4 software will be employed in Mobile Ultra

4.0 SERVICE STRATEGY

4.1 SERVICE RESPONSIBILITIES

Reference 2318669-100 On-Site Mobile set up procedure

4.2 MAINTENANCE

Same as fixed-site

4.3 PREVENTIVE MAINTENANCE

- *Maintenance Frequency depends on Site activity.*

Source pole recommendations are:

- *In warranty: 4 PM visits/year*
- *Out of warranty: 1PM per 250 Kslices*

Maintenance Time:

- *In warranty: 13 hours/year.*
- *Out of warranty:*

- **Required Tool: Items supplied by Field (Preliminary List)**

Service Tool requirements will be the same as for the LightSpeed Ultra fixed site product. Any special tools (dollies, etc) that are required to support the scanner in a non-fixed environment will be supplied by the environment manufacturer.

- **PM Manual: 2141834**

4.4 CORRECTIVE MAINTENANCE

Correct maintenance strategy (tools, process, times, etc) for the non-fixed site version will be the same as for the fixed site LightSpeed Ultra.

Major Field Replaceable Units (FRU) are:

- *X-Ray Tube*
- *High Voltage Tank*
- *Detector*
- *Slip Ring*
- *DAS*

4.5 SPARE PARTS – ADDITIONS FOR MOBILE APPLICATION

Maintenance policy is stock at FRU level.

Part #	Description
GANTRY	
2303589	<i>Harness, tilt switch Assy.</i>
2303601	<i>Harness, ct mobile tilt</i>
TABLE	
46-296317p4	<i>OBC-STC & Table EMC power supply</i>
2302366	<i>Pogo Stick</i>
CONSOLE	
2306872	<i>18.1" Global Standard LCD Monitor</i>
PDU	
2310005	<i>HARDENED UPS FOR MOBILE LIGHTSPEED +</i>

5.0 TRAINING

There will be no specific training class for the LightSpeed Ultra in non-fixed site installations. There will be a T-CON with primary F.E to discuss the differences between the Fixed site and Non-Fixed site services.

Page Left Intentionally Blank



GE Medical Systems

P.O. BOX 414

MILWAUKEE WISCONSIN 53201-0414 U.S.A.