

TABLE OF CONTENTS

o.1.1 High Speed Stability (HSS)	3
o.3.1 Spike Noise (Head Axial, A/P)	4
o.3.2 Spike Noise (Head Axial, R/L)	6
o.3.3 Spike Noise (Head Sagittal)	7
o.3.4 Spike Noise (Head Coronal)	8
o.3.5 Spike Noise (Body Coronal)	9
o.3.6 Spike Noise (Body Sagittal)	10
o.3.7 Spike Noise (Body Axial, R/L)	11
o.3.8 Spike Noise (Body Axial, A/P)	12
o.6.1 TPS/RF Waveform	13
o.7.1 Head EPIWP Base Z	15
o.8.1 Head EPIWP Base Y	17
o.9.1 Head EPIWP Base X	19
o.10.1 DD/TR Board	21
o.11.1 Body T/R Check	23
o.11.2 Head T/R Check	25
o.12.1 Head Slice Thickness (2D, 3mm)	27
o.12.2 Head Slice Thickness (2D, 5mm)	29
o.12.3 Head Slice Thickness (2D, 10mm)	30
o.12.4 Head Slice Thickness (2D, Grad Echo, 5mm)	31
o.12.5 Head Slice Thickness (3D, Localizer)	32
o.12.6 Head Slice Thickness (3D, Grad Echo, 1.5mm)	33
o.13.1 SNR Check (Body, Axial)	34
o.13.2 SNR Check (Body, Sagittal)	36
o.13.3 SNR Check (Body, Coronal)	37
o.13.4 SNR Check (Head, Axial)	38
o.13.5 SNR Check (Head, Sagittal)	40
o.13.6 SNR Check (Head, Coronal)	42
o.15.1 Slice Offset (2D, 3mm, Table 50mm)	43
o.15.2 Slice Offset (2D, 3mm, Table -50mm)	45
o.15.3 Slice Offset (3D, Grad Echo Localizer)	46
o.15.4 Slice Offset (3D, Grad Table 50mm)	47
o.15.5 Slice Offset (3D, Grad Table -50mm)	48
o.16.1 RFT Body Scan	49
o.16.2 RFT Head Scan	50
o.17.1 Body F/R Quadrature	51
o.17.2 Head F/R Quadrature	53
o.18.1 Dummy Load	55
o.19.1 LVShim (Localizer)	57
o.19.2 LVShim Scan	59
o.20.1 First Image	60
o.21.1 T2 Uniformity (Body, Axial)	62
o.21.2 T2 Uniformity (Body, Sagittal)	64
o.21.3 T2 Uniformity (Body/Head, Coronal)	65
o.22.1 Isocenter Cal (DQA-III)	66
o.23.1 Head Power Monitor	68
o.23.2 Body Power Monitor	70
o.23.3 Amp Shutdown Verification	72
o.24.1 System Gain Cal (Body)	74
o.24.2 System Gain Cal (Head)	76
o.25.1 Pin Diode Test (Head)	78
o.25.2 Pin Diode Test (Body)	80
o.26.1 SST Head Full Power	82
o.26.2 SST Body Full Power	84

o.26.3 SST Multicoil Test	86
o.26.4 SST Solid-state Driver Mode	88
o.26.5 SST True Head Mode	89
o.27.1 RFS Head.....	90
o.27.2 RFS Body.....	92
o.28.1 Full Field Distortion (Sagittal)	94
o.28.2 Full Field Distortion (Coronal Localizer)	96
o.28.3 Full Field Distortion (Coronal).....	98
o.29.1 Noise Floor (Head).....	99
o.29.2 Noise Floor (Body)	101
o.30.1 Carrier Leakage.....	102
o.31.1 Respiratory Compensation.....	104
o.32.1 Gradcal-X (DQA-III)	105
o.32.2 Gradcal-Y (DQA-III)	107
o.32.3 Gradcal-Z (DQA-III)	109
o.33.1 Body EPIWP Base Z	111
o.34.1 Body EPIWP Base Y	113
o.35.1 Body EPIWP Base X	115
o.36.1 Correlated Noise (256 matrix).....	117
o.36.2 Correlated Noise (512 matrix).....	119
o.36.3 Multicoil Correlated Noise	120
o.36.4 EPI Correlated Noise	122
o.37.1 Body Slice Thickness (Sagittal, 3mm).....	124
o.37.2 Body Slice Thickness (Sagittal, 5mm).....	126
o.37.3 Body Slice Thickness (Sagittal, 10mm).....	127
o.37.4 Body Slice Thickness (Localizer).....	128
o.37.5 Body Slice Thickness (Coronal, 3mm)	129
o.37.6 Body Slice Thickness (Coronal, 5mm)	130
o.37.7 Body Slice Thickness (Coronal, 10 mm)	131
o.38.1 Geometry Verification (Coronal)	132
o.38.2 Geometry Verification (Axial)	134
o.39.1 FID Tuning.....	135
o.40.1 Gradient Waveforms	137
o.41.1 Body APB Cal.....	139
o.41.2 Head APB Cal.....	141
o.42.1 Bandpass Asymmetry Correction Compensation (BACC).....	143
o.43.1 Center Frequency	144
o.46.1 Spike Noise (C Class).....	146
o.47.1 ECMT	148

o.1.1 High Speed Stability (HSS)

SCAN OPTION INPUT

PATIENT REGISTER . **[New Pt]**

PATIENT INFORMATION

Patient Id **geservice**

Patient Name **hss**

Weight (Lb) **111**

[Landmark]

Landmark **[>] [Nasion]**

PATIENT PROTOCOLS **[Patient Position]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Head] [Accept]**

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [Spin Echo] [Accept]**

Imaging Options none (default)

Psd Name **hss**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES

0:45 Hz 1:4± Line Freq **0** (0 to 1)

Number of Scan Passes **5** (5 to 10)

[Accept]

(lowest window) **[Save Series]**

o.3.1 Spike Noise (Head Axial, A/P)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	spike noise test
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Gradient Echo] [Accept]
Imaging Options	[...] [Sequential]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[10]
TR.	[20]
Flip Angle	[10]
Bandwidth	[15.63]

SCANNING RANGE

FOV	[24]
Slice Thickness	[5]
Spacing	[1.5]
S/I Start	[I60]
S/I End	[S60]
# of Slices	20 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1] (1:45)
# of Locs Before Pause	0
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
(lowest window)	[Save Series]

[Research Operations]	[Setup Params]
------------------------------	-----------------------

R1	11
R2	15
TG	0
Number of Frames	2
	[Done]

o.3.2 Spike Noise (Head Axial, R/L)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note

ACQUISITION TIMING

Freq Dir	[>] [R/L]
----------	---------------------

(lowest window)	[Save Series]
-----------------	----------------------

<u>RX MANAGER</u>	[Prepare to Scan]
-------------------	--------------------------

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.3.3 Spike Noise (Head Sagittal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series]. Double-click left mouse button on new series to activate it. See Note

IMAGING PARAMETERS

Plane **[>] [Sagittal]**

SCANNING RANGE

R/L Start **[R60]**

R/L End **[L60]**

P/A Center 0 (default)

I/S Center 0 (default)

ACQUISITION TIMING

Freq Dir **[>] [A/P]**

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.3.4 Spike Noise (Head Coronal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note

IMAGING PARAMETERS

Plane **[>] [Coronal]**

SCANNING RANGE

A/P Start **[P60]**

A/P End **[A60]**

S/I Center 0 (default)

R/L Center 0 (default)

ACQUISITION TIMING

Freq Dir **[>] [R/L]**

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.3.5 Spike Noise (Body Coronal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series]. Double-click left mouse button on new series to activate it. See Note
<u>PATIENT POSITION</u>	
Coil	[...] [Bodyd] [Accept]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.3.6 Spike Noise (Body Sagittal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
<u>SCANNING RANGE</u>	
R/L Start	[R60]
R/L End	[L60]
P/A Center	0 (default)
I/S Center	0 (default)
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [S/I]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.3.7 Spike Noise (Body Axial, R/L)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
<u>SCANNING RANGE</u>	
S/I Start	[I60]
S/I End	[S60]
L/R Center	0 (default)
P/A Center	0 (default)
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [R/L]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.3.8 Spike Noise (Body Axial, A/P)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [A/P]
(lowest window)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series

o.6.1 TPS/RF Waveform

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	rf waveforms
Weight (Lb)	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	pcal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1
TE	[25]
TR	[200]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[1] (0:28)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.7.1 Head EPIWP Base Z

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	epiwp head z
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Shots	[8]
TE	[Min Full]
TR	[6000]
Band Width	[62.50]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[10]
A/P Start	[P20]
A/P End	[A20]
# of Slices	3 (default)
S/I Center	0 (default)
R/L Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[128]
Phase	[64]
NEX	[1] (0:06)
Phase FOV	1.00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Water]

<u>(lowest window)</u>	[Save Series]
------------------------	----------------------

o.8.1 Head EPIWP Base Y

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	epiwp head y
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Shots	[8]
TE	[Min Full]
TR	[6000]
Band Width	[62.50]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[10]
R/L Start	[L20]
R/L End	[R20]
# of Slices	3 (default)
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[128]
Phase	[64]
NEX	[1] (0:06)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Water]

(lowest window) **[Save Series]**

o.9.1 Head EPIWP Base X

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	epiwp head x
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Shots	[8]
TE	[Min Full]
TR	[6000]
Band Width	[62.50]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[10]
S/I Start	[120]
S/I End	[S20]
# of Slices	3 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[128]
Phase	[64]
NEX	[1] (0:06)
Phase FOV	1.00 (default)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Water]

(lowest window) **[Save Series]**

o.10.1 DD/TR Board

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	dd/tr check
Weight (Lb))	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[200]
Bandwidth	15.63 (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:52)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

o.11.1 Body T/R Check

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body tr check
Weight (Lb))	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[150]
Bandwidth	15.63 (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	0
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:39)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

<u>(lowest window)</u>	[Save Series]
------------------------	----------------------

o.11.2 Head T/R Check

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	head tr check
Weight (Lb))	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[150]
Band Width	15.63 (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	0
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:39)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

<u>(lowest window)</u>	[Save Series]
------------------------	----------------------

o.12.1 Head Slice Thickness (2D, 3mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	head slice thick
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[300]
Bandwidth	[7.81] (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[20]
Slice Thickness	[3]
Spacing	[77]
S/I Start	[180]
S/I End	[S80]
# of Slices	3 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[4]
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.12.2 Head Slice Thickness (2D, 5mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button On new series to activate it. See Note 1
<u>IMAGING PARAMETERS</u>	
Imaging Options	none
<u>SCAN TIMING</u>	
TE	[Min Full]
TR	[800]
Bandwidth	[15.63]
<u>SCANNING RANGE</u>	
Slice Thickness	[5]
Spacing	[75]
<u>ACQUISITION TIMING</u>	
NEX	[4]
(<u>lowest window</u>)	[Save Series]
RX MANAGER	[Prepare to Scan]

Note: 1. Only values different from previous series shown. All other values should be the same as the previous series.

o.12.3 Head Slice Thickness (2D, 10mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note 1
<u>SCANNING RANGE</u>	
Slice Thickness	[10]
Spacing	[70]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: 1. Only values different from previous series shown. All other values should be the same as the previous series.

o.12.4 Head Slice Thickness (2D, Grad Echo, 5mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note 1
<u>IMAGING PARAMETERS</u>	
Pulse Seq	[...] [Gradient Echo] [Accept]
<u>SCAN TIMING</u>	
TE	[9]
TR	[100]
Flip Angle	[30]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
Slice Thickness	[5]
Spacing	[77]
S/I Start	0
S/I End	0
# of Slices	1
<u>ACQUISITION TIMING</u>	
Phase	[128]
NEX	[1] (0:16)
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: 1. Only values different from previous series shown. All other values should be the same as the previous series.

o.12.5 Head Slice Thickness (3D, Localizer)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note 1
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Imaging Options	none
<u>SCAN TIMING</u>	
TR	[50]
<u>SCANNING RANGE</u>	
FOV	[24]
Spacing	[1.5]
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [S/I]
(<u>lowest window</u>)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series are shown. All other values should be the same as the previous series.

o.12.6 Head Slice Thickness (3D, Grad Echo, 1.5mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [3D]
Pulse Seq	[...] [Gradient Echo] [Accept]
<u>SCAN TIMING</u>	
TR	[34]
Bandwidth	[15.63]
TE	[9]
Flip Angle	[30]
<u>SCANNING RANGE</u>	
Slice Thickness	[1.5]
# of Slices	[28]
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [A/P]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series are shown. All other values should be the same as the previous series.

o.13.1 SNR Check (Body, Axial)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body snr
Weight (Lb)	111
	[Landmark]
Landmark	[Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[20]
TR	[2000]
<u>SCANNING RANGE</u>	
FOV.	[40]
Slice Thickness.	[3]
Spacing	[1.5]
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1] (8:52)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
 (lowest window)	 [Save Series]

o.13.2 SNR Check (Body, Sagittal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note.

IMAGING PARAMETERS

Plane **[>] [Sagittal]**

SCANNING RANGE

R/L Start **0**
R/L End **0**
P/A Center 0 (default)
I/S Center 0 (default)

ACQUISITION TIMING

Freq Dir **[>] [S/I]**

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.13.3 SNR Check (Body, Coronal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note.
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
<u>SCANNING RANGE</u>	
A/P Start	[SEE PROCEDURE]
A/P End	[SEE PROCEDURE]
S/I Center	0 (default)
R/L Center	0 (default)
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.13.4 SNR Check (Head, Axial)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	head snr
Weight (Lb)	111
	[Landmark]
Landmark	[Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[20]
TR	[2000]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[3]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1] (8.52)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.13.5 SNR Check (Head, Sagittal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note

IMAGING PARAMETERS

Plane **[>] [Sagittal]**

SCANNING RANGE

R/L Start **0**
R/L End **0**
P/A Center 0 (default)
I/S Center 0 (default)

ACQUISITION TIMING

Freq Dir **[>] [S/I]**

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.13.6 SNR Check (Head, Coronal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
<u>SCANNING RANGE</u>	
A/P Start	[SEE PROCEDURE]
A/P End	[SEE PROCEDURE]
S/I Center	0 (default)
R/L Center	0 (default)
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.15.1 Slice Offset (2D, 3mm, Table 50mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	head slice thick
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [SpinEcho] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[300]
Bandwidth	15.63 (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[20]
Slice Thickness	[3]
Spacing	[97]
I/S Start	I50
I/S End	S50
# of Slices	2 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	50

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[2] (2:35)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(<u>lowest window</u>)	[Save Series]
--------------------------	----------------------

o.15.2 Slice Offset (2D, 3mm, Table –50mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>SCANNING RANGE</u>	
Table Delta	[-50]
(<u>lowest window</u>)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.15.3 Slice Offset (3D, Grad Echo Localizer)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Pulse Seq	[...] [Gradient Echo] [Accept]
<u>SCAN TIMING</u>	
TE	[9]
TR	[50]
Flip Angle	[30]
Bandwidth	[15.63]
<u>SCANNING RANGE</u>	
Slice Thickness	[5]
Spacing	[1.5]
R/L Start	0
R/L End	0
# of Slices	1
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)
<u>ACQUISITION TIMING</u>	
Phase	[128]
NEX	[1] (0:10)
Freq Dir	[>] [S/I]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.15.4 Slice Offset (3D, Grad Table 50mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [3D]
<u>SCAN TIMING</u>	
TR	[34]
<u>SCANNING RANGE</u>	
Slice Thickness	[3]
# of Slices	[60]
Table Delta	[50]
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [A/P]
(lowest window)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.15.5 Slice Offset (3D, Grad Table –50mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	Place cursor in Series list area Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>SCANNING RANGE</u>	
Table Delta	[-50]
(<u>lowest window</u>)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.16.1 RFT Body Scan

SCAN OPTION INPUT

PATIENT REGISTER . **[New Pt]**

PATIENT INFORMATION

Patient Id **geservice**

Patient Name **rft**

Weight (Lb) **111**

[Landmark]

Landmark **[>] [Sternal Notch]**

PATIENT PROTOCOLS **[Patient Position]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Body] [Accept]**

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [SpinEcho] [Accept]**

Imaging Options none (default)

Psd Name **rft**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES (Refer to procedure for proper User CV values.)

(lowest window) **[Save Series]**

o.16.2 RFT Head Scan

SCAN OPTION INPUT

PATIENT REGISTER . **[New Pt]**

PATIENT INFORMATION

Patient Id **geservice**

Patient Name **rft**

Weight (Lb) **111**

[Landmark]

Landmark **[>] [Sternal Notch]**

PATIENT PROTOCOLS **[Patient Position]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Head] [Accept]**

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [SpinEcho] [Accept]**

Imaging Options none (default)

Psd Name **rft**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES (Refer to procedure for proper User CV values.)

(lowest window) **[Save Series]**

o.17.1 Body F/R Quadrature

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body f/r quad
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EDR]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[92]
Band Width	[15.63]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[2] (0:43)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
 (lowest window)	 [Save Series]

o.17.2 Head F/R Quadrature

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	hd f/r quad
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EDR]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[92]
Bandwidth	[15.63]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[2] (0:43)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.18.1 Dummy Load

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	dummy load cal
Weight (Lb))	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	cal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[200]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:52)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

[Research Operations], [Display CVs]

Modify the following:

Calmode	2 (trapezoid pulse)
p2_ramp	1 (1 μ sec ramptime)
t2	50000 (50 msec tr)
pismode	1 (exec service)
pmode	1 (data collection)
daqm	1 (data collection)

[Accept]

[Research Operations] [Setup Params]

R1	7
R2	14
Number of Frames	[2]
	Window 1
	Frame 1 + Frame 0
	[Done]

[Manual Prescan]

Windows (From Menu Bar)	2
Window Type	I Channel
	[Done]

o.19.1 LVShim (Localizer)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	LV shim
Weight (Lb))	300
	[Landmark]
Landmark	[>] [Sternal]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[20]
TR	[300]
Bandwidth	[15.63] (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
S/I Center	0 (default)
R/L Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1] (1:23)
Phase FOV	1.00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.19.2 LVShim Scan

SCAN OPTION INPUT

PATIENT REGISTER . **[New Pt]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Body] [Accept]**

Series Description Note 1

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [Spin Echo] [Accept]**

Imaging Options none (default)

Psd Name **LVshim**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES

No. of scan planes **[6]** [2.0 to 16.0]

Bandwidth in Hz **[Note 2]** [100.0 to 20000.0]

[Accept]

(lowest window) **[Save Series]**

Note 1: Refer to procedure for description name to use.

Note 2: Refer to procedure for proper Bandwidth to use.

o.20.1 First Image

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	first image
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	[1] (default)
TE	[20]
TR	[400]
Band Width	[15.63] (1.0T only)
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[2] (3:26)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
 (lowest window)	 [Save Series]

o.21.1 T2 Uniformity (Body, Axial)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body T2
Weight (Lb))	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	[4]
TE	[30]
TR	[1000]
BW	[15.63]
BW2	[15.63]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[1] (2:20)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.21.2 T2 Uniformity (Body, Sagittal)

RX MANAGER Place cursor in Series list area.
Press right mouse button, then
[Copy Series], [Paste Series].
Double-click left mouse button
on new series to activate it.
See Note

IMAGING PARAMETERS

Plane .[>] **[Sagittal]**

SCANNING RANGE

R/L Start **0**
R/L End **0**
P/A Center 0 (default)
I/S Center 0 (default)

ACQUISITION TIMING

Freq Dir **[>] [S/I]**

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.21.3 T2 Uniformity (Body/Head, Coronal)

RX MANAGER Place cursor in Series list area.
Press right mouse button, then
[Copy Series], [Paste Series].
Double-click left mouse button
on new series to activate it.
See Note

IMAGING PARAMETERS

Plane **[Coronal]**

SCANNING RANGE

A/P Start see procedure
A/P End see procedure
S/I Center 0 (default)
R/L Center 0 (default)

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.22.1 Isocenter Cal (DQA-III)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	iso cal
Weight (Lb))	111
	[Landmark]
Exam Description	dqa isoz cal (see Note 1)
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[300]
Bandwidth	[15.63]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[3.0]
Spacing	[1.5]
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[512]
Phase	[256]
NEX	[1] (1:20)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

Note 1: The Exam Description name must be exact because it is used by the automated analysis tool (Daily Quality).

o.23.1 Head Power Monitor

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	head pm
Weight (Lb))	300 ← IMPORTANT
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	cal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[200]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	0
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:52)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]

(<u>lowest window</u>)	Save Series]
--------------------------	---------------------

o.23.2 Body Power Monitor

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body pm
Weight (Lb))	300 ← <i>IMPORTANT</i>
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	cal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[200]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	0
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] [R/L]
Auto Center Freq	[>] [Peak]

(lowest window) **Save Series]**

[Research Operations], [Display CVs]

Modify the following:

dcset	255
t3	20000
	[Accept]

o.23.3 Amp Shutdown Verification

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	pm check
Weight (Lb)	300 ← <i>IMPORTANT</i>
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	cal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[200]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	0
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[128]
NEX	[2] (0:52)
Freq Dir	[>] [R/L]
Auto Center Freq.	[>] [Peak]

(lowest window) [Save Series]

[Research Operations], [Display CVs]

Modify the following:

dcset	255
aset	30
calmode	2
trig	1

For RF/PEN Cabinet only, modify the following additional CVs:

pwset	255
aset	120
p1	3100
p3	2400

[Accept]

o.24.1 System Gain Cal (Body)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	bd sys gain
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[21]
TR	[700]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[5]
Spacing	[1.5]
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[128]
NEX	[1] (1:38)
Phase FOV	1.00 (default)
Freq Dir	[>] [R/L]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.24.2 System Gain Cal (Head)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	hd sys gain
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[700]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[128]
NEX	[1] (1:38)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.25.1 Pin Diode Test (Head)

<u>SCAN OPTION</u>	<u>INPUT</u>		
<u>PATIENT REGISTER</u>	[New Pt]		
<u>PATIENT INFORMATION</u>			
Patient Id	geservice		
Patient Name	tlt head		
Weight (Lb)	111		
	[Landmark]		
Landmark	[>] [Nasion]		
<u>PATIENT PROTOCOLS</u>	[Patient Position]		
<u>PATIENT POSITION</u>			
Patient Position	[>] [Supine]		
Patient Entry	[>] [Head First]		
Coil	[...] [Head] [Accept]		
<u>IMAGING PARAMETERS</u>			
Plane	[>] [Axial]		
Mode	[>] [2D]		
Pulse Seq	[...] [Spin Echo] [Accept]		
Imaging Options	none (default)		
Psd Name	tlt		
Protocol	no entry		
<u>ADDITIONAL PARAMETERS [USER CVs]</u>			
<u>USER CONTROL VARIABLES</u>			
SNR/T2 Tests: (-1=noise only)	-1		(-1 to 4)
TRMap: 0=off, 1=ax, 2=sag, 3=cor	0		(0 to 3)
Stability: 1=Z, 2=X, 3=Y, 4=all	0		(0 to 4)
Run GradCal Test: 0=off, 1=on	0		(0 to 1)
ASC Analysis: 0=off, 1=immediate, 2=queued	0		(0 to 2)
	[Accept]		
<u>SCANNING RANGE</u>			
FOV	[24]		
Slice Thickness	[4]		
Spacing	0		
S/I Start	0		
S/I End	0		
# of Slices	1 (default)		
L/R Center	0 (default)		
P/A Center	0 (default)		
Table Delta	0.00 (default)		

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[1]
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
(lowest window)	[Save Series]

o.25.2 Pin Diode Test (Body)

<u>SCAN OPTION</u>	<u>INPUT</u>		
<u>PATIENT REGISTER</u>	[New Pt]		
<u>PATIENT INFORMATION</u>			
Patient Id	geservice		
Patient Name	tlr body		
Weight (Lb)	111		
	[Landmark]		
Landmark	[>] [Sternal Notch]		
<u>PATIENT PROTOCOLS</u>	[Patient Position]		
<u>PATIENT POSITION</u>			
Patient Position	[>] [Supine]		
Patient Entry	[>] [Head First]		
Coil	[...] [Body] [Accept]		
<u>IMAGING PARAMETERS</u>			
Plane	[>] [Axial]		
Mode	[>] [2D]		
Pulse Seq	[...] [Spin Echo] [Accept]		
Imaging Options	none (default)		
Psd Name	tlr		
Protocol	no entry		
<u>ADDITIONAL PARAMETERS</u> [USER CVs]			
<u>USER CONTROL VARIABLES</u>			
Shim Test: 0=off, 1=on	0		[0 to 1]
GraImage: 1=ax,2=sa,3=co,4=all	0		[0 to 4]
SNR/T2 Tests: (-1=noise only)	-1		[-1 to 4]
TRMap: 0=off,1=ax,2=sag,3=cor	0		[0 to 3]
Stability: 1=Z,2=X,3=Y,4=all	0		[0 to 4]
Run GradCal Test: 0=off, 1=on	0		[0 to 1]
ASC Analysis: 0=off, 1=immediate, 2=queued	0		[0 to 2]
	[Accept]		
<u>SCANNING RANGE</u>			
FOV	[48]		
Slice Thickness	[4]		
Spacing	0		
Start	0		
End	0		
# of Slices	1 (default)		
L/R Center	0 (default)		
P/A Center	0 (default)		
Table Delta	0.00 (default)		

ACQUISITION TIMING

Freq.	[256]
Phase	[128]
NEX	[1]
Freq Dir	[>] [R/L]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.26.1 SST Head Full Power

<u>SCAN OPTION</u>	<u>INPUT</u>		
<u>PATIENT REGISTER</u>	[New Pt]		
<u>PATIENT INFORMATION</u>			
Patient Id	geservice		
Patient Name	sst head		
Weight (Lb)	111		
	[Landmark]		
Landmark	[>] [Sternal Notch]		
<u>PATIENT PROTOCOLS</u>	[Patient Position]		
<u>PATIENT POSITION</u>			
Patient Position	[>] [Supine]		
Patient Entry	[>] [Head First]		
Coil	[...] [Body] [Accept]		
<u>IMAGING PARAMETERS</u>			
Plane	[>] [Axial]		
Mode	[>] [2D]		
Pulse Seq	[...] [Spin Echo] [Accept]		
Imaging Options	none (default)		
Psd Name	sst		
Protocol	no entry		
<u>ADDITIONAL PARAMETERS</u> [USER CVs]			
<u>USER CONTROL VARIABLES</u>			
Stab: -1=RF, <1,2,3>=<x,y,z>w, w/o Grad, 4=all	[4]		[-1 to 4]
RF S/N: 1=on, -1=Noise only	[1]		[-1 to 1]
RF Linearity: 1=on, 0=off	[1]		[0 or 1]
Grad Bits: 1=on	[1]		[0 or 1]
RF Bits: 1=on	[1]		[0 or 1]
Coil/Sample: 2=O/N, 3=U/N	[See Note 1]		[2 or 3]
Rcv Coil: 1=h, 2=b, 3=s, 4=t	[1]		[1 to 4]
Configcode:	[0]		[-3.40282e+38 to 3.40282e+38]
ASC Analysis: 0=off, 1=immediate, 2=queued	[0]		[0 to 2]
Auto GradAmp Ctrl: 1=on 0=off	[1]		[0 or 1]
	[Accept]		

(lowest window)

[**Save Series**]

Note 1: You must select proper *Coil/Sample* value as follows: **2** = O/N (original sst coil/NiCl₂ sample), **3** = U/N (universal sst coil/NiCl₂ sample).

o.26.2 SST Body Full Power

SCAN OPTION INPUT

PATIENT REGISTER **[New Pt]**

PATIENT INFORMATION

Patient Id **geservice**

Patient Name **sst body**

Weight (Lb) **111**

[Landmark]

Landmark **[>] [Sternal Notch]**

PATIENT PROTOCOLS **[Patient Position]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Body] [Accept]**

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [Spin Echo] [Accept]**

Imaging Options none (default)

Psd Name **sst**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES

Stab: -1=RF, <1,2,3>=<x,y,z>w, w/o Grad, 4=all **[4]** [-1 to 4]

RF S/N: 1=on, -1=Noise only **[1]** [-1 to 1]

RF Linearity: 1=on, 0=off **[1]** [0 or 1]

Grad Bits: 1=on **[1]** [0 or 1]

RF Bits: 1=on **[1]** [0 or 1]

Coil/Sample: 2=O/N, 3=U/N **[See Note 1]** [2 or 3]

Rcv Coil: 1=h, 2=b, 3=s, 4=t **[2]** [1 to 4]

Configcode: **[0]** [-3.40282e+38 to 3.40282e+38]

ASC Analysis: 0=off, 1=immediate, 2=queued **[0]** [0 to 2]

Auto GradAmp Ctrl: 1=on 0=off **[1]** [0 or 1]

[Accept]

(lowest window)

[Save Series]

Note 1: You must select proper *Coil/Sample* value as follows: **2** = O/N (original sst coil/NiCl₂ sample), **3** = U/N (universal sst coil/NiCl₂ sample).

o.26.3 SST Multicoil Test

SCAN OPTION INPUT

PATIENT REGISTER **[New Pt]**

PATIENT INFORMATION

Patient Id **geservice**

Patient Name **sst mc**

Weight (Lb) **111**

[Landmark]

Landmark **[>] [Sternal Notch]**

PATIENT PROTOCOLS **[Patient Position]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Phased Array]]**

See *Coil Choices* below **[Accept]**

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [Spin Echo] [Accept]**

Imaging Options none (default)

Psd Name **sst**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES

Stab: -1=RF, <1,2,3>=<x,y,z>w, w/o Grad, 4=all **[0]** [-1 to 4]

RF S/N: 1=on, -1=Noise only **[1]** [-1 to 1]

RF Linearity: 1=on, 0=off **[0]** [0 or 1]

Grad Bits: 1=on **[0]** [0 or 1]

RF Bits: 1=on **[1]** [0 or 1]

Coil/Sample: 2=O/N, 3=U/N **[3]** [2 or 3]

Rcv Coil: 1=h, 2=b, 3=s, 4=t **[1]** [1 to 4]

Configcode: **[0]** [-3.40282e+38 to 3.40282e+38]

ASC Analysis: 0=off, 1=immediate, 2=queued **[0]** [0 to 2]

Auto GradAmp Ctrl: 1=on 0=off **[0]** [0 or 1]

[Accept]

(lowest window)

[Save Series]

COIL CHOICES

CTL TOP

[For coil channels 2, 3, 4, 5]

CTL BOT

[For coil channels 6, 7]

o.26.4 SST Solid-state Driver Mode

<u>SCAN OPTION</u>	<u>INPUT</u>	
<u>PATIENT REGISTER</u>	[New Pt]	
<u>PATIENT INFORMATION</u>		
Patient Id	geservice	
Patient Name	sst ss drvr	
Weight (Lb)	111	
	[Landmark]	
Landmark	[>] [Sternal Notch]	
<u>PATIENT PROTOCOLS</u>	[Patient Position]	
<u>PATIENT POSITION</u>		
Patient Position	[>] [Supine]	
Patient Entry	[>] [Head First]	
Coil	[...] [Body] [Accept]	
<u>IMAGING PARAMETERS</u>		
Plane	[>] [Axial]	
Mode	[>] [2D]	
Pulse Seq	[...] [Spin Echo] [Accept]	
Imaging Options	none (default)	
Psd Name	sst	
Protocol	no entry	
<u>ADDITIONAL PARAMETERS</u> [USER CVs]		
<u>USER CONTROL VARIABLES</u>		
Stab: -1=RF, <1,2,3>=<x,y,z>w, w/o Grad, 4=all	[4]	[-1 to 4]
RF S/N: 1=on, -1=Noise only	[1]	[-1 to 1]
RF Linearity: 1=on, 0=off	[1]	[0 or 1]
Grad Bits: 1=on	[0]	[0 or 1]
RF Bits: 1=on	[0]	[0 or 1]
Coil/Sample: 2=O/N, 3=U/N	[3]	[2 or 3]
Rcv Coil: 1=h, 2=b, 3=s, 4=t	[1]	[1 to 4]
Configcode:	[2]	[-3.40282e+38 to 3.40282e+38]
ASC Analysis: 0=off, 1=immediate, 2=queued	[0]	[0 to 2]
Auto GradAmp Ctrl: 1=on 0=off	[0]	[0 or 1]
	[Accept]	
(lowest window)	[Save Series]	

o.26.5 SST True Head Mode

SCAN OPTION INPUT

PATIENT REGISTER **[New Pt]**

PATIENT INFORMATION

Patient Id **geservice**

Patient Name **sst true head**

Weight (Lb) **111**

[Landmark]

Landmark **[>] [Sternal Notch]**

PATIENT PROTOCOLS **[Patient Position]**

PATIENT POSITION

Patient Position **[>] [Supine]**

Patient Entry **[>] [Head First]**

Coil **[...] [Head] [Accept]**

IMAGING PARAMETERS

Plane **[>] [Axial]**

Mode **[>] [2D]**

Pulse Seq **[...] [Spin Echo] [Accept]**

Imaging Options none (default)

Psd Name **sst**

Protocol no entry

ADDITIONAL PARAMETERS **[USER CVs]**

USER CONTROL VARIABLES

Stab: -1=RF, <1,2,3>=<x,y,z>w, w/o Grad, 4=all	[4]	[-1 to 4]
RF S/N: 1=on, -1=Noise only	[1]	[-1 to 1]
RF Linearity: 1=on, 0=off	[1]	[0 or 1]
Grad Bits: 1=on	[0]	[0 or 1]
RF Bits: 1=on	[0]	[0 or 1]
Coil/Sample: 2=O/N, 3=U/N	[3]	[2 or 3]
Rcv Coil: 1=h, 2=b, 3=s, 4=t	[1]	[1 to 4]
Configcode:	[0]	[-3.40282e+38 to 3.40282e+38]
ASC Analysis: 0=off, 1=immediate, 2=queued	[0]	[0 to 2]
Auto GradAmp Ctrl: 1=on 0=off	[0]	[0 or 1]
	[Accept]	

(lowest window)

[Save Series]

o.27.1 RFS Head

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	rfs
Weight (Lb)	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	pcal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[500]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[256]
NEX	[2] (1:43)
Freq Dir	[>] [A/P]
Auto Center Freq.	[>] [Peak]
(<u>lowest window</u>)	[Save Series]

o.27.2 RFS Body

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	rfs
Weight (Lb)	300
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	pcal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[500]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[256]
NEX	[2] (1:43)
Freq Dir	[>] [R/L]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.28.1 Full Field Distortion (Sagittal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	ffd
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[25]
TR.	[300]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[20]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[256]
NEX	[1] (1:23)
Phase FOV	1:00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.28.2 Full Field Distortion (Coronal Localizer)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	[New Series] (If this procedure is Being performed immediately after The Sagittal series, otherwise click [New Pt]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[20]
TR.	[150]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[20]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
R/L Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[128]
NEX	[1] (0:24)
Phase FOV	1:00 (default)
Freq Dir	[>] [R/L]
Auto Center Freq.	[>] [Peak]

(<u>lowest window</u>)	[Save Series]
--------------------------	----------------------

<u>RX MANAGER</u>	[Prepare to Scan]
-------------------	--------------------------

o.28.3 Full Field Distortion (Coronal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u>	[New Series]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[25]
TR.	[300]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[20]
Spacing	[1.5]
Start	See procedure
End	See procedure
# of Slices	1 (default)
S/I Center	0 (default)
R/L Center	0 (default)
Table Delta	0.00 (default)
<u>ACQUISITION TIMING</u>	
Freq.	[256]
Phase	[256]
NEX	[1] (1:23)
Phase FOV	1:00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

o.29.1 Noise Floor (Head)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	noise floor
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Gradient Echo] [Accept]
Imaging Options	[...] [Sequential]
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[9]
TR.	[20]
Flip Angle	[10]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[256]
NEX	[1] (0:05)
Phase FOV	1:00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.29.2 Noise Floor (Body)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button On new series to activate it See Note
<u>PATIENT POSITION</u>	
Coil	[...] [Body] [Accept]
(lowest window	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.30.1 Carrier Leakage

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	carrier leak
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Gradient Echo] [Accept]
Imaging Options	[...] [Sequential]
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[16]
TR.	[35]
Flip Angle	[30]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[40]
Slice Thickness	[5]
Spacing	[2.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq.	[256]
Phase	[128]
NEX	[1] (0:05)
Phase FOV	1:00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq.	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.31.1 Respiratory Compensation

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	resp bellows
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [Respiratory Compensation] [Accept]
Psd Name	none
Protocol	no entry

It is not necessary to enter the remainder of scan protocol for this procedure.

o.32.1 Gradcal-X (DQA-III)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	gradcal
Weight (Lb)	111
Extra Description	dqa xfield cal (See Note 1)
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[25]
TR.	[300]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[3]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[512]
Phase	[256]
NEX	[1] (1:20)
Phase FOV	1 (default)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

Note 1: The Exam Description name must be exact because it is used by the automated analysis tool (Daily Quality).

o.32.2 Gradcal-Y (DQA-III)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	gradcal
Weight (Lb)	111
Exam Description	dqa yfield cal (See Note 1)
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[25]
TR.	[300]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[3]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[512]
Phase	[256]
NEX	[1] (1:20)
Phase FOV	1 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

Note 1: The Exam Description name must be exact because it is used by the automated analysis tool (Daily Quality).

o.32.3 Gradcal-Z (DQA-III)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	gradcal
Weight (Lb)	111
Exam Description	dqa zfield cal (See Note 1)
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[25]
TR.	[300]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[3]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[512]
Phase	[256]
NEX	[1] (1:20)
Phase FOV	1.00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

Note 1: The Exam Description name must be exact because it is used by the automated analysis tool (Daily Quality).

o.33.1 Body EPIWP Base Z

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	epiwp body z
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Shots	[8]
TE.	[Min Full]
TR.	[6000]
Bandwidth	[62.50]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[10]
A/P Start	[P20]
A/P End	[A20]
# of Slices	3 (default)
S/I Center	0 (default)
R/L Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[128]
Phase	[64]
NEX	[1] (0:06)
Phase FOV	[1]
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Water]
<u>(lowest window)</u>	[Save Series]

o.34.1 Body EPIWP Base Y

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	epiwp body y
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Shots	[8]
TE.	[Min Full]
TR.	[6000]
Bandwidth	[62.50]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[10]
R/L Start	[L20]
R/L End	[R20]
# of Slices	3 (default)
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[128]
Phase	[64]
NEX	[1] (0:06)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Water]

<u>(lowest window)</u>	[Save Series]
------------------------	----------------------

o.35.1 Body EPIWP Base X

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	epiwp body x
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Shots	[8]
TE.	[Min Full]
TR.	[6000]
Bandwidth	[62.50]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[10]
Spacing	[10]
S/I Start	[120]
S/I End	[S20]
# of Slices	3 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[128]
Phase	[64]
NEX	[1] (0:06)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Water]
<u>(lowest window)</u>	[Save Series]

o.36.1 Correlated Noise (256 matrix)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	corr noise
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Gradient Echo] [Accept]
Imaging Options	[...] [Sequential]
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[10]
TR	[20]
Flip Angle	[10]
Bandwidth	[>] 15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1] (0:05)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
 (lowest window)	 [Save Series]

[Research Operations]	[Setup Params]
R1	11
R2	15
TG	0
Number of Frames	[2]
	[Done]

o.36.2 Correlated Noise (512 matrix)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>SCAN TIMING</u>	
Bandwidth	[>] 31.25
<u>SCANNING RANGE</u>	
FOV	[20]
<u>ACQUISITION TIMING</u>	
Freq	[512]
<u>(lowest window)</u>	[Save Series]
<u>RX MANGER</u>	[Prepare to Scan]

Note: Only values different from previous series shown. All other values should be the same as the previous series.

o.36.3 Multicoil Correlated Noise

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	mc corr noise
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Phased Array] [CTLMID] or [Pelvic] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Gradient Echo] [Accept]
Imaging Options	[...] [Sequential]
Psd Name	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[10]
TR	[31]
Flip Angle	[10]
Bandwidth	[>] 31.25
<u>SCANNING RANGE</u>	
FOV	[20]
Slice Thickness	[5]
Spacing	[1.5]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[512]
Phase	[256]
NEX	[1] (0:06)
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(lowest window) [Save Series]

[Research Operations] [Setup Params]

R1	11
R2	15
TG	0
Number of Frames	[2]

[Done]

[Research Operations] [Display CVs]

CV Name: saveinter <Enter>

Set value to: 1 <Enter> (Note 1)

[Accept]

Note 1: Setting this CV to 1 generates five images for each single multi-coil scan. Images 1 to 4 will correspond to data from Receivers 0 to 3. Therefore, Image 1 is reconstructed from Receiver 0 data. Image 5 is the combined and processed data from all 4 Receivers. This is the same image normally observed when saveinter = 0.

o.36.4 EPI Correlated Noise

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	fast rcvr corr noise
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	[...] [EPI]
Psd Name	no entry
Protocol	default
<u>ADDITIONAL PARAMETERS</u>	[USER CVs] (1.5T only)
<u>USER CONTROL VARIABLES</u>	
	Ramp Sampling: 1=on, 0=off 0 (0 or 1)
<u>SCAN TIMING</u>	
# of Shots	1
TE	[Min Full]
TR	[2000]
Bandwidth	[15.63]
<u>SCANNING RANGE</u>	
FOV	[99]
Slice Thickness	[10]
Spacing	[10]
Start	0
End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[1] (0:02)
Phase FOV	1.00 (default)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]
Autoshim	no (Verify)
Phase Correct	no (Verify)
<u>(lowest window)</u>	[Save Series]

o.37.1 Body Slice Thickness (Sagittal, 3mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body slice thick
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[300]
Bandwidth	[7.81]
<u>SCANNING RANGE</u>	
FOV	[20]
Slice Thickness	[3]
Spacing	[77]
R/L Start	[R80]
R/L End	[L80]
# of Slices	3 (default)
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[4]
Phase FOV	1.00 (default)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]
 (<u>lowest window</u>)	 [Save Series]

o.37.2 Body Slice Thickness (Sagittal, 5mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it See Note 1
<u>SCANNING RANGE</u>	
Slice Thickness	[5]
Spacing	[75]
(<u>lowest window</u>)	[Save Series]
<u>RX MANGER</u>	[Prepare to Scan]

Note: 1. Only values different from previous series are shown. All other values Should be the same as the previous series.

o.37.3 Body Slice Thickness (Sagittal, 10mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button On new series to activate it See Note 1
<u>SCANNING RANGE</u>	
Slice Thickness	[10]
Spacing	[70]
(<u>lowest window</u>)	[Save Series]
<u>RX MANGER</u>	[Prepare to Scan]

Note: 1. Only values different from previous series are shown. All other values Should be the same as the previous series.

o.37.4 Body Slice Thickness (Localizer)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note
<u>SCANNING RANGE</u>	
Slice Thickness	[3]
Spacing	[1.5]
Start	0
End	0
# of Slices	1
<u>AQUISTION TIMING</u>	
Freq	[256]
Phase	[128]
NEX	[1]
Phase FOV	1.00 (default)
Freq Dir	[>] [S/I]
(lowest window)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series are shown. All other values Should be the same as the previous series.

o.37.5 Body Slice Thickness (Coronal, 3mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note.

IMAGING PARAMETERS

Plane **[>] [Coronal]**

SCANNING RANGE

Spacing **[77]**

AQUISITION TIMING

Freq **[256]**

Phase **[256]**

NEX **[4]**

Phase FOV 1.00 (default)

Freq Dir **[>] [S/I]**

Auto Center Freq **[>] {Peak}**

(lowest window) **[Save Series]**

RX MANAGER **[Prepare to Scan]**

Note: Only values different from previous series are shown. All other values
Should be the same as the previous series.

o.37.6 Body Slice Thickness (Coronal, 5mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note1.
<u>SCANNING RANGE</u>	
Slice Thickness	[5]
Spacing	[75]
(lowest window)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note 1: Only values different from previous series are shown. All other values Should be the same as the previous series.

o.37.7 Body Slice Thickness (Coronal, 10 mm)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it See note 1.
<u>SCANNING RANGE</u>	
Slice Thickness	[10]
Spacing	[70]
(lowest window)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note 1: Only values different from previous series are shown. All other values
Should be the same as the previous series.

o.38.1 Geometry Verification (Coronal)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	gradient polarity (See Note 1)
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Coronal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE.	[20]
TR.	[300]
Bandwidth	15.63
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[10]
Spacing	[1.5]
A/P Start	0
A/P End	0
# of Slices	1 (default)
S/I Center	0 (default)
R/L Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1]
Phase FOV	1:00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Peak]

(lowest window) **[Save Series]**

Note 1: The Exam Description name must be exact because it is used by the automated analysis tool (Daily Quality).

o.38.2 Geometry Verification (Axial)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>RX MANAGER</u> .	Place cursor in Series list area. Press right mouse button, then [Copy Series], [Paste Series] . Double-click left mouse button on new series to activate it. See Note 1.
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
<u>SCANNING RANGE</u>	
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
<u>ACQUISITION TIMING</u>	
Freq Dir	[>] [R/L]
(lowest window)	[Save Series]
<u>RX MANAGER</u>	[Prepare to Scan]

Note: Only values different from previous series are shown. All other values should be the same as the previous series.

o.39.1 FID Tuning

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	fid
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	4
TE	[20]
TR	[250]
Band Width	[15.63]
Band Width 2	[15.63]
<u>SCANNING RANGE</u>	
FOV	[40]
Slice Thickness	[15]
Spacing	[1.5]
R/L Start	0
R/L End	0
# of Slices	1 (default)
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[2]
Phase FOV	1.00 (default)
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.40.1 Gradient Waveforms

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	grad
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	[4]
TE	[25]
TR	[1000]
Bandwidth	[15.63]
Bandwidth 2	[15.63]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[5]
Spacing	[1.5]
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[1] (4:28)
Phase FOV	1.00 (default)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]
<u>(lowest window)</u>	[Save Series]

o.41.1 Body APB Cal

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	body apb check
Weight (Lb)	300 ← IMPORTANT
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	cal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[55]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:52)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(lowest window)	[Save Series]
-----------------	---------------

[Research Operations]	[Display CVs]
-----------------------	---------------

Modify the following:

Calmode	5 (normal rf sinc normal)
1a_rf1	32766 (sets 90° pulse full-scale)
1a_rf2	0 (turns off 180° pulse)

[Accept]

[Research Operations]	[Setup Params]
-----------------------	----------------

TG	50
----	----

[Done]

o.41.2 Head APB Cal

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	head apb check
Weight (Lb)	300 ← <i>IMPORTANT</i>
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Head] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	cal
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	1 (default)
TE	[25]
TR	[55]
<u>SCANNING RANGE</u>	
FOV	[24]
Slice Thickness	[5]
Spacing	0
S/I Start	0
S/I End	0
# of Slices	1 (default)
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[2] (0:52)
Freq Dir	[>] [A/P]
Auto Center Freq	[>] [Peak]

(lowest window)	[Save Series]
-----------------	---------------

[Research Operations]	[Display CVs]
-----------------------	---------------

Modify the following:

calmode	5 (normal rf sinc normal)
1a_rf1	32766 (sets 90° pulse full-scale)
1a_rf2	0 (sets 180° pulse to small-scale)

[Accept]

[Research Operations]	[Setup Params]
-----------------------	----------------

TG	50
----	----

[Done]

o.42.1 Bandpass Asymmetry Correction Compensation (BACC)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	bacc
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none (default)
Psd Name	bat
Protocol	no entry
<u>(lowest window)</u>	[Save Series]

o.43.1 Center Frequency

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	center frequency
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Sagittal]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none
Psd Name	none
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	[4]
TR	[250]
Bandwidth	[15.00]
Bandwidth2	no entry
<u>SCANNING RANGE</u>	
FOV	[40]
Slice Thickness	[15]
Spacing	[1.5]
R/L Start	0
R/L End	0
# of Slices	1 (default)
P/A Center	0 (default)
I/S Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[256]
NEX	[2]
Phase FOV	1 (default)
Freq Dir	[>] [S/I]
Auto Center Freq	[>] [Peak]

<u>(lowest window)</u>	[Save Series]
------------------------	----------------------

o.46.1 Spike Noise (C Class)

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER .</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	spike noise
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Sternal Notch]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Body] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Gradient Echo] [Accept]
Imaging Options	[...] [Sequential] [Accept]
Psd Name	no entry
Protocol	no entry
<u>SCAN TIMING</u>	
# of Echoes	[1]
TE	[9]
TR	[33]
Flip Angle	[30]
Bandwidth	[15.63]
<u>SCANNING RANGE</u>	
FOV	[48]
Slice Thickness	[3]
Spacing	[1.5]
S/I Start	[I60]
S/I End	[S60]
# of Slices	[60]
L/R Center	0 (default)
P/A Center	0 (default)
Table Delta	0.00 (default)

ACQUISITION TIMING

Freq	[256]
Phase	[128]
NEX	[1]
Phase FOV	1 (default)
Freq Dir	[>] [R/L]
Auto Center Freq	[>] [Peak]

<u>(lowest window)</u>	[Save Series]
------------------------	----------------------

o.47.1 ECMT

<u>SCAN OPTION</u>	<u>INPUT</u>
<u>PATIENT REGISTER</u>	[New Pt]
<u>PATIENT INFORMATION</u>	
Patient Id	geservice
Patient Name	ecmt
Weight (Lb)	111
	[Landmark]
Landmark	[>] [Nasion]
<u>PATIENT PROTOCOLS</u>	[Patient Position]
<u>PATIENT POSITION</u>	
Patient Position	[>] [Supine]
Patient Entry	[>] [Head First]
Coil	[...] [Surface] [geserviceRcv] [Accept]
<u>IMAGING PARAMETERS</u>	
Plane	[>] [Axial]
Mode	[>] [2D]
Pulse Seq	[...] [Spin Echo] [Accept]
Imaging Options	none
Psd Name	[ecmt1]
Protocol	no entry

It is not necessary to enter the remainder of scan protocol for this procedure.

REVISION HISTORY

REV	DATE	AUTHOR	PRIMARY REASONS FOR CHANGE
0	August 9, 2000	R. Liu	Update protocols from 8x version to OpenSpeed per bay and target validation
1	Jan 8, 2001	K. L-P	Updated Body SNR (o.13) to FOV=40 for 0.7T