State: RELEASE - Document is released and under formal Change Control. Changes are subject to the ECR/ECO Process.
See the GEMS ePDM System to determine the status of this document.

DCB (TX) — FIBER OPTIC P/N 2293408 — HSDCD (TX) J3 — HARNESS, FO-XMTR P/N 2260621-2 — HSDCCD TX PS +15VDC

FUSE BOX P/N 2331963-4: LINE, NEUTRAL, FUSE 20A, SW, LINE FILTER, LINE, NEUTRAL, 3A 3A 3A 3A 3A 3A 6A, HARNESS, FUSE INPUT P/N 2335331

J22 DAS FAN

HARNESS, FUSE BOX INPUT P/N 2260623-2

J5 — LEFT — DETECTOR CENTER — RIGHT — J4

DMB P/N 2285540 J1

JEDI 60 DC P/N 2371333-x

POWER UNIT: HVTANK (HV-, HV+); FAN — 120VAC SUPPLY P/N 2368968; J5 — KV/MA SIGNAL P/N 2377649; J3 — RCIB2 (JEDI - ORP) P/N 2368769; J7 — FAN/PUMP AND SMT ID P/N 2368768; PI FILTER (U1 +, U2 -) — HVDC PLUS P/N 2368449, HVDC MINUS P/N2368628

HARNESS, DAS/FUSE PWR SPLY P/N 2335330

P/N 2334457: 24VDCPS; -12VDCPS; -12VDCPS; P/N 2334459: +12VDCPS; +12VDCPS; P/N 2334459; -5VDCPS P/N 2334455; +5VDCPS P/N 2334455; 5VDCPS LOGIC P/N 2334458

HARNESS, MDAS DC POWER P/N 2335326

DAS CENTER BACKPLANE P/N 2318700 OR 2388723: J96, J104

HARNESS, LOGIC POWER P/N 2335328

J1 DHCB P/N 2285538 OR 5122006 J3 J2 — DCB/DHCB RS-232 P/N 2288183 — J33

DAS RIGHT BACKPLANE P/N 2371627: J30, J31, J32

HARNESS, DET/DHCB/COLL P/N 2358011

RFEFER TO 2371333-X SCH FOR GENERATOR INTERCONNECTS.

CATHODE HV CABLE P/N2201185-3 OR 5128609-5

ANODE HV CABLE P/N2368445 OR 5128609-3

AUX UNIT

HEMIT: HV+, ST+ — ANODE HV CABLE P/N2241313-2 OR 5128609

J6,7

H-POWER I/F BOARD P/N 2320973: J8, J7, J6, J2, J5, J1

ANODE, CATHODE — X-RAY TUBE — J1, J2

THERMAL/PRESS SWITCH P/N 2377474

120V SUPPLY P/N 2368770

FUSE 8A — TB1

120VAC TERMIN AL ASSY P/N 2368450

ORP P/N 2290254-4 OR 2389885-2: J5, J1, J8, J7, J6, J2

ENGINEERING USE ONLY: SERVICE TERMINAL

RCIB1 (ORP - DAS) P/N 2333309

LSCOM (ORP - S/R) P/N 2333308

HARNESS, LASER ALIGNMENT LIGHT P/N 2333306

LASER LIGHT POWER

SLIP RING: S/R#3: 120VAC LINE; S/R#4: 0VAC NEUTRAL; S/R#1: HVDC+; S/R#2: HVDC-; J2; S/R#3: 120VAC LINE; S/R#4: 0VAC NEUTRAL

COLLIMATOR P/N 2214768-4: J1, J2, J3 — TERMINATER P/N 2215315

HARNESS DET/DHCB/ORP/COLL P/N 2358011

CAN (COLLIMETER - DAS) P/N 2333314

ROTATING SIDE (16 SLICE)

REV.	ZONE		DATE	REV.BY
2	全	Remove detail generator interconnect references and adding new HV cables.	'05-Feb-02	有山
1	全	D-1 Cathode HV cable 変更に伴い書替え新図 106	'04－1－09	有山
		FZ2 - NPH -		

DR. 有山 | Elec eng

CH. 粂川

APP. 薩た | Jan/09/2004

H-POWER60 GANTRY INTERCONNECT(MDAS)

2378661SCH

PAGE 3

SIZE A3 技術部 MICRO

Form Y-13A-1

GE Yokogawa Medical Systems