


Detector Plenum Fan Control

RS232 to DHC
(Detector Heater
Controller on the
Plenum)

RS232 to
Backplane




Common Fan Controller (Rotating Side)

Isolator

Processor



All Temp.
Sensors
moved into
the same A/D

A/D




Thermistor 1 (In Plenum)
Thermistor 2 (In Plenum)
Thermistor 3 (In Plenum)
NC

Reference Diagram showing the Left and
Right when looking at Plenum




Fan Controller




To 48V Power
Supply

+48V Return




5 Fans on Plenum

Crtl Volt
Tach
GND
48V




Left Plenum Fan near 48V PS


Crtl Volt
Tach
GND
48V




Crtl Volt
Tach
GND
48V



Crtl Volt
Tach
GND
48V



Crtl Volt
Tach
GND
48V



Right Plenum Fan near ORP

Physical location of the connectors

