

De-Install Form CT

1850 Town and Country, Norco, CA 92860 USA
Main: 951.734.3737, Fax: 951-735-3373



OEM ID _____
Site Name _____
Site Contact _____
SMR Rep _____

Van ID _____
City, State _____
Phone Number _____
Inspection Date _____

TWO WEEKS BEFORE

Site Verifications

Start Date Start Time Helium Level Pressure

Rigger Verifications

Start Date Start Time Estimated time
☐ Pallet Jack
☐ Fork Lift for Gantry
☐ Tow truck for Gantry

Shipping Verifications

☐ U-line shipment
Arrive Date Arrive Time

☐ Equipment truck
Equipment delivery Location
Start Date Start Time Street Address
City State Zip Code

Reservations

☐ Flight Reservations
Airline
Confirmation Number
Flight departure Flight Return
☐ Car Reservations
Company
☐ Hotel Reservations
Hotel Name

Personel

Helper 1 Helper 2 Helper 3
Contact Contact Contact

Directions

☐ Site Directions ☐ Hotel Directions

De-Install Form

1850 Town and Country, Norco, CA 92860 USA
Main: 951.734.3737, Fax: 951-735-3373



OEM ID _____
Site Name _____
Site Contact _____
SMR Rep _____

Van ID _____
City, State _____
Phone Number _____
Inspection Date _____

TRAVEL DAY

Paper work

- | | |
|---|--|
| ☐ Pre site Inspection Report | ☐ Flight reservation information |
| ☐ Man-Power information | ☐ Car reservation information |
| ☐ system specs, extra Equipment list | ☐ Hotel reservation information, directions |
| ☐ De-install Procedure | ☐ Site Information, directions |

Tools Needed

- ☐ Engineer Needs to take following tools
- ☐ Basic tool set; Crescent wrench, pliers, screw drivers, Allen wenches

ARRIVAL DAY

- ☐ Introduce - Talk to Radiology Manager / Director; Get Emergency / Contact Number; After hours access.

Radiology Manager: Phone Number:

☐ After Hours Access Key After Hours Access Code:

☐ Get Tools and Supplies

☐ Uline shipment; pallets, bubble wrap, shrink wrap

Walk Site

- ☐ Identify Gantry exit path, any problems, steel beams, window cracks, height, width, corners, and or soft ground

☐ Verify with Rigger Gantry Opening Width; min ft Gantry Opening Height; min ft

- ☐ Identify Equipment exit path, doorways, raised floor, loading dock, pallet jack access, staging area

☐ Verify with Rigger Door Height; min in. Door Width; min in.

Check System

Gantry Serial number Gantry Type Rotation count

De-Install Form

1850 Town and Country, Norco, CA 92860 USA
Main: 951.734.3737, Fax: 951-735-3373



OEM ID _____ Van ID _____
Site Name _____ City, State _____
Site Contact _____ Phone Number _____
SMR Rep _____ Inspection Date _____

System Functional Checks and Test Scans

- ☐ Setup and run test scan
- ☐ Verify Options Against Site Evaluation - Discrepancies, Call Don or Russ
- ☐ Verify Equipment List Against Site Evaluation - Discrepancies Call Don or Russ
- ☐ Review Error Log - Go back two weeks, any major issues, Call Don or Russ
- ☐ Run save state
- ☐ Make sure to remove all CD's, MOD's, and DAT's
- ☐ Shut down the system, Turn off the breaker at the A1 for the CT PDU

De-Cable and Package

Disconnect cables and equipment

- ☐ Separate out any equipment and cables customer is keeping (injectors, stereos, ect....)
- ☐ Disconnect all cables from Console area
- ☐ Disconnect all cables from Scan Room
- ☐ Disconnect all cables from equipment room (if applicable)

Packaging

- ☐ Wrap cabinet corners with large bubble wrap and shrink wrap over them (remove keys and tape to back of cabinets)
- ☐ Try to keep cables separate in boxes and wrapped neatly
- ☐ make sure that monitors and computers are wrapped and secured with software and books

Rigging DAY

- ☐ Supervise Equipment Removal, Gantry Rig Out.
- ☐ Verify all pallets and equipment are in Shipping Container.
- ☐ Verify Gantry and table are Secure and lowered on to floor
- ☐ Verify everything is strapped down and checked before truck leaves

☐ Sign off Rigger

Delivery Location

Delivery Arrival Date

Delivery Arrival Time

☐ Sign off Shipper

Street Address

City

State

Zip Code