

MRI INSTALL COMPLETE REPORT

1290 Elm st. Suite B Ontario, Ca. 91761
Main: 951.734.3737, Fax: 951-735-3373



OEM ID	_____	Van ID	_____
Site Name	_____	City, State	_____
Site Contact	_____	Phone Number	_____
SMR Rep	_____	Inspection Date	_____

New Site Location

Site Name	_____	Contact:	_____
Street Address	_____	Phone Number:	_____
Shipping Address	_____	Loading Dock	_____
City	_____	State	_____
		Zip Code	_____

Magnet	_____	Serial #	_____	Helium	_____	Pres.	_____	Magnet Bore (cm)	_____
Last upgrade date	_____	Software level	_____						

Extras and Options (Specify who they are provided by)

Main Disconnect	_____	DC Lighting	_____	Stereo System	_____
Injector system	_____	Chiller unit	_____	Other	_____
Coils	_____				

S/C Shim Currents

AX1	_____	T1-1	_____	T2-1	_____
AX2	_____	T1-2	_____	T2-2	_____
AX3	_____	T1-3	_____	T2-3	_____
AX4	_____	T1-4	_____	T2-4	_____
AX5	_____	T1-5	_____	T2-5	_____
AX6	_____	T1-6	_____	T2-6	_____

Ramp Info

Magnet S/N _____

Current	_____	Center Frequency	_____
Config	_____	Field Strength	_____

Notes or Follow up items

N
o
t
e
s

MRI INSTALL COMPLETE REPORT

1290 Elm st. Suite B Ontario, Ca. 91761
Main: 951.734.3737, Fax: 951-735-3373



OEM ID _____
Site Name _____
Site Contact _____
SMR Rep _____

Van ID _____
City, State _____
Phone Number _____
Inspection Date _____

System Performance

Does System Start-up Without Errors? ☐ Yes ☐ No ☐ Action Required

Does Table Function (Up/Dn/In/Out/Alignment) ☐ Yes ☐ No ☐ Action Required

Does System Scan Head? ☐ Yes ☐ No ☐ Action Required

R1 R2 TG AX

Manuel Pre Scan
(Scan TR)

R1% R2%

Does System Scan Body? ☐ Yes ☐ No ☐ Action Required

R1 R2 TG AX

Manuel Pre Scan
(Scan TR)

R1% R2%

Does System Scan Multi-coil? ☐ Yes ☐ No ☐ Action Required

CTLTOP R1 R2 TG AX

RC1 R1/R2		RC2 R1/R2		RC3 R1/R2		RC4 R1/R2	
RC5 R1/R2		RC6 R1/R2		RC7 R1/R2		RC8 R1/R2	

CTLBOT R1 R2 TG AX

RC1 R1/R2		RC2 R1/R2		RC3 R1/R2		RC4 R1/R2	
RC5 R1/R2		RC6 R1/R2		RC7 R1/R2		RC8 R1/R2	

☐ Log Book and all documentation for calibrations left on site

Room Layout

Check console cable run, magnet exit path, equipment exit path and flooring