

GE Healthcare

Optima MR450w
Field Replaceable Units (FRU)

Technical Publication

© 2010 – 2011 General Electric Company
GE Medical Systems, a General Electric Company, going to market as GE Healthcare
3000 N. Grandview Boulevard
Waukesha, Wisconsin 53188
USA

www.gehealthcare.com



5670005
Rev 3

**COVER
BOTTOM
CROP LINE**

**COVER
BOTTOM
CROP LINE**

**COVER
FOLD
LINE** | **(2) CENTER STAPLES
FOR BINDING**

**COVER
RIGHT
CROP LINE**

**5/16" RADIUS
CORNER TRIM
TYPICAL**

**5/16" RADIUS
CORNER TRIM
TYPICAL**

**COVER
LEFT
CROP LINE**

SEND COMMENTS ABOUT THIS MANUAL TO:

GE MEDICAL SYSTEMS
MANAGER, SERVICE ENGINEERING
W-808
3200 N. GRANDVIEW BLVD.
WAUKESHA, WISCONSIN 53188
USA

This manual belongs to:

INTRODUCTION TO FIELD REPLACEABLE UNIT (FRU)

This direction is a guide to the FRU's for MR450w and MR450w GEM systems. Many of the FRU's are used on both systems. System specific FRU's will be designated as Non-GEM or GEM only.

Parts designated as a FRU are critical to maintain the safety and/or functionality of the system. FRU's are available on a Priority 1 basis from Service Parts Operation.

As a reference aid, the FRU's are listed under an upper-level part number where they are used. In most cases the upper-level part number is not a FRU. In some cases the parts listed under an upper-level assembly may have been nested in intermediate sub-assemblies which are not shown.

Where possible, FRU's which are part of Option Catalogs are shown under the area where they would be installed, rather than being listed in the Coils or Options Sections (SEE TABLE OF CONTENTS).

LIST OF EFFECTIVE PAGES


<u>PAGE</u>	<u>REV</u>	<u>PAGE</u>	<u>REV</u>	<u>PAGE</u>	<u>REV</u>
Cover	3	1 to 64	3		

[illegible]

TITLE	PAGE
INTRODUCTION TO FIELD REPLACEABLE UNIT(FRU)	1
LIST OF EFFECTIVE PAGES	1
TABLE OF CONTENTS	3
WARNING SIGNS & LABELS	4
MAGNET ENCLOSURE	
COMMON ENCLOSURE COMPONENTS	5
END BELL COMPONENTS	6
ENCLOSURE ELECTRONIC COMPONENTS	7
ENCLOSURE COMPONENTS FOR IN-ROOM DISPLAY	8
ENCLOSURE COMPONENTS FOR LED DISPLAY	9
ENCLOSURE & SHIM COMPONENTS	10
1.5T GRADIENT and RF BODY COILS	11
REAR PEDESTAL	15
LPCA	16
RF HUB	17
DOCK	18
SHIELD COOLER	19
MRU and MONITORS	20
PATIENT TABLE	22 thru 41
HEAD COIL and PATIENT HANDLING	42
HEC CABINET	45
PGR CABINET	46 & 47
CRONUS MODULE	48
CAM MODULE	49
PEN CABINET	50
PEN CABINET PENETRATION PANEL	51
SPW	52
MAIN DISCONNECT PANEL	54
OPERATOR WORKSPACE	55
GOC with HP 8400	56
GOC with HP Z400	57
SCIM AND KEYBOARD	58
SERVICE PHANTOMS & ACCESSORIES	59
CABLES	61
REVISION HISTORY	63

46-255325P1	Security Zone Warning Sign
46-255814P1	Magnetic Warning Label, ~2.5 x ~1.5
46-301671P1	Warning Placard
2289812	Authorized Personnel Only Warning Sign

This image shows a blank sheet of white paper with horizontal ruling lines. The lines are evenly spaced and run across the width of the page. There are no margins, text, or other markings on the paper.

ENCLOSURE COMPONENTS – MR450w only**MAGNET ENCLOSURE**

END BELL COMPONENTS


5328924 FRONT END BELL (non-GEM0)
5404650 FRONT END BELL (GEM, not shown)

5183293
MICROPHONE CABLE ASM

5344830
FRONT BORE LIGHT ASM

5344941
ENDBELL BORE LIGHT CABLE

5328925 REAR END BELL(non-GEM0)
5404651 REAR END BELL (GEM, not shown)



ENCLOSURE ELECTRONIC COMPONENTS**MAGNET FRU's**

- 46-260852G3** Emergency Evacuation Release Tool
5366120 Stud Replacement Kit

CABLE FRU's


- 2128699** Run 606, Magnet to ERU/MRU
5183293 Microphone Cable, Run M3345
5137339-6 E-Stop Cable
5180907 Run 3350, MAG-SRI-J10 to MAG-BORE, Bore Temp,
5304950-3 RF Cable Kit, Runs M1322 and M1323

SCAN ROOM FRU KIT's

- 5196918** MRU Assembly
2394952 Magnet Monitor 3
2219341 Ruthenium Oxide Sensor Amplifier
5155291-2 Magnet Service Platform


- 5328962**
 DVw PAC Asm
46-317903G5
 PPG CABLE
 OR

- 5341318-3**
 PAC-5 ASM
5393066-2
 FIBER OPTIC
 PPG CABLE
5194965-2
 Y-CABLE FOR
 POWER AND
 ONE-WIRE
 OR
5422053
 PAC2Bv ASM
 (GEM only)

**MAGNET ENCLOSURE**

ENCLOSURE COMPONENTS for IN-ROOM DISPLAY (IRD)


(MR450w only)




ENCLOSURE COMPONENTS for LED DISPLAY

(MR450w only)

5351731
LED DISPLAY COLLECTOR FRU

**MAGNET ENCLOSURE**

ENCLOSURE AND SHIM COMPONENTS

5329393
XRMW SHIM TRAY ASM

5342951
FIELD PASSIVE SHIM FRU KIT contains:

5343857	Shim, Silicon Stl 2.1mm thick, M4 Thru (40)
5343857-2	Shim, Silicon Stl 4.9mm thick, 4.5 Dia (40)
5343857-3	Shim, Silicon Stl .35mm thick, 4.5 Dia (250)
5343857-4	Shim, Silicon Stl Half .35mm thick, 4.5 Dia (120)
5343857-5	Shim, Silicon Stl 3.5mm thick, M4 Thru (40)
5343857-6	Shim, Silicon Stl 3.5mm thick, 4.5 Dia (50)
5343857-7	Shim, Silicon Stl 4.9mm thick, M4 Thru (50)
5385706-5	M4 x 10mm Brass flat Head Screw (75)
5212343	Passive Shim Square Nut (150)
5214479	Passive Shim Screw (100)

MR450w 1.5T GRADIENT AND RF BODY COIL**2208999-15**

1.5T RF BODY COIL Includes
BLUE & BLACK FRU CASE
AND SKID

5342791

XRMW GRADIENT COIL
Includes GRAY COIL CART
WITH CRADLE


5342778

COOLANT MANIFOLD COLLECTOR FRU
contains:

- 5339471 Coolant Supply Manifold – BLU
- 5339472 Coolant Supply Manifold – BLK
- 46-252065P155 Hose Clamps (25)
- 46-294167P91 Electrical Tape
- 5351851 Hose Clamp, 12mm x 20-32
- 5352839 Straight Hose Connector, 12mm
- 5350791-10 XRMW Manifold Tube
- 5340312 Reduction Barb Adaptor
- 5352243 Tooling-Temp Sensor Protector

5406803

Secondary Anti-Whistle Seal (front and rear)

**5342776**

XRMW GRADIENT MOUNTING FRU KIT contains:
(For mounting front and rear of coil)


- 5352690 Mounting Wedge (20)
- 5352691 Mounting Stop Block (4)
- 5352696 Mounting Pad (2)
- 5352697 Mounting Block Shim (16)
- 5363536-20 M10 x 30mm Hex Head Cap Screw (8)
- 5110638-65 M10 X 35mm Brass Hex Head CS (20)
- 5146169 M10 x 24mm Washer (28)
- 46-294151P11 Locktite 271

GRADIENT & RF BODY COILS

[illegible]


ENCLOSURE COMPONENTS – MR450w GEM only


- 5404650** FRONT END BELL ASM (See page 6)
5404651 REAR END BELL ASM (See page 6)
5415480 GEM ENCLOSURE MOUNTING KIT
 (Contains varieties of HSH Cap Screws and Flat Washers)
5418532 CABLE CONCEALMENT KIT
 REAR PEDESTAL COVERS (See page 14)
 IRD COMPONENTS (See page 14)


**MAGNET ENCLOSURE**

IRD & ENCLOSURE COMPONENTS

MR450w GEM only



REAR PEDESTAL

LPCA

RF HUB for 16/0, 32/0 and 32/32**6250162** (GEM only)

(2 Req'd)

RF HUB SWITCH BD (RFSB2)

OR**6250052** (non-GEM)

(16/0, 1 Req'd) & (32/0, 2 Req'd)

RF HUB SWITCH BD (RFSBv)

OR**5250162** (non-GEM)

(32/32, 2 Req'd)


RF HUB SWITCH BD (RFSB)


5250080-2


(16/0, 1 Req'd) & (32/0, 2 Req'd)

RF HUB MULTICOIL DRIVER BD
with 4 DRIVER MODULES**OR****5250080-3**

(32/32, 2 Req'd)

RF HUB MULTICOIL DRIVER BD
with 8 DRIVER MODULES**6250054**RF HUB POWER
DISTRIBUTION BD**5159474-32**
RF HUB CHASSIS AND BACKPLANE
with EMC GASKETING**5250098** (1.5T)
RF HUB
CONTROL BD**REAR PEDESTAL****17**

MAGNET DOCKING PARTS:

SHIELD COOLER & COLD HEAD

5314887
Sumitomo F-50
Shield Cooler Compressor

SHIELD COOLER FRU's Not Shown:


- 2154502-2** Run 621, Supply Gas Line, 65.6 ft or 20M long
- 2154505-2** Run 622, Return Gas Line, 65.6 ft or 20M long
- 5271637** Return Gas Line Adaptor, F-50 Compressor
- 5265687** Supply Gas Line Adaptor, 180 Degrees, F-50 Compressor
- 2155316-2** Run623, Compressor to Pen Panel Cable, 49.21 ft or 15M long
- 2218192** Run624, Cold Head Control Cable
- 2214062** Run 832, LCC Sleeve and Cold Head Cable
- 2259837** LCC Cold Head Control Cable Asm
- 2200835** Power Cable for Sumitomo Shield Cooler Compressor
- 5335002** Flow Control Adapter, used with F-50 Compressor only

MAGNET RUNDOWN UNIT (MRU) FRU'S:

5196918	MAGNET RUNDOWN UNIT (Primary)
5196919	MAGNET RUNDOWN UNIT (Remote)
5363631	SPARE FUSE KIT includes: 1.25A, 250V Fuse; 2.5A, 250V Fuse; 400ma, 120V Fuse; 200ma, 250V fuse
5199490	BATTERY PACK ASM
5180324	ADAPTER PLATE
5196921	REMOTE TO PEN PANEL CABLE
5196920	PRIMARY TO PEN PANEL CABLE
2128699	MAGNET TO MRU CABLE, RUN 606

OXYGEN MONITOR (O2) FRU'S:

5390676	OXYGEN MONITOR FRU – Contains all of the parts included with the O2 monitor, box, sensor and cables
5399042	OXYGEN MONITOR CALIBRATION KIT – Contains all of the parts needed to calibrate either the nw or old O2 monitor
5390678-3	REMOTE 5-YEAR SENSOR – Sensor for the ISA-60M O2 monitor

MAGNET MONITOR:

2394952
MAGMON3 (MON)

[illegible]

PATIENT TABLE

FRU part names, listed in numerical order, are found on pages 30 thru 37.

5368689 and **5367857** are for the oncology flat table only.

5368689
Non-GEM FLAT
TABLE INSERT

5367857
Non-GEM
TABLE FILLER

CRADLE DETAILS
Non-GEM, See page
NO TAG


GEM, See page NO TAG

TABLE TOP DETAILS
See pages 23 and
NO TAG

5173360 or 5173360-2
TABLE BELLOWS COVER


**TABLE LOWER
MECHANISMS DETAILS**
See pages 23 thru
NO TAG


PATIENT TABLE

PATIENT TABLE TOP & BOTTOM

PATIENT TABLE PEDAL ASM'S


FRU part names, listed in numerical order, are found on pages 30 thru 37.



PATIENT TABLE PEDAL ASM'S


PATIENT TABLE PEDAL ASM

FRU part names, listed in numerical order, are found on pages 30 thru 37.




PATIENT TABLE LIFT & FRAME


FRU part names, listed in numerical order, are found on pages 30 thru 37.


**PATIENT TABLE**

PATIENT TABLE DOCK AREA & CASTORS


FRU part names, listed in numerical order, are found on pages 30 thru 37.



PATIENT TABLE CRADLE, CABLE TRACKS, & COVERS**PATIENT TABLE**

PATIENT TABLE TOP FRU's


GEM PATIENT TABLE PEDAL ASM



FRU part names, listed in numerical order, are found on pages 34 thru 41.


GEM CRADLE ASM

FRU part names, listed in numerical order, are found on pages 34 thru 41.



GEM CRADLE PANELS AND PADS

FRU part names, listed in numerical order, are found on pages 34 thru 41.



2409533 Rop3 Flat Insert Asm (Oncology Option)
5410549 Rop3 Locating Tab Kit (Oncology Option)

PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
2119594	Anti-seize Lubricant; Permatex 133A
2297496	Spacer, Alum Bar 1/2~ X 3/4~ X 3.5~
2297499	Spring Drum 1.96~Dia, 1.2~W Aluminum
2297515	Round Spacer, 1 IN OD, .390 ID
2386538	Velcro Strap Type 889
5136198	Caster Wheel
5142223	Up Sensor Cable
5142223~2	Uplimit Dock
5142784	Reciprocating Pump Asm
5142784~2	Reciprocating Pump Asm – High Pressure GEM
5142785~2	Cylinder Asm
5147698~3	Clevis Pin
5147699	Rod Link Connector
5147699~2	Rod Link Connector
5147700	Shaft Rocker
5147700~2	Shaft Rocker
5147701	Rocker Clamp
5151137~2	Down Pedal Asm
5151139~2	Pump Pedal Asm
5151142~2	Undock Pedal Asm
5151146~2	Dock Pedal Asm
5154883	Lateral Wheel Housing
5165231	Block Hook Undock
5167237	Cradle Asm
5167237~2	16 Channel Cradle Asm
5167918	Link Undock
5168057	Front LH Side Bumper
5168057~2	Front RH Side Bumper
5168193	Arm Board Bumper
5168326	Rear Left Hand Side Bumper
5168326~2	Rear Right Hand Side Bumper
5168363~2	Cotton Rod
5168363~3	Cotton Rod
5168456	Compression Spring 10mm, 4.5mm ID, 18mm L
5168584~2	Rod Cap
5172109	Pin Undock Link
5172109~2	Pin Undock Link
5172715~2	Pedal Connector Asm
5172732	Front Latch Asm
5172856	Cradle Stop Polyurethane

PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
5179600-2	Front RH T & G Asm
5180380	Wheel Housing Asm
5173088	Rear Armboard Hinge Asm
5173088-2	Front Armboard Hinge Asm
5173360	Table Bellows Cover
5173360-2	Table Bellows Cover – GEM
5173376-2	Pedal, Brake Steer
5173590	Front Base Cover
5173863	Rear Base Cover
5174010	LH Handle Hinge
5174186	Armboard Asm
5174245	Bracket
5174518	Link Bar Connector Pedal
5176295	Pin
5176314	Link Connector Undock
5176477	Shaft Connector Rocker
5176477-2	Shaft Connector Rocker – GEM
5176644	Link Down
5177236-2	Lever
5177506-2	Release Bar
5177611	Binding Head Slotted Screw, IN 10-32x0.5 SS 18-8
5177750	Pedal Stop Rubber
5178055	Connector Cover
5178055-2	Connector Cover
5178063	Shaft-linkage-right
5178101	Link, Left Front
5178144	Link, Right Front
5178145	Hexagonal Socket Head Screw, IN 8-32 X 3/8, Brass
5178145-2	Hexagonal Socket Head Screw, IN 8-32 X 1/2, Brass
5178210	Shaft-linkage-left
5178255	Link, left Rear
5178291	Cover Hinge
5178320	Connecting Rod
5178483	Link, Right Rear
5178611	Block-mounting
5178635	Extension-shaft
5178738	Pin-actuator
5178915	Link-reversing
5179251	Rear LH T & G Asm
5179251-2	Rear RH T & G Asm
5179376	Arm Board T & G Asm

PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
5392173	Sled Weldment GEM
5392175	Linkage Arm Shaft GEM
5394052	Cradle Mid-Section GEM
5394890	Base Plate Slider
5395819	Hard Top GEM
5395821	Strap Rail
5179376-2	Arm Board T & G Asm
5179600	Front LH T & G Asm
5181517	1.5T Table Cable Track
5181517-2	1.5T Table Cable Track
5181751	Reel
5183907	Decal Pedal Set
5184060	Lever
5189212	ODU Harness Cable
5189213	Flat Head Slotted Screw IN 1/4-20 x 3/4 Brass
5191326	Bushing
5193793	Harness Jumper Touch and GOC
5198459	Connector Return Spring
5198829	Shaft Spring Connection
5199333	Pedal Top Rubber
5199333-2	Pedal Top Rubber
5213295	Guide Strip
5213295-2	Guide Strip
5220240	Shoulder Screw, IN 1/4-20 X 3/8, SS18-8
5220240-2	Shoulder Screw, IN 1/4-20 X 5/8, SS 18-8
5220240-3	Shoulder Screw, IN 1/4-20 X 1/2, SS 18-8
5220579	Nylon Washer
5220580	Rubber Washer
5263096	Bearing IGUS MCI-12-02
5263182	Rear Cradle Cover
5306407	Table Bumper FRU Kit
5338208	Wheel, Caster
5338208-2	Wheel, Caster – GEM
5341331	Hydraulic Rotary Pump Asm
5341568	Hydraulic Rotary Pump Upgrade Kit
5344577	Service Filler Pad
5367814	Connecting Rod Block
5390324-6	Table-Side Docking Connector Carrier Frame Only – GEM
5390489-3	1.5T "Touch & Go" PA GEM Coil Table Harness
5391126	Shaft Link Actuator – Left
5391126-2	Shaft Link Actuator – Right
5392170	Rocker Block GEM


PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
5392171	Rocker Clamp Support GEM
5392172	Slide Sled Block GEM
5395169	Slide Plate
5395509	P1 Only Cable Track Asm
5395828	Adaptor Panel Asm GEM
5395829	Filler Panel Asm GEM
5395949	PA Hard Top Pad GEM
5395951	Filler Hard Top pad GEM
5396470-2	Cradle GEM coil
5396758	Front Base Cover GEM
5396759	Side Base Cover GEM
5396760	Rear Base Cover GEM
5396877	Cable Guide Bushing
5397223	Cradle Front Section GEM
5397224	Cradle Rear Section GEM
5397342	Cradle Mid Rib GEM
5397813	Cradle End Rib GEM
5398699	Up-Limint Switch Mounting Bracket
5398700	Up-Limit Switch Adapter
5398701	Up-Limit Switch Link
5398898	Cradle Handle Bar GEM
5398900	Cradle Support Block GEM
5398901	Cradle Connecting Block GEM
5399212	Cradle Adapter Plate GEM
5399408	Cradle Spacer GEM
5400366	Egress upper Cover GEM
5400367	Egress Lower Cover GEM
5400489	Upper Front Right Cover GEM
5400489-2	Upper Front Left Cover GEM
5400492	Upper Front Center Cover GEM
5400493	Upper Rear Right Cover GEM
5400493-2	Upper Rear Left Cover GEM
5400495	Upper Rear Center Cover GEM
5400956	1.5T PA Mux Box Asm with Cable GEM
5401787	16W16 and 25 Pin Sub-D Spacer Bracket GEM
5401788	16W16 and 8W8 Spacer Bracket GEM
5405298	Front Cable Shield
5405847	T-Block Slider
5406112	Cable Clamp
5406819	P-Port Cover GEM
5407124	P2 Table Docking Connector Cable FRU w/Replacement Track Lids
5407124-2	P4 Table Docking Connector Cable FRU w/Replacement Track Lids

PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
5407279	PA Filler GEM
5407280	60 Micron Stainless Steel Body Oil Filter
5407808	Handle Cover
5409760	Cover Mounting Block Insert – Right
5409760–2	Cover Mounting Block Insert – Left
5410992	Front Cover Mounting Block GEM
5411051	Rear Cover Mounting Block GEM
5411702	Pedal Pad Replacement Kit GEM
5412152	Hex Socket Head Shoulder Screw
5412197	Support
5412198	Bell Crank
5412974	Pump Reverse Lever GEM
5413479	Spring – Free Length 1.50 Inch, 1.19 INLB Rate
5414813	Screw – #10–32 x 1 in SS PH
5417597	Screw – #10–32 x 1.5 in SS PH
1000904P41	0.266 ID, 0.75 OD, .064 THK Brass Washer
1000904P61	ID 0.391 OD 0.750 THK .031 Phosphor Bronze Washer
1000904P419	ID 0.328, OD 0.5, THK .016, Washer
1000904P425	0.2 ID, 0.5 OD, .04 THK Washer
1000904P460	0.201 ID, 0.375 OD, .05 Thick, Brass Washer
1000904P477	ID 0.531 OD 1.062 THK .095 ALUM 6061 Washer
1000904P478	ID 0.531 OD 1.062 THK .095 Brass
1000904P498	0.635 ID, 1~ OD, .125 Thick Aluminum Washer
1000904P502	0.406 ID, 1~ OD, .062 THK Nylon Washer
46–136342P23	Bearing Oil–imp Bronze, Sleeve, 0.365 ID, .503 OD. .313 Lg
46–136342P51	Bearing Oil–imp Brnz, Sleeve, 0.377 ID, .503 OD. .875 Lg (non–GEM)
46–136342P60	Bearing Flange: OD 0.500~, THK .046~ OIL–IMP Bronze
46–136342P128	Bearing Flange: OD 0.625~, THK .3625~ OIL–IMP Bronze
46–136342P131	Bearing Oil–imp Bronze, Sleeve, 0.3155 ID, .4395 OD. .25 Lg
46–136353P105	Spring–inch Free LG 2.000 RATE LB/IN 2.1 (non–GEM)
46–136393P147	Groove Pin, .156 OD, .5 Lg
46–170024P1	Screw–inch 10–32 SLOT HD #303 SST
46–170024P12	Screw 8–32 SLOT HD #303 SST
46–170073P5	.002 Thk x 2.0 Wide x 60 Yd Roll
46–170490P38	Spring–inch Extension 2.5 Inches .25 Free Lgth 2.500 SST 316
46–170490P59	Spring–inch Extension 2.75 Inches .2 Free Lgth 2.75 SST
46–170490P60	Spring–inch Extension 4.5 Inches .5 Free Lgth 4.50 SST
46–170686P1	Loctite 242, 0.5CC Tube
46–208560P96	Screw N7715010 (Passivated)
46–208563P88	Hex Socket Set Screw SST 316, 10–32 x 5/16
46–208563P90	Hex Socket Set Screw SST 316, 8–32 x 3/16

PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
46-208601P25	3/8- Screw, Internal Tooth Phos Brz
46-208695P75	Heyco No.2210 SB-1375-16 Black
46-208695P89	Heyco Part No.2017 SB-375-281
46-208735P58	Split Lengthwise .187 Dia X 1.000 Lg .040 Wall 420 SST
46-208735P76	Pin, Spring, Dowel
46-208737P3	Hex Socket Head Cap Screw 6-32 x 3/8, SST
46-208737P12	Hex Socket Head Cap Screw,10-32 X .50, SST
46-208737P13	Hex Socket Head Cap Screw,10-32 X 5/8, SST
46-208737P20	Hex Socket Head Cap Screw,10-32 X 1.0, SST
46-208737P27	Hex Socket Head Cap Screw,1/4-20 X 3/4, SST
46-208737P28	Hex Socket Head Cap Screw 10-32 X 3/4 SST
46-208737P47	Hex Socket Head Cap Screw, 3/8-16 x 3/4, SST
46-208746P9	Cable Tie Panduit TM3510C
46-208747P2	Adhesive Mount - Cable Tie
46-208758P1	3.62~ X .094~ Cable Tie
46-208758P5	Self Locking Cable Tie
46-208910P4	Screw 18-8 SST (Passivated)
46-208913P9	Brass Slotted Flt Hd Scr 10-32x3/4
46-208913P24	Brass Slotted Flt Hd Scr 1/4-20x1/2
46-208932P7	Hex Head Cap Screw, 3/8-16 x 7/8, SST Passivated
46-208932P10	Hex Head Cap Screw, 1/4-20 X .50, SST Passivated
46-208932P16	Hex Head Cap Screw, 1/4-20 X 2.0, SST Passivated
46-208935P5	6-32, .313 Hex X .109 Thk, Brass
46-208938P5	10-32, 3/8- Hex X .130 Thk, 18-8 SST
46-208945P3	1/4-28, .438 Hex X .219 Thk, Brass
46-208946P4	Hexgon Type 18-8 SST Jam Nut (non-GEM)
46-220184P18	Round Head Slotted Brass Screw, 6-32 x 1 Lg
46-220184P26	Round Head Slotted Brass Screw, 8-32 x 1/2 Lg
46-220196P3	Washer, #10 Screw Size (0.190), SST
46-220270P9	E-Clip 0.750 Shaft Dia, BE-CU
46-220277P18	E-Clip 0.500 Shaft Dia, BE-CU
46-220277P19	E-Clip 0.250 Shaft Dia, BE-CU
46-220277P22	E-Clip 0.312 Shaft Dia, BE-CU
46-220277P23	E-Clip 0.375 Shaft Dia, BE-CU
46-220277P24	E-Clip 0.625 Shaft Dia, BE-CU
46-220277P27	E-Clip 0.188 Shaft Dia, BE-CU
46-221304P7	0.188 DIA X 0.65 Length, 303 SST
46-221560P1	Hex Socket Oval Point Set Screw 1/4-20 x 3/4, SST Passivated
46-221560P3	Hex Socket Oval Point Set Screw 8-32 x 3/8, SST Passivated
46-221560P4	Hex Socket Oval Point Set Screw 10-32 x 1/2, SST Passivated
46-221806P2	Pan Head Phillips Screw, 10-32 x 1/2 SST Passivated

PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
46-221828P4	Cotter Pin
46-221922P7	Hex Head Screw, 10-32 x 1-1/4 Lg SST Passivated
46-230006P2	Hydraulic Lubricating Oil Red Colored Oil, 32oz.
46-230434P2	0.047 Wire Dia X 0.594 Lg 302 SST Hairpin Type
46-230947P1	.312~ I.D., .5~ O.D., .093~ Wall.tygon F-4040-A Tubing
46-230957P1	Loctite 680, 50ml Bottle
46-243178P3	Bearing, ID 0.500~, OD 1.375~, Width 0.375~ Open, Turcite Plastic
46-244996P1	Bushing, Dock Guide Roller, 1~ Dia. Nylatron
46-256047P1	Mobiltemp Shc 32 14 Oz Cartridge
46-256049P16	Hole Dia 1.000, Head Dia 1.2187 O~all Ht .4062 Nylon White
46-256080P1	Cable,Push Pull,Nylon,Molybdenum Disulfite Impregnated
46-256081P6	3/8~ Screw Size, Hi-collar
46-260789P88	Screw,.375-16 X 1.25 Soc Hd Capscrew 304 Stainless Steel
46-256193P9	3/8-24 Hex 0.564 Ht .453 Ring Type Nylon Locking Element
46-260797P82	Screw Binding Head
46-265887P2	Upper Scissors Roller Shaft, 3/8 Dia
46-265887P6	Lower Scissors Roller Shaft, 3/8 Dia
46-265888P1	Roller, 2 OD, .50 ID, .31 Lg
46-265892P1	Dock Hook, 3/8-24 Thd, SST, Hardened
46-265895P1	Pin, .187 OD, 4.23 Lg SST
46-265896P1	Brass Bushing, .837 ID
46-265897P1	Link Lever, Polycarbonate Plastic
46-265898P1	Link Lever, Polycarbonate Plastic
46-265899P1	6.7~ Long, .081 SST Spring Wire
46-265905P1	Foot Pedal Pad, Rubber
46-265906P1	Docking Guide Roller Shaft
46-265907P1	Casting
46-265908P1	Shaft, .625 OD, 6.56 Lg SST
46-271028P1	Hydraulic Cylinder To Table Top
46-271127P1	Molded Plastic W/Cap & Rubber Gasket
46-271129P1	Molded Polycarbonate
46-271134P1	1~ & .562~OD, .316~ID X.5~, Brass
46-271161P1	.375~& .312~ DIA, 1.17~LG, SST
46-271162P1	2.12~X1.19~, .5~ Thk Alum Bar
46-271178P1	Secondary Stop Pin, Brass (non-GEM)
46-271179P1	Secondary Stop Pin Housing, Delrin (non-GEM)
46-271192P1	.312~DIA, 6.45~LG, .125 Keyway, SST
46-271194P1	.312~DIA 6.88~ Long, SST
46-271195P2	Shaft, Hook Stop, .312 OD, 7.75 Lg, SST
46-271195P4	Shaft, Link Pivot, .312 OD, 0.89 Lg \SST
46-271195P7	Shaft, Dock Pedal, .312OD, 1.38 Lg SST


PATIENT TABLE PARTS <i>(NOTE: Patient table parts are listed in numerical order)</i>	
46-271195P8	Shaft, Down, .312 OD, 1.46 Lg SST
46-271195P11	Shaft, Dock Hook Pivot, .312 OD, 2.97 Lg SST
46-271232P1	Slide, White Delrin
46-271235P2	Key, 3/16~ SQ, .44 Long, Brass
46-271235P3	Key, 3/16~ SQ, .62 Long, Brass
46-271408P1	I.V. Bushing, 1.38~OD, .635~ Hole, 2.5~LG, Alum
46-271438P1	1~OD, .66~ID, White Velcro Hooks
46-271510P1	Shaft For Pedals
46-271511P1	Cable, 71~ (Trim Ar); Swaged Ball One End (non-GEM)
46-271512P1	Adjustment Screw, 1/4-28 x 1.38 Lg
46-271551G1	Detachable, Telescoping IV Support
46-271554P1	Roller Wear Strip, 1.93 x 8.63 SST
46-271555P1	BNC Right Angle Adaptor
46-271562P1	Spacer, .39 ID, 1.00 OD, .40 Lg Nylon
46-271576G1	Mechanical Linkage
46-271587P1	2.19~ X 1.5~ Gusset, .081 Alum
46-271913P1	Bell Crank, Brass
46-271914P1	Pivot Stud, 1/4-28 Thread, .194 Hole, Brass
46-271946P1	Oil Filter, Reservoir Internal Cap Poly Foam
46-271966P1	Cylinder Mounting Pin .437~OD X 4.34
46-282216P1	Cable Clevis, 1~X.44~X.38~ .194 Hole, Brass
46-282440P1	Bushing, .442~OD .314~ID, .25~Thk, Bronze
46-307077P1	Flat Head Solid Rivet, 3/16 DIA X 3/8 3/8 Dia Head 18-8 SST
46-307878G1	Hydraulic Oil Filter And Gasket
46-317179P1	Squeeze-shut Ratchet-latch, 1/2~ Dia

SPLIT HEAD COIL:**Note: The following parts are not shown:****5328732 (1.5T) QHTR Circuit Board Asm Kit****5267144** Head Coil Cushion – Flat**5271434** Head Coil Cushion – Head**2355094** Replacement Spring Contacts Kit**PATIENT HANDLING FRU KIT's:****46-306574P7** 0.5 x 9 x 14.5 Gray Rectangular Pad, Nylon Covered**46-306574P8** 1.0 x 9 x 14.5 Gray Rectangular Pad, Nylon Covered**46-306574P9** 0.5 x 9 x 36 Gray Rectangular Pad, Nylon Covered**46-306574P10** 1.0 x 9 x 36 Gray Rectangular Pad, Nylon Covered**46-317758P1** Pneumatic Patient Alert System**5122500** ECG Lead Kit

GEM COIL FRU'S

<u>5391440</u>	HNU Mirror
<u>5391446</u>	16 Channel AA
<u>5391451</u>	HNU Tilt
<u>5391452</u>	PV Ramp Pad
<u>5397413</u>	HNU Head Pad
<u>5413906</u>	Legacy Head Coils Transition Pad
<u>5415766</u>	Rectangular Patient Pad, 0.88 x 3.00 x 5.00
<u>5415767</u>	Triangular Patient Pad, 3.00 x 10.00 x 10.00
<u>5415769</u>	Knee Rest Patient Pad
<u>5404900</u>	Coil Phantom Foam Support
<u>5405748</u>	PA Phantom Positioner
<u>5405749</u>	PV Phantom Positioner
<u>5405750</u>	HNU Phantom Positioner
<u>5407408</u>	HNU Coil
<u>5407409</u>	HNU System Cable
<u>5407411</u>	PA Coil
<u>5407412</u>	PV Coil
<u>5407413</u>	PV System Cable
<u>5407414</u>	PV Latch Asm
<u>5408884</u>	Spare Parts Kit
<u>5415766-2</u>	Rectangular Patient Pad, 0.50 x 9.00 x 36.00
<u>5415766-3</u>	Rectangular Patient Pad, 1.00 x 3.00 x 14.00
<u>5415766-4</u>	Rectangular Patient Pad, 0.50 x 9.00 x 19.75
<u>5415767-2</u>	Triangular Patient Pad, 1.75 x 4.00 x 14.50
<u>5430000-2</u>	Flex Coil 16-L Array, 1.5T Receive Only
<u>5430000-3</u>	Flex Coil 16-M Array, 1.5T Receive Only
<u>5430000-4</u>	Flex Coil 16-S Array, 1.5T Receive Only
<u>5430000-5</u>	Flex Interface 16-Channel Fixed, 1.5T P-Connector
<u>5430000-7</u>	Flex Knee Stabilization for Flat MRI Tables
<u>5430000-9</u>	Flex General Purpose Strap and Interface Cover Kit
<u>5430000-10</u>	Flex Coil 16-L Covering Replacement Kit
<u>5430000-11</u>	Flex Coil 16-M Covering Replacement Kit
<u>5430000-12</u>	Flex Coil 16-S Covering Replacement Kit
<u>5430000-13</u>	Flex Stabilization Device for Curved and Flat Tables
<u>5430000-14</u>	Flex System Cable Take Up Pad
<u>5430000-15</u>	Flex 16 Channel System Cable P-Connector

[illegible]

HEC CABINET**HEC (Heat Exchanger Cabinet) FRU KIT's**

5264701	Deionization Filter Cartridges Kit
5156315	3hp Blower Kit
5156316	Cryogen Compressor Flow Meter Kit
5156318	Mixing Valve Kit for Gradient Coil and PGR Cooling Loops
5156319	Power Box Kit (Contains breakers, red LED digital displays)
5156320	Mixing Valve Kit for Chilled Air Blower
5156321	5hp Pump Kit
5262029	Equipment Room Hoist Support Tube
5266827	Signal Box Fuse Kit
5267643	Liquid Pressure Sensor Kit
5267644	Level Switch Kit
5267649	Resistivity Sensor Kit
5268665	Liquid Temperature Sensor Kit
5269683	Coolant Removal Kit (Contains pump, hoses, flask, etc.)
5269685	Cryogen Compressor Filter Kit (Contains both filter types)
5269689	Air Temperature Sensor Kit
5269691	Signal Box Kit (Contains LCD display)
5357155	Cryo Compressor Flare Fitting FRU Kit
5372868	Ball Valve LOTO Device
5376555	Cryo Supply Header (Contains single valve, filter; for use with all HEC's)
5400660	Facility Tee with QD (For use on PE or GC Heat Exchanger)
5402967	Pump Suction Hose FRU


HEC COOLING SYSTEM FRU's

5267596	Liquid Cooling Valve Kit (Contains ball valves and check valves)
5174313-4	Gradient Coil Coolant, 4 Gallons – non-copper corrosive, algae resistant

PGR CABINET(5148810 Series)
(5373011 Series)**5150275**
DC CURRENT
TRANSDUCER**46-208718P5**
1/2~ ID LOOP**5157422**
XGA SLIDE
RAILS**5183212**
FAN TRAY
REMOVABLE**5183573**
XGD FAN
BOX ASSY**5157426**
SLIDE RAILS**5194812**
XPS MOD
RETAINING
BRKT**5370799**
PGR BASE
CABINET
w/PDU FRU**5215252**
POWER
DIST.
BLOCK**5184566-2**
XGD AMPLIFIER
FRU (3)**5183329-2**
SFP F/O XCEIVER
MODULEXGD CRONUS MODULE
See Page 48**5196748**
HOIST BRACKET**5260695**
PGR POWER
DISTRIBUTION
MODULE
(5148810 Series)
(PGR Cab Only)**5371251**
POWER STRIP
(5373011 Series)
(PGR Cab Only)IMAGE
COMPUTE
NODE
See Page 47**5139592**
GIGABIT
ETHERNET
SWITCH**5161071**
XGD AUX
PWB ENCL
ASM
46-312346P4
NUT M6 X 1CAM
MODULE
See Page 49**5266756**
HOIST
BEAM**2351573**
1.5T SRFDII RF Amp FRU
(Includes Hoist Bracket)**5148804-2**
XGD POWER
SUPPLY FRU (3)**5183329-2**
SFP F/O XCEIVER
MODULE**XGD PDU FRU'S (Not Shown)****5343114-2** XGD PDU Control Board Asm – AUX (5373011 Series)**2301200-3** TVSS Module**5343114-4** 3-Pole, 65 Amp Contactor (5373011 Series)**5140621-2** XGD PDU Control Board Asm (5148810 Series)**5140621-3** 3 Pole Power Contactor for E-Stop**5140621-4** 48 Volt 1000W Power Supply**5140621-5** Fan with Tach Output**OTHER CABINET FRU'S (Not Shown)****5266125** Fluid Sensor Kit**46-306565P3** Aluminum Mesh Air filter, 16 x 10**2109873-25** M16 X 16mm L Pan Head Screw**5263794** Lifting Bracket – XGA and XPS FRU

PGR CABINET

5168052
 7-16 FEMALE TO 7-16
 FEMALE BULKHEAD
5263113
 ICN 1 TO ICN 2
 CONNECT CABLE
5257687-2
 ICN 2 POWER CABLE
5127452-6 or **5337894**
 SUN 4100 ICN SUN 4170 ICN
 (16 Channel)
 or
5337894-2
 SUN 4170 ICN
 (32 Channel)
5337894-5
 4170 HDD
5183329-2
 SFP F/O TRANSCEIVER MODULE
 or
5911000-3
 GEN 4.1 DV 32 Ch 72GB ICN (GEM)
2379186
 TERM SERVER POWER SUPPLY/CORD

**REAR VIEW****4170 ICN DV Gen 3 FRU's**

5411351-2 ICN Conversion Kit
5374808 Hardware FRU Kit

Internal Transmit Cable FRU's

5167234-3 1.5T Body TX cable, LMR600
5167234-4 1.5T Body TX cable, LMR600
5192690-3 1.5T Head TX cable, LMR600
5192690-4 1.5T Head TX cable, LMR600

5304065 (5148810 Series) & 5388782 (5373011 Series),**Internal Signal Cable FRU Kit contains:**

5178931	5165870	5165970	5176566	5182238	5182238-2
5182238-3	5182238-4	5184303	5257682	5257683	5257683-2
5257684	5257685	5257685-2	5257685-3	5257685-4	5257685-5
5257685-6	5257685-10	5167344	5167005	5257680	5257681
5166354	5166222-13	5166222-18	5257685-11	5257685-12	5257685-13
5176990	5176262	5176262-2	5176262-3		
5174739 (5148810 Series Only)			5376191 (5373011 Series Only)		


5304386, Internal Fiber Optic Cable FRU Kit contains:



5182888	5182317	5182317-2	5182317-3	5182317-4	5182317-5
5182317-6	5257686				


5304635, Internal Power Cable FRU Kit contains:

5173399	5164904	5165800	5172391	5174242	5174242-2
5183977					

CRONUS MODULE



CAM MODULE (FRONT)**CAM MODULE (REAR)**

PEN CABINET**INTERNAL CABLE FRU KIT'S****5308001-2, Power Cable FRU Kit contains:**


5167143	5167143-2	5172838	5155030	5172849	5171737
5154994					

5308001-3, RF Cable FRU Kit contains:


5181464	5166512-3	5166512-2	5172955-2	5172955	5166749
5166749-9	5172679-10	5172679-11	5172679		

5308001-4, Power Distribution Cable FRU Kit contains:


2322434	5168641	5168043	5168317	5167882	5166461
---------	---------	---------	---------	---------	---------


PEN CABINET PENTRATION PANEL**REAR VIEW****INTERNAL CABLE FRU KIT'S (Continued from previous page)****5308001-5, Signal Cable FRU Kit contains:**

5165866	5166997	5169443	5172726	5166354-2	5171626
5170521	5172868	5167365	5171583	5172959	5169381
5168461	5168597	5172812	5172595-2	5169345	5175777
5172595					

SPW (Secondary PEN Wall)

[illegible]

MAIN DISCONNECT PANEL (MDP)**MDP FRU's Not Shown****5161287-2** 1.5A, 380/480V Fuse**5161287-3** 60A, 380/480V Fuse**5161287-4** Green PB Lens**5316807** PULS ML 15-121 Power Supply**5316808** BVUPS 12v FRU Asm**5340703-2** 1A 500V Fast Acting Fuse**5323547** Power Sonic 12V 7.0AH Battery

OPERATOR WORKSPACE**OPERATOR WORKSPACE FRU's Not Shown**

5165713 DVD Media for Linux

2156158-80 DV Apps Software Collector M3 (22.x only)

2156158-87 DV Apps Software Collector M4 (22.x only)

GOC COMPUTER CABINET WITH HP8400;**5267342**

GOC CableCollector
(See details below)

5133768-2
FIBER OPTIC
CONVERTER
BOX

5139480-2
8-PORT SWITCH

5262750
PDU ASM

2395043 or 2395043-2
GOC AUDIO ASM

5183547-30

HOSTPC

5183547-15
73GB SAS HARD
DRIVE FOR PC

5183547-20
SINGLE 1GB DIMM

5183547-36
SINGLE 2GB DIMM

5183547-22
DVD-ROM DRIVE

5184500
USB to RS422
CONVERTER
(Connected to
PC USB Port)

46-271382P6
1 POLE, 15 AMP
CIRCUIT BREAKER

5265358


One-Wire Collector
(See details below)

5267342 GOC Cable Collector FRU consists of;

- 5270774** Host Power Cable
- 5270803** Main LCD Monitor Power Cable
- 5263413** Run E3046 Main Monitor DVI-D Cable
- 5264533** Run E3047 SCIM-GOCOA Cable
- 5267826** Audio Cable Asm
- 5270723** GOCOA-PDU DC Power Cable
- 5267389** Run E3045 Modem Cable
- 5272839** Host Hub Ethernet Cable
- 5272839-2** Hub-Front Port Ethernet Cable
- 5264521** Run E3048 USB Extension Cable
- 5169105** Run E3016 Y6 Cable
- 5265059** Host to IRDC DVI-D Cable
- 5265784** Host to IRDC USB Cable
- 5270246** IRDC Power Cable

5265358 One-Wire Collector FRU consists of:

Humidity & Temperature iButton, iButton Holder, USB to RS422 Converter, USB and Sensor cables, Run E3019, and Dual Locking Tape

GOC COMPUTER CABINET WITH HP Z400;

5364704 One-Wire Collector FRU consists of:
 Humidity & Temperature iButton, iButton Holder, USB to RS422 Converter, USB and
 Sensor cables, Run E3019, and Dual Locking Tape


* The 5370415 Z400 Host computer only compatible with **2156158-80** (M3) or
2156158-87 (M4) DV Apps software collector.

GOC COMPUTER CABINET WITH HP Z400 (CONTINUED):

5389531	GOC Conversion Kit to DV with Z400 FRU consists of;
5389626	GOC Host PC Mounting Asm for Conversion Kit
5364484	GOC PC Holder
5272817	Screw for Host (2)
5389627	GOC Host PC Cover Asm for Conversion Kit
5330295-2	Host to USB–Serial Converter USB Cable
5330301	Audio Cable Asm
5330309	Host–Hub Ethernet Cable
5192248-4	Dust Filter
U0202AC	Clamp CV 150 (10)

SCIM & KEYBOARD:**NOTE:**

If Keyboard is to be replaced, also order **46-317475P2**, 2 inch x 4 inch adhesive backed velcro to attach to bottom of new keyboard.

**KEYBOARDS:**

5143798-7	Standard USB Keyboard – English
5143798-8	Standard USB Keyboard – French
5143798-9	Standard USB Keyboard – German
5143798-10	Standard USB Keyboard – Spanish
5143798-11	Standard USB Keyboard – Italian
5143798-12	Standard USB Keyboard – Portuguese
5143798-13	Standard USB Keyboard – Scandinavian
5183547-46	Standard USB Keyboard – Danish
5183547-47	Standard USB Keyboard – Dutch
5183547-48	Standard USB Keyboard – Norwegian

SCIM OVERLAYS:

5148389-2	MR SCIM – Black with English Overlay
5148389-9	MR SCIM – Black Coordinated Global Overlay Set

SERVICE KITS

CABLE FRU KIT's

5306523 Service Cable Hardware Kit
(Includes BNC components, studs, loopbacks, gender changers)


MRE HARDWARE

5408001

MR Elastography option kit
contains all MRE hardware




UNIFIED SURFACE COIL PHANTOM SET



5342679

Large Cylindrical
Unified Phantom




5342680

Small Cylindrical
Unified Phantom






5342681

Cubical
Unified Phantom




5343347



TL Unified Phantom

1.5T SERVICE PHANTOMS AND ACCESSORIES**2274476**LV SHIM PHANTOM
INNER SPHERE**2306642**LV PHANTOM
SMALL HALF ASM**2306645**LV PHANTOM
BIG HALF ASM**2134213**SPT PHANTOM
CART**2125244**SPT SHORT LOADER
w/SPHERE ASM**2131027-2**

DQA III ASM

5121115LOCATION
BRACKET
w/SCREWS**2321556**DQA PHANTOM
WITH LOADER

For Field Replacement Order:

2135650-2 – SPT SPHERE**2135652-2** – SPT LOADER**2291292**EPI HEAD
POSITIONER**5110241**HEAD LOADER
POSITIONER**2170481**EPI FOAM
POSITIONER**46-317586G1**10cm NiCl
PHANTOM**2258366**10cm SILICON
FILLED PHANTOM

CABLE FRU'S**5331721, Scan Room Base SHORT Cable FRU Kit contains:**

Run M1302	Run M1303	Run M1304	Run M1310	Run M1311	Run M1312
Run M2025	Run M3030	Run M3031	Run M3300	Run M3301	Run M3302
Run M3303	Run M3304	Run M3306	Run M3307	Run M3308	Run M3309
Run M3310	Run M3311	Run M3312	Run M3313	Run M3315	Run M3316
Run M4005	13W6 Shunt Connector for 16 & 32 Channel Only				

5331721-2, Scan Room Base LONG Cable FRU Kit contains:

Run M1302	Run M1303	Run M1304	Run M1310	Run M1311	Run M1312
Run M2025	Run M3030	Run M3031	Run M3300	Run M3301	Run M3302
Run M3303	Run M3304	Run M3306	Run M3307	Run M3308	Run M3309
Run M3310	Run M3311	Run M3312	Run M3313	Run M3315	Run M3316
Run M4005	13W6 Shunt Connector for 16 & 32 Channel Only				

5304950-3, 1.5T RF Cables, Hybrid to Body Coil:

Run M1322	Run M1323
-----------	-----------

5343119, Fiber Optic Cable FRU Kit contains:

Run P2003	Run P2004
-----------	-----------

5343120, In-Room Display (IRD) Cable FRU Kit contains:

Run P2026	Run M3366
-----------	-----------

5352518, Display Cable FRU Kit contains:

Run P2026	Run M3366	Run M3314	Run M3363	Run M33349
-----------	-----------	-----------	-----------	------------

5169566	Run 3367, MAG-TDM-J1 to MAG-RX1, 44-pin HD sub-D 560mm
5169566-2	Run 3368, MAG-TDM-J2 to MAG-RX2, 44-pin HD sub-D 1360mm
5180907	Run 3350, MAG-SRI-J10 to MAG-BORE, Bore Temp, 3000mm
5194965-2	Run M3360, MAG-J6 to MAG-PAC-J9, J2, 5600mm
5329119-2	Run M1002, PEN PNL J28 to RF-SW1 J121, 990mm
5329340-7	Run M1305, SPW-J122 to PED-IO-J1, 14306mm
5329340-8	Run M1305, SPW-J122 to PED-IO-J1, 18393mm
5329340-9	Run M1306, SPW-J120 to PED-IO-J2, 13965mm
5329340-10	Run M1306, SPW-J120 to PED-IO-J2, 18052mm
5329340-11	Run E1002, SPW-J128 to PGR-J18, 10218mm
5329340-12	Run E1002, SPW-J128 to PGR-J18, 16348mm
5330729-5	Run E1001, PGR-J19 to SPW-J106, 10327mm
5330729-6	Run E1001, PGR-J19 to SPW-J106, 16523mm
5330927-16	Run M3378, PEN-J111 to MAG-J50, 14000mm
5330932-5	Run M1310, PEN-J72 to MAG-J9, 14000mm
5330932-26	Run M1312, PEN-J74 to MAG-HYB-J6, 17000mm
5330932-27	Run M1313, MAG-DDSPLIT-J20 to MAG-BORE-J6, 6250mm
5330932-28	Run M1314, MAG-DDSPLIT-J21 to MAG-BORE-J7, 6250mm
5344701	Run M1001, SPW-J106 to MAG-HYB-J1, 12394mm
5344701-2	Run M1001, SPW-J106 to MAG-HYB-J1, 16525mm
5395724	Run M3341, MAG-SRI-J20 to MAG-DCK-L-J1, J4 "Touch & Go"

[illegible]

REVISION HISTORY

REV/DATE REASON FOR CHANGE

- | | | |
|---|------------|---|
| 1 | Feb 3, 10 | Initial release |
| 2 | Aug 2, 10 | Added new top laser mounting bracket for IRD (MRIhc49551). Added 5357155, 5376555 & 5400660 to HEC (MRIhcc49551). Added new 4170 ICN's to PGR (ECO 2090162). Revised OW monitor part to 5245466 (MRIhc49551). Added O2 monitor FRU's (2098349). Added MRU FRU's (ECR 2080462). Added SFP fiber optic tansceiver module to PGR section (MRIhc5145). Added expanded FRU parts to Patient Table (ECR 2102683). Added 9–Pin filter adapter to SPW (MRIhc51545). Added HP Z400 computer (ECO 2056587). Added MDP battery and charger FRU's (ECO 2097255). Added Unified Phantom set (MRIhc49551). |
| 3 | Jul 27, 11 | Updated the entire manual with new parts related to the intial release of MR450w GEM product. Updated title of manual to Optima MR450w & MR450w GEM FRU's. Main areas of updates pertaining to the new GEM product are the magnet enclosure, patient table & cradle, and GEM coils. For Magnet: Added PAC–5 parts (eNPI 20147). Added gradient coil secondary seal (ECO 2108006). For PGR Cabinet: Added XGA/XPS lifting bracket and 4170 ICN conversion kit (eNPI 20147). For OW: Added DV Apps M4 s/w collector and conversion kit to DV with Z400 (eNPI 20147). Added 1.5T I & Q RF cable FRU kit (ECO 2115667). Added SPW Air Inlet Filter (MRIhc49141). Updates and correctoins applied via HII Review (DOC0090238). |

[illegible]