

GE Healthcare

Signa HDx 3.0T Series  
Field Replaceable Units (FRU)

Technical Publication

© 2006 – 2008 General Electric Company  
GE Medical Systems, a General Electric Company, going to market as GE Healthcare  
3000 N. Grandview Boulevard  
Waukesha, Wisconsin 53188  
USA

[www.gehealthcare.com](http://www.gehealthcare.com)



5166013  
Rev 11

**COVER  
BOTTOM  
CROP LINE**

**COVER  
BOTTOM  
CROP LINE**

**COVER  
FOLD  
LINE** | **(2) CENTER STAPLES  
FOR BINDING**

**COVER  
RIGHT  
CROP LINE**

**5/16" RADIUS  
CORNER TRIM  
TYPICAL**

**5/16" RADIUS  
CORNER TRIM  
TYPICAL**

**COVER  
LEFT  
CROP LINE**

**SEND COMMENTS ABOUT THIS MANUAL TO:**

GE MEDICAL SYSTEMS  
MANAGER, SERVICE ENGINEERING  
W-808  
3200 N. GRANDVIEW BLVD.  
WAUKESHA, WISCONSIN 53188  
USA

This manual belongs to:

---

---

---

---

## INTRODUCTION TO FIELD REPLACEABLE UNIT (FRU)

This direction is a guide to the FRU's for HDxt and HDx.

Parts designated as a FRU are critical to maintain the safety and/or functionality of the system. FRU's are available on a Priority 1 basis from Service Parts Operation.

As a reference aid, the FRU's are listed under an upper-level part number where they are used. In most cases the upper-level part number is not a FRU. In some cases the parts listed under an upper-level assembly may have been nested in intermediate sub-assemblies which are not shown.

Where possible, FRU's which are part of Option Catalogs are shown under the area where they would be installed, rather than being listed in the Coils or Options Sections (SEE TABLE OF CONTENTS).

### LIST OF EFFECTIVE PAGES

| <u>PAGE</u> | <u>REV</u> | <u>PAGE</u> | <u>REV</u> | <u>PAGE</u> | <u>REV</u> |
|-------------|------------|-------------|------------|-------------|------------|
|-------------|------------|-------------|------------|-------------|------------|

|                    |    |  |  |  |  |
|--------------------|----|--|--|--|--|
| Cover Page . . . . | 11 |  |  |  |  |
|--------------------|----|--|--|--|--|

|                   |    |  |  |  |  |
|-------------------|----|--|--|--|--|
| 1 to 60 . . . . . | 11 |  |  |  |  |
|-------------------|----|--|--|--|--|

| <b>TITLE</b>                                             | <b>PAGE</b>               |
|----------------------------------------------------------|---------------------------|
| <b>INTRODUCTION TO FIELD REPLACEABLE UNIT(FRU)</b> ..... | <b>1</b>                  |
| <b>LIST OF EFFECTIVE PAGES</b> .....                     | <b>1</b>                  |
| <b>TABLE OF CONTENTS</b> .....                           | <b>2</b>                  |
| <b>PICTORIAL TABLE OF CONTENTS</b> .....                 | <b>3</b>                  |
| <b>MAGNET AND ENCLOSURE (ASSEMBLIES)</b>                 |                           |
| FRONT COVER, MISC. PARTS, & TRM .....                    | 5                         |
| BRM & TRM LIST (OF PARTS NOT SHOWN) .....                | 5                         |
| ENCLOSURE & VACUUM RELATED .....                         | 6                         |
| FRONT & REAR ENDBELL, BRIDGE SUPPORT .....               | 6                         |
| 3T94 ENCLOSURE COMPONENTS .....                          | 7 & 8                     |
| REAR PEDESTAL PARTS .....                                | 9                         |
| SPLIT BRIDGE & DOCKING PARTS .....                       | 10                        |
| LOW PROFILE CARRIAGE ASSEMBLY .....                      | 11                        |
| SRI/PAC & TOUCH-UP PAINTS .....                          | 12                        |
| MRU, MONITORS, & TRANSFORMER .....                       | 13                        |
| HDx MAGNET ENCLOSURE .....                               | 14 & 15                   |
| HDx BRIDGE PARTS .....                                   | 16                        |
| HDx LPCA .....                                           | 17                        |
| <b>PATIENT SUPPORT TABLE</b> .....                       | <b>18 &amp; 19</b>        |
| <b>CHILLER</b> .....                                     | <b>20 &amp; 21</b>        |
| <b>PENETRATION PANEL</b> .....                           | <b>22</b>                 |
| <b>OPERATOR WORKSPACE</b> .....                          | <b>24</b>                 |
| <b>GOC COMPUTER CABINET, SCIM AND KEYBOARD</b> .....     | <b>25</b>                 |
| <b>MAIN DISCONNECT PANEL</b> .....                       | <b>28</b>                 |
| <b>HFD-S/PDU &amp; HFD/PDU CABINET</b> .....             | <b>29, 30</b>             |
| <b>RFS CABINET</b> .....                                 | <b>32, 33, &amp; 34</b>   |
| <b>SYSTEM CABINET</b> .....                              | <b>35</b>                 |
| <b>CAM, MGD, RRF, &amp; ASC MODULES</b> .....            | <b>36, 37, 38, 39, 40</b> |
| <b>HDx 32 CHANNEL CABINET (OPTION)</b> .....             | <b>41</b>                 |
| <b>RF ACCESSORY CABINET</b> .....                        | <b>43</b>                 |
| <b>NB RF AMP CABINET</b> .....                           | <b>44 &amp; 45</b>        |
| <b>8KW MNS BB RF AMP CABINET (OPTIONAL)</b> .....        | <b>46</b>                 |
| <b>TWIN ACCESSORY CABINET (TAC)</b> .....                | <b>47, 48, 49</b>         |
| <b>REPLACEMENT CABLES</b> .....                          | <b>50 &amp; 51</b>        |
| <b>UNIFIED SURFACE COIL PHANTOM SET</b> .....            | <b>52</b>                 |
| <b>SERVICE TOOLS</b> .....                               | <b>53</b>                 |
| <b>TOOLS, REPAIR KITS, AND HEAD COILS</b> .....          | <b>54</b>                 |
| <b>COILS</b> .....                                       | <b>55, 56 &amp; 57</b>    |
| <b>BrainWave OPTIONS</b> .....                           | <b>57</b>                 |
| <b>MNS OPTIONS</b> .....                                 | <b>58</b>                 |
| <b>SOFTWARE OPTION KEYS</b> .....                        | <b>59</b>                 |
| <b>REVISION HISTORY</b> .....                            | <b>60</b>                 |

|                |         |
|----------------|---------|
| INTRO TO FRU'S | PAGE 1  |
| CHILLER PARTS  | PAGE 20 |
| CABLES         | PAGE 50 |
| COILS          | PAGE 56 |
| MNS OPTIONS    | PAGE 58 |
| S/W OPTIONS    | PAGE 57 |
| BrainWave      | PAGE 57 |




**2241999**  
PATIENT AIR SUPPLY  
BLOWER CABINET (MG6)



PENETRATION  
PANEL  
PAGE 22

PATIENT  
SUPPORT  
TABLE  
PAGE 18




**MAGNET ENCLOSURE: PGS 5 to 12**

**HDx MAGNET ENCLOSURE: PGS 14 to 17**

TRM COIL,  
MOUNTING  
BRACKETS, CART,  
AND CRADLE,  
SEE PAGE 5

## MAGNET ROOM

### OPERATOR/EQUIPMENT ROOM




SERVICE TOOLS  
PAGE 53



**2219400**  
MAGNET  
MONITOR  
(MSM)




OPERATOR  
WORKSPACE  
PAGES 24 & 25




ACCESSORY  
CABINET  
PAGES  
47 THRU 49




MAIN  
DISCONNECT  
PANEL  
PAGE 28




32 CHANNEL  
CABINET  
PAGE 41



HFD-S/PDU  
PAGE 29  
HFD/PDU  
PAGE 30




RFS &  
SYSTEM  
CABINET  
PAGES  
33, THRU 40



RF ACCESSORY  
CABINET  
PAGE 43




NB RF AMP  
CABINET  
PAGES  
44 & 45



8KW BB RF  
AMP CABINET  
PAGE 46

## PICTORIAL TABLE OF CONTENTS

[illegible]

**WideOpen FRONT COSMETIC COVER & PARTS (BACK VIEW):**

**NOTE:** The following LCC 3.0T FIXED W/TRM & Enclosure related parts **Not Shown**;  
**MISCELLANEOUS:**

|                    |                                          |
|--------------------|------------------------------------------|
| <b>2203803</b>     | SPLIT O-RING                             |
| <b>2242756</b>     | RELIEF VALVE, 3/8 NPT X 7 PSIG, GEN      |
| <b>2299843</b>     | MAGNET PRESSURE TRANSDUCER, SUNX DP2-41N |
| <b>46-294231G1</b> | MAGNET RUNDOWN UNIT                      |
| <b>46-294231P1</b> | BUSSMAN, F1, 110VAC .400 GDC             |

**TRM:**

|                           |                                                     |
|---------------------------|-----------------------------------------------------|
| <b>5113878*</b>           | RF SHIELD FOR TRM COIL                              |
| <b>2144093-9</b>          | TRM WITH MOUNTING BRACKETS, CART, AND CRADLE        |
| <b>2208999-7</b>          | 3.0T RF BODY COIL AND CASE                          |
| <b>2208999-12</b>         | HDx new color 3.0T RF BODY COIL & Purple/Black CASE |
| <b>2218148</b>            | COLD SHRINK (FOR REAR ENDBELL AND TRM COIL)         |
| <b>2299498</b>            | COOLING CIRCUIT REPAIR KIT                          |
| <b>2383070 or 5250104</b> | DIODE SLOT BOARD                                    |
| <b>5123812</b>            | (1260) BORE TEMPERATURE SENSOR CABLE                |
| <b>5123813</b>            | (1261) OPERATOR DISPLAY TO SRI CABLE                |
| <b>5127441</b>            | GRADIENT CABLE CLAMP BLOCK KIT                      |
| <b>2293612</b>            | MALE CONNECTOR 3/4" HOSE TO 3/4" NPT                |
| <b>2324086-2</b>          | 5 GALLON PAIL w/ETHYLENE GLYCOL WARNING LABEL       |

**\*2389049** RF Shield Change Out Kit – MUST be ordered for the RF Shield replacement

**ENCLOSURE & VACUUM RELATED PARTS:**

2293771  
RADIAL SHIM SPACER, 1.5MM  
2293771-2  
RADIAL SHIM SPACER, 0.8MM  
2241176  
RUBBER ISOLATION PAD

**FRONT ENDBELL PARTS:**  
(TRM ONLY)


2162554-4  
FRONT COVER LASER

2295639  
O-RING,  
ACCESS  
PANELS

GASKET,  
ENDBELL  
TO RF TUBE  
2294358, 4MM  
2294358-2, 6MM

FRONT  
ENDBELL ASM  
2296177

2297761  
O-RING, WATERLINE  
FEEDTHROUGH


**BRIDGE SUPPORT PARTS:**  
(TRM ONLY)

2293539  
BRIDGE  
SUPPORT  
BUSHING

2293956  
NYLON HEX NUT,  
M10 X 1.5MM  
X 9.5MM HIGH

**REAR ENDBELL ASSEMBLY:**  
(TRM ONLY)

2284170  
REAR ENDBELL  
VACUUM ASSEMBLY;  
& 2218148  
COLD SHRINK  
(FOR REAR ENDBELL &  
TRM COIL)  
& 2218149  
COMPRESSION  
SPLICER  
(FOR REAR ENDBELL)



2358470  
COMFORT MODULE  
ASSEMBLY  
2206437  
M4 X 12mm NYLON  
SCREW FOR  
SPEAKER GRILL

**GRADIENT GASKETS:**  
(TRM ONLY)



2297540  
(4MM)  
2297540-2  
(6MM)



**NOTE: The following parts are Not Shown.**

|         |                                 |
|---------|---------------------------------|
| 2354050 | SPECTRO T/R ASM (OPTION)        |
| 2381249 | MNS LPCA CABLE HARNESS (OPTION) |
| 2297760 | VACUUM GREASE, MSDS# 8366442    |
| 2297672 | GLYCOL, 1 GALLON/3.8 LITERS     |




**3T94 FRONT COVER & PARTS (BACK VIEW):****FRONT END BELL (BACK VIEW):**

**NOTE:** The following parts **Not Shown**:

|                 |                                    |
|-----------------|------------------------------------|
| <b>5113881*</b> | 3T CRM RF Shield and FRU Container |
| <b>5109260</b>  | 3T Eclipse Body Coil FRU (in case) |
| <b>2319599</b>  | 3T/94 RT Shim Kit                  |
| <b>2370499</b>  | 3T Eclipse CRM B0 Coil FRU         |


**\*2389049** RF Shield Change Out Kit – MUST be ordered for the RF Shield replacement

**3T94 COMPONENTS:****BLOWER CABINET**


**2121199-2**  
COMFORT MODULE  
BLOWER



**5106787**  
BODY COIL  
BLOWER

**SRI MODULE**

**5121793**  
SRI-3 MODULE  
**5123812**  
RUN 1260  
**5123813**  
RUN 1261  
**5115910**  
SRI-2 MODULE


**COMFORT MODULE**

**2220079-2**  
COMFORT  
MODULE

**2206437**  
M4 X 12mm NYLON  
SCREW FOR  
SPEAKER GRILL

**MAGNET INTERFACE BOARD**

**46-226552G2**  
LIMIT SWITCH  
BOARD



REVISION 11


DIRECTION 5166013

REAR PEDESTAL

5224109

LONGITUDINAL ENCODER

5250136 or 6250012  
REMOTE INTERCOM  
BOARD, MG3 A2 A1

3T94 3.0T HYBRID SPLITTER:

2364795

RUN 893  
COAX CABLE

2214448

3T PREAMP

2302305

3T BODY  
HYBRID ASM

G3 3.0T HYBRID SPLITTER:

2370221

3T DUAL BODY  
HYBRID


2369611-2

IMAGE REJECT  
FILTER

2377703


3T BODY PREAMP  
& COMBINER  
MODULE


MAGNET AND ENCLOSURE (ASSEMBLIES)

**WideOpen BRIDGE PARTS:**

**NOTE: The following part is Not Shown.**


|             |                             |
|-------------|-----------------------------|
| 2386348     | FRONT BRIDGE FIBERGLASS     |
| 5117304     | FRONT BRIDGE PULLEY ASM     |
| 2293538     | BRIDGE EDGE FOAM            |
| 46-282436P1 | CARRIAGE TRAVEL SENSOR FLAG |

**MAGNET DOCKING PARTS:**


**HD LOW PROFILE CARRIAGE ASM****MAGNET AND ENCLOSURE (ASSEMBLIES)**

**SRI/PAC PLATFORM:**

**5112668**  
NON-LCC  
PAC-2B MODULE



**5122500**  
ECG LEAD KIT



**5105270**  
SRI-3  
MODULE

**5121793**  
SRI-3 U/G KIT  
w/CABLES


**5112668**  
PAC-2B MODULE

**TOUCH-UP PAINTS:**


**2183491-3** GRAY #1 0.6 OZ. NON-AEROSOL

**46-303460P3** STEEL BLUE 0.6 OZ. NON-AEROSOL

**46-303460P4** DARK STEEL BLUE 0.6 OZ. NON-AEROSOL




**NOTE:** To obtain enclosure part numbers not found in this manual, refer to Direction 2138247, *Signa Illustrated Parts List*, for enclosure covers, endbells, and other parts.

**CRYOGEN PANEL STEP-DOWN TRANSFORMER FRU'S:****MAGNET RUNDOWN UNIT (MRU) FRU'S:**


|                |                                                                                                |
|----------------|------------------------------------------------------------------------------------------------|
| <b>5196918</b> | <b>MAGNET RUNDOWN UNIT (Primary)</b>                                                           |
| <b>5196919</b> | <b>MAGNET RUNDOWN UNIT (Remote)</b>                                                            |
| <b>5363631</b> | SPARE FUSE KIT includes: 1.25A, 250V Fuse; 2.5A, 250V Fuse; 400ma, 120V Fuse; 200ma, 250V fuse |
| <b>5199490</b> | BATTERY PACK ASM                                                                               |
| <b>5180324</b> | ADAPTER PLATE                                                                                  |
| <b>5196921</b> | REMOTE TO PEN PANEL CABLE                                                                      |
| <b>5196920</b> | PRIMARY TO PEN PANEL CABLE                                                                     |
| <b>2128699</b> | MAGNET TO MRU CABLE, RUN 606                                                                   |

**OXYGEN MONITOR (O2) FRU'S:**

|                  |                                                                                                                |
|------------------|----------------------------------------------------------------------------------------------------------------|
| <b>5390676</b>   | OXYGEN MONITOR FRU – Contains all of the parts included with the O2 monitor, box, sensor and cables            |
| <b>5399042</b>   | OXYGEN MONITOR CALIBRATION KIT – Contains all of the parts needed to calibrate either the nw or old O2 monitor |
| <b>5390678-3</b> | REMOTE 5-YEAR SENSOR – Sensor for the ISA-60M O2 monitor                                                       |

**MAGNET MONITOR:**


**2219400**  
MAGNET MONITOR (MSM)

**HDx MAGNET ENCLOSURE****3.0T HDx ENCLOSURE PARTS**

|                  |                                                  |
|------------------|--------------------------------------------------|
| <b>5164786</b>   | HDx Enclosure Hardware Spares Kit                |
| <b>5161262</b>   | Turtle Bracket Gasket TRM                        |
| <b>5150552</b>   | Rear Turtle Arm, TRM Single Boss                 |
| <b>5150552-2</b> | Rear Turtle Arm, TRM Double Boss                 |
| <b>5150552-3</b> | Rear Turtle Arm, TRM Single Boss with Relief Cut |
| <b>5160997</b>   | Front End Bell TRM Assembly                      |
| <b>5160999</b>   | Filler Cover Left, Front End Bell TRM            |
| <b>5160999-2</b> | Filler Cover Right, Front End Bell TRM           |
| <b>5161000</b>   | Rear End Bell TRM Assembly                       |
| <b>5161003</b>   | Filler Cover Left, Rear End Bell TRM             |
| <b>5161003-2</b> | Filler Cover Right, Rear End Bell TRM            |
| <b>5143732-2</b> | Rear Cover Assembly, TRM                         |
| <b>5146696</b>   | TRM Comfort Module Assembly                      |
| <b>2222549-4</b> | Twin Lumitex Light Panel, 1085mm                 |




| <b>3.0T HDx ENCLOSURE PARTS</b> |                                         |
|---------------------------------|-----------------------------------------|
| 5151721-2                       | Front Right Arc Cover Assembly          |
| 5151722-2                       | Front Left Arc Cover Assembly           |
| 5151723-2                       | Rear Right Arc Cover Assembly           |
| 5151725-2                       | Rear Left Arc Cover Assembly            |
| 5151718-2                       | Right Side Top Cover, White             |
| 5151771-2                       | Right Side Bottom Cover, White          |
| 5161378-2                       | Right Latch Hook                        |
| 5151720-2                       | Left Side Bottom Cover, White           |
| 5161378                         | Left Latch Hook                         |
| 5145172                         | Left Side Top Cover, White              |
| 5146812                         | Left Side Top Cover, Rear Corner 3.0T   |
| 5151729                         | Front Left Skirt Cover                  |
| 5151727                         | Front Right Skirt Cover                 |
| 5151731                         | Rear Right Skirt Cover                  |
| 5151732                         | Rear Left Skirt Cover                   |
| 5143554                         | Left Rear Pedestal Cover                |
| 5143555                         | Right Rear Pedestal Cover               |
| 2215716-51                      | Center Display Panel With Overlay       |
| 2215716-10                      | Center Display Panel                    |
| 2215716-60                      | Left Control Panel with all Overlays    |
| 2215716-11                      | Left Control Panel                      |
| 2215716-70                      | Right Control Panel with all Overlays   |
| 2215716-12                      | Right Control Panel                     |
| 5160276                         | Bracket I and Q Lines                   |
| 5150097                         | Turtle Bracket Front Arm                |
| 5160991                         | Dock Bracket Assembly                   |
| 5150410                         | Front Trim Ring Assembly                |
| 5150411                         | Rear Trim Ring Assembly                 |
| 5143642                         | Front Cover Assembly                    |
| 5141500                         | Microphone Block and Cable Assembly     |
| 5142985                         | Vacuum Port Cover Assembly              |
| 5149066                         | Remote PAC Base                         |
| 5149068                         | Remote PAC Inner Mounting Tray          |
| 5149065                         | Remote PAC Accessory Hook               |
| 2344509-2                       | Bore Light End Cap                      |
| 5160206                         | Side Cover Mounting Bracket Assembly RH |
| 5160206-2                       | Side Cover Mounting Bracket Assembly LH |
| 5160992                         | Rear Bracket Assembly, Left             |
| 5160992-2                       | Rear Bracket Assembly, Right            |
| 2206437                         | M4 x 12mm Nylon Screw for Speaker Grill |

**HDx BRIDGE PARTS:**


**NOTE: The following part is Not Shown.**


|                    |                             |
|--------------------|-----------------------------|
| <b>5117304</b>     | FRONT BRIDGE PULLEY ASM     |
| <b>2293538-5</b>   | BRIDGE EDGE FOAM            |
| <b>46-282436P1</b> | CARRIAGE TRAVEL SENSOR FLAG |

**HDx LOW PROFILE CARRIAGE ASM****MAGNET AND ENCLOSURE (ASSEMBLIES)**

**PATIENT TABLE – INFINITY and HDx**

**NOTE:** Refer to page 19 for information on bumper FRU's for HDx, Infinity, and Pre-Infinity patient tables.



**PATIENT TABLE – BUMPERS AND PANELS**






| HDx EXISTING PART                               | HDx REPLACEMENT FRU                                                |
|-------------------------------------------------|--------------------------------------------------------------------|
| 5145946, Panel – Left<br>5145947, Panel – Right | <b>5179590</b> , New panels and side bumpers FRU kit for HDx table |
| 5145939, 22 Side Bumper                         | <b>5161172</b> , Side Bumper 22, Color P17                         |
| 5145939–2, 16.25 Side Bumper                    | <b>5161172–2</b> , Side Bumper 16.25, P17                          |
| 5145941, Foot End Bumper                        | <b>5145941</b> , Foot End Bumper                                   |
| 5145935, Head End Bumper                        | <b>5145935</b> , Head End Bumper                                   |

| INFINITY EXISTING PART                                                      | INFINITY REPLACEMENT FRU                                                |
|-----------------------------------------------------------------------------|-------------------------------------------------------------------------|
| 46–271022G1, 2393106, Panel – Left<br>46–271021G1, 2393106–2, Panel – Right | <b>5179706</b> , New panels and side bumpers FRU kit for Infinity table |
| 5154908–4, Panel – Left<br>5154908–3, Panel – Right                         | <b>5179706</b> , New panels and side bumpers FRU kit for Infinity table |
| 46–271025P5, 22 Side Bumper                                                 | <b>5161172–3</b> , Side Bumper 22, Color P17                            |
| 46–271025P6, 16.25 Side Bumper                                              | <b>5161172–4</b> , Side Bumper 22, Color P17                            |
| 46–271027P3, Foot End Bumper                                                | <b>46–271027P3</b> , Foot End Bumper                                    |
| 46–271026P3, Head End Bumper                                                | <b>46–271026P3</b> , Head End Bumper                                    |


| Pre-INFINITY EXISTING PART                                        | Pre-INFINITY REPLACEMENT FRU                                            |
|-------------------------------------------------------------------|-------------------------------------------------------------------------|
| 5154908–4, Panel – Left (Grey)<br>5154908–3, Panel – Right (Grey) | <b>5179706</b> , New panels and side bumpers FRU kit for Infinity table |
| 22 Side Bumper (Accent Grey or Blue)                              | <b>5161172–3</b> , Side Bumper 22, Color P8                             |
| 16.25 Side Bumper (Accent Grey or Blue)                           | <b>5161172–4</b> , Side Bumper 16.25, Color P8                          |
| Foot End Bumper (Accent Grey or Blue)                             | <b>46–271027P3</b> , Foot End Bumper                                    |
| Head End Bumper (Accent Grey or Blue)                             | <b>46–271026P3</b> , Head End Bumper                                    |

**G3 SYSTEM COOLING CABINETS;**


REFER TO ELLIS &amp; WATTS SERVICE MANUALS FOR FRU INFORMATION

|                                                                                                                                                                                                                       |                                                                                                                                                                                                                                   |                                                                                                                    |
|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------------------------------------------------------------------------------------|
| <b>OUTDOOR TSCC</b><br><br><br>REMOTE CONTROL PANEL  | <b>INDOOR WATER COOLED TSCC</b><br>                                                                                                              | <b>INDOOR AIR COOLED TSCC</b><br> |
| <b>OUTDOOR TGWC</b><br><br><br>REMOTE CONTROL PANEL | <b>INDOOR WATER COOLED TGWC</b><br><br><br>REMOTE CONTROL PANEL | <b>INDOOR AIR COOLED TGWC</b><br> |

REFER TO AIRSYS SERVICE MANUALS FOR FRU INFORMATION

|                                                                                                  |                                                                                                                                                                                                               |
|--------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>GWHX</b><br> | <b>MRCC</b><br><br><br>REMOTE CONTROL PANEL |
|--------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|

**2343325** AIRSYS HEAT EXCHANGER (GWHX)**2341077-2** AIRSYS QUICK DISCONNECT KIT**2341077-3** AIRSYS REMOTE CONTROL PANEL**2298100-4** E & W MODEL LTL10 HEAT EXCHANGER CABINET**2298100-5** CONTROL PANEL FOR LTL10 ONLY**2297672** GLYCOL USED WITH E & W OR AIRSYS

**3T94 GRADIENT WATER CHILLER**


**2222564-2**  
60 Hz GRADIENT CHILLER

OR

**2222564-21**  
50 Hz GRADIENT CHILLER


**NOTE:** The following **WATER CHILLER** parts are **Not Shown**:

|                          |                                    |
|--------------------------|------------------------------------|
| <b><u>2222564-6</u></b>  | <u>LYTRON 50/60 Hz PUMP MOTOR</u>  |
| <b><u>2222564-8</u></b>  | <u>LYTRON 0.25A/250V FUSE</u>      |
| <b><u>2222564-18</u></b> | <u>LYTRON 60Hz FAN MOTOR</u>       |
| <b><u>2222564-19</u></b> | <u>LYTRON 50Hz FAN MOTOR</u>       |
| <b><u>2222564-31</u></b> | <u>LYTRON NYLON MOTOR COUPLING</u> |
| <b><u>2138791</u></b>    | <u>BODY COIL COOLANT, 1 GALLON</u> |

G3 & 3T94 PENETRATION PANEL






**OPERATOR WORKSPACE – HDx**

\* When ordering a monitor replacement, be aware that models may be updated without notice.

**NOTE: The following parts are Not Shown.**

|                   |                            |
|-------------------|----------------------------|
| <b>2156158-84</b> | S/W COLLECTOR FOR 15.0 M4B |
| <b>2389540</b>    | SCAN TOOLS MOD             |


**GOC COMPUTER CABINET WITH HP Z400;**



**NOTE:** The following parts are **Not Shown**.

- 5373029** Cable FRU kit for GOC contains:  
Run E3016, Run E3046, Run E3047, Run E3048, Host ot IRDC DVI-D cable, Host to IRDC USB cable, IRDC Power cable
- 5376172** Dust Filter FRU Kit for Z400
- 5364704** One Wire Collector for Z400

**SCIM & KEYBOARD;**




**NOTE:**  
If Keyboard is to be replaced, also order **46-317475P2**, 2 inch x 4 inch adhesive backed velcro to attach to bottom of new keyboard.

**NOTE:** The following parts are **Not Shown**.

|                  |                                            |
|------------------|--------------------------------------------|
| <b>2262158-9</b> | CABLE, 50 PIN SCIM KEYBOARD                |
| <b>5148389-2</b> | SCIM <sup>W</sup> /ENGLISH OVERLAY (Black) |
| <b>5148389-9</b> | MR SCIM GLOBAL OVERLAY SET (Black)         |

**KEYBOARDS ONLY;**

- 5143798-7** (English)  
**5143798-8** (French)  
**5143798-9** (German)  
**5143798-10** (Spanish)  
**5143798-11** (Italian)  
**5143798-12** (Portuguese)  
**5143798-13** (Scandinavian)

**GOC COMPUTER CABINET WITH HP9300 OR HP8400:**

\* **5143798-14** is only compatible with 14M5 software or later

\*\* Sites with SCSI Tower MUST order 5183547-41 as a replacement.  
Sites without a SCSI Tower can order 5183547-41 or 5183547-60 as a replacement.


**NOTE: The following parts are Not Shown.**

**5151612** USB CABLE FOR OPERATOR WORKSPACE

**5141564** AUTOVOICE CABLE



**NOTE: Refer to the Directions on applicable MR Service Methods DVD for additional product information.**



**NOTE: The following items are Not Shown.**

**5338648** Upgrade kit for MDP Battery and Charger


**5375947** Power Sonic Battery FRU

**5316807** Power Supply FRU

| LOCATION | AMPS | VOLTS | GEMS<br>PART<br>NUMBER | MANUFACTURER |           |
|----------|------|-------|------------------------|--------------|-----------|
|          |      |       |                        | BUSS         | G-S       |
| FUSE F1  | 1.25 | 480   | 2276547-2              | SNQ-R-1-1/4  | ATQR1-1/4 |
| FUSE F2  | 3    | 120   | 2276547-3              | FNQ-R-3      | ATQR3     |
| FUSE F3  | 1    | 12VDC | 2276547-8              | LP-CC-1      | ATDR1     |
| FUSE F4  | 0.12 | 480   | 2276547-9              | SNQ-R-8      | ATQR1/8   |
| FUSE F5  | 125  | 600   | 2276547-19             | LPJ125SP     | AJT125    |
| FUSE F6  | 50   | 600   | 2276547-20             | LPJ-50SP     | AJT50     |
| FUSE F7  | 20   | 600   | 2276547-22             | LPJ-20SP     | AJT20     |
| FUSE F8  | 5    | 600   | 2276547-21             | FRS-R-5      | TRS5R     |
| FUSE F9  | 8    | 400   | 2276547-6              | FRS-R-8      | TRS8R     |
| FUSE F10 | 10   | 120   | 2276547-11             | FNQ-R-10     | ATQR10    |
| FUSE F11 | 10   | 120   | 2276547-11             | FNQ-R-10     | ATQR10    |

**2276547-14** 12 VDC BATTERY

**2276547-16** 12 VDC SUPPLY

**HFD-S/PDU CABINET ASSEMBLY W/SGA/SGAPS/GP**

**NOTE: The following Part is Not Shown.**

**2303317** CAN Terminator

HFD/PDU CABINET

**5116803**  
HFD CABINET  
FAN ASM

**2284074**  
ALUMINUM  
AIR FILTER

**5118444**  
HFD CABINET  
WITH PDU

**2331400-2**  
GP3 MODULE

**2393906-4**  
HFA ASSEMBLY  
**2396050-2**  
HFA CONTROL BD

**2334862-2**  
SGA PS ASSEMBLY

**46-170021P87**  
SGA PS FUSES


**2301200-3**  
TVSS MODULE  
**2374930-2**  
PDU CONTROL  
BOARD



[illegible]

**HDx RFS CABINET with CAM Module**

\* If a cabinet contains a 5127452 ICN, a replacement 5127452 must be ordered.  
If a 5127452-3 is in the cabinet, a 5127542-3 must be ordered.




**NOTE: The following items are Not Shown.**

**5160683**      RRF SERVICE DIAGNOSTIC KIT

**46-296817P1**      50 OHM BNC TERMINATOR

**HDx RFS CABINET**


\* If a cabinet contains a 5127452 ICN, a replacement 5127452 must be ordered.  
If a 5127452-3 is in the cabinet, a 5127542-3 must be ordered.



**NOTE: The following items are Not Shown.**

**5160683** RRF SERVICE DIAGNOSTIC KIT

**46-296817P1** 50 OHM BNC TERMINATOR

**HD RFS CABINET UPGRADED TO HDx**

\* If a cabinet contains a 5127452 ICN, a replacement 5127452 must be ordered.  
If a 5127452-3 is in the cabinet, a 5127452-3 must be ordered.

**MGD MODULE**  
SEE PAGE 37

**REMOTE RF MODULE**  
SEE PAGE 38

**5110612**  
DRIVER  
MODULE

**5162073**  
24 PORT  
ETHERNET  
SWITCH

**511201**  
AC 120V  
DISTRIBUTION  
MODULE

**5127452\***  
IMAGE COMPUTE NODE

OR

**5127452-3\***  
SUN 4100 ICN

**5127452-4**  
SUN HARD DISK

OR

**5337894-3** OR **5337894-4**  
SUN 4170 ICN  
(16 Channel)  
**5337894-5**  
4170 HDD

**5139347**  
AC DISTRIBUTION  
BOX

**5115474**  
UTNS SPLITTER BOX

**2363041-2**  
ASC CHASSIS  
SEE PAGE 40

**2379186**  
TERMINAL  
SERVER

**2272740**  
ETHERNET HUB  
SERVICE KIT


**5120780**  
TERM SERVER  
CABLE FRU KIT  
(4 CABLES)

**5115759**  
SINGLE PORT  
TERM SERVER  
(OPTION)

**NOTE: The following items are Not Shown.**

**5160683** RRF SERVICE DIAGNOSTIC KIT



**46-296817P1** 50 OHM BNC TERMINATOR

**EXISTING SYSTEM CABINET UPGRADED TO HDx**

**NOTE:** The following items are **Not Shown**.


**5160683** RRF SERVICE DIAGNOSTIC KIT

**46-296817P1** 50 OHM BNC TERMINATOR




**NOTE:** The following CAM Module parts are Not Shown.

**5181495-4** CAM CHASSIS with MIDPLANE

**MGD MODULE (FRONT)**

\* For 32 Channel systems, use **2358122** IRF2 Bd.

**MGD MODULE (REAR)**

**NOTE:** The following MGD Module parts are Not Shown.

**2294300-5** BRIDGE BOARD



**2294300-6** MGD CHASSIS W/BACKPLANE & AC ENTRY (No Boards)

**2294300-7** MGD FAN

**2294300-8 or**

**5181495-3** POWER SUPPLIES

**2294300-9** INDICATOR BOARD

**2305884 REMOTE RF MOD ASM (CHASSIS W/MIDPLANE) (FRONT)****2305884 REMOTE RF MOD ASM (CHASSIS W/MIDPLANE) (REAR)**

**NOTE: Additional REMOTE RF MOD ASM Parts/Cables are located on page 39.**

\* A part that does not have a "-2" can be replaced with either the same part number or the same number ending with a "-2". If replacing the "-2" version, then only the part number with a "-2" can be used.

\*\* 2274188 and 2274188-2 are replaced by 2274188-3. Also, all 2274188 series MUX boards can be replaced by backwards compatible MUX2 board, 5250106.

\*\*\* For 32 Channel systems, use 5250144 Exciter 2.1, in main RFS Cabinet only (not in 32 Channel cabinet)





**NOTE: The following RRF MODULE (page 38) Parts/Cables are Not Shown.**


|                    |                                                        |
|--------------------|--------------------------------------------------------|
| <b>46-306565P1</b> | CABINET FILTER                                         |
| <b>2294728</b>     | COMPACT PCI FRONT PANEL BLANK (8HP) (2 SLOT)           |
| <b>2319152</b>     | COMPACT PCI FRONT PANEL BLANK (4HP) (1 SLOT)           |
| <b>2308729</b>     | BNC TO SMB CABLE (MUX J8 TO CAB I/F J1)                |
| <b>2308728</b>     | 8W8 CABLE ASSEMBLY (MUX J5)                            |
| <b>2322578</b>     | J15/23/25/26 DATA INPUT CABLE (CAB I/F)                |
| <b>2322580</b>     | CABLE J18, UNBLANK (CAB I/F)                           |
| <b>2322581</b>     | CABLE J24, MAGNET MONITOR (CAB I/F)                    |
| <b>2322582</b>     | CABLE J30/31, SERIAL EXP (CAB I/F)                     |
| <b>2322583</b>     | CABLE J62 (15 PIN SUB-D) (CAB I/F)                     |
| <b>2322586</b>     | CABLE J61 (37 PIN SUB-D) (CAB I/F)                     |
| <b>2308729-2</b>   | BNC TO SMB CABLE (MUXJ13 TO CAB I/F J2)                |
| <b>2308729-3</b>   | BNC TO SMB CABLE (MUXJ14 TO CAB I/F J3)                |
| <b>2394870-3</b>   | RRF slot14R-J5 I/O J200                                |
| <b>5110796</b>     | RRF slot14R-J13 I/O J202                               |
| <b>5110795</b>     | RRF slot 14R-J13 I/O J202                              |
| <b>5115417-3</b>   | RRF slot14R-J8 RF Amp J14                              |
| <b>5115417</b>     | RRF slot12R-J31 Splitter-1                             |
| <b>5113193</b>     | RRF slot12R-J7 RRF slot11R-J6/7                        |
| <b>2394870-2</b>   | RRF slot5R-J5 I/O J201                                 |
| <b>5113193-2</b>   | RRF slot3R-J7 RRF slot2R-J6/7                          |
| <b>5115417-2</b>   | RRF slot3F-J31 Splitter-2                              |
| <b>5115630</b>     | J7 MUX (2) TO J6 MUX (1) CABLE                         |
| <b>5115630-2</b>   | RRF slot14R-J11 RRF slot9R-J6                          |
| <b>5116213-2</b>   | RF Extension cable to Run 1238                         |
| <b>5121269</b>     | RF Extension cable to Run 1259                         |
| <b>5145157</b>     | Fiber Optic cable, MGD Slot 11R-J20 to RF Slot 10F-J21 |
| <b>5145157-2</b>   | Fiber Optic cable, MGD Slot 11F-J7 to RF Slot 1F-J21   |
| <b>5150527</b>     | Y Power cable, 10 feet                                 |
| <b>5160775</b>     | HDx, FRU ICN Ethernet Cable Collector                  |

**NOTE: The following ASC MODULE (page 40) Parts/Cables are Not Shown.**


|                  |                         |
|------------------|-------------------------|
| <b>5115417-4</b> | ASC slot14-J1 RF Amp J8 |
| <b>5115417-5</b> | ASC slot14-J4 RF Amp J7 |
| <b>5115417-6</b> | ASC slot11-J1 RF Amp J6 |
| <b>5120955</b>   | ASC POWER SUPPLY        |
| <b>2391148</b>   | BB2 AMP I/F ASM         |

**2363041 or 2363041-2\* ASC CHASSIS ASSEMBLY****2363041 or 2363041-2\* ASC CHASSIS ASSEMBLY WITH MNS**

**NOTE:** Additional ASC MODULE Parts/Cables are located on page 39.

**HDx 32 CHANNEL CABINET (Option)**

[illegible]



NOTE: The following RF ACCESSORY CABINET parts Not Shown:

**46-306565P1** 10" FRONT COVER FILTER

**2365797** CABLE - I/O J127 TO ASC SLOT 11/14 J1/2/4/5

**2368552** CABLE - ASC SLOT 14-J3 TO I/O J100

**2369552-8** CABLE - ASC SLOT 13-J4 TO I/O J102

**2369552-9** CABLE - ASC SLOT 13-J5 TO I/O J103


**2369552-10** CABLE - ASC SLOT 10-J4 TO I/O J104

**2369552-11** CABLE - ASC SLOT 10-J5 TO I/O J105

**2379507** DRSM JUMPER PLUG

**2383066** DRSM JUMPER CABLE (ASC-DRSM-I/O)

This page for G3 MAGNET (Short Bore)  
35KW NARROW BAND RF AMP



NOTE: The following NB RF AMP CABINET parts are Not Shown:


|                   |                                       |
|-------------------|---------------------------------------|
| <u>2336064-8</u>  | <u>35KW RF AMP FUSE KIT</u>           |
| <u>2336064-15</u> | <u>35KW RF AMP TEMPERATURE SENSOR</u> |

NOTE: The following RF TRANSMIT FLEX CABLES are Not Shown:

|                  |                                                     |
|------------------|-----------------------------------------------------|
| <u>2384850</u>   | <u>RUN 1156 – MR8-J4 TO CONNECT TO RUN 1125</u>     |
| <u>2384850-2</u> | <u>RUN 1160 – MG3-A3-J1 TO CONNECT TO RUN 1127</u>  |
| <u>2384851</u>   | <u>RUN 1157 – MR8-J3 TO CONNECT TO RUN 1126</u>     |
| <u>2384851-2</u> | <u>RUN 1158 – MR1-CPC-J2 TO CONNECT TO RUN 1137</u> |
| <u>2384851-3</u> | <u>RUN 1159 – MR9-J4 TO CONNECT TO RUN 1146</u>     |

This page for 3T94 MAGNET (Long Bore)\*

**25kW NB RF AMPLIFIER CABINET (for sites WITHOUT LCD Display)**




**2278810-4**  
NB RF AMP  
CABINET

NOTE: The following 25kW NB RF AMP CABINET (without LCD) parts are Not Shown:

|                  |                                                 |
|------------------|-------------------------------------------------|
| <b>2336064-4</b> | 35KW RF AMP IPA TUBE & TUBE PULLER KIT          |
| <b>2336064-5</b> | 35KW RF AMP PA TUBE & HEX WRENCH KIT            |
| <b>5111003</b>   | 55-50dB DUAL DIRECTIONAL BODY RF COUPLER FOR E2 |
| <b>5111004</b>   | 45-40dB DUAL DIRECTIONAL HEAD RF COUPLER FOR E2 |


**25/35kW NB RF AMPLIFIER CABINET (for sites WITH LCD Display)**



NOTE: The following 25/35kW NB RF AMP CABINET parts are Not Shown:

|                   |                                |
|-------------------|--------------------------------|
| <b>2336064-8</b>  | 35KW RF AMP FUSE KIT           |
| <b>2336064-15</b> | 35KW RF AMP TEMPERATURE SENSOR |


\* **Note:** If the site has a 3T94 Long Bore magnet, it MUST use a 25KW NB RF Amplifier on this page.



NOTE: The following RF POWER AMP internal parts are **Not Shown**:

|                     |                                   |
|---------------------|-----------------------------------|
| <b>2206456-2</b>    | RF POWER AMP PREAMP/SPLITTER      |
| <b>2206456-3</b>    | RF AMPLIFIER (5)                  |
| <b>2206456-4</b>    | RF POWER AMP COMBINER/COUPLER     |
| <b>2206456-5</b>    | RF POWER AMP COUPLER (4)          |
| <b>2206456-6</b>    | RF POWER AMP BIAS TEE             |
| <b>2206456-13</b>   | CABLE – SPLITTER TO AMPLIFIER (5) |
| <b>2206456-20</b>   | COUPLER-BIAS TEE                  |
| <b>2206456-21</b>   | POWER BIAS TEE                    |
| <b>2206456-22</b>   | PMI INTERFACE KIT                 |
| <b>46-170021P29</b> | 5 AMP FUSES (3)                   |





**2239500 TAC**

**NOTE: The following parts are Not Shown.**

|                  |                               |
|------------------|-------------------------------|
| <b>2282045-2</b> | VACUUM PUMP TIP SEAL KIT      |
| <b>2293608</b>   | QF25 CENTERING RING & O-RING  |
| <b>2293608-2</b> | QF16 CENTERING RING & O-RING  |
| <b>2293609</b>   | QF25 VACUUM ELBOW             |
| <b>2293616</b>   | QF25 ADAPTER TO ONE INCH HOSE |
| <b>2293617</b>   | VACUUM REDUCING 'T'           |
| <b>2293618</b>   | QF16 CLAMP                    |
| <b>2293618-2</b> | QF25 CHAIN CLAMP              |
| <b>2294010-3</b> | HOSE CLAMP, SST ONE INCH      |
| <b>2297760</b>   | VACUUM GREASE, MSDS# 8366442  |
| <b>2329620-4</b> | VACUUM GAUGE CONTROLLER CABLE |
| <b>2284301</b>   | VACUUM GAUGE TO TAC J2 CABLE  |

**TWIN ACCESSORY CABINET (TAC)**


**5118162 TAC**



**NOTE: The following parts are Not Shown.**

|                  |                              |
|------------------|------------------------------|
| <b>2282045-2</b> | VACUUM PUMP TIP SEAL KIT     |
| <b>5118818</b>   | CLAMP NW 20-25               |
| <b>2297760</b>   | VACUUM GREASE, MSDS# 8366442 |

**5127147**  
SERVICE KIT FOR TAC VACUUM  
SYSTEM UPGRADE  
(Used to replace old style vacuum  
system, with solenoid)



**TAC VACUUM SYSTEM**

The table below provides information for generic replacement cables. If any of the cables are ordered, make sure to order write-on labels to be used for identifying cable part number and connection designations at the end of each cable.

#### GENERIC REPLACEMENT CABLES

| Description                                      | Length         |                |
|--------------------------------------------------|----------------|----------------|
|                                                  | 65 Feet        | 100 Feet       |
| 9 Pin Sub-D Male/Female                          | 46-244714G30   | 46-244714G37   |
| 9 Pin Sub-D Female/Female                        | 46-244713G30   | 46-244713G37   |
| 15 Pin Sub-D Male/Female                         | 46-244712G30   | 46-244712G37   |
| 15 Pin Sub-D Female/Female                       | 46-244711G30   | 46-244711G37   |
| 25 Pin Sub-D Male/Female                         | 46-244710G30   | 46-244710G37   |
| 25 Pin Sub-D Female/Female                       | 46-244709G30   | 46-244709G37   |
| 37 Pin Sub-D Male/Female                         | 46-258877G30   | 46-258877G37   |
|                                                  | <b>50 Feet</b> | <b>65 Feet</b> |
| BNC RG 223                                       | 46-251710G1    | 46-251710G18   |
| Cable Write-on Labels<br>(60 labels per booklet) | 46-208781P3    |                |

#### DESIGNATED CABLE REPLACEMENTS

| CABLE PART # | CABLE RUN                        |
|--------------|----------------------------------|
| 46-244710G30 | Run 1227<br>Run 1236             |
| 46-244712G30 | Run 1188                         |
| 46-244713G30 | Run 1181                         |
| 46-244714G30 | Run 1180<br>Run 1181<br>Run 1233 |
| 46-258877G30 | Run 1189<br>Run 1190             |

**CABLE KITS**

| <b>FRU Kit Description</b>                            | <b>Part #</b> |
|-------------------------------------------------------|---------------|
| Receive Cables for Equipment Room, 34.5 ft (10.56m)   | 5119953       |
| Receive Cables for Equipment Room, 65.6 ft (20m)      | 5119953-2     |
| Receive Cables for Magnet Room, 28.2 ft (8.6m)        | 5119954       |
| Receive Cables for Magnet Room, 49.2 ft (15m)         | 5119954-2     |
| Transportable/Relocatable MC Rcv Cable Kit (Eqmt. Rm) | 5119953-3     |
| Body Receive Cables (both rooms), 34.5 ft (10.52m)    | 5119956       |
| Body Receive Cables (both rooms), 65.6 ft (20m)       | 5119956-2     |
| Term Server Cables (4 cables)                         | 5120780       |

**FIBER OPTIC CABLE KIT – FIXED SITE**


| <b>Description</b>                                                | <b>Part #</b> |
|-------------------------------------------------------------------|---------------|
| Includes clamps, cable ties, mounts, and Runs 1045, 1083, & 1179. | 5115713       |

**TIMES MICROWAVE CABLE PARTS**


| <b>Description</b>                                 | <b>Part #</b> |
|----------------------------------------------------|---------------|
| Straight Male N Connector for LMR 600              | 2352195       |
| 7/8 EIA Right Angle Plug for LMR 1200 & Center Pin | 2384859       |

**VRE CABLE KITS**

| <b>Description</b>                                  | <b>Part #</b> |
|-----------------------------------------------------|---------------|
| Infiniband cables – Infiniband to MGD and ICN       | 5160774       |
| Ethernet cables – ICN to Ethernet, Runs 1371 & 1345 | 5160775       |

**UNIFIED SURFACE COIL PHANTOM SET**


**5342679-2**  
Large Cylindrical  
Unified Phantom




**5342680**  
Small Cylindrical  
Unified Phantom



**5342681**  
Cubical  
Unified Phantom



**5343347-2**  
TL Unified Phantom


**SERVICE TOOLS**

**MCR3 TOOL:**

For use with systems that have Hypertronics connectors only in LPCA.

**5146813**  
3.0T MCR3  
ENCLOSURE TOOL

**5146208**  
MCR3 CABLE ASM  
w/CONNECTOR

**RF TRANSMIT TERMINATION KIT:**

**5131924** 1.5T AND 3.0T RF TRANSMIT CABLE SERVICE KIT

**GRAFIDY KIT:**

**2270695**  
CuSO4 SAMPLE VIAL

**HDx TEMPERATURE AND SENSOR KIT:**

**5147136-2** ENVIRONMENTAL REPAIR KIT

**HEAD COILS:****HDx SINGLE CHANNEL SURFACE COIL ADAPTOR**

**5145746** SURFACE COIL SINGLE CHANNEL ASSEMBLY

**3T SPLIT TOP HEAD COIL**

**2376114** SPLIT HEAD COIL ASM

**5191687** HEAD COIL LATCHING FRU KIT

**2145391** MIRROR ASM

**2376104** QD BOX ASM

**2355094** REPLACEMENT SPRING CONTACTS KIT

**5140887** HDx HEAD COIL QHTR ASSEMBLY

**3T94 HEAD COIL**

**46-282118G8** 3T/94 HEAD COIL

**2376104-2** HEAD QUICK DISCONNECT CABLE FOR ECLIPSE 2



**G3 & 3T94 HDx COILS:****3T 8 CHANNEL BODY BREAST BIOPSY COIL ..... M3335LA**

|                |                       |
|----------------|-----------------------|
| <b>2417367</b> | BREAST COIL w/o CABLE |
| <b>2417366</b> | OUTPUT CABLE ASM      |
| <b>2417234</b> | PHANTOM ASM           |
| <b>2414541</b> | BIOPSY MOUNT          |
| <b>2412285</b> | BIOPSY ARM            |
| <b>2412281</b> | BREAST COIL HEADREST  |

**3T 8 CHANNEL TORSO COIL (HD CONNECTOR) ..... M3335LN**

|                |                |
|----------------|----------------|
| <b>2418063</b> | ANTERIOR COIL  |
| <b>2418064</b> | POSTERIOR COIL |
| <b>2416958</b> | CABLE ASM      |

**3T 8 CHANNEL CTL COIL (HD CONNECTOR) ..... M3335LM**

|                   |                |
|-------------------|----------------|
| <b>2417540</b>    | COIL ASM       |
| <b>2417542</b>    | CABLE ASM      |
| <b>2414538</b>    | EAR PAD SET    |
| <b>2381682-9</b>  | CTL CT PHANTOM |
| <b>2381682-10</b> | TL PHANTOM     |

**3T 3 CHANNEL SHOULDER COIL ..... M3335LR**

|                  |                    |
|------------------|--------------------|
| <b>2416878</b>   | COIL ASM           |
| <b>2416877</b>   | CABLE ASM          |
| <b>2375136-5</b> | PATIENT PAD        |
| <b>2375136-3</b> | PHANTOM POSITIONER |

**3T NV ARRAY COIL**

|                |                                 |
|----------------|---------------------------------|
| <b>2414383</b> | E3 NV ARRAY W/O CABLE           |
| <b>2414385</b> | E3 NV SYSTEM COIL               |
| <b>2414386</b> | ANTERIOR CHEST WITH CHEST CABLE |
| <b>2414389</b> | NV PHANTOM SET                  |
| <b>2414391</b> | E2 NV SYSTEM CABLE              |

**3T BRAIN/SPINE COIL ..... M3335LC (FP) & M3335LD (U/G)**

|                |                                       |
|----------------|---------------------------------------|
| <b>2417351</b> | HEAD NECK UNIT (A & P)                |
| <b>2418075</b> | HNU ADAPTOR BLOCK                     |
| <b>2417350</b> | NECK CHEST UNIT                       |
| <b>2417349</b> | TL UNIT                               |
| <b>2417352</b> | OUTPUT CABLE – HYPERTRONICS (UPGRADE) |
| <b>2417353</b> | OUTPUT CABLE – BENDIX (UPGRADE)       |
| <b>2417355</b> | OUTPUT CABLE (FORWARD PRODUCTION)     |
| <b>2417354</b> | TL UNIT CABLE                         |
| <b>2418080</b> | MIRROR                                |

**1.5T BRAIN/SPINE ACCESSORIES**

|                |                                 |
|----------------|---------------------------------|
| <b>2418077</b> | TL PHANTOM KIT                  |
| <b>2418079</b> | HN PHANTOM KIT                  |
| <b>2418081</b> | PAD KIT                         |
| <b>5344844</b> | PHANTOM POSITIONER, HNS-UNIFIED |

COILS CONTINUED ON NEXT PAGE.

**G3 & 3T94 HD COILS:****3T 8 CHANNEL BRAIN COIL**

|                  |                                 |
|------------------|---------------------------------|
| <b>2380637-2</b> | BRAIN COIL/MIRROR/BASEPLATE ASM |
|------------------|---------------------------------|

|                  |                    |
|------------------|--------------------|
| <b>2317112-3</b> | PHANTOM POSITIONER |
|------------------|--------------------|

|                  |           |
|------------------|-----------|
| <b>2380637-5</b> | CABLE ASM |
|------------------|-----------|

|                  |            |
|------------------|------------|
| <b>5146634-6</b> | MIRROR ASM |
|------------------|------------|

**3T QUAD EXTREMITY COIL**

|                  |                       |
|------------------|-----------------------|
| <b>2380635-2</b> | KNEE/FOOT ASM W/CABLE |
|------------------|-----------------------|

|                  |                     |
|------------------|---------------------|
| <b>2380635-5</b> | KNEE/FOOT CABLE ASM |
|------------------|---------------------|

|                   |                      |
|-------------------|----------------------|
| <b>2357819-10</b> | EXTREMITY REPAIR KIT |
|-------------------|----------------------|

|                   |                    |
|-------------------|--------------------|
| <b>2357819-11</b> | PHANTOM POSITIONER |
|-------------------|--------------------|

|                   |         |
|-------------------|---------|
| <b>2357819-12</b> | PHANTOM |
|-------------------|---------|

**3T GP FLEX COIL**

|                |              |
|----------------|--------------|
| <b>2382817</b> | GP FLEX COIL |
|----------------|--------------|

|                |                        |
|----------------|------------------------|
| <b>2389979</b> | GP FLEX COIL CABLE KIT |
|----------------|------------------------|

**G3 HD COILS:****3T 8 CHANNEL TORSO ARRAY COIL**

|                  |                 |
|------------------|-----------------|
| <b>2381683-2</b> | COIL WITH CABLE |
|------------------|-----------------|

|                  |         |
|------------------|---------|
| <b>2381683-3</b> | PHANTOM |
|------------------|---------|

|                  |                    |
|------------------|--------------------|
| <b>2382683-4</b> | PHANTOM POSITIONER |
|------------------|--------------------|

**3T CTL ARRAY COIL**

|                  |                   |
|------------------|-------------------|
| <b>2413107-1</b> | CTL ARRAY W/CABLE |
|------------------|-------------------|

|                  |                 |
|------------------|-----------------|
| <b>2413107-4</b> | CTL ARRAY CABLE |
|------------------|-----------------|

|                  |                        |
|------------------|------------------------|
| <b>2381682-9</b> | CTL PHANTOM CT SECTION |
|------------------|------------------------|

|                   |                        |
|-------------------|------------------------|
| <b>2381682-10</b> | CTL PHANTOM TL SECTION |
|-------------------|------------------------|

**3T94 HD COILS:****3.0T 4 CHANNEL TORSO ARRAY COIL**

|                |                         |
|----------------|-------------------------|
| <b>2366942</b> | 3T TORSO ARRAY ASSEMBLY |
|----------------|-------------------------|

|                      |                                         |
|----------------------|-----------------------------------------|
| <b>5110897 or -2</b> | 3T/94 EXCITE 4 CHANNEL TORSO COIL CABLE |
|----------------------|-----------------------------------------|

**3.0T NV ARRAY COIL**

|                  |                 |
|------------------|-----------------|
| <b>2383478-2</b> | NV ARRAY W/COIL |
|------------------|-----------------|

|                  |                 |
|------------------|-----------------|
| <b>2383477-3</b> | NV COIL PHANTOM |
|------------------|-----------------|

|                  |                            |
|------------------|----------------------------|
| <b>2383477-4</b> | NV COIL PHANTOM POSITIONER |
|------------------|----------------------------|

|                  |              |
|------------------|--------------|
| <b>2383478-5</b> | NV CABLE ASM |
|------------------|--------------|

**3.0T 4 CHANNEL CTL COIL**

|                  |                 |
|------------------|-----------------|
| <b>2380634-2</b> | 3T/94 CTL ARRAY |
|------------------|-----------------|

|                  |                             |
|------------------|-----------------------------|
| <b>2377326-3</b> | USA1 8 CHANNEL COIL PHANTOM |
|------------------|-----------------------------|

|                      |                                        |
|----------------------|----------------------------------------|
| <b>5110896 or -2</b> | 3T/94 EXCITE 4 CHANNEL CTL ARRAY CABLE |
|----------------------|----------------------------------------|

**3.0T 8 CHANNEL CTL ARRAY COIL**

|                  |                   |
|------------------|-------------------|
| <b>2413107-1</b> | CTL ARRAY W/CABLE |
|------------------|-------------------|

|                  |                  |
|------------------|------------------|
| <b>2413107-4</b> | CTL OUTPUT CABLE |
|------------------|------------------|

COILS CONTINUED ON NEXT PAGE.

|                                       |                                               |                |
|---------------------------------------|-----------------------------------------------|----------------|
| <b>3.0T 8 CHANNEL WRIST COIL</b>      |                                               | <b>M3340CA</b> |
| <b>5268024-2</b>                      | 3T HD 8 CH WRIST COIL W/ BASEPLATE & CABLE    |                |
| <b>5268024-3</b>                      | 3T HD 8 CH WRIST ARRAY COIL CABLE             |                |
| <b>3.0T 8 CHANNEL FOOT ANKLE COIL</b> |                                               | <b>M3340CB</b> |
| <b>5273235-2</b>                      | 3T HD 8CH FOOT ANKLE COIL W/ CABLE. ELEC ASSY |                |
| <b>5273235-3</b>                      | 3T HD 8CH FOOT ANKLE COIL CABLE               |                |
| <b>5273235-5</b>                      | HD 8 CH FAC PHANTOM ASSEMBLY                  |                |
| <b>5273235-6</b>                      | 3T HD 8 CH FOOT ANKLE COIL BASEPLATE ASSY     |                |
| <b>3.0T 8 CHANNEL SHOULDER COIL</b>   |                                               | <b>M3340CC</b> |
| <b>5271265-2</b>                      | 3T HD 8 CH SHOULDER COIL ASSY WITH CABLE      |                |
| <b>5271265-3</b>                      | HD 8 CH SHOULDER COIL CABLE ASSEMBLY          |                |
| <b>5271265-5</b>                      | 8 CH SHOULDER COIL PHANTOM HOLDER ASSY        |                |

**BrainWave OPTIONS:**

|                   |                                                                                                                                                 |
|-------------------|-------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>2363727</b>    | 15" LCD Display with 5 foot VGA Cable for Computer connection                                                                                   |
| <b>2394034-5</b>  | (Cedrus) Response Pad Asm. Left & right hand units with two buttons for each hand. Includes fiber optic cable (20 ft) and optical converter box |
| <b>2394034-6</b>  | (Cedrus) "Lumina" response controller box                                                                                                       |
| <b>2394034-7</b>  | (Cedrus) AC Power module for Lumina Controller Box                                                                                              |
| <b>2394034-9</b>  | FRU usage only Kit containing all cables [Cedrus-4, -5, -6, -7, 2xxxxxx-6, 46-265505P1 (qty 2)]                                                 |
| <b>2394034-10</b> | Penetration Sub-Panel Asm                                                                                                                       |
| <b>2394034-16</b> | PC compatible Keyboard (narrow)                                                                                                                 |
| <b>2394034-17</b> | PC compatible Mouse                                                                                                                             |
| <b>5160492</b>    | BrainWave 2.0 Application Software CD-ROM                                                                                                       |
| <b>5110736</b>    | EXCITE II Scanner License Key MOD for the BrainWave H/W Option                                                                                  |
| <b>5173371-2</b>  | PC and Platform Software Asm (For HDx Only)                                                                                                     |
| <b>5110968</b>    | Patch CD for BrainWave Stimulus PC                                                                                                              |

MNS OPTIONS:

| Description           | Part #    |
|-----------------------|-----------|
| 3.0T 13C TR Switch    | 2354050-6 |
| 3.0T 13C MNS Preampl  | 5114799-2 |
| MNS QA Probe 3.0T 13C | 5114403-3 |
| MNS QA Probe 3.0T 31P | 5114978-4 |
| 3.0T 13C Sample Vial  | 5114978-3 |
| 3.0T 31P Sample Vial  | 5114978-4 |

OPTIONS:

|                         |                                                |
|-------------------------|------------------------------------------------|
| CADSTREAM 4.1 w/SURELOC | M3335KL                                        |
| 5164701                 | CONFIRMA 4.1 SERVER PC & S/W w/MONITOR & MOUSE |
| 5160894                 | CADSTREAM S/W & DOCS CD-ROM                    |
| 5154402                 | CADSTREAM 3.X TO 4.1 UPGRADE                   |
| CONFIRMA MAMMO CAD      | M3033NF                                        |
| 5117875                 | COMPLETE PC WITH SOFTWARE                      |
| BAR CODE READER KIT     | M1090PM                                        |
| 2180690-2               | READER_ONLY                                    |
| 2180690-3               | Y CABLE_ONLY                                   |
| ReportCARD HARDWARE     |                                                |
| 5117866-26              | ReportCARD COMPUTER                            |
| 5160890                 | ReportCARD SOFTWARE COLLECTOR                  |
| 2390369                 | ReportCARD OPTON KEY                           |
| FUNCTOOL                |                                                |
| 5160880                 | FUNCTOOL 4.0 APPLICATIONS CD                   |

**EXCITE SOFTWARE OPTION KEYS:**

|                   |                                           |
|-------------------|-------------------------------------------|
| <b>2156158-64</b> | Release 15.x Software Collector HDxt (M3) |
| <b>2156158-65</b> | Release 14.x Software Collector HDx (M5)  |
| <b>2383494</b>    | Fiesta C                                  |
| <b>2390037</b>    | Asset                                     |
| <b>2390040</b>    | ConnectPro                                |
| <b>2390075</b>    | 3D FRFSE                                  |
| <b>2390078</b>    | EchoPlus                                  |
| <b>2390079</b>    | Diffusion Tensor                          |
| <b>2390084</b>    | Probe 2000                                |
| <b>2390085</b>    | Probe 200 Upgrade                         |
| <b>2390086</b>    | Sage 7                                    |
| <b>2390091</b>    | 2D Delayed Enhancement                    |
| <b>2390095</b>    | iDrive ProPlus                            |
| <b>2390097</b>    | SmartPrep 2000                            |
| <b>2390098</b>    | Flouro Trigger MRA                        |
| <b>2390103</b>    | Performed Proc Step                       |
| <b>5110228</b>    | Single Voxel Probe-P                      |
| <b>5110229</b>    | Single Voxel Probe-S and Probe-P          |
| <b>5110230</b>    | Probe 2D – CSI                            |
| <b>5110231</b>    | Probe 3D – CSI                            |
| <b>5110232</b>    | MNS, SAGE, Probe-P                        |
| <b>5110319</b>    | Research Proton Spectro                   |
| <b>5120544</b>    | Release 12.0 M4 Apps DVD                  |
| <b>5121548</b>    | ReportCard on AW                          |
| <b>5121550</b>    | Tractography on AW                        |
| <b>5122363</b>    | LAVA                                      |
| <b>5122364</b>    | PFO Evaluation                            |
| <b>5122365</b>    | PROPELLOR Brain Suite                     |
| <b>5122366</b>    | Tractography                              |
| <b>5122367</b>    | 2D FatSat FIESTA                          |
| <b>5122368</b>    | 3D Flair                                  |
| <b>5122370</b>    | RealTime FIESTA                           |
| <b>5122371</b>    | VIBRANT-II                                |
| <b>5122372</b>    | 8 Channel HDx                             |
| <b>5122373</b>    | 16 Channel HDx                            |
| <b>5122375</b>    | 3.0T HDx ScanTools 14.0                   |

**SOFTWARE OPTIONS**

## REVISION HISTORY

### REV/DATE REASON FOR CHANGE

- 1 Mar 03, 06 Initial release
- 2 Jun 26, 06 Added HDx new color RF Coil. Updated ethernet switch part number to 5162073 for 24 port.
- 3 Jan 26, 07 Added RF Shield Change Out Kit, Speaker Grill nylon screw, LPCA Wheel Housing FRU kit (MRIhc21558), & Dock pedal FRU parts (MRIhc17629). Corrected Linux 3-button mouse part number. For RFS cabinet, added Term Server cable FRU kit for and RRF Module MUX2 board. Removed Infiniband Switch and MCR2 Tool.
- 4 Mar 13, 07 Added front cover laser, 5 gallon coolant pail, and LPCA LED cable. Added note for referring to 2138247 for parts not in this manual. Updated Ethernet cable info. Added center pin to description of Times Microwave right angle connector (CQA 13099582). Removed ECMT kit (CQA 13097585). Added Head Coil Latch Kit. Updated BrainWave s/w asm part # to 5173371.
- 5 Jun 7, 07 Removed erroneous LED cable from LPCA. Updated APS & APG Module parts numbers in MGD to 2294300-16.
- 6 Sep 17, 07 Moved Revision History to last page. Updated HFA-S Asm part # in HFD-S cabinet (ECO 20428110). Added RFS Cabinet with CAM Module and CAM Module board details. Revised MGD & ASC Module part #'s. Added LPCA bracket, pillar, and screws FRU kit (PQR 13121160), Patient Table bumper and panel FRU kit (PQR 13127363), Grafidy CuSO4 vial (PQR 13098795), and Brain Spine accessories. Added NV coil FRU parts (PQR 13113173)
- 7 May 28, 08 Changed name of manual to *Signa HDx 3.0T Series FRU's*. Rear Pedestal – updated Longitudinal Encoder part number (PQR 13175398) and added Encoder board part # 46-226552G2. Added Lytron pump coupler part # (PQR 13136000). Corrected 25kw NB RF Amp part # (PQR13142203). HFD: Updated FHA Control Bd part #. RFS Cab: Added Sun ICN part # and replacement note. Updated SCP2 Bd part #. Added 5181495-3 MGD P.S. NB RF Amp: Corrected MKS 25kw Amp part # (PQR 13142203). Tools: Added RF Transmit Termination kit. Coils: Added Mirror assembly to Brain Coil (ECR 203882). Options: Updated BrainWave software part #.
- 8 Sep 23, 08 Added HDxt parts to OW, coils, and software sections. Updated Remote Intercom Board part number. Added EMI Filter Asm to Pen Panel (ECO 2060894). Updated PHPS FRU's on Pen Panel (ECO 2040374). Added Linux PC module for 14M5 s/w. Added Sun hard drive FRU. Updated part # for Environmental Repair Kit.
- 9 Jun 15, 09 Added Flexible Coupling (PQR 13206931) and Depopulated 16-CH Switch to Rear Pedestal (ECO 2063089). Updated Filter FRU's on PEN Panel (ECO 2084029). Updated Monitor offerings on OW (ECO 2047560). Also added HP9300 DIMM (ECR 2078642), HP8400 DIMM (ECR 2038518) & DVD-RW (ECR 2059074). In RRF Chassis, added 5250028-2 RF-DIF (ECO 2080729).
- 10 Nov 3, 09 Updated OW LCD monitor to 5343103 (ECO 2083909) and added new GOC AA 2395043-2 (ECO 2084000). Added 2294300-22 SCP3 Bd. to CAM and MGD modules (ECO 2064203, MRIhc42663). Added 2294300-23 APS2 bd to CAM (ECO 2064898).
- 11 Jun 28, 10 Added Rear Bridge Plate Cover (ECO 20635630). Added O2 Monitor and MRU (ECO 2076814) and Step-Down Transformer (ECO 2091712). Updated Host Monitor to 5245466. Added HP Z400 computer (ECO 2056587). Added HP8400 alternate replacement (ECO 2101300). Added MDP battery and charger FRU's (ECO 2097255). Added new 16 & 32 channel ICN's (ECO 2101939). Added Unified Surface Coil Phantom set (ECO 2081133).