Measurement protocol

IAP RAS

September 23, 2025

Abstract

This is the debug template for measurements reports of detector calibration process.

1 Parameters

This is the parameters section.

RF frequencies	[2450000000.0, 0.0, 1]
RF levels	[-5.0, 15.0, 51]
SPAN wide	2000000.0
SPAN narrow	100000.0
RBW wide	10000.0
RBW narrow	1000.0
VBW wide	10000.0
VBW narrow	1000.0
REF level	15.0
SWEEP points	1001.0
SWEEP time	AUTO ON
Precise	True
HOR scale	0.01
High res	True
Impedance 50Ohm	True
Coupling DC	True
Channel	4

Table 1: Parameters