

# National Health and Nutrition Examination Survey

## 2017-March 2020 Data Documentation, Codebook, and Frequencies

### Chromium - Urine (P\_UCM)

**Data File:** P\_UCM.xpt

**First Published:** November 2021

**Last Revised:** NA

---

## Component Description

The NHANES program suspended field operations in March 2020 due to the coronavirus disease 2019 (COVID-19) pandemic. As a result, data collection for the NHANES 2019-2020 cycle was not completed and the collected data are not nationally representative. Therefore, data collected from 2019 to March 2020 were combined with data from the NHANES 2017-2018 cycle to form a nationally representative sample of NHANES 2017-March 2020 pre-pandemic data. These data are available to the public. Please refer to the Analytic Notes section for more details on the use of the data.

### Chromium

Chromium is considered both an essential element for carbohydrate metabolism and toxic (Cefalu et al., 2004). Chromium can be introduced into the human body via consumption of plants or supplements, degradation of metal-on-metal implants, and industrial exposure. Inhalation of chromium(VI) can damage the nose and cause lung cancer, while chromium(VI) in drinking water has led to increased stomach ulcers (ATSDR, 2012). This method is used to achieve rapid and accurate quantification of chromium. Use this method to screen urine when people are suspected to be acutely exposed to these elements or to evaluate chronic environmental or other non-occupational exposure (Jones, 2021).

## Eligible Sample

All examined participants aged 3 to 5 years and a one-third subsample of examined participants aged 6 years and older in the NHANES 2017-March 2020 pre-pandemic sample were eligible.

## Description of Laboratory Methodology

This method (Jones, 2021) directly measures the chromium, arsenic, and nickel content of urine specimens using inductively coupled plasma mass spectrometry (ICP-MS) after a simple dilution sample preparation step. Liquid samples are reduced to small droplets in an argon aerosol via a nebulizer, and then the droplets enter the ICP where the thermal energy atomizes the sample and then ionizes the atoms. The ions are pulled into the mass spectrometer where they are focused through the universal cell technology (UCT), the quadrupole mass filter, and finally are selectively counted in rapid sequence at the detector according to their mass-to-charge (m/z) ratio. When detecting chromium, the UCT is filled with ammonia gas to eliminate polyatomic interferences. Quantification of chromium is accomplished by comparing the blank-subtracted counting rate of chromium from a sample, ratioed to an internal standard, to the counting rate, ratioed to an internal standard, of blank-subtracted, matrix-matched, external calibrators, tested within the same analytical run.

Refer to the Laboratory Method Files section for a detailed description of the laboratory methods used.

## Laboratory Method Files

[Arsenic, Chromium, and Nickel Urine Lab Procedure Manual](#) (August 2021)

[Arsenic, Chromium, and Nickel Urine Lab Procedure Manual](#) (November 2021)

## Laboratory Quality Assurance and Monitoring

Urine samples are processed, stored, and shipped to the Division of Laboratory Sciences, National Center for Environmental Health, Centers for Disease Control and Prevention, Atlanta, GA for analysis.

Detailed instructions on specimen collection and processing are discussed in the [2017-2018](#) and [2019-2020](#) NHANES Laboratory Procedures Manuals (LPMs). Vials are stored under appropriate frozen ( $-30^{\circ}\text{C}$ ) conditions until they are shipped to National Center for Environmental Health for testing.

The NHANES quality assurance and quality control (QA/QC) protocols meet the 1988 Clinical Laboratory Improvement Amendments mandates. Detailed QA/QC instructions are discussed in the NHANES LPMs.

### Mobile Examination Centers (MECs)

Laboratory team performance is monitored using several techniques. NCHS and contract consultants use a structured competency assessment evaluation during visits to evaluate both the quality of the laboratory work and the QC procedures. Each laboratory staff member is observed for equipment operation, specimen collection and preparation; testing procedures and constructive feedback are given to each staff member. Formal retraining sessions are conducted annually to ensure that required skill levels were maintained.

### Analytical Laboratories

NHANES uses several methods to monitor the quality of the analyses performed by the contract laboratories. In the MEC, these methods include performing blind split samples collected on “dry run” sessions. In addition, contract laboratories randomly perform repeat testing on 2% of all specimens.

NCHS developed and distributed a QC protocol for all CDC and contract laboratories, which outlined the use of Westgard rules (Westgard et al, 1981) when running NHANES specimens. Progress reports containing any problems encountered during shipping or receipt of specimens, summary statistics for each control pool, QC graphs, instrument calibration, reagents, and any special considerations are submitted to NCHS quarterly. The reports are reviewed for trends or shifts in the data. The laboratories are required to explain any identified areas of concern.

All QC procedures recommended by the manufacturers were followed. Reported results for all assays meet the Division of Laboratory Sciences' QA/QC criteria for accuracy and precision, similar to the Westgard rules (Caudill et al, 2008).

## Data Processing and Editing

The data were reviewed. Incomplete data or improbable values were sent to the performing laboratory for confirmation.

## Analytic Notes

The COVID-19 pandemic required suspension of NHANES 2019-2020 field operations in March 2020 after data were collected in 18 of the 30 survey locations in the 2019-2020 sample. Data collection was cancelled for the remaining 12 locations. Because the collected data from 18 locations were not nationally representative, these data were combined with data from the previous cycle (2017-2018) to create a 2017-March 2020 pre-pandemic data file. A special weighting process was applied to the 2017-March 2020 pre-pandemic data file. The resulting sample weights in the demographic data file should be used to calculate estimates from the combined cycles. These sample weights are not appropriate for independent analyses of the 2019-2020 data and will not yield nationally representative results for either the 2017-2018 data alone or the 2019-March 2020 data alone. Please refer to the NHANES website for additional information for the NHANES 2017-March 2020 pre-pandemic data, and for the previous 2017-2018 public use data file with specific weights for that 2-year cycle.

Refer to the [2017-2018](#) and [2019-2020 Laboratory Data Overview](#) documents for general information on NHANES laboratory data.

There are over 800 laboratory tests performed on NHANES participants. However, not all participants provided biospecimens or enough volume for all the tests to be performed. The specimen availability can also vary by age or other population characteristics. Analysts should evaluate the extent of missing data in the dataset related to the outcome of interest as well as any predictor variables used in the analyses to determine whether additional re-weighting for item non-response is necessary.

Please refer to the NHANES [Analytic Guidelines](#) and the on-line NHANES [Tutorial](#) for further details on the use of sample weights and other analytic issues.

## Subsample Weights

The analytes included in this dataset were measured in all examined participants aged 3-5 years, and in a one-third subsample of participants 6 years and older. Special sample weights are required to analyze these data properly. Variable (WTSAPRP) encoding of the specific sample weights for this subsample is included in this data file and should be used when analyzing these data. These special sample weights were created to account for the subsample selection probability, as well as the additional nonresponse to these lab tests. Therefore, if participants were eligible for the subsample, but did not provide a urine specimen, they would have the sample weight value assigned as "0" in their records.

## Demographic and Other Related Variables

The analysis of NHANES laboratory data must be conducted using the appropriate survey design and demographic variables. The [NHANES 2017- March 2020 Pre-Pandemic Demographics File](#) contains demographic data, health indicators, and other related information collected during household interviews as well as the sample design variables. The recommended procedure for variance estimation requires use of stratum and PSU variables (SDMVSTRA and SDMVPSU, respectively) in the demographic data file.

This laboratory data file can be linked to the other NHANES data files using the unique survey participant identifier (i.e., SEQN).

Starting in the 2015-2016 NHANES cycle, the variable URXUCR (urine creatinine) will not be reported in this file. URXUCR can be found in the data file titled Albumin & Creatinine – Urine.

## Detection Limits

The detection limits were constant for all of the analytes in the data set. Two variables are provided for each of these analytes. The variable name ending in "LC" (ex., URDUCMLC) indicates whether the result was below the limit of detection: "0" means that the result was at or above the limit of detection, "1" indicates that the result was below the limit of detection. The other variable prefixed URX (ex., URXUCM) provides the analytic result for that analyte. For analytes with analytic results below the lower limit of detection (ex., URDUCMLC=1), an imputed fill value was placed in the analyte results field. This value is the lower limit of detection divided by square root of 2 (LLOD/sqrt [2]).

The lower limit of detection (LLOD, in µg/L) for urinary chromium is:

Variable Name	Analyte Description	LLOD
URXUCM	Urinary Chromium	0.19

## References

- Agency for Toxic Substances and Disease Registry (ATSDR), 2012. Toxicological profile for chromium. Atlanta, GA: U.S. Department of Health and Human Services, Public Health Service.
- Caudill S.P., Schleicher R.L., Pirkle J.L. Multi-rule quality control for the age-related eye disease study. Statist. Med. (2008) 27(20:4094-4106.
- Cefalu, W.T. and F.B. Hu, Role of chromium in human health and in diabetes. Diab. Care, 2004. 27: p. 2741-2751.
- Jones, D.R., et al. "Biomonitoring method for measurement of As, Cr, and Ni in human urine samples by ICP-UCT-MS." International Journal of Hygiene and Environmental Health. Volume 234, May 2021, 113713.
- Westgard J.O., Barry P.L., Hunt M.R., Groth T. A multi-rule Shewhart chart for quality control in clinical chemistry. Clin Chem (1981) 27:493-501.

## Codebook and Frequencies

### SEQN - Respondent sequence number

<b>Variable Name:</b>	SEQN
<b>SAS Label:</b>	Respondent sequence number
<b>English Text:</b>	Respondent sequence number
<b>Target:</b>	Both males and females 3 YEARS - 150 YEARS

## WTSAPRP - Subsample A Weights Pre-Pandemic

**Variable Name:** WTSAPRP  
**SAS Label:** Subsample A Weights Pre-Pandemic  
**English Text:** Subsample A Weights Pre-Pandemic  
**Target:** Both males and females 3 YEARS - 150 YEARS

Code or Value	Value Description	Count	Cumulative	Skip to Item
2395.195359 to 955677.30961	Range of Values	4727	4727	
0	No Lab Specimen	163	4890	
.	Missing	0	4890	

## URXUCM - Chromium, Urine (ug/L)

**Variable Name:** URXUCM  
**SAS Label:** Chromium, Urine (ug/L)  
**English Text:** Chromium, Urine (ug/L)  
**Target:** Both males and females 3 YEARS - 150 YEARS

Code or Value	Value Description	Count	Cumulative	Skip to Item
0.13 to 106.56	Range of Values	4569	4569	
.	Missing	321	4890	

## URDUCMLC - Chromium, Urine Comment Code

**Variable Name:** URDUCMLC  
**SAS Label:** Chromium, Urine Comment Code  
**English Text:** Chromium, Urine Comment Code  
**Target:** Both males and females 3 YEARS - 150 YEARS

Code or Value	Value Description	Count	Cumulative	Skip to Item
0	At or above detection limit	1870	1870	
1	below lower detection limit	2699	4569	
.	Missing	321	4890	