

Frankfurt University of Applied Science
Faculty 2: Computer Science and Engineering
Vietnamese - German University
Department of Electrical Engineering and Information Technology

DESIGN AND IMPLEMENTATION OF A WIRELESS FALL DETECTION NETWORK PROTOTYPE USING MEMS SENSORS

by

Ho Ngoc Khang Minh

Matriculation No.: 1148602

First Supervisor: Prof. Dr.-Ing. Kira Kastell
Second Supervisor: Dipl. -Ing Mohammad Reza Mansooji

Submitted in partial fulfillment of the requirements
for the degree of Bachelor Engineering in study program
Electrical Engineering & Information Technology,
Vietnamese - German University

Frankfurt am Main, Germany 2018

Design and Implementation of a Wireless Fall Detection Network prototype using MEMS sensors

Approved by

Prof. Dr.-Ing Kira Kastell, First Supervisor

Dipl.-Ing Mohammad Reza Mansooji

Disclaimer

I hereby declare that this thesis is a product of my own work, unless otherwise referenced. I also declare that all opinions, results, conclusions and recommendations are my own and may not represent the policies or opinions of Vietnamese - German University and Frankfurt University of Applied Science.

Ho Ngoc Khang Minh

©2018 by Ngoc Khang Minh, Ho. All right reserved.

Abstract

Falling is becoming a serious problem to elderly people with the age of over 65, often causing unpredictable injuries such as hip fractures, head traumas, etc. More seriously, falls may lead to disability or even death on the victim if assistances from caregivers are not received in time. From this situation, there is a need for a wireless communication network which can automatically detect the fall and send an alarm to the caregivers if there is no safe signal from the victim within 10 minutes. There have been several algorithms such as detection of body orientation after a fall, image processing to detect the fall or applying machine learning techniques (Support Vector Machine (SVM), Markov model) in classifying between falling and other activities of daily living (ADL). However, these complex techniques require a huge amount of computations which can result in the system being overloaded or heavily delayed.

Addressing this problem calls for less computation-intensive techniques while retaining the accuracy and robustness of the system. One such approach is the combination of data from both accelerometer and gyroscope. The focus of this thesis is developing a wireless fall detection network which can combines the data from above sensors to detect falls and distinguish them from fall-like activities. A comparison on the performance of this network with other existing works is also included to evaluate the robustness of the system.

Keywords: *elderly people, fall detection, accelerometer, gyroscope, wireless communication network*

Acknowledgements

I would like to express my sincere appreciation to Prof. Dr.-Ing. Kira Kastell, Vice President of Frankfurt University of Applied Science (FRA-UAS) for supervising, supporting and encouraging me during the time I conduct my senior project and final thesis. Also, I would like to thank Mr. Dipl.-Ing Mansooji - communication lab engineer at FRA-UAS, Dr. Vo Bich Hien - lecturer and Mr. Duong Huynh Bao - former control lab engineer at Vietnamese German University for guiding me on technical and programming aspect. I also would like to thank my family for their supports during my bachelor degree time. This project will not be completed without their grateful guidances and supports. Thanks to their encouragements, I can improve myself, not only on the fundamental background in my major, but also necessary skills for my future career. I would like to mention all of my seniors and classmates who have helped me through this thesis.

Table of Contents

List of Tables	viii
List of Figures	ix
1 Introduction	2
1.1 Introduction to Fall	2
1.2 Definition of a Fall	2
1.3 Cause of a Fall	2
1.3.1 Intrinsic risk factors	3
1.3.2 Extrinsic risk factors	4
1.4 Consequence of Falls	4
1.5 Challenges in Detecting Falls	6
1.6 Objectives	7
1.7 Thesis Organization	7
2 Literature Review	8
2.1 Existing Commercial Devices	8
2.2 Existing Products from Other Research Groups	9
3 Statement of Problem and Methodology	11
3.1 Problem	11
3.2 Methodology	11
3.2.1 Algorithm	12
3.2.2 Position of sensor	15
4 System overview	16
4.1 System Components	16
4.1.1 Hardware Components	16
4.1.1.1 Sensor	16

4.1.1.2	Microcontroller	16
4.1.1.3	Wireless Module	16
4.1.2	Software	16
4.1.2.1	Embeddded Software on Sensor Node	16
4.1.2.2	Communication System	16
4.1.2.3	Data Acquisition Software	16
4.1.2.4	Data Analysis Software	16
4.1.3	System Intergration	16
5	Experimental and Procedure	17
5.1	Data Collection	17
5.2	Data Analysis	17
5.3	Experiment on Real Subjects to Evaluate the Performance of The Network	17
6	Results and Discussion	18
6.1	Results Assessment	18
6.2	Comparing the performance with existing works	18
7	Conclusion and Future Works	19
7.1	Summary	19
7.2	Limitations	19
7.3	Future works	19
	APPENDICES	19
A	C/C++ Code for Reading Data from Sensor	20
B	Python Code for Data Acquisition Program	25
	References	27

List of Tables

List of Figures

2.1	Medical Guardian Fall Alert System	8
2.2	myHalo Medical Alert System from MobileHelp TM	9
2.3	TEMPO 3.0 sensor node	10
2.4	The final design of device from group research of Binh Nguyen and Jonathan Tomkun	10
3.1	Fall detection algorithm	14

Acronyms

NCOA the National Council of Aging. 4

U.S. United State of America. 5, 8

WHO World Health Organization. 2, 5

1. Introduction

1.1 Introduction to Fall

Nowadays, fall is becoming a dangerous issue which mostly causes the injuries and even leads to disability or fatal death on human, especially the elderly. According to Hwang *et.al.* [1], in the United State, one-third to one-half of elderly people over 65 years old fall at least one time each year and two-third of them will do so again within the next 6 months. Every 11 seconds, there is an elderly person is treated in the hospital's emergency area due to fall-related injuries and every 19 minutes, one faller die [2]. It is also reported that one out every 200 falls results in a hip fracture on people with age among 65 and 69 and increase to one out of 10 for those aged 85 and more [3]. The most profound effect of falling is the loss of functioning associated with the dependency of the elderly for the rest of their life. Besides, a great amount time and money have been spent on the medical treatments for the falls. Approximately \$179 million were used as direct medical costs for treating fatal falls and \$19 billion for non-fatal fall injuries within 2000 [4].

1.2 Definition of a Fall

Before starting a research about falls, it is necessary to understand about the meaning of the term *falls*. According to World Health Organization (WHO), a fall is defined as an event which results in a person coming to rest inadvertently on the ground or floor or other lower level with or without loss of consciousness or injury [5].

1.3 Cause of a Fall

In recent years, a lot of researches have been done by different groups to find out the causes of falls. There have been different ways to classify the causes of a fall such as age and sex, drugs, cognitive functions, postural control, etc. However, because falling is an unintentional action, many causes usually combine to produce a fall. Therefore, these causes can be divided into two main categories, intrinsic and extrinsic, to ease the complication of research activities.

1.3.1 Intrinsic risk factors

Intrinsic factors are the factors which are within the body. It is also the physical aspect of the body that can cause injuries. Intrinsic factors include physical diseases such as cognitive impairment, postural hypotension, cardiovascular problems, etc.

Cognitive Impairment

It is well recognized that cognitive impairment is the common cause of fall on human, especially the elderly. Due to the research of Prudheim *et al.* in 1981, fallers have in general been found to have a higher prevalence of cognitive impairment than non-fallers [6]. It is also reported that human with dementia are approximately 3 times more likely to fall than non-demented [7]. The reason why patients with cognitive impairment are likely to fall is that they have increased reaction time, increased postural sway and increased leaning balance which result in the decrease on muscle strength, worse balance and poorer mobility.

Postural Control

Postural control is a complex motor skill that requires interaction of multiple body systems which results in the ability to maintain postural orientation or postural stability [8][9]. The impairment of postural control is indicated as one of significant causes of the fall during any activities at any age. Impaired control of gait and balance are two main aspects of the postural control that have been considered in many studies about falls. Subjectively assessed gait has been reported to be abnormal in many fall activities and other studies using more objective measurements have found some relations between the impaired gait and balance and risk of falling [10].

Cardiovascular problem

Cardiovascular disorders are responsible for approximately 77% of patients with unexplained or recurrent falls and falls associated with unexplained loss of consciousness [11]. There is a fact that fallers with cardiovascular disorders have a greater mortality than those with non-cardiovascular or unknown causes [12].

1.3.2 Extrinsic risk factors

Extrinsic factors are those related to the environment such as lighting, walking surface, loose carpets, and high or narrow steps.

Dim lighting or glare

Dim lighting is one significant cause of fall on human, especially on the elderly. While walking on the low light condition, the patient's visibility is reduced which prevents them from detecting the obstacles on the walking path and causing stumble and fall. Too bright lighting can also cause fall on human because of creating glares and distorting the way object look.

Bad staircase design

Bad staircase design is also another extrinsic risk that cause the fall on human. Too high or too low rise of staircases have caused a number of falls because people fail to perceive the abnormal elevation change or incur a misstep on descent. Besides, slippery surface of the treads can cause slip on patients while walking up or down the stairs. One more mistake of staircase design that can lead to the fall is lighting on the stair. Poor visibility and inadequate lighting can cause a user to misread the stair edge, resulting in faulty foot placement and falls.

1.4 Consequence of Falls

The consequences of falls are serious to human at any age in any circumstance, but to the elderly, they have significance beyond that in younger people. There are different consequences related to the falls on human including physical effects, psychological effects and other consequences such as dependency, hospital admission or economic consequences.

Physical consequences

Physical consequences of falls are always a significant aspect that many scientists have focused on in recent years. According to the the National Council of Aging (NCOA) of the U.S., falls are the leading cause of fatal injuries and the most common cause of non-fatal traumas [2]. There are several physical injuries such as bruise, fracture or head injury

reported to happen with patients who suffer from falls. In such injuries, hip fractures and head traumas are two common problems which usually analyzed by researchers.

Hip fractures cause the greatest health problems and greatest number of death. It is reported that a quarter million hip fractures occur each year among people older than 50 years in the U.S. but more common in woman than men and increase in frequency with increasing age [10] [13]. Most patients with the hip fractures after falls are hospitalized, but about half of them cannot return home or live independently after that. In 1986, it costs for more than \$3 billions for direct medical treatment of hip fractures [13].

Head injuries also usually happen to patients after suffering a fall. It is reported that people with the age of over 65 years old make up for 10-15% of admission to the hospital because of head injuries while three-quarters of them are caused by fall [10]. In California from 1996 to 1999, 71% of fall-related head injuries occurred in adults aged 65 and more [14]. The head trauma from previous fall can increase the risk of repeated fall and head injuries in the near future.

Finally, the most serious impacts of fall is fatal injuries causing death to patient. According to data from the WHO, the fall morality rate of people with the age of over 65 in the United State of America (U.S.) (in 2003) was 36.8 per 100 000 citizens [15]. The main reason of these accidental death is that many patients have lie on the ground for several hours before receiving the help from caregivers.

Psychological consequences

Fall can result in long lasting psychological impact, called fear of falling again, more than just short term injuries. For those people who have fallen one time before seem to have a feeling of vulnerability which limits them from their normal activities. They may be worried about falling and hurting themselves again, thus stop going out on their own or reduce their level of physical activities. Unfortunately, in attempting to reduce their risk of falling, they may increase risk of falling because of the attenuation of physical functions and mobility.

Other consequences

Some research indicated that there are other consequences such as dependency, hospital admission or economic consequence. As already mentioned, the fear of falling again prevent patients from living alone and start depending on the long-term nursing care. This consequence directly leads to increased dependency, which require more time and financial resources from both government and family.

Falls are also the main reason for fallers to be admitted to hospital as an emergency case. Due to the information provided by Accident Department at a sample of hospitals in England and Wales, in 1998, approximately 200 000 people over 60 are treated at this hospital per year because of a fall at home [16].

Finally, falls are also a economic burden of every countries around the world, especially for low and lower middle countries. According to the U.S. Center for Disease Control and Prevention, in 2015, there was \$50 billion being used for fall injuries and expected to increase as the population ages and may reach \$67.7 billion by 2020 [2]. These costs are spent for hospitalization, medical, pharmaceutical, nursing home and other costs which related to the medical treatments for fall patients.

1.5 Challenges in Detecting Falls

As mentioned in previous section, most of patient's death is caused by the late assistance from caregivers after fall accidents. Timely detection can minimize the negative impacts of falls on patients. However, depend on age and physical condition, different people have different gaits, different balance ability that result in several types of falls. These falls can happen in any direction at any speed. Therefore, it is still a challenge to design a system which can detect successfully all the falls of patients.

Another challenge in detecting the fall is that it is difficult for a system to distinguish a real fall from other fall-likes activities such as sitting, jumping or running, which also result in high acceleration. Therefore, it is required to collect a great amount of data from different subjects; and, then, there should be some analysis to find out the most common pattern of every fall. From that, researchers can design an algorithm to detect falls based on this pattern. However, it is not easy to collect real data for the system in the daily life

on real human body, since it is an accidental, dangerous action which can happen randomly at any time. Therefore, it is a overwhelming task for researcher to built a dataset for their system.

1.6 Objectives

From above challenges, it is necessary to study and design a wireless sensor network which can automatically detect the fall of patient and send alarm to the caregiver instantly to reduce their time of arrival, and hence reduce mortality rate. The sensor node is portable and attached on the chest of user. This sensor node contains one accelerometer and one gyroscope to provide information about the body's acceleration and angular velocity during the fall event. This sensor node can connect to the host computer via Wi-Fi with the help of Wi-Fi module called ESP8266 from ESPRESSIF®company and send data directly to a monitoring software on the computer. Upon detecting a fall, sensor node will wait for the safe signal from patient within next 10 minutes. If there is no response, an emergency email will be sent immediately to the caregiver. There is also a super bright red led on the sensor node to indicate the fall and help caregiver to easily find out the patient.

1.7 Thesis Organization

In chapter 1, basic knowledge about falls as well as the causes and consequences of them are introduced. It is also provides the information about current challenges in detecting fall and motivation for this thesis. In chapter 2, a review of existing commercial fall detection devices in the market and researches related to this topic from other research group is included. Chapter 3 describes the problem of fall detection and method to solve this problem. Also in this chapter is the description of algorithm used in this thesis. Chapter 4 describes the system with both hardware and software aspects. Chapter 5 and 6 describe the data collection and analysis procedure as well as result assessment of the experiments on real human body. Finally, chapter 7 gives the conclusion for this thesis, limitation of current solution and recommendation for future works.

2. Literature Review

Nowadays, there are a lot of companies focusing on developing portable fall detection devices to suffice the need of current society where there are more and more elderly people fall every day. These products are used widely by many people around the world and have positive effects on health-care systems of these countries. Beside that, numerous universities and research groups carried out studies on the topic of fall detection from various perspectives and have great contributions to the evolution of health monitoring services.

2.1 Existing Commercial Devices

One popular fall detection device, which is highly evaluated by customers, is the Medical Fall Alert in the *Home Guardian* option from Medical Guardian®. It is designed in form of a lightweight wearable neck pendant. This pendant is waterproof and small enough that have no disturbance on user's daily activity. Medical Guardian® provides a home system including one base unit connecting to the monitoring station through landline and one auto-alert pendant on user body. Whenever detecting a fall, the wearable pendant will send a signal to the base unit and the base station will announce "*Fall Detected, Press or Hold Button to Cancel*". If the patient actual need help, do not press the button. Then, the system will connect to operators at the monitoring center, they will ask if patient need help or not. If the patient cannot speak because of unconsciousness, the operators will automatically send emergency services to patient's home. The monthly fee for this service is around \$44.95 per month and it is used inside the U.S..



Figure 2.1: Medical Guardian Fall Alert System

Another device on the market is myHaloTM from MobileHelpTM. This device has more advantages compared to Medical Guardian one because of full body monitoring functions. It can detect a fall, monitoring heart rate, skin temperature, sleep/wake pattern, etc. and send alert signal to the authorized contacts. This product supports the medical alert outside the home powered by nationwide cellular network and do not require landline phones. It also provides precise location of patients when they go outside. This enable emergency call center to pinpoint patients' location and direct medical support to them. This service costs from \$41.95 per month and support 24/7. However, because this system offers so many features, it seems to less focus on the fall detection.



Figure 2.2: myHalo Medical Alert System from MobileHelpTM

2.2 Existing Products from Other Research Groups

Beside many commercial fall detection devices being available on the market, there are still many products from different research groups surrounding the fall detection topic. A group of Qiang Li and colleagues have used TEMPO (Technology-Enabled Medical Precision Observation) 3.0 sensor node for their project [17]. This sensor node contains one tri-axial accelerometer and one tri-axial gyroscope and controlled by an TI MSP430F1611 microcontroller. This group proposed an algorithm which combines the data from both accelerometer and gyroscope to reduce both false positive and false negative detection and improving the fall detection accuracy. This system is low computational requirements and

real-time response. The authors stated that their method has difficulties in differentiating jumping into bed and falling against wall with a seated posture.




Figure 2.3: TEMPO 3.0 sensor node

Binh Nguyen and Jonathan Tomkun designed and created a fall detection system for the elderly as a wearable monitoring device which can distinguish between fall and non-fall events [18]. This device can link wirelessly with a pre-programmed laptop computer or Bluetooth-compatible mobile phone. Upon detecting a fall, the device communicates wirelessly with the laptop or cellphone to call 911 or issued emergency contacts. This device can also detect abnormal tilt and warns the user to correct their posture to minimize the risk of falling. In addition to visual LED to alert the fall, there are also audio and tactile alert options for people with hearing or seeing disabilities. Regarding the performance of system, some actions have not been distinguished by one of four proposed algorithm to be falls or non-falls.



Figure 2.4: The final design of device from group research of Binh Nguyen and Jonathan Tomkun

3. Statement of Problem and Methodology

3.1 Problem

Nowadays, there have been various research groups studying on the academic area about fall detection on human. The majority of them focus on designing a new algorithm to successfully distinguish between fall and fall-like activities. From the first time, most of research groups used the algorithm mainly based on the data from only one accelerometer such as [19][20][21]. However, focusing only on the data from the accelerometer can result in many false positives as other activities such as sitting, running and jumping which may also cause large peak accelerations. After that, there are other algorithms rely on the detection of body orientation after the fall such as [22][23]. The main drawback of these strategies is that the fall can be confused by activities with similar postures such as sleeping, reclining, etc. It is also less effective when a person's fall posture is not horizontal. Then, there are some groups have nominated such complex algorithms that used Support Vector Machine (SVM)[24] and Markov model [25] to detect the fall. These techniques requires huge amount of computational resources which cause the delay on system which effect the robustness and accuracy of the system.

3.2 Methodology

From the disadvantages of prior works as mentioned above, this project proposes using an algorithm which combines data from one accelerometer and one gyroscope to detect the fall. The data from accelerometer provides valuable information regarding body inertial change due to impact, while the gyroscope's data provides information about the body's rotational velocity during a fall event. This method helps improving the accuracy of falling detection system without the need of complex computations.

3.2.1 Algorithm

This algorithm is nominated by Quoc T. Huynh *et al.* [26], which used data from both accelerometer and gyroscope sensors with predefined critical thresholds, to detect a fall with maximum sensitivity and specificity. There are two parameters used to analyze in this algorithm are normalized acceleration and normalized angular velocity.

Normalized acceleration is the total sum acceleration vector, named ***Acc_{norm}*** containing both static and dynamic acceleration components in different directions. This parameter is calculated with the formula below

$$\mathbf{Acc}_{norm} = \sqrt{(A_x)^2 + (A_y)^2 + (A_z)^2} \quad (3.1)$$

where A_x , A_y and A_z represent the acceleration of three directions x , y , z .

Normalized angular velocity is the total sum angular velocity vector, named ***ω_{norm}***, containing the rotational velocity components in different direction. This parameter is calculated with the formula below

$$\omega_{norm} = \sqrt{(\omega_x)^2 + (\omega_y)^2 + (\omega_z)^2} \quad (3.2)$$

where ω_x , ω_y and ω_z represent the angular velocity of three directions x , y , z .

When stay in stationary condition, the body of user gives the acceleration magnitude of a constant at $+1g$ and angular velocity at $0^\circ/s$. When user falls, there is a rapid change in body's acceleration while angular velocity produces a variety of signals in the fall direction. To detect the fall, the acceleration and angular velocity are compared with critical thresholds. These thresholds are defined as follows by Quoc T. Huynh *et al.*

- (a) *Lower fall threshold(LFT)*: local minima for the Acc resultant of each recorded activity are referred to as the signal lower peak values (LPVs). The LFT_{acc} for the acceleration signals is set at the level of the smallest magnitude lower fall peak (LFP) recorded.

(b) *Upper Fall Threshold(UFT)*: local maxima for the Acc and ω resultant of each recorded activity are referred to as the signal upper peak values (UPV_s). The UFT for each of the acceleration (UFT_{acc}) and the angular velocity (UFT_{gyro}) signals are set at the level of the lowest upper fall peak (UFP) recorded for the Acc and ω , respectively. The UFT_{acc} is related to the peak impact force experienced by the body segment during the impact phase of the fall.

When a fall happens, the normalized acceleration first falls below the LFT_{Acc} threshold which indicates the start of a fall event. In the next 0.5 seconds, usually called fall windows, data from both accelerometer and gyroscope are compared to UFT_{Acc} and UFT_{Gyro} . The fall windows was obtained from the literature [27][28]. If the magnitude of both parameters exceed the two upper thresholds, a fall is detected. However, if only one of these conditions is not satisfied, there is no fall detection. Quoc T. Huynh et al. [26] have proven, with practical experiments on real subjects, that the three conditions above just happen simultaneously only with a fall but not other activities like standing, running, jumping or sitting up. Figure 3.1 summarizes the main steps of this algorithm.




Figure 3.1: Fall detection algorithm

3.2.2 Position of sensor

Until now, there have been a lot of works on developing new algorithms for the falls detection system, with the help of advance techniques in machine learning and computational areas. However, these researches still lack a study that determines the best sensor placement on the patient's body. When the most optimized place is found, fall detection device can be better designed to archive more accuracy.

4. System overview

4.1 System Components

4.1.1 Hardware Components

4.1.1.1 Sensor

4.1.1.2 Microcontroller

4.1.1.3 Wireless Module

4.1.2 Software

4.1.2.1 Embeddded Software on Sensor Node

4.1.2.2 Communication System

4.1.2.3 Data Acquisition Software

4.1.2.4 Data Analysis Software

4.1.3 System Intergration

5. Experimental and Procedure

5.1 Data Collection

5.2 Data Analysis

5.3 Experiment on Real Subjects to Evaluate the Performance of The Network

6. Results and Discussion

6.1 Results Assessment

6.2 Comparing the performance with existing works

7. Conclusion and Future Works

7.1 Summary

7.2 Limitations

7.3 Future works

A. C/C++ Code for Reading Data from Sensor

```
#include "Arduino.h"
#include "MPU6050.h"
#include "Wire.h"
#include "WiFiClient.h"
#include "ESP8266WiFiMulti.h"

ESP8266WiFiMulti WiFiMulti ;

#if I2CDEV_IMPLEMENTATION == I2CDEV_ARDUINO_WIRE
#include "Wire.h"
#endif

MPU6050 accelgyro;
//MPU6050 accelgyro(0x69); // <-- use for AD0 high

int16_t ax, ay, az;
int16_t gx, gy, gz;
float rotX, rotY, rotZ, normAcc;
float gyroX, gyroY, gyroZ, normGyro;
#define OUTPUT_READABLE_ACCELGYRO

#define LED_PIN 2
bool blinkState = false;

void Print_IMU()
{
#ifdef SERIAL_DEBUG
Serial.println("Accelerometer:");
Serial.print(rotX);
```

```

Serial.print("\t");
Serial.print(rotY);
Serial.print("\t");
Serial.println(rotZ);
Serial.println("Gyroscope: ");
Serial.print(gyroX);
Serial.print("\t");
Serial.print(gyroY);
Serial.print("\t");
Serial.println(gyroZ);
#endif
Serial.print(normAcc);
Serial.print(" ");
Serial.println(normGyro);
}

void init_IMU() {
// join I2C bus (I2Cdev library doesn't do this automatically)
Wire.begin();
Serial.begin(115200);

// initialize device
Serial.println("Initializing I2C devices...");
accelgyro.initialize();

// verify connection
Serial.println("Testing device connections...");
Serial.println(accelgyro.testConnection() ? "MPU6050 connection successful"
: "MPU6050 connection failed");

// use the code below to change accel/gyro offset values

```

```

Serial.println("Updating internal sensor offsets...");
//reading current sensor's offset
Serial.print("\n");
accelgyro.setXGyroOffset(55);
accelgyro.setYGyroOffset(-28);
accelgyro.setZGyroOffset(-2);
accelgyro.setXAccelOffset(-2581);
accelgyro.setYAccelOffset(-3937);
accelgyro.setZAccelOffset(1199);
}

```

```

void Init_Wifi()
{
// Initialize Wifi_connection
WiFiMulti.addAP("Connectify-me", "hazeduh5");
Serial.println();
Serial.println();
Serial.print("Wait for WiFi... ");
while(WiFiMulti.run() != WL_CONNECTED) {
Serial.print(".");
delay(500);
}
Serial.println("");
Serial.println("WiFi connected");
Serial.println("IP address: ");
Serial.println(WiFi.localIP());
delay(500);
}

```

```

void sendData()
{

```



```

const uint16_t port = 8000;
const char * host = "192.168.85.1"; // ip or dns

// Use WiFiClient class to create TCP connections
WiFiClient client;

if (!client.connect(host, port)) {
Serial.println("connection failed");
client.stop();
return;
}

client.print(normAcc);
//      client.print(" , ");
//      client.println(normGyro);
//Serial.println("closing connection");
client.stop();
}

void setup(){
Serial.begin(115200);
Wire.begin();
init_IMU();
delay(10);
Init_Wifi();
}

void loop() {
accelgyro.getMotion6(&ax, &ay, &az, &gx, &gy, &gz);
rotX = ax / 16384.0;
rotY = ay / 16384.0;
rotZ = az /16384.0;
normAcc = sqrt(rotX*rotX + rotY*rotY + rotZ*rotZ);
gyroX = gx /131;

```

```
gyroY = gy /131;  
gyroZ = gz /131;  
normGyro = sqrt(gyroX*gyroX + gyroY*gyroY + gyroZ*gyroZ);  
Print_IMU();  
sendData();  
}
```

B. Python Code for Data Acquisition Program

```
import matplotlib.pyplot as plt
import matplotlib.animation as animation
from drawnow import *
import socketserver

accel = []
gyro = []

plt.ion() #tell the matplotlib that we want to draw live data
def makeFig():
    plt.subplot(2,1,1)
    plt.plot(accel, 'r', label= 'Normalized Acceleration')
    plt.legend(loc='upper left')
    plt.ylim(0,5)
    plt.subplot(2,1,2)
    plt.plot(gyro, 'b', label = 'Normalized Angular Velocity')
    plt.legend(loc='upper left')
    plt.savefig('testplot.png')

class myTCPServer(socketserver.StreamRequestHandler):
    def handle(self):
        data = self.rfile.readline()
        dataArray = data.decode().split(',')
        normAcc = float(dataArray[0])
        normGyro = float(dataArray[1])
        accel.append(normAcc)
        gyro.append(normGyro)
        if len(accel) > 100:
            accel.pop(0)
```

```
if len(gyro) > 100:
    gyro.pop(0)
    drawnow(makeFig)
#create TCP server
serv = socketserver.TCPServer(("",8000),myTCPServer)
serv.serve_forever()
```

References

- [1] J. Y. Hwang, J. M. Kang, and H. C. Kim, “Development of novel algorithm and real-time monitoring ambulatory system using bluetooth module for fall detection in the elderly,” *Proceedings of the 26th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (IEMBS 04)*, vol. 1, pp. 2204–2207, September 2004.
- [2] Nation Council of Aging, “Falls prevention facts.” [Online]. Available: <https://www.ncoa.org/news/resources-for-reporters/get-the-facts/falls-prevention-facts/>
- [3] “10 shocking statistics about elderly falls,” Aug. 201. [Online]. Available: <https://www.shellpoint.org/blog/2012/08/13/10-shocking-statistics-about-elderly-falls/>
- [4] E. A. F. T. R. M. J. A. Stevens, P. S. Corso, “The costs of fatal and non-fatal falls among older adults,” Oct. 2006.
- [5] World Health Organiztion, “Falls.” [Online]. Available: <http://www.who.int/en/news-room/fact-sheets/detail/falls>
- [6] E. J. Prudham D, “Factors associated with falls in the elderly: a community study.” in *Age Ageing*, 1981.
- [7] J. C. M. E. H. R. E. J. M. S. A. Mandel, “Senile dementia of the alzheimer’s type: An important risk factor for serious falls,” *Journal of Gerontology*, vol. 42, no. 4, Jul.
- [8] D. M. Asir John Samuel, John Solomon, “A critical review on the normal postural control,” *Physiotherapy and occupational Journal*, vol. 8, no. 2, pp. 71–75, 2015.
- [9] R. C. W. N. K. L. P. N. A. M. J. Poonam K. Pardasaney, Mary D. Slavin, “Conceptual limitations of balance measures for community-dwelling older adults,” *Physical Therapy*, vol. 93, pp. 1351–1368, 2013.

- [10] J. H. Downtow, *Falls in the Elderly*. Edward Arnold, 1993, ch. 1, p. 5.
- [11] K. R. Davies AJ, “Falls presenting to the accident and emergency department: types of presentation and risk factor profile,” *Age Aging*, vol. 25, p. 3626, 1996.
- [12] M. V. P. R. J. D. M. J. G. H. P. F. A. J. H. J. P. F. J. A. V. R. F. C. F. W. D. K. J. J. Juan J. Carrero, Dinanda J. de Jager, “Cardiovascular and noncardiovascular mortality among men and women starting dialysis,” *Clinical Journal of the American Society of Nephrology*.
- [13] MedicineNet.com, “Falls and fractures in seniors,” 2004. [Online]. Available: <https://www.medicinenet.com/script/main/art.asp?articlekey=7774>
- [14] M. M. M. W. Rep, “Nonfatal fall-related traumatic brain injury among older adults—california, 1996–1999,” *Morbidity and Mortality Weekly Report*, 2003.
- [15] J. A. Stevens, G. Ryan, and M. Kresnow, “Fatalities and Injuries from Falls Among Older Adults—United States, from 1993 to 2003 and from 2001 to 2005,” *Journal of the American Medical Association*, pp. 32–33, 2007.
- [16] Consumer Safety Unit, “Home and leisure accident research,” 1988.
- [17] Q. Li, J. A. Stankovic, M. Hanson, A. Barth, and J. Lach, “Accurate, fast fall detection using gyroscopes and accelerometer-derived posture information,” in *Wearable and Implantable Body Sensor Networks*, Berkeley, CA, USA, Jun. 2009.
- [18] J. Tomkun and B. Nguyen, “Design of a Fall Detection and Prevention System for the Elderly,” Bachelor Thesis, McMaster University Hamilton, Ontario, Canada, 2010.
- [19] A. W. W. X. Weihao Qu, Feng Lin, “Evaluation of a low-complexity fall detection algorithm on wearable sensor towards falls and fall-alike activities,” *Gait Posture*, Aug. 2008.
- [20] A. Bourke, J. O’Brien, and G. Laighin, “Evaluation of a threshold-based tri-axial accelerometer,” vol. 26, pp. 194–199, 01 2006.
- [21] N. H. L. Dean M. Karantonis, “Implementation of a real-time human movement classifier using a triaxial accelerometer for ambulatory monitoring,” *IEEE TRANSAC-*

TIONS ON INFORMATION TECHNOLOGY IN BIOMEDICINE, vol. 10, no. 1, Jan. 2006.

- [22] L. H. Nair and Ragimol, “Ahrr based body orientation estimation for real time fall detection,” in *2017 International Conference on Innovations in Information, Embedded and Communication Systems (ICIECS)*, March 2017, pp. 1–4.
- [23] D. C. J. L. R. B. Jay Chen, Karris Kwong, “Wearable sensors for reliable fall detection,” in *Proceedings of the 2005 IEEE. Shanghai, China: Engineering in Medicine and Biology 27th Annual Conference*, Sep. 2005.
- [24] T. Zhang, J. Wang, L. Xu, and P. Liu, “Fall detection by wearable sensor and one-class SVM algorithm,” *Intelligent Computing in Signal Processing and Pattern Recognition: International Conference on Intelligent Computing, ICIC 2006 Kunming, China*, vol. 345, August 1619, 2006.
- [25] P. Jayachandran, T. F. Abdelzaher, J. A. Stankovic, and R. K. Ganti, “Satire: a software architecture for smart attire,” *Proceedings of the 4th International Conference on Mobile Systems, Applications and Services (MobiSys 06)*, pp. 110–123, June 2016.
- [26] Q. T. Huynh, U. D. Nguyen, L. B. Irazabal, N. Ghassemian, and B. Q. Tran, “Optimization of an accelerometer and gyroscope-based fall detection algorithm,” *Journal of Sensor*, no. 452078, 2015.
- [27] M. I. Y. Alwathiqbellah Ibrahim, “Simple fall criteria for mems sensors: Data analysis and sensor concept,” *Sensor*, vol. 14, no. 7, p. 1214912173, 2014.
- [28] M. M. M. S. Mihail Popescu, Benjapon Hotrabhavananda, “V ampir- an automatic fall detection system u sing a vertical pir sensor array,” in *Proceedings of the 6th International Conference on Pervasive Computing Technologies for Healthcare*, San Diago, California, USA, May 2012, pp. 163–166.