

REPORT OF PERCENTAGE OF RETAKES FOR DIGITAL RADIOGRAPHY

FACILITY : Pusat Kesihatan Universiti.
 ROOM NO. / LOCATION : Universiti Malaysia Terengganu
 YEAR : 2020
 STANDARD : ≤ 2.5%

TYPE OF ERROR	MONTHS												TOTAL	PERCENTAGE (%)
	Jan.	Feb.	Mar.	Apr.	May	June	July	Aug.	Sept.	Oct.	Nov.	Dec.		
1. Over Exposure/ High Index														
2. Under Exposure/ Low Index														
3. Double Exposure														
4. Wrong Technique			1										1	0.1%
5. Wrong Patient/ Exam														
6. No Primary/ Wrong Marker		1											1	0.1%
7. Collimation Error														
8. Patient Movement														
9. Patient Related Artifact	11	1	11										6	0.6%
10. Equipment Fault (X-Ray Tube/ Grid/ Bucky)														
11. Detector/ Imaging Plate														
12. Image Artifact														
13. Processing Fault														
14. Miscellaneous. Please specify:.....														
OTHERS														
Total number of error	2	3	3	-	-	-	-	-	-	-	-	-	6	0.9%

∴ Total x-ray taken in 2020 = 863

Total rejected = 8

IRWAN ISKANDAR BIN JUSOFF
 JURU X-RAY U32
 PUSAT KESIHATAN UNIVERSITI
 UNIVERSITI MALAYSIA TERENGGANU (UMT)

13 JUL 2021