AI-driven ventilator management in ICU

"How can Ai support timely and personalized ventilator setting adjusments for mechanically ventilated ICU patients ? what predective models and decision support approches have been studied and what challenges remain regarding patient safety,clinical validation,and clinician supervision?" *Generated by LeapSpace*

AI-Supported Personalized Ventilator Adjustment at a Glance

Key Findings Table

Key finding	Population / setting	Evidence
AI is currently most mature as advisory support for ventilator management: prediction of gas exchange/mechanics, weaning readiness, extubation success, and asynchrony detection, with clinicians usually retaining control of final setting changes.	Adult mechanically ventilated ICU patients; mixed ICU cohorts	[1] [2] [3] [4] [5] [6] [7] [8]
Predictive models have been studied for short-horizon PaO2, PaCO2, compliance, mechanical power, and weaning/extubation timing, often with good discrimination but limited calibration and external validation.	Adult ICU datasets including MIMIC-III, eICU, MIMIC-IV, single-center cohorts	[1] [2] [4] [5] [9] [10] [11]
Decision-support approaches studied include open-loop physiological simulators (for example BEACON), rule-based weaning systems (SmartCare/PS), adaptive/closed-loop ventilation (ASV, INTELLiVENT-ASV), reinforcement-learning policies, and real-time alerts/dashboards.	Adult ICU, postoperative ICU, ARDS-focused studies, mixed medical ICU	[6] [7] [12] [13] [14] [15] [16] [17]

Key finding	Population / setting	Evidence
Closed-loop modes more consistently improve process measures—fewer manual adjustments, faster weaning steps, more time in target ranges, fewer unacceptable ventilation episodes—than hard outcomes such as mortality or ICU length of stay.	Adult ICU trials and reviews	[18] [19] [20] [21] [22] [23] [24] [25] [26]
Human oversight is treated as a core safety requirement: clinician-set targets, pressure/PEEP limits, review workflows, and override capability are standard.	Open-loop and closed-loop ventilator DSS studies	[12] [13] [14] [27] [28] [29] [30] [31] [32]
The biggest bottleneck is clinical validation: many studies are retrospective, single-center, internally validated only, and report discrimination more often than calibration, workflow outcomes, or patient-centered benefit.	Cross-cutting across ventilator AI literature	[11] [33] [34] [35] [36] [37] [38]
Safety reporting is incomplete: many papers describe systems as safe, but few report AI-specific malfunctions, override rates, false alarms, drift, or harms from wrong recommendations.	Reviews and implementation-focused studies	[14] [25] [39] [40] [41] [42] [43]

Direct Answer

AI can support timely and personalized ventilator adjustment in three practical ways. First, it can **predict near-term physiology**—such as PaO2, PaCO2, compliance, mechanical power, or weaning/extubation readiness—so clinicians can act earlier and tailor settings to likely patient response rather than waiting for delayed deterioration signals [1] [2] [9] [44]. Second, it can **estimate the likely effect of a proposed setting change** using model-based simulation or virtual-patient methods, especially for PEEP, tidal volume, respiratory rate, and pressure support [5] [45] [46] [47]. Third, it can provide **decision support or constrained automation** through systems such as BEACON, SmartCare/PS, ASV, and INTELLiVENT-ASV, as well as real-time alerts for unsafe ventilation or asynchrony [6] [12] [13] [14] [16] [17].

The predictive models studied include multitask neural networks for short-horizon gas exchange/compliance prediction, XGBoost/CatBoost/logistic/CNN models for weaning and extubation timing, random-forest models for ventilator-associated complications, reinforcement-learning and offline policy-learning systems for PEEP/FiO2/Vt strategy, and waveform or computer-vision systems for patient-ventilator asynchrony detection [1] [2] [4] [10] [15] [48] [49] [50] [51].

The main remaining challenges are **patient safety assurance**, **sparse external/prospective validation**, **uncertain benefit on hard outcomes**, **limited reporting of alert burden and overrides**, and the need to preserve **clinician accountability through guardrails, transparency, and supervision** [11] [30] [31] [38] [40] [52]. In practice, the evidence supports a **human-in-the-loop "AI co-pilot" model**, not unsupervised autonomous ventilator titration [8] [29].

Confidence level: High - The overall pattern is strongly consistent across predictive, automation, validation, and implementation studies [1] [21] [29] [38].

Study Scope

This report covers adult ICU mechanical ventilation research spanning predictive modeling, ventilator decision support, automation, and implementation studies, with evidence ranging from foundational weaning and protocolization work to recent AI/ML and reinforcement-learning studies, including retrospective database analyses, prospective observational studies, randomized trials of automated modes, and reviews/meta-analyses [24] [53] [54] [55]. Subgroups addressed in the evidence include ARDS, COPD, postoperative/cardiac surgery, elderly patients, and mixed medical ICU cohorts [39] [56] [57] [58].

Assumptions & Limitations

The evidence base is uneven: many studies are retrospective or single-center, direct setting-by-setting AI-versus-clinician agreement data are sparse, calibration is infrequently reported, and multicountry generalizability, fairness, uncertainty quantification, drift monitoring, and AI-specific harm reporting remain limited [5] [11] [38] [41] [59] [60]. Hard-outcome superiority of any single AI strategy for bedside ventilator titration is therefore not established [22] [52].

Suggested Further Research

- Prospective multicenter trials comparing AI-guided ventilator adjustment against protocolized standard care on ventilator-free days, reintubation, ICU stay, and mortality [38] [55].
- Routine reporting of calibration, uncertainty, subgroup performance, and dataset-shift robustness, not just AUROC or accuracy [11] [60].
- Bedside studies of uptake, override frequency, time-to-action, and alert burden for ventilator AI tools [43] [61].
- Prospective validation of digital-twin and counterfactual models for PEEP, Vt, and pressure-support titration [47] [62].
- Human-factors trials testing whether explanation style and uncertainty display improve correct clinician use and trust [34] [63].

Full Report

Foundational Context

Mechanical ventilation requires repeated adjustment of settings such as **FiO2, PEEP, tidal volume, respiratory rate, inspiratory pressure, and pressure support** to balance oxygenation, CO2 clearance, work of breathing, and avoidance of ventilator-induced lung injury [64] [65] [66]. The clinical problem is dynamic: lung mechanics, gas exchange, sedation state, patient effort, and weaning readiness change over time, often faster than intermittent clinician review or arterial blood gas sampling can capture [1] [67] [68].

This creates three distinct roles for AI:

1. **Prediction**: forecast what will happen soon if current support is maintained, such as future PaO2, PaCO2, compliance, or extubation success [1] [2].
2. **Counterfactual estimation**: estimate what may happen if a clinician changes PEEP, Vt, RR, or pressure support before actually applying that change [5] [47].
3. **Decision support / automation**: either recommend settings for clinician approval or directly adjust settings within pre-specified target zones and limits [6] [12] [14].

Taken together, these functions suggest that AI is best viewed not as a single technology, but as a stack ranging from forecasting to constrained control [5] [14] [29].

Confidence level: High - Core concepts are repeatedly supported across predictive-model, closed-loop, and implementation literatures [1] [6] [14].

Overview and Clinical Use Cases

Why ventilator decision support matters

The strongest practical rationale is the need for **timely individualized changes** in ICU patients whose respiratory state evolves continuously, while clinician attention is divided across many competing priorities [6] [17] [53]. Existing protocolized care improved liberation and safety in earlier weaning research, establishing that structured ventilator decision-making can improve care consistency even before AI entered the field [69] [70] [71].

AI studies focus on several recurring ICU scenarios:

- **PEEP and oxygenation titration** in ARDS or hypoxemic respiratory failure [5] [72].
- **Prevention of injurious ventilation** via driving pressure, plateau pressure, or mechanical power surveillance [68] [73].
- **Weaning and extubation timing** [2] [44].

- **Detection of patient-ventilator asynchrony** [50] [74].
- **Closed-loop support** to reduce manual workload and standardize care [13] [14].

Main decision targets

The ventilator parameters most often targeted are **PEEP, FiO2, tidal volume, respiratory rate, inspiratory pressure, and pressure support** [5] [66]. BEACON advises FiO2, inspiratory pressure or tidal volume, PEEP, and respiratory frequency using individualized physiologic models [7] [66]. SmartCare/PS automates pressure-support reduction and SBT progression using a “zone of comfort” logic [12] [75]. ASV and INTELLiVENT-ASV optimize rate-volume-pressure tradeoffs and, in the latter, oxygenation targets as well [13] [14].

Confidence level: High - Decision targets and use cases are directly documented across multiple system descriptions and trials [12] [14] [66].

Predictive Models for Personalized Ventilator Adjustments

Short-horizon prediction of gas exchange and mechanics

A central line of work forecasts near-term respiratory physiology to support earlier adjustment. A 2024 multitask neural network trained on MIMIC-III and eICU predicted **PaO2, PaCO2, and respiratory-system compliance about 45 minutes ahead** and then simulated patient-specific PEEP changes from those predicted values [1] [5]. This is one of the clearest examples of AI directly linked to setting adjustment rather than generic risk prediction [5].

Other work has forecast **mechanical power in real time**, aiming to warn clinicians before lung-damaging ventilation patterns persist [9] [68]. However, across these studies, reporting of **calibration, uncertainty, and external validation** is limited; the short-horizon gas exchange/compliance papers mainly emphasize test-set performance rather than deployment-ready validation [1] [5] [11].

Weaning readiness and extubation prediction

Weaning/extubation is the best-developed predictive use case. An XGBoost EHR model predicted invasive MV weaning readiness within 24/48/72 hours with external-validation AUROCs of **0.847/0.795/0.789**, sensitivity above 0.75 across horizons, but weaker performance in older and neurosurgical patients [2]. A CatBoost extubation-timing model in MIMIC-IV achieved internal AUROC **0.835** and prospective AUROC **0.803** in a cardiac surgical ICU, showing one of the stronger examples of temporal or prospective validation in this literature [4].

Other models used continuous ventilator waveforms and SBT numerics. A CNN predicted weaning success with AUROC **0.912** and AUPR **0.767**, though without external validation [76]. A multicenter cough/diaphragm/SBT model predicted extubation success far better than SBT success alone, with extubation AUROC **0.91**, sensitivity **96.8%**, and specificity **78.4%** at the reported cutoff [44]. Ultrasound-based models and serial RSBI strategies also showed useful discrimination in some subgroups, but performance varied by population [57] [77] [78].

Dynamic respiratory mechanics and individualized titration

Mechanics-oriented personalization often relies on ventilator-derived features such as **airway pressure, flow, volume, PEEP, plateau pressure, respiratory rate, compliance/elastance, driving pressure, work of breathing, and minute ventilation** [65] [72] [79] [80]. The literature repeatedly treats **driving pressure** and related stress/strain markers as more informative summaries of injurious load than tidal volume or PEEP considered in isolation [64] [73].

One notable mechanics-derived metric, **mechanical power normalized to dynamic compliance**, predicted weaning failure better than mechanical power alone in ICU patients using pre-SBT respiratory mechanics from medical records [81]. This suggests that combining global load with patient-specific mechanics may be more useful than generic thresholds alone [81].

Digital-twin, simulation, and counterfactual approaches

Emerging work moves beyond prediction toward **counterfactual estimation**. Virtual-patient and digital-twin models have been used to forecast the effect of higher PEEP using prior lower-PEEP states and respiratory mechanics, estimating **elastance, distension, functional residual capacity, and peak inspiratory pressure** across stepwise PEEP changes [45] [47] [62] [82]. These studies report good accuracy for predicting peak pressure and distension-related injury risk before bedside application [62] [82].

Simulation has also been applied to optimize **tidal volume and respiratory rate** while preserving PaCO2 in ARDS, requiring patient-specific response data to ventilator inputs and gas-exchange targets [46]. Pressure-support response models used pressure, flow, end-tidal CO2, and esophageal pressure to estimate how patients altered work of breathing and tidal volume when support changed [83].

Inferred: Taken together, these studies suggest digital-twin methods are more directly aligned with personalized ventilator titration than static risk models, but they remain early-stage and physiologically data-hungry [47] [62].
Confidence level: Medium - The direction is clear, but prospective bedside validation remains sparse [34] [62].

Closed-Loop and Semi-Automated Ventilator Control

Adaptive support ventilation and SmartCare/PS

ASV was designed to maintain a preset minute ventilation by automatically adjusting inspiratory pressure and rate to the patient's mechanics, while still using predefined transitions and extubation criteria [13] [84]. In a medical ICU RCT, ASV shortened total ventilation and weaning duration, reduced manual ventilator adjustments, and did not change day-28 mortality [23] [85]. In COPD weaning, ASV also shortened weaning time versus PSV [86].

SmartCare/PS automates pressure-support reduction, performs SBTs, and signals separation potential [12]. However, its benefit appears context-dependent: one Australian ICU found no advantage over experienced 1:1 nurse-led care, suggesting workflow and staffing moderate value [12] [87].

INTELLiVENT-ASV and oxygenation control

INTELLiVENT-ASV extends ASV to closed-loop control of both ventilation and oxygenation using predefined optimal and acceptable zones with continuous monitoring [14] [32]. Trials in sedated or postoperative ICU patients reported fewer clinician interventions, lower FiO2 or airway pressures in some settings, and fewer episodes of unacceptable ventilation, without major safety interruptions [14] [25] [39].

Closed-loop oxygen-control studies in adult critical care mainly show improved **time in SpO2 target range**, reduced hypoxemia burden, lower FiO2 exposure, and fewer manual adjustments, rather than hard clinical-outcome gains [88] [89] [90]. The adult ICU literature retrieved here did not show completed randomized trials of automated FiO2/PEEP control with definitive patient-centered benefits [91].

Clinical effects of automation

Across trials and reviews, the most consistent benefits are:

- fewer manual ventilator changes [23] [92]
- shorter weaning duration in some settings [24] [93]
- improved ventilation quality and target-zone adherence [21] [25]
- reduced unacceptable ventilation episodes [25]

By contrast, effects on **mortality, ICU length of stay, ventilator-free days, and overall survival** are weak, inconsistent, or absent [18] [19] [20] [21] [22] [26]. This process-versus-outcome gap is one of the field's most stable findings [24] [52].

Confidence level: High - Multiple RCTs, observational studies, and meta-analytic summaries converge on this pattern [20] [21] [24].

Multimodal ICU Decision Support and Real-Time Workflow Integration

Multimodal prediction and early warning

Beyond direct ventilator control, AI has been used to predict **ventilator-associated complications, decompensation, ARDS, and unsafe ventilation** from EHR plus physiologic streams [16] [17] [48]. A random-forest model using EHR and physiologic vitals reached AUROC **0.83** for ventilator-associated complication prediction using high-frequency time-series features [48]. A rule-based surveillance algorithm combined EHR data and ventilator settings to flag unsafe ventilation in near real time via text-page alerts [16].

A research CDSS integrating streaming ventilator and device data used machine learning to detect potentially harmful ventilation and ARDS and delivered alerts through a mobile app [17]. These systems are less about picking a single setting and more about ensuring earlier recognition of trajectories that should trigger setting reassessment [16] [17].

Bedside dashboards, mobile alerts, and visualization

Clinician-facing systems have presented trends, color-coded abnormalities, respiratory metrics, and actionable prompts via bedside terminals, dashboards, or mobile alerts [94] [95] [96] [97]. However, usability evaluation remains thin. Some dashboard studies report perceived value and suggestions for better training, but few measure how dashboards change ventilator decisions or outcomes [98] [99].

Waveform and screen-image analytics

Waveform analytics is a comparatively strong niche. NexoVent used ventilator-screen computer vision plus OCR and waveform analysis to detect modes, parameters, and six PVA types in real time, achieving **95.4%** parameter accuracy, **94.0%** mode accuracy, and **81.6% to 97.8%** accuracy across asynchrony classes versus expert consensus [50]. INTELLI-PVA reported average **F1 0.849** and therapist-level agreement on unseen configurations and demographics [51].

Automated real-time waveform detection of ineffective triggering showed much higher sensitivity than clinicians (**83.2% vs 41.1%**) with similarly high specificity [74]. Because clinician waveform interpretation is often limited, this is an especially plausible area for AI assistance [100] [101] [102].

Confidence level: High - The evidence strongly supports feasibility of multimodal surveillance and especially waveform-based asynchrony detection, though not full end-to-end recommendation engines [17] [74] [103].

Clinical Validation, Safety, and Remaining Challenges

Validation design and evidence quality

The dominant limitation is that model development still outpaces clinical validation. Reviews repeatedly describe the literature as **retrospective, single-center, internally validated**, with sparse calibration reporting and few multicenter prospective studies [11] [33] [34] [35] [37] [38]. Even among more advanced studies, calibration and decision-curve analysis are exceptions rather than standard practice [10] [60].

Prospective bedside evidence does exist, but it is limited and often underpowered or process-oriented. BEACON advice was applied in **88%** of bedside opportunities and was associated with fewer adverse or hypoxemic events, yet a multicenter trial found no significant difference in ventilation duration or survival [8]. A computerized ventilator DSS trial reported **94%** adherence to instructions and reduced morbidity/overdistension injury, without significant survival or ICU length-of-stay differences [104].

Safety, error modes, and clinician supervision

Human supervision is not an optional extra; it is the default safety architecture. Open-loop systems such as BEACON require clinicians to review and implement recommendations themselves [8]. Closed-loop systems still rely on clinician-defined targets and limits for SpO2, end-tidal CO2, maximum airway pressure, and PEEP [27] [28] [31] [32].

Reported guardrails include:

- user-defined target zones [27] [31]
- pressure and PEEP caps [28] [32]
- passive or sedated-patient use in some protocols [14] [105]
- alerting when the controller reaches limits or oxygen adjustment is turned off [106]

Yet published safety reporting is incomplete. Reviews note that included studies rarely document **AI-specific malfunctions, fairness issues, override frequency, delayed alarms, or harm from incorrect recommendations** [40] [41] [42]. Thus, the absence of reported harm should not be interpreted as comprehensive proof of safety [41] [43].

Explainability, trust, and acceptance

Explainability is consistently described as necessary for adoption [63] [107] [108]. A mechanical-ventilation explainable user interface evaluated layered explanations: first-level outlier warning and output certainty, then optional input availability, feature importance, and rule-based explanations [63]. Nurses and physicians both preferred concise first-level explanations, while physicians valued deeper detail more than nurses, indicating role-specific transparency needs [63].

Older rule-based expert systems were often well accepted: ESNIC advice was consulted for **76%** of adjustments and accepted **83%** of the time [109]. However, very few ventilator AI studies directly test whether explanations improve trust, decision accuracy, or correct use [34] [63].

Workflow fit and implementation barriers

Implementation studies emphasize integration with **existing ICU routines, bedside review, EHR or ventilator-protocol workflows, training, and preservation of decision ownership** [30] [43] [110]. A ventilator CDS alert tool for possible ARDS implemented across 13 ICUs reduced expected alert burden by **41%**, and **48%** of alert recommendations were implemented within 2 hours, showing that careful workflow design can produce measurable uptake [61].

Still, a 2024 review found that across 29 invasive ventilation CDSS studies, **no process measures, quality metrics, or clinical workflow descriptions were reported**, highlighting a major implementation-evidence gap [43]. Common barriers include documentation burden, inconsistent training, communication problems, and lack of technical support [43].

Confidence level: High - The challenges of validation, safety reporting, and workflow integration recur across nearly all evidence streams [11] [43].

Practical Synthesis: What AI Can Realistically Do Now

The evidence supports a staged view of current bedside usefulness:

What is ready or closest to readiness

- **Weaning/extubation prediction** with external or prospective validation in selected settings [2] [4].
- **Waveform/asynchrony detection** that can outperform routine clinician recognition [74].
- **Closed-loop support modes** that reduce manual workload and improve time in target zones under predefined safety constraints [14] [25].

What is promising but not yet fully validated

- **Short-horizon gas exchange/compliance prediction** for direct PEEP or FiO2 adjustment [1] [5].
- **Digital-twin/counterfactual titration** for PEEP, Vt, and RR [45] [47].
- **Reinforcement-learning policies** for PEEP/FiO2/Vt strategy [15] [49].

What remains unsupported for autonomous use

- Unsupervised end-to-end AI that integrates waveforms, vitals, labs, imaging, and notes to directly control all major ventilator settings in real ICU practice [8] [17] [103].

Inferred: A pragmatic near-term deployment model is an **AI co-pilot** that displays predicted physiologic response, injury-risk tradeoffs, and a small set of candidate setting changes, while leaving final action to the clinical team [5] [8].
Confidence level: Medium - This is a synthesis across advisory, digital-twin, and implementation findings rather than a direct single-study conclusion [5] [29].

Conclusion

AI can support timely and personalized ventilator adjustment primarily by **forecasting near-term respiratory state**, **estimating response to candidate setting changes**, and **providing advisory or constrained automated control** for settings such as PEEP, FiO2, Vt, RR, and pressure support [1] [5] [6] [14]. The best-supported applications today are weaning/extubation prediction, asynchrony detection, and closed-loop systems that improve process quality and reduce manual workload [2] [24] [74].

What the field still lacks is not more isolated retrospective accuracy studies, but stronger evidence on **calibration, generalizability, workflow fit, clinician behavior, and patient safety under real-world supervision** [11] [38] [43]. Until that evidence matures, the most evidence-based role for ventilator AI is **clinician-supervised decision support with explicit guardrails, transparent rationale, and easy override**, rather than fully autonomous control [29] [31] [63].

Overall confidence level: High - The literature is highly consistent that AI is useful as supervised support, promising for personalization, and not yet validated for broad unsupervised ventilator titration [21] [29] [36] [52].

References

1. Händel, C., Frerichs, I., Weiler, N., Bergh, B. (2024) Prediction and simulation of PEEP setting effects with machine learning models. *Medicina Intensiva*. https://doi.org/10.1016/j.medin.2023.09.009
2. Zappalà, S., Scaravilli, V., Rovati, L., (...), Grasselli, G. (2025) Development and validation of a machine learning model for real-time prediction of invasive mechanical ventilation weaning readiness. *Journal of Critical Care*. https://doi.org/10.1016/j.jcrc.2025.155105
3. Yan, Y., Xie, Y., Luo, J., (...), Li, X. (2023) Development and validation of a mechanical power-oriented nomogram model for predicting the risk of weaning failure in mechanically ventilated patients: an analysis using the data from MIMIC-IV. *Chinese Critical Care Medicine*. https://doi.org/10.3760/cma.j.cn121430-20221115-00997
4. Zhao, Q.-Y., Wang, H., Luo, J.-C., (...), Luo, Z. (2021) Development and Validation of a Machine-Learning Model for Prediction of Extubation Failure in Intensive Care Units. *Frontiers in Medicine*. https://doi.org/10.3389/fmed.2021.676343

5. Händel, C., Frerichs, I., Weiler, N., Bergh, B. (2024) Prediction and simulation of PEEP setting effects with machine learning models. *Medicina Intensiva*. https://doi.org/10.1016/j.medin.2023.09.009
6. Vizcaychipi, Marcela P., Karbing, Dan S., Martins, Laura, (...), Rees, Stephen E. (2024) Evaluation of decision support to wean patients from mechanical ventilation in intensive care: a prospective study reporting clinical and physiological outcomes. *Journal of Clinical Monitoring and Computing*. https://doi.org/10.1007/s10877-024-01231-5
7. Karbing, DS, Spadaro, S, Rees, SE, Volta, CA (2015) Prospective evaluation of a decision support system providing advice on ventilator settings of: inspiratory oxygen, delivered pressure or volume, frequency and peep. *Intensive Care Medicine Experimental*. https://doi.org/10.1186/2197-425X-3-S1-A672
8. Vizcaychipi, M.P., Karbing, D.S., Martins, L., (...), Rees, S.E. (2025) Evaluation of decision support to wean patients from mechanical ventilation in intensive care: a prospective study reporting clinical and physiological outcomes. *Journal of Clinical Monitoring and Computing*. https://doi.org/10.1007/s10877-024-01231-5
9. Ruiz-Botella, M., Manrique, S., Gómez, J., Bodí, M. (2024) Advancing ICU patient care with a Real-Time predictive model for mechanical Power to mitigate VILI. *International Journal of Medical Informatics*. https://doi.org/10.1016/j.ijmedinf.2024.105511
10. Pai, K.-C., Su, S.-A., Chan, M.-C., (...), Chao, W.-C. (2022) Explainable machine learning approach to predict extubation in critically ill ventilated patients: a retrospective study in central Taiwan. *BMC Anesthesiology*. https://doi.org/10.1186/s12871-022-01888-y
11. Gallifant, J., Zhang, J., Arias López, M., (...), Formenti, F. (2022) Artificial intelligence for mechanical ventilation: systematic review of design, reporting standards, and bias. *British Journal of Anaesthesia*. https://doi.org/10.1016/j.bja.2021.09.025
12. Rose, L., Presneill, J.J., Johnston, L., Cade, J.F. (2008) A randomised, controlled trial of conventional versus automated weaning from mechanical ventilation using SmartCare™/PS. *Intensive Care Medicine*. https://doi.org/10.1007/s00134-008-1179-4
13. Sulzer, C.F., Chioléro, R., Chassot, P.-G., (...), Revelly, J.-P. (2001) Adaptive support ventilation for fast tracheal extubation after cardiac surgery: A randomized controlled study. *Anesthesiology*. https://doi.org/10.1097/00000542-200112000-00010
14. Arnal, J.-M., Wysocki, M., Novotni, D., (...), Jacques, D.-G. (2012) Safety and efficacy of a fully closed-loop control ventilation (IntelliVent-ASV) in sedated ICU patients with acute respiratory failure: A prospective randomized crossover study. *Intensive Care Medicine*. https://doi.org/10.1007/s00134-012-2548-6
15. Peine, Arne, Hallawa, Ahmed, Bickenbach, Johannes, (...), Martin, Lukas (2021) Development and validation of a reinforcement learning algorithm to dynamically optimize mechanical ventilation in critical care. *npj Digital Medicine*. https://doi.org/10.1038/s41746-021-00388-6

16. Herasevich, V., Tsapenko, M., Vukoja, M., (...), Gajić, O. (2011) Limiting ventilator-induced lung injury through individual electronic medical record surveillance. *Critical Care Medicine*. https://doi.org/10.1097/CCM.0b013e3181fa4184

17. Rehm, G.B., Woo, S.H., Chen, X.L., (...), Chuah, C.-N. (2020) Leveraging IoTs and Machine Learning for Patient Diagnosis and Ventilation Management in the Intensive Care Unit. *IEEE Pervasive Computing*. https://doi.org/10.1109/MPRV.2020.2986767

18. Rose, L., Schultz, M.J., Cardwell, C.R., (...), Blackwood, B. (2025) Automated versus non-automated weaning for reducing the duration of mechanical ventilation for critically ill adults and children. *Cochrane Database of Systematic Reviews*. https://doi.org/10.1002/14651858.CD009235.pub4

19. Rose, L., Schultz, M.J., Cardwell, C.R., (...), Blackwood, B. (2013) Automated versus non-automated weaning for reducing the duration of mechanical ventilation for critically ill adults and children. *Cochrane Database of Systematic Reviews*. https://doi.org/10.1002/14651858.CD009235.pub2

20. Rose, L., Schultz, M.J., Cardwell, C.R., (...), Blackwood, B. (2014) Automated versus non-automated weaning for reducing the duration of mechanical ventilation for critically ill adults and children. *Cochrane Database of Systematic Reviews*. https://doi.org/10.1002/14651858.CD009235.pub3

21. Sinnige, J.S., Buiteman-Kruizinga, L.A., Horn, J., (...), Neto, A. (2026) Effect of Automated Closed-Loop Ventilation vs Protocolized Conventional Ventilation on Ventilator-Free Days in Critically Ill Adults: A Randomized Clinical Trial. *JAMA*. https://doi.org/10.1001/jama.2025.24384

22. Kampolis, C.F., Mermiri, M., Mavrovounis, G., (...), Pantazopoulos, I. (2022) Comparison of advanced closed-loop ventilation modes with pressure support ventilation for weaning from mechanical ventilation in adults: A systematic review and meta-analysis. *Journal of Critical Care*. https://doi.org/10.1016/j.jcrc.2021.11.010

23. Kirakli, C., Naz, I., Ediboğlu, O., (...), Tellioğlu, E. (2015) A randomized controlled trial comparing the ventilation duration between adaptive support ventilation and pressure assist/control ventilation in medical patients in the ICU. *Chest*. https://doi.org/10.1378/chest.14-2599

24. Cui, Nianqi, Yan, Xiaoli, Zhang, Yuping, (...), Jin, Jingfen (2022) Non-Pharmacological Interventions for Minimizing Physical Restraints Use in Intensive Care Units: An Umbrella Review. *Frontiers in Medicine*. https://doi.org/10.3389/fmed.2022.806945

25. Fot, Evgenia V., Izotova, Natalia N., Yudina, Angelika S., (...), Kirov, Mikhail Y. (2017) Automated Weaning from Mechanical Ventilation after Off-Pump Coronary Artery Bypass Grafting. *Frontiers in Medicine*. https://doi.org/10.3389/fmed.2017.00031

26. Burns, K.E.A., Lellouche, F., Lessard, M.R., Friedrich, J.O. (2014) Automated weaning and spontaneous breathing trial systems versus non-automated weaning strategies for discontinuation time in invasively ventilated postoperative adults. *Cochrane Database of Systematic Reviews*. https://doi.org/10.1002/14651858.CD008639.pub2

27. Buiteman-Kruizinga, Laura A., van Meenen, David M. P., Bos, Lieuwe D. J., (...), Schultz, Marcus J. (2023) A closed-loop ventilation mode that targets the lowest work and force of breathing reduces the transpulmonary driving pressure in patients with moderate-to-severe ARDS. *Intensive Care Medicine Experimental*. https://doi.org/10.1186/s40635-023-00527-1

28. Tsai, Y.-C., Jhou, H.-J., Huang, C.-W., (...), Hsu, S.-D. (2024) Effectiveness of Adaptive Support Ventilation in Facilitating Weaning from Mechanical Ventilation in Postoperative Patients. *Journal of Cardiothoracic and Vascular Anesthesia*. https://doi.org/10.1053/j.jvca.2024.04.030

29. Hotz, J.C., Chang, D., Khemani, R.G., Newth, C.J.L. (2026) Decision Assist During Mechanical Ventilation. *Respiratory Care*. https://doi.org/10.1177/19433654251384240

30. Palamim, Camila Vantini Capasso, Marson, Fernando Augusto Lima (2027) Harnessing artificial intelligence to modulate mechanical power: Toward safer and outcome-oriented ventilation in critical care. *Intensive and Critical Care Nursing*. https://doi.org/10.1016/j.iccn.2026.104589

31. Schädler, D., Miestinger, G., Becher, T., (...), Hörmann, C. (2017) Automated control of mechanical ventilation during general anaesthesia: Study protocol of a bicentric observational study (AVAS). *BMJ Open*. https://doi.org/10.1136/bmjopen-2016-014742

32. Lellouche, F., Bouchard, P.-A., Simard, S., (...), Wysocki, M. (2013) Evaluation of fully automated ventilation: A randomized controlled study in post-cardiac surgery patients. *Intensive Care Medicine*. https://doi.org/10.1007/s00134-012-2799-2

33. Viderman, D., Ayazbay, A., Kalzhan, B., (...), Abdildin, Y. (2024) Artificial Intelligence in the Management of Patients with Respiratory Failure Requiring Mechanical Ventilation: A Scoping Review. *Journal of Clinical Medicine*. https://doi.org/10.3390/jcm13247535

34. Amanguli, M., Yaxiaerjiang, M., Guo, Y.-S., (...), Zhong, M. (2026) Artificial intelligence in mechanical ventilation: a narrative review of clinical applications and research gaps. *Journal of Thoracic Disease*. https://doi.org/10.21037/jtd-2026-1257

35. Muñoz, J., Ruíz-Cacho, R., Fernández-Araujo, N.J., (...), Muñoz-Visedo, J. (2025) Artificial intelligence in the management of patient-ventilator asynchronies: A scoping review. *Heart and Lung*. https://doi.org/10.1016/j.hrtlng.2025.05.003

36. Ossai, C.I., Wickramasinghe, N. (2021) Intelligent decision support with machine learning for efficient management of mechanical ventilation in the intensive care unit – A critical overview. *International Journal of Medical Informatics*. https://doi.org/10.1016/j.ijmedinf.2021.104469

37. Gallifant, J., Zhang, J., Arias López, M., (...), Formenti, F. (2022) Artificial intelligence for mechanical ventilation: systematic review of design, reporting standards, and bias. *British Journal of Anaesthesia*. https://doi.org/10.1016/j.bja.2021.09.025

38. Muñoz, J., Ruíz-Cacho, R., Fernández-Araujo, N.J., (...), Muñoz-Visedo, J. (2025) Artificial intelligence in the management of patient-ventilator asynchronies: A scoping review. *Heart and Lung*. https://doi.org/10.1016/j.hrtlng.2025.05.003

39. Yeremenko, A.A., Komnov, R.D., Koshek, E.A. (2022) The Efficacy and Safety of Automatic Modes During Respiratory Support After Cardiac Surgery. *Obshchaya Reanimatologiya*. https://doi.org/10.15360/1813-9779-2022-3-21-29
40. Gupta, P., Pearce, A.K., Pham, T., (...), Afshar, M. (2025) Artificial intelligence-driven decision support for patients with acute respiratory failure: a scoping review. *Intensive Care Medicine Experimental*. https://doi.org/10.1186/s40635-025-00791-3
41. Gupta, Preeti, Pearce, Alex K., Pham, Thaidan, (...), Afshar, Majid (2025) Artificial intelligence-driven decision support for patients with acute respiratory failure: a scoping review. *Intensive Care Medicine Experimental*. https://doi.org/10.1186/s40635-025-00791-3
42. Botta, Michela, Tsonas, Anissa M., Sinnige, Jante S., (...), Schultz, Marcus J. (2022) Effect of Automated Closed-loop ventilation versus convenTional VEntilation on duration and quality of ventilation in critically ill patients (ACTiVE) – study protocol of a randomized clinical trial. *Trials*. https://doi.org/10.1186/s13063-022-06286-w
43. Murali, M., Ni, M., Karbing, D.S., (...), Patel, B.V. (2024) Clinical practice, decision-making, and use of clinical decision support systems in invasive mechanical ventilation: a narrative review. *British Journal of Anaesthesia*. https://doi.org/10.1016/j.bja.2024.03.011
44. https://www.sciencedirect.com/science/article/pii/S136485352300217X
45. Calje-van der Klei, T.L., Sun, Q., Chase, J.G. (2023) Pulmonary Response Prediction Comparison through Three Personalized Basis Functions in a Virtual Patient Model. *IFAC-PapersOnLine*. https://doi.org/10.1016/j.ifacol.2023.10.1109
46. Wang, W., Das, A., Cole, O., (...), Bates, D.G. (2015) Computational simulation indicates that moderately high-frequency ventilation can allow safe reduction of tidal volumes and airway pressures in ARDS patients. *Intensive Care Medicine Experimental*. https://doi.org/10.1186/s40635-015-0068-8
47. Sun, Q., Zhou, C., Chase, J.G. (2020) Parameter updating of a patient-specific lung mechanics model for optimising mechanical ventilation. *Biomedical Signal Processing and Control*. https://doi.org/10.1016/j.bspc.2020.102003
48. Tan, Y., Young, M., Girish, A., (...), Bergmann, J. (2023) Predicting respiratory decompensation in mechanically ventilated adult ICU patients. *Frontiers in Physiology*. https://doi.org/10.3389/fphys.2023.1125991
49. https://www.sciencedirect.com/science/article/pii/S2047250125000327
50. de Carvalho, D., Hoffmann, K., Filho, J.R.N., Baptistella, A.R. (2026) Enhancing mechanical ventilation management with AI: Computer vision for automated detection of ventilatory modes, parameters and asynchrony. *Journal of Critical Care*. https://doi.org/10.1016/j.jcrc.2025.155238
51. Zhang, L., Feng, X., Lu, F., (...), Pan, Q. (2026) INTELLI-PVA: Informative sample annotation-based contrastive active learning for cross-domain patient-ventilator asynchrony detection. *Computer Methods and Programs in Biomedicine*. https://doi.org/10.1016/j.cmpb.2025.109203

52. Muñoz, J., Fernández-Araujo, N.J., Ruíz-Cacho, R., Muñoz-Visedo, J. (2026) Artificial intelligence and computerized decision support in adult intensive care: A systematic review of randomized controlled trials. *Journal of Critical Care*. https://doi.org/10.1016/j.jcrc.2026.155600

53. Boles, J.-M., Bion, J., Connors, A., (...), Welte, T. (2007) Weaning from mechanical ventilation. *European Respiratory Journal*. https://doi.org/10.1183/09031936.00010206

54. Saeed, M., Villarroel, M., Reisner, A.T., (...), Mark, R.G. (2011) Multiparameter intelligent monitoring in intensive care II: A public-access intensive care unit database. *Critical Care Medicine*. https://doi.org/10.1097/CCM.0b013e31820a92c6

55. Bellani, G., Laffey, J.G., Pham, T., (...), Pesenti, A. (2016) Epidemiology, patterns of care, and mortality for patients with acute respiratory distress syndrome in intensive care units in 50 countries. *JAMA - Journal of the American Medical Association*. https://doi.org/10.1001/jama.2016.0291

56. Ghiani, Alessandro, Paderewska, Joanna, Sainis, Alexandros, (...), Neurohr, Claus (2020) Variables predicting weaning outcome in prolonged mechanically ventilated tracheotomized patients: a retrospective study. *Journal of Intensive Care*. https://doi.org/10.1186/s40560-020-00437-4

57. Krieger, B.P., Isber, J., Breitenbucher, A., (...), Ershowsky, P. (1997) Serial measurements of the rapid-shallow-breathing index as a predictor of weaning outcome in elderly medical patients. *Chest*. https://doi.org/10.1378/chest.112.4.1029

58. Fialkow, Léa, Farenzena, Maurício, Wawrzeniak, Iuri Christmann, (...), Bozzetti, Mary Clarisse (2016) Mechanical ventilation in patients in the intensive care unit of a general university hospital in southern Brazil: an epidemiological study. *Clinics*. https://doi.org/10.6061/clinics/2016(03)05

59. Liu, S., Xu, Q., Xu, Z., (...), See, K.C. (2024) Reinforcement Learning to Optimize Ventilator Settings for Patients on Invasive Mechanical Ventilation: Retrospective Study. *Journal of Medical Internet Research*. https://doi.org/10.2196/44494

60. Li, F., Xin, H., Zhang, J., (...), Lian, Z. (2021) Prediction model of in-hospital mortality in intensive care unit patients with heart failure: Machine learning-based, retrospective analysis of the MIMIC-III database. *BMJ Open*. https://doi.org/10.1136/bmjopen-2020-044779

61. Knighton, Andrew J, Kuttler, Kathryn G, Ranade-Kharkar, Pallavi, (...), Grissom, Colin K (2022) An alert tool to promote lung protective ventilation for possible acute respiratory distress syndrome. *JAMIA Open*. https://doi.org/10.1093/jamiaopen/ooac050

62. Sun, Q., Chase, J., Zhou, C., (...), Shaw, G.M. (2021) Predicting pulmonary distension in a virtual patient model for mechanical ventilation. *IFAC-PapersOnLine*. https://doi.org/10.1016/j.ifacol.2021.10.237

63. Jung, I.-C., Zerlik, M., Schuler, K., (...), Sedlmayr, B. (2026) An Explanation User Interface for Artificial Intelligence–Supported Mechanical Ventilation Optimization for Clinicians: User-Centered Design and Formative Usability Study. *JMIR Formative Research*. https://doi.org/10.2196/77481

64. Ochiai, Ryoichi (2015) Mechanical ventilation of acute respiratory distress syndrome. *Journal of Intensive Care*. https://doi.org/10.1186/s40560-015-0091-6

65. Baedorf Kassis, E., Schaefer, M.S., Maley, J.H., (...), Talmor, D. (2021) Transpulmonary pressure measurements and lung mechanics in patients with early ARDS and SARS-CoV-2. *Journal of Critical Care*. https://doi.org/10.1016/j.jcrc.2021.02.005

66. Rees, Stephen Edward, Spadaro, Savino, Dalla Corte, Francesca, (...), Karbing, Dan Stieper (2022) Transparent decision support for mechanical ventilation using visualization of clinical preferences. *BioMedical Engineering Online*. https://doi.org/10.1186/s12938-021-00974-5

67. Ji, Q., Wu, Y., Yang, M., (...), Hu, W. (2026) Integration, challenges, and future of artificial intelligence in critical care medicine: comprehensive applications from predictive models to clinical integration. *Frontiers in Medicine*. https://doi.org/10.3389/fmed.2026.1874474

68. Ruiz-Botella, M., Manrique, S., Gómez, J., Bodí, M. (2024) Advancing ICU patient care with a Real-Time predictive model for mechanical Power to mitigate VILI. *International Journal of Medical Informatics*. https://doi.org/10.1016/j.ijmedinf.2024.105511

69. Esteban, A., Frutos, F., Alìa, I., (...), Blanco, J. (1995) A Comparison of Four Methods of Weaning Patients from Mechanical Ventilation. *New England Journal of Medicine*. https://doi.org/10.1056/NEJM199502093320601

70. Brochard, L., Rauss, A., Benito, S., (...), Lemaire, F. (1994) Comparison of three methods of gradual withdrawal from ventilatory support during weaning from mechanical ventilation. *American Journal of Respiratory and Critical Care Medicine*. https://doi.org/10.1164/ajrccm.150.4.7921460

71. Ely, E.W., Baker, A.M., Dunagan, D.P., (...), Haponik, E.F. (1996) Effect on the duration of mechanical ventilation of identifying patients capable of breathing spontaneously. *New England Journal of Medicine*. https://doi.org/10.1056/NEJM199612193352502

72. Chiew, Y.S., Tan, C.P., Chase, J.G., (...), Mat Nor, M.B. (2018) Assessing mechanical ventilation asynchrony through iterative airway pressure reconstruction. *Computer Methods and Programs in Biomedicine*. https://doi.org/10.1016/j.cmpb.2018.02.007

73. Schuijt, Michiel T. U., van Meenen, David M. P., Martin–Loeches, Ignacio, (...), Serpa Neto, Ary (2021) Association of Time–Varying Intensity of Ventilation With Mortality in Patients With COVID−19 ARDS: Secondary Analysis of the PRoVENT–COVID Study. *Frontiers in Medicine*. https://doi.org/10.3389/fmed.2021.725265

74. Phan, T.S., Costa, R., Haddad, W.M., (...), Gholami, B. (2020) Validation of an automated system for detecting ineffective triggering asynchronies during mechanical ventilation: a retrospective study. *Journal of Clinical Monitoring and Computing*. https://doi.org/10.1007/s10877-019-00442-5

75. Rees, S.E. (2011) The Intelligent Ventilator (INVENT) project: The role of mathematical models in translating physiological knowledge into clinical practice. *Computer Methods and Programs in Biomedicine*. https://doi.org/10.1016/S0169-2607(11)00307-5

76. Park, J.E., Kim, D.Y., Park, J.W., (...), Chung, W.Y. (2023) Development of a Machine Learning Model for Predicting Weaning Outcomes Based Solely on Continuous Ventilator Parameters during Spontaneous Breathing Trials. *Bioengineering*. https://doi.org/10.3390/bioengineering10101163

77. Tenza-Lozano, Eva, Llamas-Alvarez, Ana, Jaimez-Navarro, Enrique, Fernández-Sánchez, Javier (2018) Lung and diaphragm ultrasound as predictors of success in weaning from mechanical ventilation. *Critical Ultrasound Journal*. https://doi.org/10.1186/s13089-018-0094-3

78. Llamas-Álvarez, A.M., Tenza-Lozano, E.M., Latour-Pérez, J. (2017) Diaphragm and Lung Ultrasound to Predict Weaning Outcome: Systematic Review and Meta-Analysis. *Chest*. https://doi.org/10.1016/j.chest.2017.08.028

79. Levy, M.M., Miyasaki, A., Langston, D. (1995) Work of breathing as a weaning parameter in mechanically ventilated patients. *Chest*. https://doi.org/10.1378/chest.108.4.1018

80. Chi, Yi, Zhang, Qing, Yuan, Siyi, (...), He, Huaiwu (2021) Twenty-four-hour mechanical power variation rate is associated with mortality among critically ill patients with acute respiratory failure: a retrospective cohort study. *BMC Pulmonary Medicine*. https://doi.org/10.1186/s12890-021-01691-4

81. https://www.sciencedirect.com/science/article/pii/S1471225325001959

82. Sun, Q., Chase, J.G., Zhou, C., (...), Shaw, G.M. (2022) Over-distension prediction via hysteresis loop analysis and patient-specific basis functions in a virtual patient model. *Computers in Biology and Medicine*. https://doi.org/10.1016/j.compbiomed.2021.105022

83. Tams, C.G., Euliano, N.R., Martin, A.D., (...), Gentile, M.A. (2020) Patients alter power of breathing as the primary response to changes in pressure support ventilation. *Journal of Critical Care*. https://doi.org/10.1016/j.jcrc.2020.03.004

84. Petter, A.H., Chioléro, R.L., Cassina, T., (...), Revelly, J.-P. (2003) Automatic "Respirator/Weaning" with Adaptive Support Ventilation: The Effect on Duration of Endotracheal Intubation and Patient Management. *Anesthesia and Analgesia*. https://doi.org/10.1213/01.ANE.0000086728.36285.BE

85. Kirakli, C., Naz, I., Ediboğlu, O., (...), Tellioğlu, E. (2015) A randomized controlled trial comparing the ventilation duration between adaptive support ventilation and pressure assist/control ventilation in medical patients in the ICU. *Chest*. https://doi.org/10.1378/chest.14-2599

86. Kirakli, C., Ozdemir, I., Uçar, Z.Z., (...), Özkan, S.A. (2011) Adaptive support ventilation for faster weaning in COPD: A randomised controlled trial. *European Respiratory Journal*. https://doi.org/10.1183/09031936.00081510

87. Lellouche, F., Mancebo, J., Jolliet, P., (...), Brochard, L. (2006) A multicenter randomized trial of computer-driven protocolized weaning from mechanical ventilation. *American Journal of Respiratory and Critical Care Medicine*. https://doi.org/10.1164/rccm.200511-1780OC

88. Branson, R.D., Blakeman, T.C., Moore, L., (...), Goodman, M. (2026) Automated control of inspired oxygen in mechanically ventilated trauma and acute care surgery patients: A randomized controlled trial. *Journal of Trauma and Acute Care Surgery*. https://doi.org/10.1097/TA.0000000000005000

89. Roca, O., Caritg, O., Santafé, M., (...), Ricard, J.-D. (2022) Closed-loop oxygen control improves oxygen therapy in acute hypoxemic respiratory failure patients under high flow nasal oxygen: a randomized cross-over study (the HILOOP study). *Critical Care*. https://doi.org/10.1186/s13054-022-03970-w

90. Mol, C.G., da Silva Vieira, A.G., Garcia, B.M.S.P., (...), Nawa, R.K. (2024) Closed-loop oxygen control for critically ill patients-A systematic review and metaanalysis. *PLoS ONE*. https://doi.org/10.1371/journal.pone.0304745

91. Eimer, C., Selpien, H., Bastian, T., (...), Schädler, D. (2025) Automated control of inspired oxygen fraction in mechanically ventilated patients: A study protocol for a single-centre randomised controlled trial. *BMJ Open*. https://doi.org/10.1136/bmjopen-2024-089450

92. Bialais, E., Wittebole, X., Vignaux, L., (...), Hantson, P. (2016) Closed-loop ventilation mode (IntelliVent®-ASV) in intensive care unit: A randomized trial. *Minerva Anestesiologica*. https://www.scopus.com/pages/publications/84975865191

93. Rose, L., Schultz, M.J., Cardwell, C.R., (...), Blackwood, B. (2015) Automated versus non-automated weaning for reducing the duration of mechanical ventilation for critically ill adults and children: A cochrane systematic review and meta-analysis. *Critical Care*. https://doi.org/10.1186/s13054-015-0755-6

94. Jawa, R.S., Tharakan, M.A., Tsai, C., (...), Rubano, J.A. (2020) A reference guide to rapidly implementing an institutional dashboard for resource allocation and oversight during COVID-19 pandemic surge. *JAMIA Open*. https://doi.org/10.1093/jamiaopen/ooaa054

95. East, T.D., Morris, A.H., Wallace, C.J., (...), Sittig, D.F. (1991) A strategy for development of computerized critical care decision support systems. *International Journal of Clinical Monitoring and Computing*. https://doi.org/10.1007/BF01739127

96. Bowles, Tim, Trentino, Kevin M., Lloyd, Adam, (...), Waterer, Grant (2024) Health in a Virtual Environment (HIVE): A Novel Continuous Remote Monitoring Service for Inpatient Management. *Healthcare*. https://doi.org/10.3390/healthcare12131265

97. Goldberg, Elizabeth, Kao, Dave, Kwan, Bethany, (...), Zane, Richard (2024) UCHealth's virtual health center: How Colorado's largest health system creates and integrates technology into patient care. *npj Digital Medicine*. https://doi.org/10.1038/s41746-024-01184-8

98. https://www.sciencedirect.com/science/article/pii/S1573689X24001300

99. Rauseo, M., Cotoia, A., Padovano, F.P., (...), Cinnella, G. (2026) Weaning From Mechanical Ventilation in Cardiac Surgery Patients: Current Strategies, Monitoring Innovations, and Future Perspectives. *Journal of Cardiothoracic and Vascular Anesthesia*. https://doi.org/10.1053/j.jvca.2026.02.009

100. Colombo, D., Cammarota, G., Alemani, M., (...), Navalesi, P. (2011) Efficacy of ventilator waveforms observation in detecting patient-ventilator asynchrony. *Critical Care Medicine*. https://doi.org/10.1097/CCM.0b013e318225753c

101. Alqahtani, J.S., Alahmari, M.D., Alshamrani, K.H., (...), Arabi, Y.M. (2020) Patient-Ventilator Asynchrony in Critical Care Settings: National Outcomes of Ventilator Waveform Analysis. *Heart and Lung*. https://doi.org/10.1016/j.hrtlng.2020.04.002

102. Chelbi, R., Thabet, F., Ennouri, E., (...), Boussarsar, M. (2024) The Ability of Critical Care Physicians to Identify Patient-Ventilator Asynchrony Using Waveform Analysis: A National Survey. *Respiratory Care*. https://doi.org/10.4187/respcare.11360

103. Theilen, R., Huhle, R., Scharffenberg, M., (...), Wittenstein, J. (2026) Intelligent Lung Support in the Intensive Care Unit (IntelliLung): study protocol for an international observational, prospective, multicentre study. *BMJ open*. https://doi.org/10.1136/bmjopen-2025-114851

104. East, T.D., Heermann, L.K., Bradshaw, R.L., (...), Steinberg, K. (1999) Efficacy of computerized decision support for mechanical ventilation: results of a prospective multi-center randomized trial.. *Proceedings / AMIA ... Annual Symposium. AMIA Symposium*. https://www.scopus.com/pages/publications/0033257291

105. https://www.sciencedirect.com/science/article/pii/S1916724517000246

106. van Limpt, G.J.C., van Vliet, P., Molenaar, M.A., (...), Buiteman–Kruizinga, L.A. (2026) Alarms and alarm management with automated versus conventional ventilation in neurocritical care patients. *Intensive and Critical Care Nursing*. https://doi.org/10.1016/j.iccn.2026.104471

107. Popoff, B., Cabon, S., Cuggia, M., (...), Clavier, T. (2025) Expectations of Intensive Care Physicians Regarding an AI-Based Decision Support System for Weaning From Continuous Renal Replacement Therapy: Predevelopment Survey Study. *JMIR Medical Informatics*. https://doi.org/10.2196/63709

108. Pranata, Y., Mirwanti, R., Bahtiar, Y.A., Purwanti, H. (2026) Translational Potential and Explainability of Artificial Intelligence-Based Clinical Decision Support for Adults in Intensive Care: A Scoping Review. *Journal of Multidisciplinary Healthcare*. https://doi.org/10.2147/JMDH.S621469

109. Snowden, S., Brownlee, K.G., Dear, P.R.F. (1997) An expert system to assist neonatal intensive care. *Journal of Medical Engineering and Technology*. https://doi.org/10.3109/03091909709008406

110. https://www.sciencedirect.com/science/article/pii/S1471246626001444