Calibration Due Summary Report

Instrument: MICROPIPETTE
Ref No: MIPL/QC/215
Make: Labquest (Borosil)
Location: QC
Range: NMT 2 % RMB & %CV
Least Count: NA
Calibration: NMT 2 % RMB & %CV
Frequency: 6 Months
Performed By: INHOUSE
Cal Done: 2026-04-03
Cal Due: 2026-10-02

Instrument: AUTOPIPETTE
Ref No: MIPL/QC/214
Make: Labquest (Borosil)
Location: QC
Range: NMT 2 % RMB & %CV
Least Count: NA
Calibration: NMT 2 % RMB & %CV
Frequency: 6 Months
Performed By: INHOUSE
Cal Done: 2026-04-03
Cal Due: 2026-10-02

Instrument: AUTOPIPETTE
Ref No: MIPL/QC/216
Make: Labquest (Borosil)
Location: QC
Range: NMT 2 % RMB & %CV
Least Count: NA
Calibration: NMT 2 % RMB & %CV
Frequency: 6 Months
Performed By: INHOUSE
Cal Done: 2026-04-03
Cal Due: 2026-10-02