A METHOD FOR ESTIMATING THE RESOLUTION OF REAL IMAGES

Ryuta Mizutani*1, Rino Saiga1, Susumu Takekoshi2, Chie Inomoto2, Naoya Nakamura2, Makoto Arai³, Kenichi Oshima³, Masanari Itokawa³, Akihisa Takeuchi⁴, Kentaro Uesugi⁴, Yasuko Terada⁴, and Yoshio Suzuki⁵


¹Department of Applied Biochemistry, Tokai University, Hiratsuka, Kanagawa 259-1292, Japan; ²Tokai University School of Medicine, Isehara, Kanagawa 259-1193, Japan; ³Tokyo Metropolitan Institute of Medical Science, Setagaya, Tokyo 156-8506, Japan; ⁴Japan Synchrotron Radiation Research Institute (JASRI/SPring-8), Sayo, Hyogo 679-5198, Japan; ⁵Graduate School of Frontier Sciences, University of Tokyo, Kashiwa, Chiba 277-8561, Japan

Estimating the resolution of real images

The resolution of sample image depends on a number of factors, including X-ray optics, stage drift and sample condition. The resolvability of the sample image should therefore be determined from the sample image itself. Here we report a method for estimating the spatial resolution of real images from a logarithmic intensity plot in the Fourier domain [1].


The analyses of test images indicated that the full width at half maximum (FWHM) of a Gaussian point-spread function applied to the original image can be estimated from the logarithmic intensity plots. The spatial resolution of imaging X-ray microtomography using Fresnel zone-plate optics was estimated with this method. The FWHM of the point spread function estimated from a cross section of human brain neuron coincided with the resolution determined from the test object.


These results indicated that the logarithmic intensity plot in the Fourier domain provides an alternative measure of spatial resolution without explicitly defining a



An aluminum test object with square-wave patterns 1

An aluminum rod with approximate dimensions of 120 x 120 x 2000 µm was subjected to ion-beam milling using a focused ion-beam (FIB) apparatus [1-3]. A series of square wells was carved on the surface. Each pattern was composed of half-pitch wells and half-pitch intervals, i.e., a 500-nm well and 500-nm interval for a 1000-nm pitch.






-Microtomography with FZP optics


microtomography Imaging performed at the BL37XU beamline of SPring-8. A Fresnel zone plate with an outermost zone width of 50 nm was used as an X-ray objective lens, and an X-ray guide tube as a beam condenser. Transmission images produced by 8-keV X-rays were recorded using a CMOSbased detector placed at 27.5 m from the sample. The cross section indicated that the patterns up to 120 nm pitches were clearly resolved. Some structures of the 100-nm pitch pattern were also visualized, indicating that the resolution is about 100-120 nm.

The lower panel shows a cross section of a human pyramidal neuron taken with the same imaging optics. Cellular dendrite, and structures are visualized. This section was used for the resolution estimation with logarithmic intensity plot.



Estimation of spatial resolution from logarithmic intensity plot in the Fourier domain 1


First, original image (panel a) was subjected to Fourier transform, giving a frequency-domain pattern (b). Then, the logarithm of squared norm of the Fourier transform was plotted against squared distance from the origin (c). Spatial resolution was estimated from the linear correlation observed at the left end of the plot. The slope of the rest of this plot represents the Nyquist limit of digitization.




← Plots of test images

Test images were generated by convolving the Gaussian pointspread function and the original image (left panels). The appearance of bar patterns embedded in the original image proved that the Gaussian function applied to the original image introduced blurring corresponding to the Gaussian FWHM.


These test images were analyzed with the logarithmic intensity plot in the Fourier domain (right panels). Linear correlations corresponding to the Gaussian were observed at the left ends of the plots. These indicated that the Gaussian pointspread function can be extracted from the image.






The same slope was identified when the same Gaussian blurring was introduced. This is independent of the type of original image, including a satellite image of Tokyo Bay (@JAXA). Therefore, provide method resolution measure that is directly related to the FWHM of the point spread function. The primary advantage of this method is that the resolution can be estimated without defining a noise explicitly criterion.

FWHM = 4




←The spatial resolution of a neuron cross section (left) was estimated from the logarithmic intensity plot in the Fourier domain. The plot was calculated from the boxed area. The FWHM estimated from this plot was 118 nm, which coincides with the resolution (100-120) nm) determined from the test object. We suggest that the logarithmic intensity plot in the Fourier domain provides an alternative measure of image resolvability 0.00015 regardless of visualization modality.



0.03 0.04

|k|2 (pixel-2)

10 Brain section A PM5544 test chart Hutch photo □ Text scan Applied Gaussian FWHM (pixel)



Lab site: https://mizutanilab.github.io/ YouTube: https://www.youtube.com/user/mizutaniLab

ryuta@tokai-u.jp

