

Cytiva Xcellerex™ XDR-10 - 4.5-10 L working volume

Local source PDF | preset cytiva-xdr-10 | compiled 10 September 2026

This PDF is stored inside the GitHub project and records the evidence used by the simulator. It is not a copied vendor brochure or manual. The official manufacturer references below remain the authoritative originals.

Manufacturer: Cytiva

Product-contact material: Plastic / single-use

Working volume: 4.5 L to 10 L

Nominal vessel or bag size: 10 L

Simulation profile: mammalian

Evidence status: Direct official model evidence

Original manufacturer references

1. Cytiva - Xcellerex XDR bioreactor system catalogue

[Open official manufacturer source online](#)

<https://www.cytivalifesciences.co.jp/catalog/1754.html>

Supports: XDR family, XDR-10 vessel parameters and larger-scale model range.

Evidence basis

Volume: The 4.5-10 L working-volume range is directly supported.

Material: Disposable single-use bag is the product-contact vessel; acrylic/stainless support hardware is not the wetted vessel.

- The 1.5 H/D and single pitched-blade impeller are model-specific source-supported fields.

Record note

This source PDF is paired with the detailed HTML provenance record in the same repository. Simulator-specific defaults and derived assumptions are listed there.

Copyright and trademarks in the linked manufacturer documentation remain with the respective vendor.