

**نموذج طلب انضمام منشأة صحيه للقيام باجراءات جراحه العيون ضمن المشروع الرئاسي للقضاء علي قوائم الانتظار**

|   |  |                |
|---|--|----------------|
| <b>Name</b>   |  | <b>Comment</b> |
| <b>Address</b>  |  |                |
| <b>Number of beds</b>   |  |                |
| <b>Number of ICU</b>  |  |                |
| <b>Basic Requirements:</b>  |  |                |
| Number of operating Rooms:<br>minimum two   |  |                |
| Number of ophthalmic<br>microscopes: minimum two  |  |                |
| Number of phacoemulsification<br>machines and its types, model and<br>Maintenance schedule : minimum<br>two not older than 7 years model  |  |                |
| Sterilization types and number  |  |                |
| Anaesthesia machine equipped for<br>General Anaesthesia   |  |                |
| Safety and occupational health<br>(fire exits, automatic operated fire<br>distinguisher, backup generator):<br>schematic should be provided and<br>approved (UPS and power<br>generators.. etc.,) |  |                |
| Patient records   |  |                |
| Life support trained personnel  |  |                |
| Personnel :<br>-Registered nurses per operating<br>room<br>-Number of operations per room<br>-Anaesthesia consultant and two<br>specialists   |  |                |

|  |  |  |
|--|--|--|
| <b>Vitreoretinal Requirement :</b>   |  |  |
| Vitreoretinal machine  |  |  |
| Visualization system   |  |  |
| Vitreoretinal consultant   |  |  |
| Cryotherapy machine (optional)   |  |  |
| Endolaser apparatus  |  |  |
| Investigative machines   |  |  |
| <b>Any center that is not equipped with vitreoretinal department doesn't handle intravitreal injections or any intravitreal procedures</b> |  |  |
| <b>Outpatient: Ophthalmology clinic equipped with</b>  |  |  |
| Slit Lamp  |  |  |
| Examination lens 90D or 78D  |  |  |
| Auto refractometer   |  |  |
| Tonometer  |  |  |
| Lens trial set   |  |  |

-تتعهد ادارته كل مستشفى و المدير الطبي للمستشفى بوضع علامه صح في المربع المخصص.

- موافقه و التزام ادارته المستشفى باتباع كافة الإجراءات الخاصه بالمشروع و ذلك من خلال الالتزام بكافه البروتوكولات العلميه الموضوعه من قبل اللجنه العلميه بالمشروع فيما يتعلق بالتشخيص للمرضي و آليه علاجهم و اليه المتابعه الخاصه بهم بعد الاجراء , مع الاستعداد لاستقبال أعضاء مرور من الغرفه المركزيه لقوائم الانتظار لفحص الملفات الطبيه لمرضي قوائم الانتظار دون سابق انذار.

تحريرا في .../...../.....

مدير المستشفى

رئيس القسم