

GOVERNMENT OF INDIA • MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION

DEPARTMENT OF CONSUMER AFFAIRS — DIRECTORATE OF LEGAL METROLOGY

NATIONAL LEGAL METROLOGY TYPE EVALUATION LABORATORY (OIML ISSUING AUTHORITY)

PATTERN EVALUATION TEST REPORT • NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI)

Evaluated in accordance with OIML Recommendation R 76-1:2006 (E) & Legal Metrology Act, 2009

Report No:	OIML-R76-2026-IND-0012	Application No:	APP-NAWI-2026-0012
Testing Date:	2026-08-14	Overall Finding:	APPROVED COMPLIANT

1. INSTRUMENT IDENTIFICATION & TECHNICAL SPECIFICATIONS

Manufacturer:	Metler Precision Weighing Systems India Pvt Ltd	Instrument Type:	Electronic Retail Counter Scale
Manufacturer Address:	Plot 44-46, Electronics City Phase II, Bangalore, Karnataka - 560100	Model Name:	MPS-Retail-3000
Accuracy Class:	Class III	Serial Number:	SN-2026-MPS-04891
Max Capacity (Max):	15.0 kg	Scale Interval (d):	0.005 kg
Min Capacity (Min):	0.1 kg	Verification Interval (e):	0.005 kg
Verification Scale Intervals (n):	n = 3000	Tare Subtraction (Max T):	-15.0 kg
Operating Temperature:	-10.0°C to +40.0°C	Power Supply:	230V AC (+10% / -15%), 50Hz with 6V 4.5A
Load Cell & Indicator:	Zemic L6E3 Aluminum Single Point C3 Load Cell	Firmware / Checksum:	v2.1.0-IND (CRC32: 0x9B41F0A2)

2. LABORATORY AND ENVIRONMENTAL CONDITIONS (Clause A.3)

Testing Laboratory:	National Legal Metrology Type Evaluation Laboratory	Ambient Temperature:	23.4 °C
Laboratory Accreditation:	NABL ISO/IEC 17025 Accredited & OIML Issuing Authority	Relative Humidity:	51.0 % RH
Local Gravity Acceleration (g):	9.7803 m/s²	Barometric Pressure:	1012.8 hPa
Standard Calibration Weights:	Class M1 Calibrated Working Standards (Traceability: NP)	Testing Officer:	Er. Rajesh Sharma (Senior Metrological Officer)

3. WEIGHING PERFORMANCE TEST OBSERVATIONS & EVALUATION (Clause A.4.4)

#	Dir	Load (kg)	Indication I	ΔL (kg)	Error E _c	E _c (e)	MPE (±e)	MPE (kg)	Verdict
1	INCR	0	0	0.0025	+0.0000	+0.00 e	±0.5 e	±0.0025	PASS
2	INCR	0.1	0.1	0.0025	+0.0000	+0.00 e	±0.5 e	±0.0025	PASS
3	INCR	2.5	2.5	0.0024	+0.0001	+0.02 e	±0.5 e	±0.0025	PASS
4	INCR	5	5	0.0025	+0.0000	+0.00 e	±1.0 e	±0.0050	PASS
5	INCR	7.5	7.5	0.0026	-0.0001	-0.02 e	±1.0 e	±0.0050	PASS
6	INCR	10	10	0.0027	-0.0002	-0.04 e	±1.0 e	±0.0050	PASS
7	INCR	15	15	0.0025	+0.0000	+0.00 e	±1.5 e	±0.0075	PASS
8	DEC R	10	10	0.0026	-0.0001	-0.02 e	±1.0 e	±0.0050	PASS
9	DEC R	5	5	0.0025	+0.0000	+0.00 e	±1.0 e	±0.0050	PASS

10	DEC R	2.5	2.5	0.0025	+0.0000	+0.00 e	±0.5 e	±0.0025	PASS
11	DEC R	0.1	0.1	0.0025	+0.0000	+0.00 e	±0.5 e	±0.0025	PASS
12	DEC R	0	0	0.0025	+0.0000	+0.00 e	±0.5 e	±0.0025	PASS

4. COMPREHENSIVE OIML R 76-1 TEST COMPLIANCE SUMMARY

Clause	Test Procedure	Statutory Acceptance Criteria	Result Status
Clause A.4.4	Weighing Performance Test (Incr/Decr)	Corrected error Ec <= MPE across range; Hysteresis <= MPE	PASS
Clause A.4.10	Repeatability Test (at 0.5 Max & Max)	Max difference between weighings <= MPE for load	PASS
Clause A.4.7	Eccentricity / Off-Center Loading	Ec <= MPE at 1/3 Max at 4 corners and center	FAIL
Clause A.4.2	Zero-Setting and Zero-Tracking	Residual zero error <= +/- 0.25 e	PASS
Clause A.4.8	Discrimination Test	Extra load 1.4 d gives noticeable change of indication	PASS
Clause A.5.3	Static Temperatures (-10°C to +40°C)	Span error <= MPE; Zero drift <= 1 e / 5°C	PASS
Clause A.5.4	Voltage Variations (+10% / -15%)	Indications remain within MPE limits	PASS

STATUTORY MODEL APPROVAL VERDICT: COMPLIANT & APPROVED

The weighing instrument submitted meets all metrological and technical requirements of OIML Recommendation R 76-1:2006 (E) and the Legal Metrology (General) Rules, 2011. Type approval is hereby recommended under Section 22 of the Legal Metrology Act, 2009.

TESTED & EVALUATED BY: Er. Rajesh Sharma (Senior Metrological Officer) Senior Metrological Officer, Legal Metrology Digital Signature: VERIFIED • ID: GOV-IN-LM-0881 Timestamp: 2026-08-16T11:45:00Z SHA-256 Hash: a4d8c910398fbe8c7d9e1029348bac19...	VERIFIED & APPROVED BY: Dr. Priya V. Iyer (Director of Legal Metrology) Director of Legal Metrology, Government of India Official Stamp: [DIGITALLY SIGNED & VERIFIED] Legal Metrology Act, 2009 Statutory Repository Tamper-Evident Digital Certification Authority
--	--