

GOVERNMENT OF INDIA • MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION

DEPARTMENT OF CONSUMER AFFAIRS — DIRECTORATE OF LEGAL METROLOGY

NATIONAL LEGAL METROLOGY TYPE EVALUATION LABORATORY (OIML ISSUING AUTHORITY)

PATTERN EVALUATION TEST REPORT • NON-AUTOMATIC WEIGHING INSTRUMENT (NAWI)

Evaluated in accordance with OIML Recommendation R 76-1:2006 (E) & Legal Metrology Act, 2009

Report No:	OIML-R76-2026-IND-0027	Application No:	APP-NAWI-2026-0027
Testing Date:	2026-09-20	Overall Finding:	TESTING IN PROGRESS

1. INSTRUMENT IDENTIFICATION & TECHNICAL SPECIFICATIONS

Manufacturer:	Bharat Heavy Weighbridges & Sensorics Ltd	Instrument Type:	Pitless Electronic Weighbridge
Manufacturer Address:	Industrial Area Phase 1, Chandigarh - 160002	Model Name:	BHW-Titan-60T
Accuracy Class:	Class III	Serial Number:	SN-2026-WB-60T-004
Max Capacity (Max):	60000.0 kg	Scale Interval (d):	20.0 kg
Min Capacity (Min):	400.0 kg	Verification Interval (e):	20.0 kg
Verification Scale Intervals (n):	n = 3000	Tare Subtraction (Max T):	-60000.0 kg
Operating Temperature:	-10.0°C to +50.0°C	Power Supply:	230V AC 50Hz with 12V Heavy Industrial U
Load Cell & Indicator:	6 x Flintec RC3 30t Stainless Steel Canister	Firmware / Checksum:	v4.0.2-WB (CRC: 0xF728D1B0)

2. LABORATORY AND ENVIRONMENTAL CONDITIONS (Clause A.3)

Testing Laboratory:	Regional Legal Metrology Verification Center (Field Station)	Ambient Temperature:	29.0 °C
Laboratory Accreditation:	NABL ISO/IEC 17025 Accredited	Relative Humidity:	58.0 % RH
Local Gravity Acceleration (g):	9.7803 m/s²	Barometric Pressure:	1008.5 hPa
Standard Calibration Weights:	Mobile Test Van with 20 tonnes of Calibrated Class M1 B	Testing Officer:	Er. Rajesh Sharma (Senior Metrological Officer)

3. WEIGHING PERFORMANCE TEST OBSERVATIONS & EVALUATION (Clause A.4.4)

#	Dir	Load (kg)	Indication I	ΔL (kg)	Error E _c	E _c (e)	MPE (±e)	MPE (kg)	Verdict
1	INCR	0	0	5	+0.0000	+0.00 e	±0.5 e	±0.0025	PASS
2	INCR	1000	1000	5	+0.0000	+0.00 e	±1.5 e	±0.0075	PASS
3	INCR	10000	10000	6	-1.0000	-200.00 e	±1.5 e	±0.0075	FAIL
4	INCR	20000	20000	5	+0.0000	+0.00 e	±1.5 e	±0.0075	PASS
5	INCR	40000	40000	8	-3.0000	-600.00 e	±1.5 e	±0.0075	FAIL
6	INCR	60000	60000	5	+0.0000	+0.00 e	±1.5 e	±0.0075	PASS

4. COMPREHENSIVE OIML R 76-1 TEST COMPLIANCE SUMMARY

Clause	Test Procedure	Statutory Acceptance Criteria	Result Status
Clause A.4.4	Weighing Performance Test (Incr/Decr)	Corrected error E _c <= MPE across range; Hysteresis <= MPE	FAIL

Clause A.4.10	Repeatability Test (at 0.5 Max & Max)	Max difference between weighings $\leq$ MPE for load	PENDING
Clause A.4.7	Eccentricity / Off-Center Loading	Ec $\leq$ MPE at 1/3 Max at 4 corners and center	FAIL
Clause A.4.2	Zero-Setting and Zero-Tracking	Residual zero error $\leq$ $\pm$ 0.25 e	PENDING
Clause A.4.8	Discrimination Test	Extra load 1.4 d gives noticeable change of indication	PENDING
Clause A.5.3	Static Temperatures (-10°C to +40°C)	Span error $\leq$ MPE; Zero drift $\leq$ 1 e / 5°C	PENDING
Clause A.5.4	Voltage Variations (+10% / -15%)	Indications remain within MPE limits	PENDING

STATUTORY MODEL APPROVAL VERDICT: NON-COMPLIANT & REJECTED

The weighing instrument submitted failed one or more prescribed tests of OIML R 76-1:2006 (E). Rectification and re-evaluation are required before approval can be granted.

TESTED & EVALUATED BY: Er. Rajesh Sharma (Senior Metrological Officer) Senior Metrological Officer, Legal Metrology Digital Signature: VERIFIED • ID: GOV-IN-LM-0881 Timestamp: 2026-09-28T10:25:07.609452+00:00 SHA-256 Hash: 4b556a47d271722d89978be196a1f499...	VERIFIED & APPROVED BY: Dr. Priya V. Iyer (Director of Legal Metrology) Director of Legal Metrology, Government of India Official Stamp: [DIGITALLY SIGNED & VERIFIED] Legal Metrology Act, 2009 Statutory Repository Tamper-Evident Digital Certification Authority
---	---