

Compliance Summary

Date Range	7/8/2026 - 8/6/2026 (30 days)
Days with Device Usage	29 days
Days without Device Usage	1 day
Percent Days with Device Usage	96.7%
Cumulative Usage	9 days 21 hrs. 39 mins. 34 secs.
Maximum Usage (1 Day)	9 hrs. 23 mins. 46 secs.
Average Usage (All Days)	7 hrs. 55 mins. 19 secs.
Average Usage (Days Used)	8 hrs. 11 mins. 42 secs.
Minimum Usage (1 Day)	6 hrs. 28 mins. 51 secs.
Percent of Days with Usage >= 4 Hours	96.7%
Percent of Days with Usage < 4 Hours	3.3%
Total Blower Time	9 days 22 hrs. 28 mins. 38 secs.

CPAP Summary

Average Time in Large Leak Per Day	1 hrs. 18 mins. 4 secs.
Average AHI	4.7
CPAP	13.0 cmH2O

Device Settings as of 8/6/2026

Device Mode CPAP - None

Device Settings

Parameter	Value
Auto-Trial	Off
CPAP Pressure	13 cmH2O
Flex Lock	Off
Auto Off	Off
Auto On	Off
View Optional Screens	Off
Ramp Type	SmartRamp
Ramp Time	20 minutes
Ramp Start Pressure	4.0 cmH2O
Mask Resistance	1
Mask Resistance Lock	Off
Tubing Type	22
Tubing Type Lock	Off
EZ-Start	Disabled
Tube Temperature	3
Humidifier	3
Humidification Mode on Heated Tube Disconnect	Adaptive