


ETH MIKE - Manual



Contents

1 List of all Parts	2
2 Setup	3
2.1 ETH MIKE Full Setup	3
2.2 Cables	3
2.3 Emergency Button	3
2.4 Hand and Finger Interface	4
2.5 Placement of Tablet	5
2.6 Incline of the Device	5
2.7 How to secure the Wrist for the Experiments	5
2.8 Calibration and Switching on	6
3 Software	8
3.1 ETH MIKE Program	8
3.2 Login	8
4 Assessment Tasks	10
4.1 Task Selection Menu	10
4.2 Range of Motion (motor impairments)	10
4.3 Force (motor impairments)	11
4.4 Position Matching (proprioception impairments)	11
4.5 Fast Reaching (motor impairments)	13
4.6 Target Following (sensorimotor impairments)	13
4.7 Result Display	14
5 Data Storage	16
5.1 Raw Data	16
5.2 Database	16
6 Packing for Transportation	17
7 Troubleshooting	19
7.1 Software disconnects from Hardware	19
7.2 Software crashes	20
8 Contact Information	21

1 List of all Parts

The robot ETH MIKE comes with the parts shown in Figure 1 and Table 1 below.




Figure 1: Overview of all components of the ETH MIKE robot.

Number	Item	Quantity
1	Extension cable (power strip)	1
2	Tablet power supply	1
3	Robot main power cable	1
4	USB connection cable between tablet and robot	1
5	Emergency stop button cable	1
6	Emergency stop button	1
7	Adapter (UK/EU depending on delivery destination)	1
8	Finger module	1
9	Pins for the finger module (short) and handles (long)	1
10	Handles (left- and right-hand version)	1
11	Tablet computer and keyboard	1
12	Tablet pen	1
13	Cable to charge the tablet pen	1
14	Allen key	1
15	Incline plane	1
16	End-effector lock (not on picture)	1
17	Protection pin (not on picture)	1
19	Leukotape classic (not on picture)	1
20	Tricofix (not on picture)	1
21	Wrist splint (not on picture)	1

Table 1: List of all components of the ETH MIKE robot.

2 Setup

2.1 ETH MIKE Full Setup

Below you can see how the ETH MIKE looks like when it's fully set-up. In the following steps below, you will be guided through the process how to get to this stage.




Figure 2: ETH MIKE full setup – with and without the tablet computer placed on the dedicated rails.

2.2 Cables

Firstly, place the extension cable (#1 Table 1) within reach. Then, insert the robot main power cable (#3 Table 1) to the input on the robot (#1 Figure 3) and its other side to the extension cable. As a next step connect the emergency stop button with its cable (#5 and #6 Table 1) and insert the other side of that cable to the robot (#2 Figure 3). Next, connect the USB cable to the robot (#3 Figure 3) and to the USB input port of the tablet. Finally, connect the power supply of the tablet to the tablet and to the extension cable. Power the extension cable using the adapter (#7 Table 1).




Figure 3: How to connect power cable and emergency button to the robot.

2.3 Emergency Button

The emergency stop button (#6 Table 1) has two states: released and pressed (Figure 4). Pressing the emergency button cuts the current to the actuation of the robot and should only be used in emergencies. To re-enable the actuation, the

emergency button needs to be released by turning the red knob, and the robot restarted.



Figure 4: Emergency button – released and pressed.

2.4 Hand and Finger Interface

In order to prepare the hand and finger interface, first place the finger module (#8 Table 1) as shown in Figure 5 and secure it with the shorter pin. Then, place one of the handles and secure it with two longer pins. Check Figure 5 for demonstration of hand placement. The placement of the hand into the system is demonstrated in detail in this video.




Figure 5: Placement of the handles and finger interface. The picture on the left demonstrates the setup for the left hand and the picture on the right the setup for the right hand.

2.5 Placement of Tablet

Locate the tablet computer on the rails above the hand interface. Follow the steps indicated in Figure 6 to securely place and remove the tablet.




Figure 6: Instructions on how to securely place and release the tablet on/from its rails.

2.6 Incline of the Device

In order to avoid parallax error and ensure that participants are looking straight onto the device when performing the experiments, the robot needs to be placed at an inclined position. For that, we are using an incline plane (#15 Table 1). Place the incline plane under the robot, as shown in Figure 7. During experiments, make sure that participants are sitting straight in front of the device with their elbow rested on an arm rest (Figure 7).



Figure 7: Incline plane placement and correct positioning in front of the ETH MIKE.

2.7 How to secure the Wrist for the Experiments

To ensure that there is no compensation from wrist movements when performing the experiments (especially relevant for the range of motion task to avoid hyperextension

due to the wrist rotation). First, you need to put a protection on subject's wrist in a form of Tricofix sleeve. Then, place the wrist splint and wrap it around with the Leukotape bandage to secure (Figure 8). Throw away the sleeve and the piece of bandage after use and disinfect the wrist splint (reusable). The wrist splint is suitable for both left and right hand.




Figure 8: How to secure the wrist for experiments. It is mainly important for the range of motion task to avoid hyperextension due to wrist rotation.

2.8 Calibration and Switching on

Calibration

The robot needs to be calibrated every time you switch it on. Once it is switched on, you don't need to recalibrate it. Ensure the robot is switched on only when you are using it and that you don't leave it switched on overnight.

In order to calibrate the robot, place the end-effector lock (#16 Table 1) so that it locks the end-effector in the middle position (as shown in Figure 9).

This calibration process needs to be performed when you switch the device on. Wait a couple of seconds with the metal element inserted in the correct place until all LEDs are switched on and you hear a clicking sound (this is when calibration happens). After this, you can remove the end-effector lock and move the end-effector freely. The end-effector lock is used again in the Force Task (see instructions for individual assessments in Section 4).

Note: when you first unpack the device, the end-effector lock is already in place (protection for transportation). The end-effector lock was tightened for transportation purposes. To loosen it, use Allen key (#14 Table 1) by turning the screw on the side of the metal piece.




Figure 9: Calibration using the end-effector lock by positioning the end-effector in the centre. Ensure the end-effector is in this position every time you switch on the robot.

LEDs

When the robot is switched off, all LEDs are also off (Figure 10A). When you switch it on and the emergency button is released, 4 out of 5 LEDs will be turned on (Figure 10C). Then, after a few seconds (this is when the end-effector lock needs to be correctly inserted as in Figure 9), all LEDs will be switched on (Figure 10D). If the emergency button is pressed, not all LEDs will be turned on (Figure 10B). Ensure the emergency button is released and all LEDs are switched on when using the robot for experiments.




Figure 10: States of the LEDs on the side of the device.

3 Software

3.1 ETH MIKE Program

To start using the ETH MIKE robot software, first you need to switch on the tablet (button on the top). Then, log in to the user “ETH-MIKE-x” with the password “admin”.

On the desktop, find the following: “ETH Mike – Shortcut” and “ETH MIKE xxx” folder. The first one is the main software and the other is the folder where all data is stored.



Figure 11: Screenshot of the tablet desktop with indicated files/folders of importance.

3.2 Login

The initial view of the “ETH Mike – Shortcut” application is shown in Figure 12. First click on the person icon to create a new subject account. A Subject ID will be generated (Figure 13). Write down that ID as you will need it to identify stored data. After you create a new login, click “New session” to start the assessments. In case you log-out during an assessment session and would like to resume, click “Resume last session”.




Figure 12: Login view of the ETH MIKE software application.

New Patient

Subject ID
slgv

Gender:
 Female Male

Age:
XX

Impaired Hand:
 Left Hand Right Hand

Handedness:
 Left handed Right handed Ambidextrous

Register

Back

Figure 13: Window for creating a new subject. Subject ID is automatically generated, under which all data for this subject will be stored – write down this subject ID to be able to identify your files later.

4 Assessment Tasks

For a video explanation of all tasks, watch this video.

4.1 Task Selection Menu

Task selection menu is shown in Figure 14. It allows you to select the hand (use the slider to switch between right and left hand) and the task you want to perform. Question marks contain explanation of the procedure of each task, as well as a possibility to conduct practice trials. A task completion during a session is indicated by a tick next to it, see Force task in Figure 14.




Figure 14: Assessment task selection menu.

4.2 Range of Motion (motor impairments)

Active Range of Motion

First flex (bend) your finger as far as possible and then extend (stretch) it. The blue needle indicates your finger position. Before each trial the finger is first automatically moved to the starting position by the robot (*Get ready*). Once you have performed the movement, click *Finish* to proceed to the next trial. There are 3 trials.

Passive Range of Motion

The tablet needs to be removed for this part of the experiment to allow the experimenter to access the subject's finger. The experimenter moves the participant's finger to maximum flexion and then maximum extension. The finger is automatically moved to the starting position by the robot before each trial (*Get ready*). Once you have performed the movement, click *Finish* to proceed to the next trial. There are 3 trials. The participant needs to relax the finger.

Automatic Passive Movement

The tablet remains removed for this part of the experiment. Participant's finger will be automatically moved 3 times within the passive range of motion. Please relax the finger.




Figure 15: Instructions of the active range of motion task.

4.3 Force (motor impairments)

The end-effector lock needs to be inserted for this task. Remove the tablet and insert the end effector lock as demonstrated in Figure 16 (at 45 degrees). After securing the end-effector, place the tablet above the hand on the designated rails again.

The instruction of this task is to generate maximum fingertip force, first in flexion, then in extension. When prompted on the screen (as shown in Figure 17), flex (bend) your finger with as much force as possible. This is repeated 3 times, followed by 3 times extension (stretch). There is a 3 seconds preparation phase before the 3 seconds force generation phase.

Remove the end-effector lock after completion of the task.

4.4 Position Matching (proprioception impairments)

The objective of this task is to measure the finger position perceived by the subject without visual feedback (due to the placement of the tablet), thereby relying on proprioception.

The robot first moves the finger to a starting position (indicated in red) and then immediately to a new position. Drag the green gauge needle to indicate the currently perceived finger position (Figure 18). Tap *Validate* to confirm your selection.




Figure 16: Location of the end-effector lock for the left hand (left side) and the right hand (right side) for the maximum fingertip force generation task.




Figure 17: Interface of the Force Task indicating to bend the finger as strong as possible.

In case the participant is unable to indicate on the screen by him/herself with the other hand, the experimenter moves the gauge needle for him/her by asking “is the finger above/below” and dragging the needle until the participant says stop. There are 11 repetitions.

This task is normally the most challenging to understand, therefore it is recommended to perform practice round before, where visual feedback is provided for 3 trials.



Figure 18: Position matching task – the finger is automatically moved to the starting position by the robot (red gauge needle). Then, the finger is moved to another position and the subject needs to indicate, by dragging the green needle, the perceived finger position.

4.5 Fast Reaching (motor impairments)

The objective of this task is to move as fast as possible from a starting position (red gauge needle) to the target (green gauge needle), as shown in Figure 19. The finger is first automatically moved to the starting position. As soon as the green needle appears, move towards it as fast as possible and keep your finger at it until the needle disappears again. Each trial lasts 4 seconds. The target will appear either in flexion or in extension direction. There are 10 repetitions.




Figure 19: In the fast target reaching task, move your finger as fast as possible from the starting position (red needle) to the target (green needle).

4.6 Target Following (sensorimotor impairments)

The objective of the task is to follow the target displayed on the screen as accurately as possible with your finger. The finger is first automatically moved to the starting position (red needle). Then, the red needle will disappear, and a green needle will appear, which will move slowly (see Figure 20). Follow this trajectory as accurately as possible. Each trial lasts 30 seconds. There are 3 slow movement trials and 3 fast movement trials.




Figure 20: Target following task – follow the green gauge needle as accurately as possible.

4.7 Result Display

4.7.1 Results Summary per Hand

In the task selection menu, in the bottom you can click *Results Summary*, where all results are displayed (as in Figure 21). These individual results are also displayed directly after a task is performed. In the Results Summary view, when you click on an individual window, the task display is enlarged and there you can see more details (e.g. Force Flexion or Force Extension). When you click on the small hand on the bottom of the display, you can switch between the display of the results for the left and the right hand.




Figure 21: Results summary per hand – click on an individual window to enlarge and see details.

4.7.2 Longitudinal Results Display

With this view, you can see progress over time, in case you are performing the measurements in a longitudinal study (Figure 22). The x-axis displays assessment session number and y-axis the assessment task metric. Click on an individual window to enlarge and see details. The higher the values on the y-axis located, the better the task performance (i.e. direction of improvement is pointing upwards).



Figure 22: Longitudinal results summary. Click on an individual window to enlarge.

5 Data Storage

5.1 Raw Data

The data is always stored under the subject ID that was automatically assigned to a new user that you created. The data storage follows the structure indicated in Figure 23, that is: ETH MIKE xxx folder → Subject ID → Task Name (note: Motor Task = Fast Reaching, Sensorimotor Task = Target Following Task) → Hand → raw data (time stamp of when the measurement was started, practice rounds are marked with practice). TDMS file is preferred for data analysis, as this is coming directly from the sensors. The CSV files is what has been sent to the front-end over UDP, therefore there is a risk of small data package loss. Additionally, Position Matching task contain .txt file with target positions and indicated positions (this should be used for this task's post-processing, as TDMS file doesn't contain information on position indicated on the tablet).




Figure 23: Raw data saving – structure of the raw data storage folder.

5.2 Database

Alternative way to access already processed data will be to open the SQL database, which is created as a basis for the longitudinal results display you saw in section 4.7.2. The SQL database will allow you to easily access already processed data without the need to run post-processing on the raw data. However, the SQL database only contains the outcome measures that are displayed in the front end (one metric per task). If you would like to explore alternative metrics, you would need to work with the raw data.

To extract metrics from the database, first locate the file *db.db* as indicated in Figure 24. You can then open it with *DB Browser for SQLite*. This Software is already downloaded on the tablet, so it will automatically open when you double click on the *db.db* file. To export results of task of interest (e.g. Force Task), follow the steps indicated in Figure 25.




Figure 24: Location of the database with calculated metrics for each task.

1) Select this first to browse through the SQL database

2) Select TaskName_Aggregate to get the summary metric per task for all subjects in the database (example here: average maximum force over 3 trials in flexion/extension)

3) Export the SQL database as CSV for further analysis (e.g. in R), click save below.

SubjectNr	StartTime	LeftHand	Flexion	AvgMaxForce	AvgMaxForceKg
1	20201104_15...	1	0	0.0	0.0
2	20201104_15...	1	1	5.489070976...	0.5597293772...
3	20201104_14...	0	0	9.1649303436...	0.9345628062...
4	20201104_14...	0	1	39.124436696...	3.9895822422...
5	20201103_09...	1	0	0.0	0.0
6	20201103_09...	1	1	3.4905941486...	0.3559415446...
7	20201019_10...	1	0	0.0	0.0
8	20201019_10...	1	1	2.1428943077...	0.2185144068...
9	20201007_15...	1	0	0.0	0.0
10	20201007_15...	1	1	0.4421215554...	0.0450838518...
11	20201007_15...	0	0	14.412576039...	1.4696737458...
12	20201007_15...	0	1	40.083155314...	4.0873443341...
13	20200921_13...	1	0	0.0	0.0
14	20200921_13...	1	1	0.6475043495...	0.0660270683...
15	20200910_10...	1	0	0.0	0.0
16	20200910_10...	1	1	0.6642877658...	0.0677385004...
17	20200910_09...	0	0	11.972222328...	1.2208269213...
18	20200910_09...	0	1	34.830268859...	3.5516989858...

Figure 25: How to browse a SQL database and export results of task of interest in CSV.

6 Packing for Transportation

Follow the steps indicated in Figure 26 to pack or unpack the robot safely. The white foam is durable and can be reused for packaging and transporting the robot multiple times.

The robot is placed a box of 50 cm × 50 cm × 70 cm. First, place the robot, with its end-effector fixated with the end-effector lock, inside the box (#1 and #2). The screw can be slightly tightened for transportation using an Allen key. Then, cover the end-effector with the bubble wrap. Cover larger components (handles and safety stop) in bubble wrap and place it under the tablet rails (#3). Tablet and its keyboard, placed inside a box, should be covered in the bubble wrap and placed at the back of the robot (#3). Then, place a foam protection above the tablet rails (#4). Next, place all cables and small components into their designated box and cover it with bubble wrap. Also cover the incline plane in bubble wrap. Place the box with cables and the incline plane on top of the foam (#5). Place two protective pieces of foam (#6). Close the box, then wrap it in bubble wrap and place inside a bigger box for transportation (57 cm × 57 cm × 77 cm). This is step only necessary for shipment. The total weight of the system is around 20 kg.




Figure 26: How to pack the robot safely.

7 Troubleshooting

Below you will find a description of possible issues and how to solve them.

7.1 Software disconnects from Hardware

Follow the steps below to address the issue:

1. Make sure that the USB cable is connected (connecting the robot with the tablet).
2. Restart the robot hardware (turn the switch on the side of the robot on and off). Don't forget to place the end-effector in the middle when turning the robot on again.
3. Software: If you are in *Task Selection*, click on the *Main Menu* button and there select *Logout*. Wait for the connection icon to turn green. Once it runs green and if you were not finished with last session, click *Resume last session* instead of starting a new session (*New Session*) with the subject code you were using when the robot disconnected.
4. If the robot disconnects during an assessment, go back to Task Selection (click *Abort*) and wait until the icon turns green. Then repeat the assessment that was interrupted.
5. If the icon doesn't turn green after you have tried steps 1-4, close the application (click *Quit* in *Main Menu*).
6. Restart the robot again.
7. Open the application. Insert subject code and click *Resume last session*.




Figure 27: In case connection between software (tablet) and hardware (robot) gets lost/disturbed, you will see the icon on the right appearing “broken” and in red colour.

7.2 Software crashes

In case the software application suddenly closes, open it again and you will see the pop-up as in Figure 28. Confirm with *Yes* that you would like to continue, and you will get back to where you were before the shutdown.




Figure 28: Pop-up that appears when you reopen the application after a software crash.

8 Contact Information

Monika Zbytniewska
Rehabilitation Engineering Laboratory
ETH Zurich
BAA, Lengthalde 5
8008 Zurich
+41 779 615 277
monika.zbytniewska@hest.ethz.ch
<http://www.relab.ethz.ch>

Olivier Lambercy
Rehabilitation Engineering Laboratory
ETH Zurich
BAA, Lengthalde 5
8008 Zurich
+41 44 510 7246
olivier.lambercy@hest.ethz.ch
<http://www.relab.ethz.ch>