



Dutch Lab Pvt Ltd

Address Line 1, WB, Kolkata-700053,
Mobile: 9804806599/ Website: dutch.com
Email: anirbankreative22@gmail.com

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER: 1710931612331							DATE OF ISSUE: 2024-03-20					
ULR NUMBER: TC-24200000004F							RECEIVED DATE: 2024-03-20					
CUSTOMER ADDRESS: Customer Address Line 1							CAL.DATE: --					
							DUE.DATE: 19/3/2024					
							CONDITION: Good					
							CALIBRATED AT: LAB					
DUC DETAILS							STANDARD DETAILS					
DESCRIPTION: Vernier Caliper 2-2 mm LC 4 mm Size 5 mm							DESCRIPTION: Good					
MAKE: MK-001							MAKE: MK001					
SL.NO: SN-001							SL.NO: SN001					
ID.NO: AN-001							CERTIFICATE.NO: CC001					
RANGE: 2 - 2 mm							VALIDITY: 19/3/2024					
L.C: 4 mm							TRACEABILITY:					
CALIBRATION PROCEDURE & REF.STD: CP002 & IS 6468												
ENVIRONMENTAL CONDITION:												
TEMPERATURE (°C): 12												
HUMIDITY (RH %): 13												
CALIBRATION RESULT (All Values are in mm):												
READING in mm	0	2.5	5.1	7.7	10.3	12.9	15	17.6	20.2	22.6	28.3	
DEVIATION in µm	Set	2.4	5.2	7.3	10.2	12.6	15.22	17.3	20.6	22.2	28.3	

SPECIFIED VALUE	MEASURED VALUE	SPECIFIED VALUE	MEASURED VALUE
mm	mm	mm	mm
25	25.55	150	150.55
50	50.22	175	175.55
75	75.33	200	200.66
100	100.22	225	225.99
125	125.11	250	250.66

The Calibration Certificate is valid only for the condition of the received DUC at the time under the stated condition of calibration. The calibration certificate shall not be reproduced in full without written approval of TCS Head. DUC: Device Under Calibration. Calibration Measurement are traceable to SI Units through unbroken chain of calibration from competent laboratory.



Dutch Lab Pvt Ltd

Address Line 1, WB, Kolkata-700053,
Mobile: 9804806599/ Website: dutch.com
Email: anirbankreative22@gmail.com

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER: 1710931612331	DATE OF ISSUE: 2024-03-20
-----------------------------------	---------------------------

REMARKS:

- 1. The Expanded Uncertainty $\pm 8.3 \mu\text{m}$ Uncertainty at 95.45 % confidence level with coverage factor $k = 2$.
- 2. The Calibration Certificate is valid for scientific & industrial purpose only.
- 3. This remarks 3rd



Raj Deep
Calibrated By



Mina Das
Approved by

The Calibration Certificate is valid only for the condition of the received DUC at the time under the stated condition of calibration. The calibration certificate shall not be reproduced in full without written approval of TCS Head. DUC: Device Under Calibration. Calibration Measurement are traceable to SI Units through unbroken chain of calibration from competent laboratory.