



Dutch Lab Pvt Ltd

Address Line 1, WB, Kolkata-700053,
Mobile: 9804806599/ Website: dutch.com
Email: anirbankreative22@gmail.com

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER: 1710933013388							DATE OF ISSUE: 2024-03-20				
ULR NUMBER: TC-24200000004F							RECEIVED DATE: 2024-03-20				
CUSTOMER ADDRESS: Customer Address Line 1							CAL.DATE: --				
							DUE.DATE: 19/3/2024				
							CONDITION: Good				
							CALIBRATED AT: LAB				
DUC DETAILS							STANDARD DETAILS				
DESCRIPTION: Vernier Caliper 2-2 mm LC 4 mm Size 5 mm							DESCRIPTION: Good				
MAKE: MK-001							MAKE: MK001				
SL.NO: SN-001							SL.NO: SN001				
ID.NO: AN-001							CERTIFICATE.NO: CC001				
RANGE: 2 - 2 mm							VALIDITY: 19/3/2024				
L.C: 4 mm							TRACEABILITY:				
CALIBRATION PROCEDURE & REF.STD: CP002 & IS 6468											
ENVIRONMENTAL CONDITION:											
TEMPERATURE (°C): 12											
HUMIDITY (RH %): 13											
CALIBRATION RESULT (All Values are in mm):											
READIN G in mm	0	2.5	5.1	7.7	10.3	12.9	15	17.6	20.2	22.6	28.3
DEVIAT ION in µm	Set	2.4	5.2	7.3	10.2	12.6	15.22	17.3	20.6	22.2	28.3

SPECIFIED VALUE	MEASURED VALUE	SPECIFIED VALUE	MEASURED VALUE
mm	mm	mm	mm
25	25.55	150	150.55
50	50.22	175	175.55
75	75.33	200	200.66
100	100.22	225	225.99
125	125.11	250	250.66

The Calibration Certificate is valid only for the condition of the received DUC at the time under the stated condition of calibration. The calibration certificate shall not be reproduced in full without written approval of TCS Head. DUC: Device Under Calibration. Calibration Measurement are traceable to SI Units through unbroken chain of calibration from competent laboratory.



Dutch Lab Pvt Ltd

Address Line 1, WB, Kolkata-700053,
Mobile: 9804806599/ Website: dutch.com
Email: anirbankreative22@gmail.com

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER: 1710933013388	DATE OF ISSUE: 2024-03-20
-----------------------------------	---------------------------

REMARKS:

- 1. The Expanded Uncertainty $\pm 8.3 \mu\text{m}$ Uncertainty at 95.45 % confidence level with coverage factor $k = 2$.
- 2. The Calibration Certificate is valid for scientific & industrial purpose only.
- 3. This remarks 3rd



Raj Deep
Calibrated By



Mina Das
Approved by

The Calibration Certificate is valid only for the condition of the received DUC at the time under the stated condition of calibration. The calibration certificate shall not be reproduced in full without written approval of TCS Head. DUC: Device Under Calibration. Calibration Measurement are traceable to SI Units through unbroken chain of calibration from competent laboratory.