



TRACEMETRICS

WORLD CLASS TEST AND CALIBRATION LAB

No. 19, 1st Floor, Peenya Industrial Area, 1st Stage, TVS Cross, Bangalore - 560 058.

www.tracemetrics.in, Email- tracemetrics@gmail.com

Phone No: +91 9986588228, 9035304667



CC-2067

NABL Accredited calibration Lab as per ISO/IEC 17025:2017

CALIBRATION CERTIFICATE

TMS/FF-22

Name of the Customer : M/s.	GLOBE Tech Engineering Service	Page No.
Address	No.18,3rd Floor,VLS Nilaya,1st House, 2nd Main,Hebbal Sarovara Layout, Kodigehalli,Bengaluru-560092	1 of 2

Customer Reference:

SRF No.	: 124	SRF Date	: 26 March 2022
Certificate No.	: TMS / 22 / 124 01	Calibrated On	: 26 March 2022
ULR No.	: CC-2067220000000512F	Recommended Cal. Due	: 26 March 2023

Details of device under calibration (DUC):

Description	: AC Leakage Current Tester	Cal. Procedure	: TMS / CAL - 125
Make	: Kyoritsu	DUC received on	: 26 March 2022
Model / Type	: KEW SNAP 2434	Status on receipt	: Satisfactory
Sl.No.	: 8272499	Cal. At	: In House
ID.No.	: ---	Total No. of Page	: 02

Environmental Conditions:

Temperature	: 25 $\pm$ 4°C	Humidity	: 30% RH to 75% RH
-------------	----------------	----------	--------------------

Standards used for Calibration and Traceability Details:

Sl. No.	Nomenclature	Make / Model	Sl. No	Traceable to	Validity
1	Multi Product Calibrator	Transmille	J1380H21	National STD	07 Oct 2022

Note:


1. The Calibration Certificate relates only to the above DUC.
2. Calibration Certificate Shall not be reproduced except in full, without written approval of the Tracemetrics
3. The usage of NABL Symbol is as per NABL guidelines given in NABL-133.
4. Standard Maintained are traceable to National / International Standard through accredited laboratories.

Calibrated By


Akash Kumar J
(Calibration Engineer)



Authorised Signatory


Ganesan C
(Technical Manager)



TRACEMETRICS

WORLD CLASS TEST AND CALIBRATION LAB

No. 19, 1st Floor, Peenya Industrial Area, 1st Stage, TVS Cross, Bangalore - 560 058.

www.tracemetrics.in, Email- tracemetrics@gmail.com

Phone No: +91 9986588228, 9035304667

NABL Accredited calibration Lab as per ISO/IEC 17025:2017

Certificate No. : TMS / 22 / 124 01

ULR No: CC-2067220000000512F

Page No: 2 of 2

Results:

Sl. No.	Parameter	Range	STD Input in	DUC Reading in	Error Claimed in $\pm$	Observed Deviation in	Measurement Uncertainty in $\pm$
1	AC Current @ 50Hz	0 - 400 mA	40.0000 mA	40.0 mA	1.2 mA	0.0 mA	0.168 mA
2			200.0000 mA	200.4 mA	4.4 mA	0.4 mA	0.840 mA
3			360.000 mA	360.9 mA	7.6 mA	0.9 mA	1.512 mA
4		0 - 4 A	0.400000 A	0.400 A	0.012 A	0.000 A	0.002 A
5			2.000000 A	2.006 A	0.044 A	0.006 A	0.008 A
6			3.600000 A	3.610 A	0.076 A	0.010 A	0.015 A
7		0 - 100 A	10.00000 A	10.0 A	0.6 A	0.0 A	0.042 A
8			50.0000 A	50.2 A	1.4 A	0.2 A	0.105 A
9			90.0000 A	90.2 A	2.2 A	0.2 A	0.189 A

Conclusion Remarks:-

1. All readings are within the accuracy limits.
2. Accuracy Mentioned as per Manufacturer.
3. Measurement Uncertainty is reported at 95% confidence level. With k=2

*****End of Calibration Certificate*****

Calibrated By

Akash Kumar J
(Calibration Engineer)



Authorised Signatory

Ganesan C
(Technical Manager)