

logo

Testing Lab
34, XX Street

Sample Calibration Certificate

Certificate No: 34

Certificate Date: 09-Jun-2023

SRF No: 2023/TL2/I00003
SRF Date: 2023-06-09
Customer DC: 2345
Customer DC Date:2023-06-07

Company Name:ABC Limited
Company Address:100 Fort Street, 3rd Cross, BTM layout
Contact Person:Ajay Kumar
Contact Number:4747847478
Department:Quality
Report Company Name:ABC Limited
Report Company Address:100 Fort Street, 3rd Cross, BTM layout

Instrument Name : RANGE: 50 – 75 mm L.C. : 0.01 mm

Make	: MITUTOYO	Identification No.	: ERTT/345
Sr. No.	: 96041228	LAB ID No.	: 34555XX

CALIBRATION RESULTS

**A. INSTRUMENTAL
ERROR**

All readings are in mm

REFERENCE STD.	50	52.5	55.1	57.7	60.3	62.9	65	67.6	70.2	72.8	75
READINGS											
ERROR in μm	0	0	0	-1	-1	0	0	-1	-2	-1	-2

Parallelism of Measuring Faces: 0.001 mm

Flatness of Measuring Face: 0.0008 mm

***** End of certificate *****