



SIZE MASTERS CALIBRATION SERVICES

488-B, 2nd Block, 2nd Floor, 14th Cross, Peenya 2nd Stage
Bengaluru - 560 058. Mob : 99026 54445
E-mail : sizemasterblr@gmail.com



CC-3254

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE OF: External Micrometer	ISSUE DATE : 11/05/2023	CERTIFICATE NO. 2023/602/001	PAGE 1 OF 2
DATE OF CALIBRATION: 11/05/2023	NEXT DATE OF CALIBRATION: 10/05/2024 (as per customer request)	ULR: CC325423000000731F	DISCIPLINE: MECHANICAL
CUSTOMER'S REF. D.C.NO. : 034 DATED: 09/05/2023 LOCATION OF CALIBRATION : In Lab	DATE OF RECEIPT : 09/05/2023 CONDITION OF ITEM : Satisfactory	GROUP: DIMENSION (BASIC-MEASURING INSTRUMENT, GAUGES ETC.)	

Name of Customer & Address : M/s. DOWIN HYDRAULICS.
No.16, Sy.No.112/2, 12th Cross, Doddanna Industrial. Estate,
Near Peenya II stage, Bangalore-560 091.

Reference Document No. : IS 2967-1983

Calibration Procedure No. : WI/M/06

Environmental Conditions : Temperature : 20 °C (±1 °C) Humidity : 50 ± 10 % Rh

EQUIPMENT USED FOR CALIBRATION & IT'S TRACEABILITY :

Name of Equipment	Identification No. / Sr. No.	Calibration Certificate No	Traceability With	Calibration Date	Valid up to
Micrometer Check Set	SMCS-B/SG-02	TCS/22-23/14127-1	CC-2231	22/12/2022	22/12/2023
Comparator Dial	SMCS-B/CD/01	25020223	CC-2460	04/02/2023	03/02/2024
Comparator with Comparator Stand	SMCS-B/CB/01	2022/56043/002	CC-2460	21/12/2022	20/12/2024

All Masters used for calibration are Traceable to National Physical Laboratory India through NABL accredited lab as per the above path.

Uncertainty of Measurement : ± 2.0 µm

The uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor $k = 2$ corresponds to confidence level of 95.45%.
Note:1) Calibration Results are given on Page No.2 onwards.

Ext. Micrometer : RANGE: 50-150 mm L.C. : 0.01 mm

Make : ---	Identification No. : DH/MM/01
Sr. No. : ---	LAB ID No. : 602/001

Calibrated by



Authority Signatory

Note : This certificate is issued subject to conditions stated overleaf.

FORMAT NO. CL05-A



SIZE MASTERS CALIBRATION SERVICES

488-B, 2nd Block, 2nd Floor, 14th Cross, Peenya 2nd Stage
Bengaluru - 560 058. Mob : 99026 54445
E-mail : sizemasterblr@gmail.com



CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE OF: External Micrometer	ISSUE DATE : 11/05/2023	CERTIFICATE NO. 2023/602/001	PAGE 2OF2
DATE OF CALIBRATION: 11/05/2023	NEXT DATE OF CALIBRATION: 10/05/2024 (as per customer request)	ULR: CC325423000000731F	DISCIPLINE: MECHANICAL
CUSTOMER'S REF. D.C.NO. : 034 DATED: 09/05/2023 LOCATION OF CALIBRATION : In Lab	DATE OF RECEIPT : 09/05/2023 CONDITION OF ITEM : Satisfactory	GROUP: DIMENSION (BASIC-MEASURING INSTRUMENT, GAUGES ETC.)	

CALIBRATION RESULTS

A. INSTRUMENTAL ERROR

All readings are in mm

REFERENCE STD. READINGS	125	127.5	130.1	132.7	135.3	137.9	140	142.6	145.2	147.8	150
ERROR in μm	0	0	0	-1	-1	-1	-2	-2	-3	-3	-4

B. Actual length of micrometer setting standard

Sr. No.	Nominal Length (mm)	Error in μm
1	50.000	9
2	75.000	7
3	100.000	19
4	125.000	16

Parallelism of Measuring Faces: 0.002 mm


Specified Accuracies as per I.S. 2967 (1983):-Clause No. 4.2 & 4.3.

Measuring Range of Micrometer (mm)	0-50	50 - 100	100 - 150	150 - 200	200 - 250	250 - 300	300 - 350	350 - 400	400 - 450	450 - 500
Tolerance on zero setting (μm)	± 2	± 3	± 4	± 5	± 6	± 7	± 8	± 9	± 10	± 11
Tolerance on Parallelism of measuring faces (μm)	2	3	4	5	6	7	8	9	10	11
Error of measurement (μm)	± 4	± 5	± 6	± 7	± 8	± 9	± 10	± 11	± 12	± 13

x-x-x-x-x- End of certificate -x-x-x-x-x-


Calibrated by




Authority Signatory

Note : This certificate is issued subject to conditions stated overleaf.

FORMAT NO. CL05-A