

Calibration On	Next Calibration Due	Certificate No.:	Certi Issue Date	Sheet No.
22 Jul 2026	22 Aug 2026	CAL/CERT/2026/00011	22 Jul 2026	1 of 1

M&M Plant Calibration Customer / Address Details	Location of Calibration : Permanent Laboratory
---	--

Description & Identification	
Instrument (UUC) : SNAP GAUGE Make : LPP Serial No. : WPA1030 ID No. : SNG934/03 Calibration Range : 8	Model No. : - Range : 8 Least Count : - Instrument Cond. : SATISFACTORY Location : M&M Plant
Procedure reference : AE/CAL-SOP/01 Environmental Conditions : Temperature at 20° C RH 55 % Standard Reference : IS / ISO Standard Calibration Guidelines Discipline : DIMENSION (Basic Measuring Instrument, Gauge etc)	

TRACEABILITY OF MASTER USED :						
Instrument Desc.	Make	Sr No / Id. No.	Cert.No.	Dt.of Cal	Due Dt.	Cal.Agency
Vernier Caliper	-	VC-001	AE/CC/REF/01	22 Jul 2026	10 Jul 2024	NABL Lab
All the measurements performed are traceable to National/Int. standards through NABL accredited cal.lab.						

Calibration Result (ALL VALUES ARE IN mm)						
Sr No.	Description	STANDARD VALUE	AVG OBS VALUE UUC	test	ERROR	Status
01	GO	20.0000	20.0100	test	0.0100	PASS
02	NOGO	30.0000	30.0100	test	0.0100	PASS
03	GO	40.0000	40.0100	-	0.0100	PASS
04	NOGO	50.0000	50.0100	-	0.0100	PASS
Uncertainty of Measurement at coverage factor k = 2 at 95.45 % of confidence Level = ±0.00mm						

Calibrated By Calibration Engineer ULR : ULR/2026/00005	CALIBRATION SEAL / STAMP	Authorized By Quality Manager
---	-----------------------------	----------------------------------

CALIBRATION CENTER : Laboratory Address, Behind Main Road, Industrial Zone, State - 440024. Website: www.gaugemaster.com Email: info@gaugemaster.com Phone: +91 98222 23948
