

CALIBRATION CERTIFICATE

Certificate No: CAL/CERT/2026/00011
Date of Calibration: 22 Jul 2026

ULR No: ULR/2026/00005
Next Due Date: 22 Aug 2026

Instrument Name	SNAP GAUGE
ID Code	SNG934/03
Make	LPP
Range	8
Least Count	-
Serial No	WPA1030
Location	M&M Plant

Name	Vernier Caliper
ID / Serial No	VC-001
Traceable To	NABL
Range	0-150mm
Least Count	0.02mm
Valid Until	10 Jul 2024

Temperature: 20

Humidity: 55

Pt	Nominal	Ascending	Descending	Error	Tolerance	Status
1	20	20.01	-	0.01	±0.01	PASS
2	30	30.01	-	0.01	±0.02	PASS
3	40	40.01	-	0.01	±0.01	PASS
4	50	50.01	-	0.01	±0.01	PASS

Measurement Uncertainty: -
Verdict: **PASS**

(Name)
Calibrated By

(Name)
Reviewed By

(Name)
Approved By

This certificate is issued based on calibration conducted as per standard procedures. The results relate only to the item calibrated.