



0~25MM MICROMETER CALIBRATION REPORT

Serial no:		Cali. place:		Standard room		Calibration Date:			
Make:		Temperature:	20° ±2°			Next cal due date:			
Range:	0 - 25mm	Least Count				Report No:			
Type:	Analog / Digital	Instrument used:				Used instr. L.C			
Humidity	RH 50±10%					Location of Instrument:			
INSTRUMENT RECEIPT CONDITION.									
Dent & Damages:-									
Working Condition:-									
Calibration measuring anvils									
Sl.No.	Std. Spec	1	2	3	4	5	Avg	Error	Judgement
1	2.510mm								
2	5.120mm								
3	7.700mm								
4	10.380mm								
5	12.900mm								
6	15.000mm								
7	17.600mm								
8	20.200mm								
9	22.800mm								
10	25.000mm								
Flatness of the anvils				Parallelism of anvils					
Std spec.	Actual	Error	Judge.			Std spec.	Actual	Error	
						Corner1			
						Corner 2			
						Corner 3			
						Corner 4			
				Measuring anvil should be free from dent and damages.					
Acceptance critiria									
Sl.No.	Actual mesured		Error						
1	Total error		0.005						
2	Parallelity		0.003						
3	Flatness		<0.001						
Remarks:									
		Approved By:		Checked By:		Calibrated By:			