



VERNIER CALIPER CALIBRATION REPORT

Serial no:		Cali. place:	Standard Room	Calibration Date:	
Make:		Temperature:	20°±2°C	Next cal due date:	
Range:		Least Count:		Report No:	
Type:	Analog / Digital	Instrument used:	VCT	Used instr. L.C	0.005
Humidity	RH 50±10%				

INSTRUMENT RECEIPT CONDITION.

Dent & Damages:-

Dial working Condition:-

Calibration of external jaws

Sl.No.	Std. Spec	1	2	3	4	5	Avg	Error	Judgement
1	20.00mm								
2	50.00mm								
3	100.00mm								
4	150.00mm								
5	200.00mm								
6	250.00mm								
7	300.00mm								

Calibration of internal jaws

Sl.No.	Std. Spec	1	2	3	4	5	Avg	Error	Judgement
1	20.00mm								
2	50.00mm								
3	100.00mm								
4	150.00mm								
5	200.00mm								
6	250.00mm								
7	300.00mm								

Depth blade calibration

Std spec.	Observed	Error	Judge.
20mm			
50mm			
100mm			

Parallelism calibration

	Std spec.	Observed	Error	Judge.
Internal jaws	25.00mm			
External jaws	25.00mm			

All the jaws are free from dent and damages.

Acceptance criteria

LEAST COUNT	0.01mm		0.02mm		0.05mm	
Measured length(mm)	Maximum permissible error in measurement					
	New	Recalib	New	Recalib	New	Recalib
0	±0.010	±0.020	±0.020	±0.030	±0.050	±0.060
100	±0.020	±0.030	±0.030	±0.040	±0.060	±0.070
200	±0.020	±0.030	±0.030	±0.040	±0.070	±0.080
300	±0.030	±0.040	±0.040	±0.050	±0.080	±0.090
400	±0.030	±0.040	±0.040	±0.050	±0.090	±0.100
500	±0.030	±0.050	±0.050	±0.060	±0.100	±0.110
600	±0.030	±0.050	±0.050	±0.060	±0.110	±0.120

Reference: IS 3651 (Based on)

Remarks:

Approved By:

Checked By:

Calibrated By:

