

CALIBRATION CERTIFICATE

Certificate No: CAL/CERT/2026/00002
Date of Calibration: 20 Jul 2026

ULR No: ULR/2026/00001
Next Due Date: 20 Aug 2026

Instrument Name	SNAP GAUGE
ID Code	SNG934/03
Make	LPP
Range	8
Least Count	-
Serial No	WPA1030
Location	M&M Plant

Name	Vernier Caliper
ID / Serial No	VC-001
Traceable To	NABL
Range	0-150mm
Least Count	0.02mm
Valid Until	10 Jul 2024

Temperature: 20

Humidity: 45

Pt	Nominal	Ascending	Descending	Error	Tolerance	Status
1	30	30.001	-	0.001	±0.01	PASS
2	40	40.001	-	0.001	±0.01	PASS
3	50	50.001	-	0.001	±0.01	PASS
4	60	60.001	-	0.001	±0.01	PASS
5	70	70.001	-	0.001	±0.01	PASS
6	80	60.001	-	-19.999	±0.01	FAIL

Measurement Uncertainty: -

Verdict: **PASS**

(Name)
Calibrated By

(Name)
Reviewed By

(Name)
Approved By

This certificate is issued based on calibration conducted as per standard procedures. The results relate only to the item calibrated.