



CALIBRATION CERTIFICATE

SANSERA
ideas@work

PLANT-2

CERTIFICATE NUMBER:	SAN/26/06/0013	DATE OF ISSUE:	09-06-2026
ULR NUMBER:	CC-239926500001151F	RECEIVED DATE:	03-06-2026
CUSTOMER NAME & ADDRESS: Sansera Engineering Ltd Plant-7 # 143 A Bommasandra-Jigini Link Road Indl. Area, Bangalore, KIADB PHASE-2, JIGANI IND 560105, Bangalore, Karanataka - 560105.		CAL.DATE:	03-06-2026
		SRF.NO:	SAN/SRF/26/06/0003
		SANSERA LAB ID NO:	IWMC03062601
Customer Reference No:	dc-02123	CALIBRATED AT:	Metrology Lab
DISCIPLINE:	MECHANICAL DIMENSION	GROUP:	BASIC INSTRUMENT & GAUGE

DUC DETAILS

DESCRIPTION:	MASTER COMPONENT	DUC RECEIPT CONDITION:	Okay
MAKE:	NIIGATA SEIKI	MODEL:	158-120
SL.NO:	-	ID.NO:	Msx213
RANGE:	0.001 mm		

CALIBRATION PROCEDURE : This Method Is Based On Standard IS 3485 As Per SWI/LAB/315

STANDARD DETAILS

DESCRIPTION	MAKE / MODEL	Calibration Agency	SL.NO / ID.NO	VALIDITY	CERTIFICATE.NO	TRACEABILITY
Multi Gauge - 10148132	Mitutoyo / DL-S3	IVIEWSENSE	IWSPG29012537 / 10148132	06/05/2027	SAN/25/01/0449	-
ENVIRONMENTAL CONDITIONS	TEMP	20 ± 2°C	RH		50 ± 10%	
	ACTUAL	20.01°C	ACTUAL		52.94%	

CALIBRATION RESULTS

DESCRIPTION	NOMINAL VALUE	OBSERVED VALUE
Diameter ID	16.04	16.0421
Diameter OD	18.05	18.0495
Height	78.003	78.0049
EXPANDED UNCERTAINTY : ± 1.6µm For Diameter ,(The uncertainty of measurement is expressed at 95.4% Confidence with coverage factor K-2)		

Calibrated By		Reviewed By		Authorized By	
Gopi V		Moorthi		Moorthi	



METROLOGY LABORATORY , SANSERA ENGINEERING LTD
261/C, Bommasandra Industrial Area, Hebbagodi Post, Anekal Taluk, Bangalore - 560099
Ph: 0987654322 / 0987654322,
E-mail id: calibmaster1@gmail.com



End of Certificate