

Tactical Innovation Document: Emergency Perfusion System

Bio-Hydraulic Engine for Ischemic Tissue and Hypoperfusion Management

Innovator / Lead Researcher: *Wasfi Ahamad Mohamed Al-shehada*

Original Archival Reference: Zenodo Archive

Final Update Date: May 20, 2026

1 Introduction and Scientific Background

Conventional therapies for acute ischemic conditions and Diabetic Foot Ulcers (DFU) rely on slow-acting biological mechanisms that stimulate microvascular growth, a process known as *Angiogenesis*. A major flaw in these traditional methods is the critical dimension of time; healthy cells perish, and ion channels degrade before the biological reconstruction can fully materialize.

This project introduces a bio-engineering solution based on **Emergency Perfusion** via a micro-mechanical network of tubes and catheters. Operating under controlled hydraulic pressure, this system instantly delivers life-saving perfusion to oxygen-deprived tissues.

2 Engineering and Mechanical System Description

The system consists of closed mechanical supply lines (micro-tubes) utilizing an active "targeted micro-injection and propulsion" strategy, completely bypassing the limitations of traditional hypodermic needle injections:

- **Propulsion and Feed Tube:** Directly infuses thrombolytic/clot-dissolving agents, vasodilators, and biological growth factors right into the core of the affected ischemic area.
- **Evacuation and Negative Pressure Tube:** Operates simultaneously to extract cellular debris resulting from chemical decomposition, effectively cleansing the newly formed "nano-vacuoles".

3 Mathematical and Physical Modeling of the System

To ensure optimal emergency perfusion without inducing mechanical rupture in the fragile microvasculature, the movement of bio-fluids and therapeutic solutions within the tubes adheres to the following hydraulic flow and continuity equations:

3.1 Volumetric Flow Rate

The volume of fluid delivered under emergency conditions through the mechanical tube is calculated using the relation:

$$Q = A \cdot v \quad (1)$$

Where Q is the volumetric flow rate, A represents the cross-sectional area of the mechanical tube, and v is the velocity of the propelled therapeutic solution.

3.2 Targeted Mechanical Pressure Formula (Poiseuille's Law)

To determine the precise high-magnitude pressure required to clear micro-vascular blockages in the diabetic foot:

$$\Delta P = \frac{8\mu LQ}{\pi r^4} \quad (2)$$

Where:

- ΔP : The differential mechanical pressure required for urgent deployment.
- μ : The dynamic viscosity of the injected bio-fluid or blood.
- L : The length of the navigating catheter tube.
- r : The inner radius of the targeted ion channel or blood vessel.

4 Conclusion and Current Applications

This innovation bridges the gap between hydraulic mechanical engineering and human cellular biology for the first time. The current documentation securely archived within the **Zenodo** platform guarantees complete intellectual priority for this advanced offensive-defensive medical tactic.