अखिल भारतीय आयुर्विज्ञान संस्थान
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
साकेत नगर भोपाल (मध्यप्रदेश) – 462020
Saket Nagar, Bhopal (M.P.) – 462020
web site: www.aiimsbhopal.edu.in

AIIMS BHOPAL
CENTRAL STORES DEPARTMENT
Dispatch No. 819
Date: 11/03/2022

Ref.: Stores/2021-22/ AIIMS Bhopal/ Ophthalmology/PAC – 166(i)/ 819 **Dated:** 11 /03 **/2022**

Subject: Purchase of **Clinical Specular Microscope, Model: Cell Chek 20** for the **Department of Ophthalmology** on Proprietary basis-Inviting comments thereon.

The request received from Department of Ophthalmology, Bhopal, for the purchase of captioned Item from **(OEM) M/s Konan Medical, Inc. 10-29 Miyanishicho, Nishinomiya , Hyogo 662-0976, Japan** and through its authorized sales and servicing distributor **M/s BioMedixoptotechnik & Devices Pvt. Ltd., No. 258, Amar Jyothi Layout, Domlur, Banglore – 560071, Karnataka, India on Proprietary basis.**

The Notice is being uploaded for general information of Aspirant Manufacturer/Dealer/Distributor to submit their objections/proposal, if any, on proprietorship of these items.

In case, the product of any Manufacturer/Authorized distributor/dealer conforms to the enclosed specifications, they may submit their proposal for the supply same accessories /Spares along with the following:-

(i) Details of the required product brochure.
(ii) Point-by-point compliance of the enclosed specifications, along with all relevant documentary evidence;

The objection/ proposal should be sent in sealed cover to **The I/c Sr. Procurement Officer, Service Building, First Floor, AIIMS, Saket Nagar, Bhopal(462020)** so as to reach on or before 25**/03/2022 upto17:00 Hrs.,** failing which it will be presumed that no any other vendor is interested to offer comments/protest and case will be decided accordingly on its merit.

The reference number: Stores/2021-22/AIIMS Bhopal/ Ophthalmology/PAC – 166(i)/ 819**, dated:** 11 **/03/2022 due on**25**/03/2022** should be super scribed on sealed envelope.

11/3/2022

Yours faithfully,

I/c Sr. Procurement cum Stores Officer.
Sr. Procurement Cum Store Officer
Central Stores Department
AIIMS, Saket Nagar, Bhopal (M.P.)

Encl: 1. Authorization Certificate
2. PAC Certificate
3. Schedule of requirement

Authorization Certificate

KONAN MEDICAL, INC.
Miyanishicho, Nishinomiya, Hyogo 662-0976, JAPAN
TEL +81-798-36-3456 FAX +81-798-26-1028

Authorization Letter

To whom it may concern,

We, Konan Medical, Inc. 10-29 Miyanishicho, Nishinomiya, Hyogo 662-0976, Japan, hereby certify that M/s. BioMedix Optotechnik & Devices Pvt. Ltd., No. 258, Amar Jyothi Layout, Domlur, Bangalore - 560 071 - Karnataka, India, is our authorized distributor for the complete product range in India.

M/s. BioMedix Optotechnik & Devices Pvt. Ltd. is authorized by us to sell, bid, negotiate, distribute, enter into sales/service contracts and to be responsible for after sales service on behalf of us in India.

This Authorization Letter shall be valid until 31st March 2022.

Yours faithfully,

IKEGAMI, Tetsuji / President & CEO
Konan Medical, Inc.

Date: 1st April, 2021

株式会社コーナン・メディカル
〒662-0976 兵庫県西宮市宮西町10-29 TEL 0798-36-3456(代表) FAX 0798-26-1028
〒169-0075 東京都新宿区高田馬場4-40-17 TEL 03-5337-2025(代表) FAX 03-5337-2009

Proprietary Article Certificate

KONAN

KONAN MEDICAL INC.

Date: 27 August, 2021

Proprietary Certificate

We hereby certify that Konan Medical, Inc. is the Manufacturer of CellChek 20 Clinical use Specular Microscope.

This Specular Microscope is built on the proprietary technology of Konan based on the experience over the last 4 decades. This device is FDA cleared, and incorporates Center method cell analysis software. It is universally known that the Center Method is considered as the gold standard. Verification of captured locate can be done on this device. No other manufacturers other than Konan Medical has this verification function.

Eiji Takaso
Sales Manager
Konan Medical, Inc.

11/03/2022

Schedule of requirement

S.No.	Name of the Equipment	Description	Qty (No.)
01	Clinical Specular Microsscope	Make : Konan Medical Inc, Japan Model: Cell Check 20	1

**** Note: Please don't quote the prices of the above schedule requirement ****

SPECULAR MICROSCOPE (CLINICAL)

1) The equipment should be reliable, advanced international standard machine with a long standing reputation.
   Corneal endothelium photography
2) Photography magnification254x (on the control panel)
3) Photography range0.25- 0.35 x0.55- 0.60 mm
4) Photographic location- center, peripheral locations (0,45,135,180,225,315)
5) Resolving powerMore than 125 line/mm
6) Fixation targetCentral and peripheral
7) Corneal thickness measurement range 0.400-0.750mm
8) Display unit: 0.001mm step display
9) Analytical method- fully automated/ semi-automated/center method with verification of captured locate
10) Display monitor– integrated touch panel
11) Dimensions 286-468mm (W) x 445-592mm (D) x 486-681mm (H)
12) Weight 15-20 Kgs
13) Power supplySource voltage 100-240V AC, 50-60 Hz
14) Power input 70-120VA
15) Motorized table -Heavy duty motorized table from original manufacturer with slots for keeping standard accessories with laser printer
16) The instrument should be with USFDA/CE or equivalent international certification.
17) Upgradable with latest and newer facilities preferable
18) Demonstration in front of technical committee of equipment along with accessories and attachment before finalization of order
19) The instrument should be with online UPS or CVT as per standard recommendation and with 5 year replaceable warranty and dust cover.
20) OEM will be responsible for equipment delivery, warranty , maintenance, service and repair if vendor, third party defaults.
21) Applicable turn-key requirement to be fulfilled for appropriate installation, functioning and patient examination for the equipment.