Canon
CANON MEXICANA, S. DE R.L. DE C.V.

TITLE: **DEVICE INSTALLATION**					
REF: HEALTHCARE-018-18	VALIDITY BEGIN			VERSION	PAGE 1 OF 3
	DAY	MONTH	YEAR		
	31	05	2018	**1 .0**	

I. OBJECTIVE

Establish the guidelines and procedures necessary to carry out the installation of Healthcare devices.

II. SCOPE

This policy will be applicable to Canon Mexicana Personnel.

III. RESPONSABILITIES

The Healthcare area, in order to ensure quality, effectiveness, efficiency and customer satisfaction during device installation, will follow these guidelines.

The Canon Healthcare Area has two product lines: radiology and ophthalmology. The Healthcare Service department will be responsible for carrying out device installation for both lines, as well as for training and certifying the external personnel who will be able to perform this task, ensuring that the specific technical and quality requirements are fulfilled according to the solution offered to the customer.

The Canon Mexicana personnel must refer the Healthcare device installation requests to the Healthcare Service Department.

IV. PROCEDURE

1) It is the responsibility of the Healthcare Manager to include all installation and start-up costs in the customer price of the devices sold by the Healthcare area, as well as to inform the Healthcare Service Department of any opposite situation.

2) Logistic coordination. The following aspects should be considered:

PREPARED BY:	REVIEWED BY:	APPROVED BY:
DANIEL PEREZ HEALTHCARE SOLUTIONS SALES AND SUPPORT ENGINEER FECHA: 31/05/2018	FEDERICO TAFOYA HEALTHCARE SOLUTIONS SALES AND SUPPORT MANAGER FECHA: 31/05/2018	MORIHIRO ETO SALES ADMINISTRATION BISG DIRECTOR FECHA: 31/05/2018

a. Coordination. The Healthcare Manager or Healthcare Specialist will provide the contact and location information of the customer that requires device installation. It is the responsibility of the Healthcare Service Department to contact and coordinate the date and conditions of installation. The following points must be verified
    i. Electrical and environmental conditions (temperature, humidity), physical location. Depending on the device technical requirements, must be ensured that these conditions are met. Otherwise, the installation will not proceed and the Healthcare Manager must be informed to carry out the appropriate corrective actions.
    ii. Schedule user training date.

b. Enter Business Trip in Lotus Notes. The travel request for the device installation must comply with the current business trip policies. In case where the policy cannot be accomplished, the Healthcare Manager will be responsible to obtain the appropriate authorizations.

3) The following topics must be considered according to the operation and service manual

    i. Unpacking
    ii. Start-up and performance tests
    iii. Configuration according to user requirements
    iv. User training
    v. Patient tests
    vi. Customer satisfaction validation.
    vii. Guarantee and service conditions.
    viii. Recommend the customer to guard the original packaging for future device shipping.
    ix. Provide contact information.
    x. Obtain the installation satisfaction survey.
    xi. Fulfill the service order using the format authorized by the Healthcare Service Supervisor.

4) Device installation is critical in the customer quality perception of the of Canon Healthcare products and services, reason why it is essential to

perform a professional and quality labor based in the knowledge acquired during the training courses offered by Canon.

5) Register the device data in the BISG HC Support data base for installed base traceability (MIF).

6) Follow up customer service regarding device good performance and offer technical support.

7) By the end of the warranty period, contact the customer to promote preventive maintenance service.