Canon
CANON MEXICANA, S. DE R.L. DE C.V.

TITLE: **PREVENTIVE MAINTENANCE**					
REF: HEALTHCARE-019/18	VALIDITY BEGIN			VERSION	PAGE 1 OF 3
	DAY	MONTH	YEAR		
	31	05	2018	**1 .0**	

I. OBJECTIVE

Establish the guidelines to provide Preventive Maintenance for the Canon Healthcare devices.

II. SCOPE

This policy is applicable to the employees of Canon Mexicana.

III. RESPONSABILITY

The Healthcare Service Department has the qualification and training to adequately attend customer service requests. In the event that any Canon representative receives a service request regarding Healthcare devices, must be referred to the Healthcare Service Department in order to assure technical quality and customer satisfaction and carry out the following points as process.

It is the responsibility of the Service Supervisor and the Healthcare Manager to establish all applicable review, service, spare parts, updates and insured shipping costs prior to the formal quotation to the customer or dealer, as well as the commercial and payment terms according to the BISG Service Management Department current policies.

IV. PROCEDURE

1) Receive customer formal request which must include name, telephone, email, location, device mode and serial number. Identify if the device was supplied directly by Canon or by a dealer, and if it is under warranty coverage.

PREPARED BY:	REVIEWED BY:	APPROVED BY:
DANIEL PEREZ HEALTHCARE SOLUTIONS SALES AND SUPPORT ENGINEER FECHA: 31/05/2018	FEDERICO TAFOYA HEALTHCARE SOLUTIONS SALES AND SUPPORT MANAGER FECHA: 31/05/2018	MORIHIRO ETO SALES ADMINISTRATION BISG DIRECTOR FECHA: 31/05/2018

2) If the device was supplied by a dealer trained and authorized to offer technical service to Healthcare devices, at first instance a professional and quality customer attention must be provided looking for a problem solution. However, the dealer will be responsible for monitoring and solving the problem according to the conditions indicated in the distribution agreement.

3) Depending on the customer geographic location and needs, the preventive maintenance service can be carried out either in the customer facilities or in Canon´s facilities. Quotations for these scenarios will be presented and must consider all costs related to shipment, personnel trips, inspections, spare parts, supplies, etc. The GP margin will be agreed between the Service Supervisor and the Healthcare Manager

4) Once the customer accepts and complies with the conditions indicated in the preventive maintenance quotation, proceed with the following points.

5) Schedule the service execution date. In case the service is to be executed in the customer's facilities, compliance with the current Business Trip policy must be observed

6) Execute the preventive maintenance service according to the indications in the service manual and in the training by Canon USA. It is the responsibility of the person who executes the service to have all the needed tools (hardware, software, manuals, support material) for the appropriate execution of the service.

7) Performance tests must be made prior to the beginning of the preventive maintenance service.

8) In the event that during the initial performance tests or during the preventive maintenance execution the service engineer finds out that spare parts, accessories or updates are necessary and were not included in the preventive maintenance routine or quotation, a corrective maintenance quotation will be presented for customer approval including all additional costs related to shipments, personnel trips, inspections, spare parts, supplies, etc.

i. In case the customer does not accept the corrective maintenance quotation, the preventive maintenance service will be finalized.
ii. In case the customer does accept the corrective maintenance quotation, proceed according to Policy HEALTHCARE 015, points 12 to 15, looking forward to finish as soon as possible the preventive maintenance service. If the execution of the corrective maintenance service entails a considerable waiting time, the preventive maintenance service must be finished and the execution of the corrective maintenance will remain pending until the conditions allow resuming.

9) Performance tests must be done at the end of the preventive maintenance service.

10) Fulfill the service order in the format authorized by the Service Supervisor, sending a copy to the customer.

11) If applicable, send back the device to the customer and confirm the correct reception thereof.

12) Register all finished support or service in the BISG HC Support Lotus Notes database, attaching the service order.

13) Follow up on the device good performance within 30 days after the service is finalized

14)Send the satisfaction survey to the customer.